

From: [Weidner, Tara](#)
To: lpwcdl@aol.com
Subject: Request for additional information
Date: Tuesday, May 07, 2013 10:46:00 AM

License No.: 37-30120-01
Docket No.: 030-33422
Control No.: 580205

Lon,

This is reference to the amendment request dated March 7, 2013 to add a location of use at the Braxton County Memorial Hospital. On April 3, 2013 we spoke and I requested the following additional information:

1. A letter signed by both parties addressing the permitted use of byproduct material at the client's address and the authority and responsibilities of the mobile service provider and the client. In addition, I asked for documentation of a clear contractual agreement concerning access to the location of use for the purpose of decontamination or removal of licensed material from the location of use in the event of disharmony between CDL and Braxton County Community Hospital.
2. Describe the scope of activities that will be performed at this location. Will patient dosages be transported to this location from your base hot lab or delivered directly to the proposed site?
3. Describe the controls that will be implemented to ensure that licensed materials are under your control and not accessible to unauthorized individuals (i.e. hospital staff). Specifically, the provided diagram indicates that the camera room is a secured area and the hot lab is a restricted area. Define what you mean by restricted vs. secured areas. Will non-licensee staff have unescorted access to these areas or will CDL have control of these areas?
4. Pertaining to Braxton County Memorial Hospital, I will need; (1) request to terminate their license, (2) a closeout survey, and (3) address what will be done with their sealed sources?

To date, I have heard from Braxton County concerning the termination of their license. They provided a closeout survey, leak test results and a request to terminate their license concurrent with the issuance of the CDL license. In addition, they stated that CDL would be taking possession of their sealed sources. Records indicate that they possess a cesium-137 (SN# 561-21-12, cobalt-57 SN# BM06057E09058104, and cobalt-57 SN# 1566-069. In addition to responding to items 1 through 3 above, please confirm that CDL will be taking possession of the above stated sealed sources.

If you have any questions feel free to give me a call or respond by e-mail.

Tara L. Weidner
Senior Health Physicist
U.S. Nuclear Regulatory Commission