

Licensee: Centro Médico del Turabo  
d/b/a Hospital HIMA San Pablo Fajardo  
License No: 52-25621-01  
Docket No: 03036225  
Control No: 580110  
Date: June 20, 2013

Dear Dr. Torres:

This correspondence is in reference to your application dated February 20, 2013, requesting the renewal of NRC License No. 52-25621-01. In order to continue our review, we need the following additional information:

1. Please specify what areas are above and below areas of use and storage of byproduct material in your nuclear medicine department.
2. Please confirm that you do not use PET radiopharmaceuticals in your facility.
3. Please confirm that you have developed and will implement and maintain written procedures for safe use of unsealed byproduct material that meets the requirements of 10 CFR 20.1101 and 10 CFR 20.1301.

We will continue our review upon receipt of this information. Please reply to my attention at the Region 1 Office and refer to Mail Control No. 580110. You may also respond via fax at (610) 337-5269 or via email to [lester.tripp@nrc.gov](mailto:lester.tripp@nrc.gov). If you have any technical questions regarding this deficiency letter, please contact me at (610) 337-5358.

In order to continue prompt review of your application, we request that you submit your response to this letter within 30 days from the date of this letter.

Thank you for your cooperation.

*Lester Tripp*  
Health Physicist  
Division of Nuclear Materials Safety  
U.S. NRC Region 1  
2100 Renaissance Blvd.  
King of Prussia, PA 19406  
(610) 337-5358 (voice)  
(610) 337-5269 (fax)  
[lester.tripp@nrc.gov](mailto:lester.tripp@nrc.gov)