



UNITED STATES
NUCLEAR REGULATORY COMMISSION

REGION III
2443 WARRENVILLE ROAD, SUITE 210
LISLE, ILLINOIS 60532-4352

JUN 07 2013

Mr. Michael Rogers
Senior Scientist
Chemir EAG
2672 Metro Blvd.
Maryland Heights, MO 63043

SUBJECT: TERMINATION OF YOUR NRC RADIOACTIVE MATERIALS LICENSE AND NON-CITED VIOLATION – EVANS ANALYTICAL GROUP, LLC – EAG LIFE SCIENCES

Dear Mr. Rogers:

In letter dated March 4, 2013, you contacted the U.S. Nuclear Regulatory Commission (NRC) and indicated that you wished to terminate your NRC radioactive material license. The NRC staff has reviewed your final status surveys. Based on its review, the staff has concluded that all licensable radioactive material has been removed from your facility and residual radioactive material attributable to licensed activities does not exceed current NRC criteria.

Based on these conclusions no further remediation or actions with respect to NRC regulated material is required. Your facility is suitable for unrestricted use and NRC license number 24-32801-01 is hereby terminated.

Based on the results of its review, the NRC has also determined that one Severity Level IV violation of NRC requirements occurred. This violation is being treated as a Non-Cited Violation (NCV), consistent with Section 2.3.2 of the Enforcement Policy. The current Enforcement Policy is included on the NRC's Web site at <http://www.nrc.gov/about-nrc/regulatory/enforcement/enforce-pol.html>. The NCV concerned the possession and transfer of 132.6 mCi of carbon-14 under your license on March 3, 2010. The possession of 132.6 mCi of carbon-14 was contrary to Item 8.B., of your license which limits the total amount of carbon-14 that may be possessed under the license to 75 mCi.

While the quantity of carbon-14 exceeded the carbon-14 possession limit listed on the license, your May 1, 2013, letter indicated that the storage containers containing the carbon-14 were never opened. As corrective action, between March 3, 2010 and March 11, 2010, you oversaw the transfer of all carbon-14 samples to an NRC licensee authorized to possess such material, and you conducted final closeout surveys for the area in which the samples were stored.

If you contest the violation or significance of this NCV, you should provide a response within 30 days of the date of this letter, with the basis for your denial, to the Nuclear Regulatory Commission, ATTN: Document Control Desk, Washington DC 20555-0001, with copies to: (1) the Regional Administrator, Region III; and (2) the Director, Office of Enforcement, United States Nuclear Regulatory Commission, Washington, DC 20555-001.


M. Rogers

- 2 -

In accordance with Title 10 of the Code of Federal Regulations 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

If you have any questions, please contact Sara A.B. Forster of my staff at (630) 829-9892.

Sincerely,

A handwritten signature in cursive script, appearing to read "Patricia J. Pelke", with a large, sweeping initial "P".

Patricia J. Pelke, Chief
Materials Licensing Branch
Division of Nuclear Materials Safety

License No. 24-32801-01
Docket No. 030-38316

Enclosure:
Amendment No. 03