

U.S. NUCLEAR REGULATORY COMMISSION**Date:** 6/25/2013**TELEPHONE CONVERSATION RECORD****Time:** 2:30 pm**Mail Control
or Report No(s).**
580082**License No(s).**

47-25622-01

Docket No(s).

030-36232

Name of Licensee: Summers County Appalachian Regional Healthcare Hospital**Name of Participant(s):** Nesta Logan, Director of Diagnostic Services
Robin Elliott, Health Physicist**Telephone No.** (304) 466-2906**Subject:** Additional information needed for license renewal**(NOTE: This will be used as the
Documents Title in ADAMS)**

Nesta was contacted to confirm/obtain the following information:

- Wes Dangerfield is the current Community CEO replacing Chris Vaught who appears on the WBL worksheet.
- The email provided for the RSO on the May 15, 2013 is a personal email and will not be added to the WBL database.
- The area below the use areas submitted in the diagrams with the May 15, 2013 letter are earth, i.e. the building is on a slab. The area above is a hallway.
- Nesta also confirmed that an L-block is used to shield doses during the assay procedure, syringe shields are used during dose administration, and all radioactive waste is shielded as well.

Action Required: Proceed with license renewal.**Document Availability:****Publicly Available****Non-Publicly Available****Non-Sensitive****Non-Sensitive Copyright****Sensitive- Proprietary****Sensitive – Privacy Act/ PII****Sensitive – Internal****Sensitive – Security-Related****Immediate Release****Normal Release Date:****Delayed Release Date:****SUNSI Review Completed By:** RElliott 6/25/13 / RA /**Document Accession #:** _____