



FROEHLING & ROBERTSON, INC.

Engineering Stability Since 1881

3015 Dumbarton Road
Richmond, Virginia 23228-5831
T 804.264.2701 | F 804.264.1202

Br. 3

June 13, 2013

Licensing Assistance Team
Division of Nuclear Materials Safety
U.S. Nuclear Regulatory Commission Region I
475 Allendale Road
King of Prussia, PA 19406-1415

SUBJECT: License No. 45-08890-02

03006580

Dear Sir,

Froehling & Robertson, Inc. would like to request an amendment to our existing radioactive materials license to change Condition 13-A regarding leakage testing interval to 12 months from the currently listed, 6 months period.

The Sealed Source and Device Registration (SS&D) on all the surface gauges in our inventory, Troxler, Humboldt, and InstronTek, require that leak test be performed every 12 months. This change would permit us to comply with the current industry standard.

Should you require any additional information, please contact us.

Respectfully submitted,
FROEHLING & ROBERTSON, INC

C. F. Starnes
Director, Technical QA/QC | Corporate RSO

REC-610628-13AM07:19

This is to acknowledge the receipt of your letter/application dated

6/13/13, and to inform you that the initial processing which includes an administrative review has been performed.

☒ Amendment (45-08890-02)
There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

☐ Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 581280.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.