

L-9
MS-16

July 9, 2013

Lester Tripp
Health Physicist
Div. of Nuclear Material Safety
USNRC, Region I
2100 Renaissance BLVD.
King of Prussia, PA 19406


Re: Request for additional information
Hospital Hima San Pablo, Fajardo, PR
Lic No. 52-25621-01
Control No. 580110

03036225

Our reply to your request is as follows:

- 1) Below the nuclear medicine department there is no other facility, only ground. Above the facility there are office space (vacant at present time) and a machinery area.
- 2) We do not use PET radiopharmaceuticals.
- 3) We have developed and will implant and maintain written procedures for safe use of unsealed byproduct materials that meets the requirements of 10CFR20.1101 and 10CFR20.1301.

Hope this reply is satisfactory. Thanks.



Heriberto Torres, Ph.D.
RSO