

Roldan, Lizette

From: Gardner, Richard [R.Gardner@MSrhc.com]
Sent: Friday, July 19, 2013 2:41 PM
To: Roldan, Lizette
Subject: RE: REQUEST FOR ADDITIONAL INFORMATION REGARDING RENEWAL CONTROL 579908
Attachments: SKMBT_60113071911050.pdf

Dear Ms. Roldan,

The requested information is included with this e-mail as an attached PDF. If I may be of further assistance, please contact me either at this e-mail address or at (907) 861-6656.

Thank you,

Richard R. Gardner, RSO

PUBLIC

- Immediate Release
- Normal Release

NON-PUBLIC

- A.3 Sensitive-Security Related
- A.7 Sensitive Internal
- Other: _____

Reviewer: *Jue* Date: *07/19/13*

1. David Mills, M.D. no longer practices in this area and was intentionally not included in the renewal application.
2. Radiation monitoring instruments will be calibrated by a person qualified to perform survey meter calibrations.
3. We reserve the right to upgrade our survey instruments as necessary as long as they are adequate to measure the type and level of radiation for which they are used.
4. We will abide by requirements listed in the most recent revision of NUREG-1556, Vol. 9 "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use License".



Richard R. Gardner, R.S.O
