

6.8

**July 3, 2013**

**HECTOR BERMUDEZ,**

**LICENSING ASSISTANCE TEAM  
DIVISION OF NUCLEAR MATERIALS SAFETY  
U.S. NUCLEAR REGULATORY COMMISSION, REGION I  
REGION I, 2100 RENAISSANCE BLVD, SUITE 100  
KING OF PRUSSIA, PA 19406-2713**

**SUBJECT: additional information requested renewal of License No.  
52-25400-01, Docket No. 030-34490.**

**Dear Bermudez:**

**1. We confirm address of use is:**

Hospital Comunitario El Buen Samaritano  
P.O. Box 3643  
Aguadilla, Puerto Rico 00605

**2. We confirm no 35.300 inpatients and revise L.C.9.C.**


**3. We confirm no PET**

**4. We add the Dose Calibrator Statement-" Equipment used to  
measu dosages will be calibrated in accordance with  
nationally recognized standards or the manufacturer's  
instructions."**

**5. We have instruments for area and wipe test.**

**6. The limit of 35.300 use is 0.6Ci.**

**Cordially yours.**



**Carmelo Pérez,  
Med Physicist,  
Hospital Comunitario El Buen Samaritano**

579687  
NMSS/RGN1 MATERIALS-002