

6.8

July 3, 2013

HECTOR BERMUDEZ,

**LICENSING ASSISTANCE TEAM
DIVISION OF NUCLEAR MATERIALS SAFETY
U.S. NUCLEAR REGULATORY COMMISSION, REGION I
REGION I, 2100 RENAISSANCE BLVD, SUITE 100
KING OF PRUSSIA, PA 19406-2713**

**SUBJECT: additional information requested renewal of License No.
52-25400-01, Docket No. 030-34490.**

Dear Bermudez:

1. We confirm address of use is:

Hospital Comunitario El Buen Samaritano
P.O. Box 3643
Aguadilla, Puerto Rico 00605

2. We confirm no 35.300 inpatients and revise L.C.9.C.


3. We confirm no PET

**4. We add the Dose Calibrator Statement-" Equipment used to
measu dosages will be calibrated in accordance with
nationally recognized standards or the manufacturer's
instructions."**

5. We have instruments for area and wipe test.

6. The limit of 35.300 use is 0.6Ci.

Cordially yours.



**Carmelo Pérez,
Med Physicist,
Hospital Comunitario El Buen Samaritano**

579687
NMSS/RGN1 MATERIALS-002