



CONVERSATION RECORD

07/24/2013

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU Pete J. Auksel		DATE OF CONTACT 07/24/2013	TYPE OF CONVERSATION <input type="checkbox"/> E-MAIL <input checked="" type="checkbox"/> TELEPHONE <input type="checkbox"/> INCOMING <input checked="" type="checkbox"/> OUTGOING
E-MAIL ADDRESS pauksel@uss.com		TELEPHONE NUMBER (219) 888-7841	

ORGANIZATION Radiation Safety Officer United State Steel-Gary Works	DOCKET NUMBER(S) 030-29318; 030-36022; 030-29080; 030-18249
---	--

LICENSE NUMBER(S) 13-23600-01; 13-26104-03; 13-07964-08; 13-07964-07	CONTROL NUMBER(S) N/A
---	--------------------------

SUBJECT
Additional Information Needed for License Amendment

SUMMARY
As discussed during our phone conversation on July 24, 2013 at 10:00 am, we will not be taking action on your request to combine multiple licenses due to the additional information being needed to complete the request.

1. You submitted a request to combine several licenses, however since the corresponding requests to terminate the licenses that you will no longer need were not submitted.

RESPONSE: Please resubmit the amendment request at the same time the license termination requests are submitted.

2. The NRC requires that for any request to combine licenses request to terminate the license(s) that you will no longer need. This request needs to be submitted at the same time the request to combine the licenses.

RESPONSE: Please submit a NRC Form 314 for each license that you would like to terminate. Include in the submission the most current leak test results for all gauges possessed under that license, even if the gauge is in storage and not being used.

Continue on Page 2

ACTION REQUIRED (IF ANY)
Please re-submit your request to combine your licenses addressing the issues in this conversation record. Be sure to include in your submission the termination request for the licenses that are to be combined with this license, the leak test results for all sources, and a statement that there is no history of leaking sources possessed under the licenses to be terminated.

For the termination of your General Licenses, after the Generally Licensed gauges have been added to your license, please send a letter to the address below notifying the NRC that of this change.

Continue on Page 3

NAME OF PERSON DOCUMENTING CONVERSATION
Jennifer L. Bishop, Materials Licensing Branch

SIGNATURE

CONVERSATION RECORD (continued)

SUMMARY: (Continued from page 1)

3. You indicated in your letter that you plan to add several gauges you currently possess under a General License to a Specific License. Please note that you will need to remove the General License labels from the gauges that are being added to your specific license. You will also want to contact the vendor who distributed the gauges to you to inform them of this change.

RESPONSE: The General Licenses you wish to terminate will need to be addressed through our Headquarters Office. Please send a letter to the address below notifying the NRC of this change.

ATTN: Document Control Desk/GTLS
Director, Office of Federal and State Materials
and Environmental Management Programs
U.S. Nuclear Regulatory Commission
11545 Rockville Pike
Rockville, MD 20852-2738