

From: Miller, John
To: JVaughn@Sikorsky.com
Subject: Application to Renew License No. 06-02269-04G
Date: Wednesday, July 10, 2013 7:18:00 AM

Dear Mr. Vaughn,

In your application dated February 25, 2013 to renew license number 06-02269-04G, you submitted Enclosure 2 which includes the IBIS Technical Manual that will be provided to your customers, the general licensees. Section 5.2.1 "Wipe Test Procedure Using the IBIS Detector" states that wipe test records shall be maintained for a minimum of two years. 10 CFR 31.5(c)(4)(i) requires general licensees to maintain leak test records for 3 years. Please submit a letter to my attention confirming that in your technical guidance provided to general licensees that you will direct them to maintain leak tests records for 3 years.

For expediency, you can send the letter as a pdf file attached to an e-mail or fax it to me. Please assure that the letter refers to the following:

License No. 06- 02269-04G
Docket No. 030-17749
Mail control No. 580142

I appreciate your cooperation look forward to receiving your response.

John J. Miller
Health Physicist
Division of Nuclear Materials Safety
Phone: 610-337-5089
Fax: 610-337-5269
E-mail: john.miller@nrc.gov