



Friday, December 23, 2011

Phil Cavendor  
AUC, LLC  
1536 Cole Blvd  
Suite 330  
Lakewood, CO 80401

RE: Reno Creek Project  
Orde

r No.: S1111043

Dear Phil Cavendor:

Inter-Mountain Laboratories received sample OM4-004-111102 on November 2, 2011.

There were no problems with the analyses and all data for associated QC met EPA or laboratory specifications except where noted in the Case Narrative.

If you have any questions regarding these tests results, please feel free to call.

Sincerely,

Lacey Ketron  
Project Manager

xc: File  
Encl.



Date: 12/23/2011

**CLIENT:** AUC LLC  
**Project:** Reno Creek Project  
**Lab Order:** S1111043

**CASE NARRATIVE**  
**Report ID:** S1111043001

Sample OM4-004-111102 was received on November 2, 2011.

All samples were received and analyzed within the EPA recommended holding times, except those noted in this case narrative. Samples were analyzed using the methods outlined in the following references:

U.S.E.P.A. 600 "Methods for Chemical Analysis of Water and Wastes", 1993  
"Standard Methods For The Examination of Water and Wastewater", 20th ed., 1998  
Test Methods for Evaluating Solid Waste, Physical/Chemical Methods, SW846, 3rd Edition  
Methods indicated with the Monday, March 12, 2007 Federal Register, 40 CFR Part 122, 136 et al.

All Quality Control parameters met the acceptance criteria defined by EPA and Inter-Mountain Laboratories except as indicated in this case narrative.

For the purposes of this report, the Reporting Limit (RL) is the Practical Quantitation Limit (PQL).

Radon 222 was analyzed through ALS Laboratory Group Certification #: CO00078.

Qualifiers by sample

S1111046-001 - General Parameters/Nitrogen, Ammonia (As N) - RPD outside accepted recovery limits  
S1111038-003 - Radiochemistry/Polonium 210 (Suspended) - RPD outside accepted recovery limits  
S1111038-003 - Radiochemistry/Lead 210 (Suspended) - RPD outside accepted recovery limits  
S1111043-001 - General Parameters/Radon-222 - Analyzed by a contract laboratory  
LCS-R77303 - General Parameters/Radon-222 - Analyzed by a contract laboratory  
MB-R77303 - General Parameters/Radon-222 - Analyzed by a contract laboratory

Reviewed by:

Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
1536 Cole Blvd
Lakewood, CO 80401

Date Reported 12/23/2011
Report ID S1111043001

ProjectName: Reno Creek Project
Lab ID: S1111043-001
ClientSample ID: OM4-004-111102
COC: RC 08399

WorkOrder: S1111043
CollectionDate: 11/2/2011 12:30:00 PM
DateReceived: 11/2/2011 4:28:00 PM
FieldSampler: TN
Matrix: Water

Comments

Table with 7 columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init. Rows include pH Reading 1-3, Conductivity Reading 1-3, Turbidity Reading 1-3, Dissolved Oxygen Reading 1-3, Temperature Reading 1-3, Oxygen Reduction Potential Reading, and Water Level Reading 1-3.

These results apply only to the samples tested.

RL - Reporting Limit

- Qualifiers: \* Value exceeds Maximum Contaminant Level
C Calculated Value
H Holding times for preparation or analysis exceeded
L Analyzed by a contract laboratory
ND Not Detected at the Reporting Limit
S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
E Value above quantitation range
J Analyte detected below quantitation limits
M Value exceeds Monthly Ave or MCL
O Outside the Range of Dilutions

Reviewed by: [Signature]
Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
1536 Cole Blvd
Lakewood, CO 80401
ProjectName: Reno Creek Project
Lab ID: S1111043-001
ClientSample ID: OM4-004-111102
COC: RC 08399

Date Reported 12/23/2011
Report ID S1111043001
WorkOrder: S1111043
CollectionDate: 11/2/2011 12:30:00 PM
DateReceived: 11/2/2011 4:28:00 PM
FieldSampler: TN
Matrix: Water

Comments

Table with 7 columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init. Rows include Anions/Cations (Alkalinity, Chloride, etc.), General Parameters (pH, Conductivity), and Data Quality (Cation Sum, Anion Sum, etc.).

These results apply only to the samples tested.

RL - Reporting Limit

- Qualifiers: \* Value exceeds Maximum Contaminant Level, C Calculated Value, H Holding times for preparation or analysis exceeded, L Analyzed by a contract laboratory, ND Not Detected at the Reporting Limit, S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank, E Value above quantitation range, J Analyte detected below quantitation limits, M Value exceeds Monthly Ave or MCL, O Outside the Range of Dilutions

Reviewed by: [Signature]
Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
1536 Cole Blvd
Lakewood, CO 80401
ProjectName: Reno Creek Project
Lab ID: S1111043-001
ClientSample ID: OM4-004-111102
COC: RC 08399

Date Reported 12/23/2011
Report ID S1111043001
WorkOrder: S1111043
CollectionDate: 11/2/2011 12:30:00 PM
DateReceived: 11/2/2011 4:28:00 PM
FieldSampler: TN
Matrix: Water

Comments

Table with 7 columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init. Rows include Metals - Dissolved (Aluminum, Arsenic, Barium, Boron, Cadmium, Chromium, Copper, Iron, Lead, Manganese, Mercury, Molybdenum, Nickel, Selenium, Uranium, Vanadium, Zinc) and Metals - Suspended (Uranium) and Metals - Total (Iron, Manganese).

These results apply only to the samples tested.

RL - Reporting Limit

- Qualifiers: \* Value exceeds Maximum Contaminant Level
C Calculated Value
H Holding times for preparation or analysis exceeded
L Analyzed by a contract laboratory
ND Not Detected at the Reporting Limit
S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
E Value above quantitation range
J Analyte detected below quantitation limits
M Value exceeds Monthly Ave or MCL
O Outside the Range of Dilutions

Reviewed by: [Signature]
Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
1536 Cole Blvd
Lakewood, CO 80401
ProjectName: Reno Creek Project
Lab ID: S1111043-001
ClientSample ID: OM4-004-111102
COC: RC 08399

Date Reported 12/23/2011
Report ID S1111043001
WorkOrder: S1111043
CollectionDate: 11/2/2011 12:30:00 PM
DateReceived: 11/2/2011 4:28:00 PM
FieldSampler: TN
Matrix: Water

Comments

Table with 7 columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init. Rows include Radionuclides - Dissolved (Gross Alpha, Lead 210, Polonium 210, Radium 226, Radium 228, Thorium 230) and Radionuclides - Suspended (Lead 210, Polonium 210, Radium 226, Thorium 230). Total radionuclides section includes Radon 222.

These results apply only to the samples tested.

RL - Reporting Limit

- Qualifiers: \* Value exceeds Maximum Contaminant Level
C Calculated Value
H Holding times for preparation or analysis exceeded
L Analyzed by a contract laboratory
ND Not Detected at the Reporting Limit
S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
E Value above quantitation range
J Analyte detected below quantitation limits
M Value exceeds Monthly Ave or MCL
O Outside the Range of Dilutions

Reviewed by: [Signature]
Lacey Ketron, Water Lab Supervisor



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1111043  
**Project:** Reno Creek Project

**Date:** 12/23/2011  
**Report ID:** S1111043001

**Alkalinity**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 76586	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	11/04/11 9:06	Alkalinity, Total (As CaCO3)	ND	5					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 76586	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	11/04/11 8:57	Alkalinity, Total (As CaCO3)	594	5	601		98.9	90 - 110	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 76586	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111084-002AD	11/04/11 12:00	Alkalinity, Bicarbonate as HCO3	785	5	783	0.282		20	
		Alkalinity, Carbonate as CO3	ND	5	ND			20	
		Alkalinity, Total (As CaCO3)	643	5	641	0.282		20	

**Conductivity**

Sample Type **MBLK** Units: µmhos/cm

Sample ID	RunNo: 76586	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	11/04/11 9:06	Electrical Conductivity	ND	5					

Sample Type **LCS** Units: µmhos/cm

Sample ID	RunNo: 76586	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	11/04/11 8:57	Electrical Conductivity	1080	5	1120		95.9	90 - 110	

Sample Type **DUP** Units: µmhos/cm

Sample ID	RunNo: 76586	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111084-002AD	11/04/11 12:00	Electrical Conductivity	8060	5	8110	0.617		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1111043  
**Project:** Reno Creek Project

**Date:** 12/23/2011  
**Report ID:** S1111043001

**Dissolved Mercury by EPA 245.1 - Water**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 76671	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LRB	11/08/11 8:06	Mercury	ND	0.001					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 76671	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS	11/08/11 8:05	Mercury	0.002	0.001	0.002		99.7	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 76671	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111038-003D	11/08/11 9:34	Mercury	0.002	0.001	0.00244	ND	96.5	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 76671	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1111038-003D	11/08/11 9:36	Mercury	0.002	0.001	0.002	0.123		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 76671	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111038-003D	11/08/11 9:32	Mercury	ND	0.001	ND			20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits



**ANALYTICAL QC SUMMARY REPORT**

**CLIENT:** AUC LLC  
**Work Order:** S1111043  
**Project:** Reno Creek Project

**Date:** 12/23/2011  
**Report ID:** S1111043001

**Dissolved Metals by ICP - EPA 200.7 - Water**Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 76543	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MBLK DISS/CAT	11/03/11 10:25	Aluminum	ND	0.1					
		Boron	ND	0.1					
		Chromium	ND	0.01					
		Iron	ND	0.05					
		Manganese	ND	0.02					
		Nickel	ND	0.01					
		Zinc	ND	0.01					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 76543	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DISS LCS Q	11/03/11 10:27	Aluminum	1.0	0.1	1		98.3	85 - 115	
		Boron	1.0	0.1	1		101	85 - 115	
		Chromium	1.04	0.01	1		104	85 - 115	
		Iron	1.04	0.05	1		104	85 - 115	
		Manganese	1.06	0.02	1		106	85 - 115	
		Nickel	1.06	0.01	1		106	85 - 115	
		Zinc	1.06	0.01	1		106	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 76543	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111044-001DS	11/03/11 19:57	Aluminum	0.5	0.1	0.5	ND	110	70 - 130	
		Boron	0.6	0.1	0.5	ND	108	70 - 130	
		Chromium	0.53	0.01	0.5	ND	106	70 - 130	
		Iron	0.52	0.05	0.5	ND	104	70 - 130	
		Manganese	0.55	0.02	0.5	0.05	102	70 - 130	
		Nickel	0.51	0.01	0.5	ND	101	70 - 130	
		Zinc	0.48	0.01	0.5	ND	96.8	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 76543	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1111044-001DSD	11/03/11 20:00	Aluminum	0.5	0.1	0.5	1.49		20	
		Boron	0.6	0.1	0.6	0.0843		20	
		Chromium	0.54	0.01	0.53	2.19		20	
		Iron	0.54	0.05	0.52	2.74		20	
		Manganese	0.56	0.02	0.55	1.37		20	
		Nickel	0.52	0.01	0.51	3.44		20	
		Zinc	0.52	0.01	0.48	6.34		20	

**Qualifiers:**

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1111043  
**Project:** Reno Creek Project

**Date:** 12/23/2011  
**Report ID:** S1111043001

**Dissolved Metals by ICP - EPA 200.7 - Water**

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 76543	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111043-001DD	11/03/11 19:53	Aluminum	ND	0.1	ND			20	
		Boron	ND	0.1	ND			20	
		Chromium	ND	0.01	ND			20	
		Iron	ND	0.05	ND			20	
		Manganese	0.16	0.02	0.16	0		20	
		Nickel	ND	0.01	ND			20	
		Zinc	ND	0.01	ND			20	

**Dissolved Metals by ICPMS EPA 200.8 - Water**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 76532	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLK	11/03/11 13:54	Arsenic	ND	0.005					
		Barium	ND	0.5					
		Cadmium	ND	0.002					
		Copper	ND	0.01					
		Lead	ND	0.02					
		Molybdenum	ND	0.02					
		Selenium	ND	0.005					
		Uranium	ND	0.001					
		Vanadium	ND	0.02					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 76532	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS	11/03/11 13:50	Arsenic	0.097	0.005	0.1		97.2	85 - 115	
		Barium	ND	0.5	0.1		96.8	85 - 115	
		Cadmium	0.097	0.002	0.1		97.2	85 - 115	
		Copper	0.10	0.01	0.1		97.7	85 - 115	
		Lead	0.10	0.02	0.1		97.7	85 - 115	
		Molybdenum	0.10	0.02	0.1		99.2	85 - 115	
		Selenium	0.099	0.005	0.1		99.1	85 - 115	
		Uranium	0.095	0.001	0.1		95.3	85 - 115	
		Vanadium	0.10	0.02	0.1		98.7	85 - 115	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1111043  
**Project:** Reno Creek Project

**Date:** 12/23/2011  
**Report ID:** S1111043001

**Dissolved Metals by ICPMS EPA 200.8 - Water**

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 76532	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111043-001DS	11/03/11 14:12	Arsenic	1.10	0.005	1.1	0.008	99.0	70 - 130	
		Barium	1.1	0.5	1.1	ND	93.1	70 - 130	
		Cadmium	1.04	0.002	1.1	ND	94.5	70 - 130	
		Copper	1.04	0.01	1.1	ND	94.5	70 - 130	
		Lead	1.06	0.02	1.1	ND	96.0	70 - 130	
		Molybdenum	1.05	0.02	1.1	ND	95.5	70 - 130	
		Selenium	1.07	0.005	1.1	ND	97.1	70 - 130	
		Uranium	1.06	0.001	1.1	ND	96.0	70 - 130	
		Vanadium	1.08	0.02	1.1	ND	98.0	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 76532	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1111043-001DMSD	11/03/11 14:16	Arsenic	1.15	0.005	1.10	4.92		20	
		Barium	1.2	0.5	1.1	13.4		20	
		Cadmium	1.16	0.002	1.04	11.5		20	
		Copper	1.15	0.01	1.04	10.1		20	
		Lead	1.09	0.02	1.06	3.22		20	
		Molybdenum	1.21	0.02	1.05	15.0		20	
		Selenium	1.15	0.005	1.07	7.44		20	
		Uranium	1.17	0.001	1.06	11.0		20	
		Vanadium	1.22	0.02	1.08	13.5		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 76532	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111043-001DD	11/03/11 14:08	Arsenic	0.008	0.005	0.008	7.45		20	
		Barium	ND	0.5	ND			20	
		Cadmium	ND	0.002	ND			20	
		Copper	ND	0.01	ND			20	
		Lead	ND	0.02	ND			20	
		Molybdenum	ND	0.02	ND			20	
		Selenium	ND	0.005	ND			20	
		Uranium	ND	0.001	ND			20	
		Vanadium	ND	0.02	ND			20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1111043  
**Project:** Reno Creek Project

**Date:** 12/23/2011  
**Report ID:** S1111043001

**Fluoride by SM 4500**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 76586	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	11/04/11 9:06	Fluoride	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 76586	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	11/04/11 8:57	Fluoride	3.0	0.1	2.96		101	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 76586	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111084-004ASPK	11/04/11 12:30	Fluoride	2.7	0.1	2.5	0.3	94.2	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 76586	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111084-002AD	11/04/11 12:00	Fluoride	0.2	0.1	0.2	2.16		20	

**Gross Alpha, Beta by SM 7110B**

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 77493	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-334	11/30/11 20:03	Gross Alpha (Dissolved)	ND	2					
		Gross Beta (Dissolved)	ND	3					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 77493	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-334	11/30/11 20:03	Gross Alpha (Dissolved)	82	2	73.6		111	83 - 133	
		Gross Beta (Dissolved)	137	3	126		109	88 - 115	

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 77493	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111043-001FMS	11/30/11 20:03	Gross Alpha (Dissolved)	120	3	147	ND	81.5	48 - 153	
		Gross Beta (Dissolved)	260	4	252	7	100	60 - 140	

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 77493	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111044-001FDUP	12/01/11 17:30	Gross Alpha (Dissolved)	263	2	243	8.23		20	
		Gross Beta (Dissolved)	33	4	30	9.19		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1111043  
**Project:** Reno Creek Project

**Date:** 12/23/2011  
**Report ID:** S1111043001

#### Anions by ION Chromatography

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 76607	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	11/04/11 9:48	Chloride	ND	1					
		Sulfate	ND	1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 76607	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DIONEX	11/04/11 10:00	Chloride	30	1	30		99.7	90 - 110	
		Sulfate	143	1	150		95.2	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 76607	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111040-003ASPK	11/04/11 19:06	Chloride	88	1	80	7	101	80 - 120	
		Sulfate	1480	1	640	735	116	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 76607	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111040-003AD	11/04/11 18:53	Chloride	7	1	7	4.17		20	
		Sulfate	743	1	735	1.09		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits

**ANALYTICAL QC SUMMARY REPORT**

**CLIENT:** AUC LLC  
**Work Order:** S1111043  
**Project:** Reno Creek Project

**Date:** 12/23/2011  
**Report ID:** S1111043001

**Cations by ICP (Method 200.7)**Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 76540	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MBLK DISS/CAT	11/03/11 10:25	Calcium	ND	1					
		Magnesium	ND	1					
		Potassium	ND	1					
		Sodium	ND	1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 76540	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
CAT LCS IML3	11/03/11 10:29	Calcium	40	1	40		99.9	85 - 115	
		Magnesium	40	1	40		99.9	85 - 115	
		Potassium	41	1	40		102	85 - 115	
		Sodium	39	1	40		97.3	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 76540	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111038-003AS	11/03/11 19:26	Calcium	91	1	50	45	93.8	70 - 130	
		Magnesium	56	1	50	5	103	70 - 130	
		Potassium	62	1	50	2	120	70 - 130	
		Sodium	77	1	50	20	112	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 76540	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1111038-003ASD	11/03/11 19:29	Calcium	90	1	91	1.16		20	
		Magnesium	56	1	56	0.366		20	
		Potassium	62	1	62	0.649		20	
		Sodium	76	1	77	0.673		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 76540	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111038-002AD	11/03/11 19:21	Calcium	40	1	39	1.69		20	
		Magnesium	4	1	4	1.48		20	
		Potassium	2	1	2	1.83		20	
		Sodium	18	1	18	0.464		20	

**Qualifiers:**

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1111043  
**Project:** Reno Creek Project

**Date:** 12/23/2011  
**Report ID:** S1111043001

**Nitrogen, Ammonia (as N)**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 76610	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	11/04/11 10:33	Nitrogen, Ammonia (As N)	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 76610	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
QC	11/04/11 10:35	Nitrogen, Ammonia (As N)	10.3	0.1	10		103	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 76610	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111046-001C	11/04/11 11:35	Nitrogen, Ammonia (As N)	10.5	0.05	10	0.30	102	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 76610	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111046-001C	11/04/11 11:34	Nitrogen, Ammonia (As N)	0.23	0.05	0.30	23.3		20	R

**Nitrogen, Nitrate-Nitrite (as N)**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 76651	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	11/07/11 10:22	Nitrogen, Nitrate-Nitrite (as N)	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 76651	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
QC	11/07/11 10:24	Nitrogen, Nitrate-Nitrite (as N)	2.1	0.1	2.31		92.6	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 76651	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111093-002B	11/07/11 11:45	Nitrogen, Nitrate-Nitrite (as N)	9.8	0.1	10	ND	97.5	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 76651	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111093-002B	11/07/11 11:44	Nitrogen, Nitrate-Nitrite (as N)	ND	0.1	ND			20	

**Qualifiers:**

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		

**ANALYTICAL QC SUMMARY REPORT**

**CLIENT:** AUC LLC  
**Work Order:** S1111043  
**Project:** Reno Creek Project

**Date:** 12/23/2011  
**Report ID:** S1111043001

**Lead 210 and Polonium 210**Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 77438	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-318	11/29/11 12:34	Lead 210 (Suspended)	ND	1					

Sample ID	RunNo: 77439	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-318	11/29/11 11:26	Polonium 210 (Suspended)	ND	1					

Sample ID	RunNo: 77557	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-318	11/21/11 14:51	Polonium 210 (Dissolved)	ND	1					

Sample ID	RunNo: 77556	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-318	11/22/11 9:19	Lead 210 (Dissolved)	ND	1					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 77438	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-318	11/29/11 12:34	Lead 210 (Suspended)	14	1	12.8		106	60.3 - 141	

Sample ID	RunNo: 77439	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-318 PO	11/29/11 11:26	Polonium 210 (Suspended)	5	1	3.89		123	48.6 - 156	

Sample ID	RunNo: 77556	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-318	11/22/11 9:19	Lead 210 (Dissolved)	12	1	12.9		95.6	67.6 - 131	

Sample ID	RunNo: 77557	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-318 PO	11/21/11 14:51	Polonium 210 (Dissolved)	3	1	2.11		129	35.1 - 175	

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 77438	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111038-001FMS	11/29/11 12:34	Lead 210 (Suspended)	14	1	12.8	ND	107	62.4 - 132	

Sample ID	RunNo: 77439	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111038-001FMS	11/29/11 11:26	Polonium 210 (Suspended)	15	1	12.8	ND	114	28 - 173	

Sample ID	RunNo: 77556	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1110428-001EMS	11/22/11 9:19	Lead 210 (Dissolved)	17	1	12.9	2	113	66 - 133	

Sample ID	RunNo: 77557	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1110428-001EMS	11/21/11 14:51	Polonium 210 (Dissolved)	5	1	12.9	ND	37.4	20.9 - 160	

**Qualifiers:**

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		





### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1111043  
**Project:** Reno Creek Project

**Date:** 12/23/2011  
**Report ID:** S1111043001

**Lead 210 and Polonium 210**

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 77438	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1111038-003FDUP 11/29/11 12:34 Lead 210 (Suspended) 1 1 ND 20

Sample ID	RunNo: 77439	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1111038-003FDUP 11/29/11 11:26 Polonium 210 (Suspended) 1 1 ND 20

Sample ID	RunNo: 77556	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1110428-003EDUP 11/22/11 9:19 Lead 210 (Dissolved) 6 1 5 11.1 20

Sample ID	RunNo: 77557	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1110428-003EDUP 11/21/11 14:51 Polonium 210 (Dissolved) ND 1 ND 20

**Radium 228**

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 78170	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB11-346 12/21/11 17:46 Radium 228 (Dissolved) ND 1

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 78170	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

LCS11-346 12/21/11 17:46 Radium 228 (Dissolved) 10 1 10 101 74.9 - 122

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 78170	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

S1111038-001FMS 12/21/11 17:46 Radium 228 (Dissolved) 12 1 10 1 108 48.4 - 167

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 78170	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1111038-003FDUP 12/21/11 17:46 Radium 228 (Dissolved) ND 1 ND 20

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1111043  
**Project:** Reno Creek Project

**Date:** 12/23/2011  
**Report ID:** S1111043001

**Radium 226 by SM 7500**

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 76990	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-311	11/11/11 12:22	Radium 226 (Dissolved)	ND	0.2					

Sample ID	RunNo: 77349	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-318	11/17/11 14:09	Radium 226 (Suspended)	ND	0.2					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 76990	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-311	11/11/11 12:22	Radium 226 (Dissolved)	4.4	0.2	5.06		86.5	62.5 - 125	

Sample ID	RunNo: 77349	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-318	11/17/11 16:38	Radium 226 (Suspended)	5.6	0.2	5.06		111	56.4 - 132	

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 76990	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1110429-001EMS	11/11/11 14:58	Radium 226 (Dissolved)	17.9	0.2	5.06	13.9	80.4	41.6 - 146	

Sample ID	RunNo: 77349	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111057-001AMS	11/17/11 14:09	Radium 226 (Suspended)	5.7	0.2	5	ND	115	51.4 - 135	

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 76990	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1110428-007EDUP	11/11/11 14:58	Radium 226 (Dissolved)	79.8	0.2	78.0	2.30		20	

Sample ID	RunNo: 77349	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111137-001EDUP	11/17/11 14:09	Radium 226 (Dissolved)	ND	0.2	ND			20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits

**ANALYTICAL QC SUMMARY REPORT**

**CLIENT:** AUC LLC  
**Work Order:** S1111043  
**Project:** Reno Creek Project

**Date:** 12/23/2011  
**Report ID:** S1111043001

**Thorium 230 in Water**Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 77305	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-312	11/21/11 8:23	Thorium 230 (Dissolved)	ND	0.2					

Sample ID	RunNo: 77377	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-313	11/23/11 8:33	Thorium 230 (Suspended)	ND	0.2					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 77305	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-312	11/21/11 8:23	Thorium 230 (Dissolved)	12	0.2	12		96.4	67.1 - 134	

Sample ID	RunNo: 77377	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-313	11/28/11 9:19	Thorium 230 (Suspended)	11	0.2	12		86.9	69.2 - 127	

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 77305	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111040-001EDUP	11/21/11 8:24	Thorium 230 (Dissolved)	ND	0.2	0.6			20	

Sample ID	RunNo: 77377	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111044-001F DUP	11/28/11 13:31	Thorium 230 (Suspended)	0.3 ± 0.2	0.2	0.3	18.3		20	

**Radon 222 by SM7500-RN**Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 77303	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB-R77303	11/03/11 0:00	Radon-222	ND	50					L

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 77303	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS-R77303	11/03/11 0:00	Radon-222	22500	100	22500		100	75 - 125	L

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 77303	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111043-001A	11/03/11 0:00	Radon-222	70	50	80	15.6		20	L

**Qualifiers:**

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1111043  
**Project:** Reno Creek Project

**Date:** 12/23/2011  
**Report ID:** S1111043001

**Silica as SiO2**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 76544	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MBLK DISS/CAT	11/03/11 10:25	Silica as SiO2	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 76544	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DISS LCS Q	11/03/11 10:27	Silica as SiO2	2.1	0.1	2.14		98.4	85 - 115	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 76544	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111043-001DD	11/03/11 19:53	Silica as SiO2	17.3	0.1	17.5	1.24		20	

**Solids By SM 2540**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 76712	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DI	11/03/11 11:40	Total Dissolved Solids (180)	ND	10					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 76712	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
CONTROL	11/03/11 11:45	Total Dissolved Solids (180)	210	10	226		94.7	90 - 110	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 76712	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111040-002A	11/03/11 12:00	Total Dissolved Solids (180)	870	10	880	1.36		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1111043  
**Project:** Reno Creek Project

**Date:** 12/23/2011  
**Report ID:** S1111043001

**Total (200.2) Metals by EPA 200.7 ICP - Water**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 76641	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB-5605	11/07/11 18:26	Iron	ND	0.05					
		Manganese	ND	0.02					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 76641	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS-5605	11/07/11 18:29	Iron	0.48	0.05	0.5		96.6	85 - 115	
		Manganese	0.20	0.02	0.2		97.6	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 76641	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111044-001ES	11/07/11 18:48	Iron	1.05	0.05	0.5	0.63	85.3	70 - 130	
		Manganese	0.25	0.02	0.2	0.06	94.6	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 76641	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1111044-001EMSD	11/07/11 18:50	Iron	1.05	0.05	1.05	0.495		20	
		Manganese	0.25	0.02	0.25	0.396		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 76641	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111043-001ED	11/07/11 18:43	Iron	0.42	0.05	0.43	2.05		20	
		Manganese	0.16	0.02	0.16	0.126		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1111043  
**Project:** Reno Creek Project

**Date:** 12/23/2011  
**Report ID:** S1111043001

**Suspended Natural Uranium**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 76729	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLK	11/08/11 13:29	Uranium Suspended	ND	0.0003					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 76729	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS	11/08/11 13:25	Uranium Suspended	0.0938	0.0003	0.1		93.8	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 76729	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111038-001FS	11/08/11 14:49	Uranium Suspended	1.08	0.0003	1.1	ND	97.8	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 76729	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1111038-001FMSD	11/08/11 14:53	Uranium Suspended	1.04	0.0003	1.08	3.39		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 76729	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111038-001FD	11/08/11 14:45	Uranium Suspended	ND	0.0003	ND			20	

**pH Water**

Sample Type **LCS** Units: s.u.

Sample ID	RunNo: 76586	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	11/04/11 8:57	pH	8.6	0.1	8.48		101	90 - 110	

Sample Type **DUP** Units: s.u.

Sample ID	RunNo: 76586	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111084-002AD	11/04/11 12:00	pH	7.6	0.1	7.6	0.112		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits





**Inter-Mountain Labs**  
 Sheridan, WY and Gillette, WY  
 1473 Terra Ave  
 Sheridan, WY 82801  
 (307) 672-9045

This is a LEGAL DOCUMENT. All shaded fields must be completed. See reverse for instructions. #RC 08399

- CHAIN OF CUSTODY RECORD -

Client Name: **AUC LLC** Project Identification: **Reno Creek Project** Sampler (Signature/Printed): *Mariel Vickerson* Telephone #: \_\_\_\_\_

Report Address: **pcavendor@auc-llc.com; mickerson@treecorp.com; jettivcr@treecorp.com** Contact Name: *Mariel Vickerson* Email: *mickerson@treecorp.com* Phone: *307-265-0696* Purchase Order #: \_\_\_\_\_

Invoice Address: **AUC LLC** *900 Weaver Ct. Ste 150 Casper WY 82401* SAMPLE IDENTIFICATION: *SM4-004-11102 CT 13* # of Containers: *1*

ITEM	LAB ID (Lab Use Only)	DATE SAMPLED	TIME SAMPLED	IDENTIFICATION	Mark	# of Containers	ANALYSES / PARAMETERS	REMARKS
1							See Attached Table 1 (4.14, Guide 8, Radon 222)	
2	<i>5111043-001</i>	<i>12/11</i>	<i>1230</i>	<i>SM4-004-11102 CT 13</i>		<i>1</i>	See Attached Table 2 (Air Particulate)	<i>Standards</i>
3							See Attached Table 3 (Vegetation)	
4							See Attached Table 4 (Sediment)	
5							See Attached Table 5 (Soil)	
6							See Attached Table 8 WYPDES Discharge	
7							See Attached Table 9 (Blanks)	
8								
9								
10								
11								
12								
13								
14								

LAB COMMENTS: *Relinquished By (Signature/Printed): [Signature] DATE TIME: 11/21/11 16:28*

*2.16°C* *Mariel Vickerson Mariel Vickerson* *11/21/11 15:30* *Rafael Bopp* *11/21/11 9:50*

SHIPPING INFO MATRIX CODES TURNAROUND TIMES COMPLIANCE INFORMATION ADDITIONAL REMARKS

UPS  Fed Express  US Mail  Hand Carried  Other \_\_\_\_\_  
 Water  Soil  Solid  Trip Blank  Other \_\_\_\_\_  
 Matrix Codes:  WT  SL  SD  TB  OT  
 Turnaround Times:  Check desired service  Standard turnaround  RUSH - 5 Working Days  URGENT - < 2 Working Days  
 Compliance Information:  Compliance Monitoring? Program (SDWA, NPDES,...)  PWSID / Permit # \_\_\_\_\_  Chlorinated?  Sample Disposal: Lab  Client