



Thursday, February 23, 2012

Phil Cavendor
AUC, LLC
1536 Cole Blvd
Suite 330
Lakewood, CO 80401

RE: Reno Creek Project
Orde

r No.: S1202019

Dear Phil Cavendor:

Inter-Mountain Laboratories received sample OM7-004-120201 on February 1, 2012.

There were no problems with the analyses and all data for associated QC met EPA or laboratory specifications except where noted in the Case Narrative.

If you have any questions regarding these tests results, please feel free to call.

Sincerely,

Lacey Ketron
Project Manager

xc: File
Encl.



Date: 2/23/2012

CLIENT: AUC LLC
Project: Reno Creek Project
Lab Order: S1202019

CASE NARRATIVE
Report ID: S1202019001

Sample OM7-004-120201 was received on February 1, 2012.

All samples were received and analyzed within the EPA recommended holding times, except those noted in this case narrative. Samples were analyzed using the methods outlined in the following references:

U.S.E.P.A. 600 "Methods for Chemical Analysis of Water and Wastes", 1993
"Standard Methods For The Examination of Water and Wastewater", 20th ed., 1998
Test Methods for Evaluating Solid Waste, Physical/Chemical Methods, SW846, 3rd Edition
Methods indicated with the Monday, March 12, 2007 Federal Register, 40 CFR Part 122, 136 et al.

All Quality Control parameters met the acceptance criteria defined by EPA and Inter-Mountain Laboratories except as indicated in this case narrative.

For the purposes of this report, the Reporting Limit (RL) is the Practical Quantitation Limit (PQL).

Radon 222 was analyzed through ALS Laboratory Group Certification #: CO00078.

Qualifiers by sample

S1201414-001 - Radiochemistry/Uranium Suspended - RPD outside accepted recovery limits
S1202019-001 - General Parameters/Radon-222 - Analyzed by a contract laboratory
LCS-R79683 - General Parameters/Radon-222 - Analyzed by a contract laboratory
MB-R79683 - General Parameters/Radon-222 - Analyzed by a contract laboratory

Reviewed by:

Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
1536 Cole Blvd
Lakewood, CO 80401

Date Reported 2/23/2012
Report ID S1202019001

ProjectName: Reno Creek Project
Lab ID: S1202019-001
ClientSample ID: OM7-004-120201
COC: RC 08183

WorkOrder: S1202019
CollectionDate: 2/1/2012 9:30:00 AM
DateReceived: 2/1/2012 4:08:00 PM
FieldSampler: TN
Matrix: Water

Comments

Table with 7 columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init

Field

Table of field readings including pH, Conductivity, Turbidity, Dissolved Oxygen, Temperature, and Oxygen Reduction Potential.

These results apply only to the samples tested.

RL - Reporting Limit

- Qualifiers: * Value exceeds Maximum Contaminant Level, C Calculated Value, H Holding times for preparation or analysis exceeded, L Analyzed by a contract laboratory, ND Not Detected at the Reporting Limit, S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank, E Value above quantitation range, J Analyte detected below quantitation limits, M Value exceeds Monthly Ave or MCL, O Outside the Range of Dilutions

Reviewed by: [Signature]
Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
 1536 Cole Blvd
 Lakewood, CO 80401

ProjectName: Reno Creek Project
Lab ID: S1202019-001
ClientSample ID: OM7-004-120201
COC: RC 08183

Date Reported: 2/23/2012
Report ID: S1202019001

WorkOrder: S1202019
CollectionDate: 2/1/2012 9:30:00 AM
DateReceived: 2/1/2012 4:08:00 PM
FieldSampler: TN
Matrix: Water

Comments

Analyses	Result	Units	Qual	RL	Method	Date Analyzed/Init	
Anions/Cations							
Alkalinity, Total (As CaCO3)	31	mg/L		5	SM 2320B	2/6/2012 4:17:33 PM	MZ
Alkalinity, Bicarbonate as HCO3	17	mg/L		5	SM 2320B	2/6/2012 4:17:33 PM	MZ
Alkalinity, Carbonate as CO3	10	mg/L		5	SM 2320B	2/6/2012 4:17:33 PM	MZ
Chloride	4	mg/L		1	EPA 300.0	2/3/2012 8:51:00 PM	AMB
Fluoride	0.1	mg/L		0.1	SM 4500FC	2/6/2012 4:17:33 PM	MZ
Nitrogen, Nitrate-Nitrite (as N)	ND	mg/L		0.1	EPA 353.2	2/6/2012 10:40:00 AM	MB
Sulfate	773	mg/L		1	EPA 300.0	2/3/2012 8:51:00 PM	AMB
Calcium	109	mg/L		1	EPA 200.7	2/6/2012 12:37:47 PM	DG
Magnesium	9	mg/L		1	EPA 200.7	2/6/2012 12:37:47 PM	DG
Potassium	35	mg/L		1	EPA 200.7	2/6/2012 12:37:47 PM	DG
Sodium	243	mg/L		1	EPA 200.7	2/6/2012 12:37:47 PM	DG
Nitrogen, Ammonia (As N)	0.8	mg/L		0.1	EPA 350.1	2/3/2012 12:43:00 PM	MB
Silica as SiO2	3	mg/L		1	EPA 200.7	2/6/2012 12:37:47 PM	DG
General Parameters							
pH	9.6	s.u.		0.1	SM 4500 H B	2/6/2012 4:17:33 PM	MZ
Electrical Conductivity	1780	µmhos/cm		5	SM 2510B	2/6/2012 4:17:33 PM	MZ
Total Dissolved Solids (180)	1250	mg/L		10	SM 2540	2/3/2012 2:54:50 PM	ARF
Data Quality							
Cation Sum	17.64	meq/L		0.01	SM 1030E	2/9/2012 8:50:04 AM	LJK
Anion Sum	16.85	meq/L		0.01	SM 1030E	2/9/2012 8:50:04 AM	LJK
Cation-Anion Balance (± 5%)	2.30	%		0.01	SM 1030E	2/9/2012 8:50:04 AM	LJK
Solids, Total Dissolved (Calc)	1200	mg/L		10	SM 1030E	2/9/2012 8:50:04 AM	LJK
Calculated TDS/TDS Ratio (0.80-1.20)	1.04	dec. %		0.01	Calculation	2/9/2012 8:53:46 AM	LJK

These results apply only to the samples tested.

RL - Reporting Limit

- Qualifiers:**
- * Value exceeds Maximum Contaminant Level
 - C Calculated Value
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - ND Not Detected at the Reporting Limit
 - S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
- E Value above quantitation range
- J Analyte detected below quantitation limits
- M Value exceeds Monthly Ave or MCL
- O Outside the Range of Dilutions

Reviewed by: 
 Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
1536 Cole Blvd
Lakewood, CO 80401
ProjectName: Reno Creek Project
Lab ID: S1202019-001
ClientSample ID: OM7-004-120201
COC: RC 08183

Date Reported 2/23/2012
Report ID S1202019001
WorkOrder: S1202019
CollectionDate: 2/1/2012 9:30:00 AM
DateReceived: 2/1/2012 4:08:00 PM
FieldSampler: TN
Matrix: Water

Comments

Table with 7 columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init. Rows include Metals - Dissolved (Aluminum, Arsenic, Barium, Boron, Cadmium, Chromium, Copper, Iron, Lead, Manganese, Mercury, Molybdenum, Nickel, Selenium, Uranium, Vanadium, Zinc) and Metals - Suspended (Uranium) and Metals - Total (Iron, Manganese).

These results apply only to the samples tested.

RL - Reporting Limit

- Qualifiers: * Value exceeds Maximum Contaminant Level
C Calculated Value
H Holding times for preparation or analysis exceeded
L Analyzed by a contract laboratory
ND Not Detected at the Reporting Limit
S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
E Value above quantitation range
J Analyte detected below quantitation limits
M Value exceeds Monthly Ave or MCL
O Outside the Range of Dilutions

Reviewed by: [Signature]
Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
 1536 Cole Blvd
 Lakewood, CO 80401

ProjectName: Reno Creek Project
Lab ID: S1202019-001
ClientSample ID: OM7-004-120201
COC: RC 08183

Date Reported: 2/23/2012
Report ID: S1202019001

WorkOrder: S1202019
CollectionDate: 2/1/2012 9:30:00 AM
DateReceived: 2/1/2012 4:08:00 PM
FieldSampler: TN
Matrix: Water

Comments

Analyses	Result	Units	Qual	RL	Method	Date Analyzed/Init	
Radionuclides - Dissolved							
Gross Alpha	ND	pCi/L		4	SM 7110B	2/16/2012 7:22:00 PM	SH
Gross Alpha Precision (±)	NA	pCi/L			SM 7110B	2/16/2012 7:22:00 PM	SH
Gross Beta	18.4	pCi/L		7	SM 7110B	2/16/2012 7:22:00 PM	SH
Gross Beta Precision (±)	4.4	pCi/L			SM 7110B	2/16/2012 7:22:00 PM	SH
Lead 210	1.3	pCi/L		1	OTW01	2/16/2012 1:28:00 PM	SH
Lead 210 Precision (±)	0.5	pCi/L			OTW01	2/16/2012 1:28:00 PM	SH
Polonium 210	ND	pCi/L		1	OTW01	2/15/2012 10:06:00 AM	SH
Polonium 210 Precision (±)	NA	pCi/L			OTW01	2/15/2012 10:06:00 AM	SH
Radium 226	0.9	pCi/L		0.2	SM 7500-Ra B	2/15/2012 2:03:00 PM	SH
Radium 226 Precision (±)	0.1	pCi/L			SM 7500-Ra B	2/15/2012 2:03:00 PM	SH
Radium 228	1.3	pCi/L		1	Ra-05	2/13/2012 5:42:00 PM	SH
Radium 228 Precision (±)	0.9	pCi/L			Ra-05	2/13/2012 5:42:00 PM	SH
Thorium 230	ND	pCi/L		0.2	ACW10	2/22/2012 9:14:00 AM	WL
Thorium 230 Precision (±)	NA	pCi/L			ACW10	2/22/2012 9:14:00 AM	WL
Radionuclides - Suspended							
Lead 210	ND	pCi/L		1	OTW01	2/16/2012 3:42:00 PM	SH
Lead 210 Precision (±)	NA	pCi/L			OTW01	2/16/2012 3:42:00 PM	SH
Polonium 210	ND	pCi/L		1	OTW01	2/15/2012 11:25:00 AM	SH
Polonium 210 Precision (±)	NA	pCi/L			OTW01	2/15/2012 11:25:00 AM	SH
Radium 226	ND	pCi/L		0.2	SM 7500-Ra B	2/13/2012 3:35:00 PM	SH
Radium 226 Precision (±)	NA	pCi/L			SM 7500-Ra B	2/13/2012 3:35:00 PM	SH
Thorium 230	ND	pCi/L		0.2	ACW10	2/14/2012 8:13:00 AM	WL
Thorium 230 Precision (±)	NA	pCi/L			ACW10	2/14/2012 8:13:00 AM	WL
Radionuclides - Total							
Radon 222	78	pCi/L	L	50	SM7500-RN	2/2/2012	LJK
Radon 222 Precision (±)	22	pCi/L	L		SM7500-RN	2/2/2012	LJK

These results apply only to the samples tested.

RL - Reporting Limit

Qualifiers:

- * Value exceeds Maximum Contaminant Level
- C Calculated Value
- H Holding times for preparation or analysis exceeded
- L Analyzed by a contract laboratory
- ND Not Detected at the Reporting Limit
- S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
- E Value above quantitation range
- J Analyte detected below quantitation limits
- M Value exceeds Monthly Ave or MCL
- O Outside the Range of Dilutions

Reviewed by: 
 Lacey Ketron, Water Lab Supervisor



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1202019
Project: Reno Creek Project

Date: 2/23/2012
Report ID: S1202019001

Alkalinity

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 79579	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	02/06/12 15:52	Alkalinity, Total (As CaCO3)	ND	5					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 79579	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	02/06/12 15:42	Alkalinity, Total (As CaCO3)	590	5	601		98.3	90 - 110	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 79579	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202043-001AD	02/06/12 20:16	Alkalinity, Bicarbonate as HCO3	81	5	84	3.91		20	
		Alkalinity, Carbonate as CO3	10	5	9	6.26		20	
		Alkalinity, Total (As CaCO3)	83	5	84	2.09		20	

Conductivity

Sample Type **MBLK** Units: µmhos/cm

Sample ID	RunNo: 79579	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	02/06/12 15:52	Electrical Conductivity	ND	5					

Sample Type **LCS** Units: µmhos/cm

Sample ID	RunNo: 79579	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	02/06/12 15:42	Electrical Conductivity	1110	5	1120		98.5	90 - 110	

Sample Type **DUP** Units: µmhos/cm

Sample ID	RunNo: 79579	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202043-001AD	02/06/12 20:16	Electrical Conductivity	547	5	548	0.182		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1202019
Project: Reno Creek Project

Date: 2/23/2012
Report ID: S1202019001

Dissolved Mercury by EPA 245.1 - Water

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 79660	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LRB	02/08/12 8:56	Mercury	ND	0.001					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 79660	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS	02/08/12 8:54	Mercury	0.002	0.001	0.002		99.9	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 79660	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1201414-001D	02/08/12 9:07	Mercury	0.002	0.001	0.00244	ND	96.0	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 79660	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1201414-001D	02/08/12 9:09	Mercury	0.002	0.001	0.002	1.53		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 79660	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1201414-001D	02/08/12 9:06	Mercury	ND	0.001	ND			20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1202019
Project: Reno Creek Project

Date: 2/23/2012
Report ID: S1202019001

Dissolved Metals by ICP - EPA 200.7 - Water

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 79576	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MBLK DISS/CAT	02/06/12 12:12	Aluminum	ND	0.1					
		Boron	ND	0.1					
		Chromium	ND	0.01					
		Iron	ND	0.05					
		Manganese	ND	0.02					
		Nickel	ND	0.01					
		Zinc	ND	0.01					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 79576	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DISS LCS Q	02/06/12 12:14	Aluminum	1.0	0.1	1		99.7	85 - 115	
		Boron	1.0	0.1	1		99.9	85 - 115	
		Chromium	0.99	0.01	1		99.2	85 - 115	
		Iron	1.02	0.05	1		102	85 - 115	
		Manganese	1.05	0.02	1		105	85 - 115	
		Nickel	1.02	0.01	1		102	85 - 115	
		Zinc	1.04	0.01	1		104	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 79576	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1202043-001DS	02/06/12 12:53	Aluminum	0.5	0.1	0.5	ND	106	70 - 130	
		Boron	0.6	0.1	0.5	ND	102	70 - 130	
		Chromium	0.50	0.01	0.5	ND	100	70 - 130	
		Iron	0.50	0.05	0.5	ND	100	70 - 130	
		Manganese	0.51	0.02	0.5	ND	101	70 - 130	
		Nickel	0.49	0.01	0.5	ND	98.2	70 - 130	
		Zinc	0.50	0.01	0.5	ND	100	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 79576	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1202043-001DSD	02/06/12 12:56	Aluminum	0.5	0.1	0.5	1.51		20	
		Boron	0.5	0.1	0.6	0.889		20	
		Chromium	0.49	0.01	0.50	1.82		20	
		Iron	0.50	0.05	0.50	1.30		20	
		Manganese	0.50	0.02	0.51	1.21		20	
		Nickel	0.48	0.01	0.49	1.41		20	
		Zinc	0.49	0.01	0.50	1.34		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1202019
Project: Reno Creek Project

Date: 2/23/2012
Report ID: S1202019001

Dissolved Metals by ICP - EPA 200.7 - Water

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 79576	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202042-001DD	02/06/12 12:49	Aluminum	ND	0.1	ND			20	
		Boron	ND	0.1	ND			20	
		Chromium	ND	0.01	ND			20	
		Iron	ND	0.05	ND			20	
		Manganese	ND	0.02	ND			20	
		Nickel	ND	0.01	ND			20	
		Zinc	ND	0.01	ND			20	

Dissolved Metals by ICPMS EPA 200.8 - Water

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 79514	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLK	02/03/12 10:22	Arsenic	ND	0.005					
		Barium	ND	0.5					
		Cadmium	ND	0.002					
		Copper	ND	0.01					
		Lead	ND	0.02					
		Molybdenum	ND	0.02					
		Selenium	ND	0.005					
		Uranium	ND	0.001					
		Vanadium	ND	0.02					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 79514	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS	02/03/12 10:17	Arsenic	0.102	0.005	0.1		102	85 - 115	
		Barium	ND	0.5	0.1		97.1	85 - 115	
		Cadmium	0.098	0.002	0.1		97.9	85 - 115	
		Copper	0.10	0.01	0.1		98.0	85 - 115	
		Lead	0.10	0.02	0.1		103	85 - 115	
		Molybdenum	0.10	0.02	0.1		98.5	85 - 115	
		Selenium	0.103	0.005	0.1		103	85 - 115	
		Uranium	0.102	0.001	0.1		102	85 - 115	
		Vanadium	0.10	0.02	0.1		98.3	85 - 115	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1202019
Project: Reno Creek Project

Date: 2/23/2012
Report ID: S1202019001

Dissolved Metals by ICPMS EPA 200.8 - Water

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 79514	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1202047-005CS	02/03/12 14:32	Arsenic	1.18	0.005	1.1	0.106	97.4	70 - 130	
		Barium	1.1	0.5	1.1	ND	93.2	70 - 130	
		Cadmium	1.08	0.002	1.1	ND	98.6	70 - 130	
		Copper	1.07	0.01	1.1	ND	97.1	70 - 130	
		Lead	1.08	0.02	1.1	ND	98.1	70 - 130	
		Molybdenum	1.09	0.02	1.1	ND	98.0	70 - 130	
		Selenium	1.10	0.005	1.1	ND	100	70 - 130	
		Uranium	1.11	0.001	1.1	ND	100	70 - 130	
		Vanadium	1.08	0.02	1.1	ND	98.1	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 79514	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1202047-005CMSD	02/03/12 14:37	Arsenic	1.21	0.005	1.18	2.82		20	
		Barium	1.0	0.5	1.1	1.30		20	
		Cadmium	1.06	0.002	1.08	2.37		20	
		Copper	1.08	0.01	1.07	0.811		20	
		Lead	1.09	0.02	1.08	0.888		20	
		Molybdenum	1.09	0.02	1.09	0.174		20	
		Selenium	1.08	0.005	1.10	2.01		20	
		Uranium	1.11	0.001	1.11	0.372		20	
		Vanadium	1.11	0.02	1.08	2.48		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 79514	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202047-005CD	02/03/12 14:27	Arsenic	0.104	0.005	0.106	1.73		20	
		Barium	ND	0.5	ND			20	
		Cadmium	ND	0.002	ND			20	
		Copper	ND	0.01	ND			20	
		Lead	ND	0.02	ND			20	
		Molybdenum	ND	0.02	ND			20	
		Selenium	ND	0.005	ND			20	
		Uranium	ND	0.001	ND			20	
		Vanadium	ND	0.02	ND			20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1202019
Project: Reno Creek Project

Date: 2/23/2012
Report ID: S1202019001

Fluoride by SM 4500

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 79579	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	02/06/12 15:52	Fluoride	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 79579	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	02/06/12 15:42	Fluoride	3.1	0.1	2.96		105	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 79579	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1202062-003ASPK	02/07/12 0:09	Fluoride	3.1	0.1	2.5	0.4	107	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 79579	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202043-001AD	02/06/12 20:16	Fluoride	1.5	0.1	1.4	9.06		20	

Gross Alpha, Beta by SM 7110B

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 79890	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB12-045	02/15/12 21:15	Gross Alpha (Dissolved)	ND	2					
		Gross Beta (Dissolved)	ND	3					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 79890	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS12-045	02/15/12 21:15	Gross Alpha (Dissolved)	83	2	73.6		112	83 - 133	
		Gross Beta (Dissolved)	131	3	126		104	88 - 115	

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 79890	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1202020-001FMS	02/15/12 21:15	Gross Alpha (Dissolved)	108	2	73.6	ND	147	48 - 153	
		Gross Beta (Dissolved)	152	3	126	15	109	60 - 140	

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 79890	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202042-001FDUP	02/15/12 21:15	Gross Alpha (Dissolved)	ND	2	ND			20	
		Gross Beta (Dissolved)	6	3	6	6.83		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1202019
Project: Reno Creek Project

Date: 2/23/2012
Report ID: S1202019001

Anions by ION Chromatography

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 79552	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLK	02/03/12 11:44	Chloride	ND	1					
		Sulfate	ND	1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 79552	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DIONEX	02/03/12 11:21	Chloride	30	1	30		101	90 - 110	
		Sulfate	142	1	150		94.5	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 79552	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1202054-002ASPK	02/03/12 22:22	Chloride	105	1	80	26	98.7	80 - 120	
		Sulfate	605	1	640	ND	94.6	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 79552	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202054-002A	02/03/12 22:10	Chloride	26	1	26	2.32		20	
		Sulfate	ND	1	ND			20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1202019
Project: Reno Creek Project

Date: 2/23/2012
Report ID: S1202019001

Cations by ICP (Method 200.7)Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 79575	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MBLK DISS/CAT	02/06/12 12:12	Calcium	ND	1					
		Magnesium	ND	1					
		Potassium	ND	1					
		Sodium	ND	1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 79575	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
CAT LCS IML3	02/06/12 12:16	Calcium	41	1	40		102	85 - 115	
		Magnesium	40	1	40		99.3	85 - 115	
		Potassium	40	1	40		98.7	85 - 115	
		Sodium	40	1	40		99.4	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 79575	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1202043-001AS	02/06/12 12:53	Calcium	58	1	50	6	104	70 - 130	
		Magnesium	52	1	50	1	102	70 - 130	
		Potassium	60	1	50	10	101	70 - 130	
		Sodium	150	1	50	92	116	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 79575	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1202043-001ASD	02/06/12 12:56	Calcium	57	1	58	1.19		20	
		Magnesium	52	1	52	0.938		20	
		Potassium	60	1	60	0.801		20	
		Sodium	148	1	150	0.716		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 79575	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202042-001AD	02/06/12 12:49	Calcium	6	1	6	0.0742		20	
		Magnesium	1	1	1	0.280		20	
		Potassium	10	1	10	0.207		20	
		Sodium	99	1	100	1.37		20	

Qualifiers:

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1202019
Project: Reno Creek Project

Date: 2/23/2012
Report ID: S1202019001

Nitrogen, Ammonia (as N)Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 79509	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	02/03/12 12:35	Nitrogen, Ammonia (As N)	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 79509	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
QC	02/03/12 12:37	Nitrogen, Ammonia (As N)	9.4	0.1	10		94.4	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 79509	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1202039-002C	02/03/12 12:50	Nitrogen, Ammonia (As N)	116	0.1	100	15.8	100	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 79509	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202039-002C	02/03/12 12:49	Nitrogen, Ammonia (As N)	18.2	0.1	15.8	15.4		20	

Nitrogen, Nitrate-Nitrite (as N)Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 79560	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	02/06/12 10:09	Nitrogen, Nitrate-Nitrite (as N)	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 79560	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
QC	02/06/12 10:11	Nitrogen, Nitrate-Nitrite (as N)	2.2	0.1	2.31		95.2	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 79560	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1202028-008B	02/06/12 10:32	Nitrogen, Nitrate-Nitrite (as N)	10.6	0.1	10	1.1	95.7	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 79560	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202028-008B	02/06/12 10:31	Nitrogen, Nitrate-Nitrite (as N)	1.0	0.1	1.1	2.30		20	

Qualifiers:

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1202019
Project: Reno Creek Project

Date: 2/23/2012
Report ID: S1202019001

Lead 210 and Polonium 210

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 79949	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB12-038	02/16/12 13:28	Lead 210 (Dissolved)	ND	1					
----------	----------------	----------------------	----	---	--	--	--	--	--

Sample ID	RunNo: 79949	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

DMB12-038	02/16/12 13:28	Lead 210 (Suspended)	ND	1					
-----------	----------------	----------------------	----	---	--	--	--	--	--

Sample ID	RunNo: 79950	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB12-038	02/15/12 10:06	Polonium 210 (Dissolved)	ND	1					
----------	----------------	--------------------------	----	---	--	--	--	--	--

Sample ID	RunNo: 79950	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

DMB12-038	02/15/12 10:06	Polonium 210 (Suspended)	ND	1					
-----------	----------------	--------------------------	----	---	--	--	--	--	--

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 79949	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

LCS12-038	02/16/12 13:28	Lead 210 (Dissolved)	15	1	12.8		115	67.6 - 131	
		Lead 210 (Suspended)	15	1	12.8		115	60.3 - 141	

Sample ID	RunNo: 79950	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

LCS12-038 PO	02/15/12 10:06	Polonium 210 (Dissolved)	2	1	1.26		148	35.1 - 175	
		Polonium 210 (Suspended)	2	1	1.26		148	48.6 - 156	

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 79949	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

S1202041-001FMS	02/16/12 15:42	Lead 210 (Suspended)	14	1	12.8	2	92.2	62.4 - 132	
-----------------	----------------	----------------------	----	---	------	---	------	------------	--

Sample ID	RunNo: 79950	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

S1202021-001FMS	02/15/12 11:25	Polonium 210 (Suspended)	11	1	11.1	7	39.0	28 - 173	
-----------------	----------------	--------------------------	----	---	------	---	------	----------	--

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 79949	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1202042-001FDUP	02/16/12 13:28	Lead 210 (Dissolved)	ND	1	ND			20	
------------------	----------------	----------------------	----	---	----	--	--	----	--

Sample ID	RunNo: 79950	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1202042-001FDUP	02/15/12 10:06	Polonium 210 (Dissolved)	ND	1	ND			20	
------------------	----------------	--------------------------	----	---	----	--	--	----	--

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1202019
Project: Reno Creek Project

Date: 2/23/2012
Report ID: S1202019001

Radium 228

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 79815	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB12-037	02/13/12 17:42	Radium 228 (Dissolved)	ND	1					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 79815	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS12-037	02/13/12 17:42	Radium 228 (Dissolved)	9	1	9.83		92.5	74.9 - 122	

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 79815	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1202021-001FMS	02/13/12 17:42	Radium 228 (Dissolved)	12	1	9.83	1	107	48.4 - 167	

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 79815	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202041-001FDUP	02/13/12 17:42	Radium 228 (Dissolved)	ND	1	ND			20	

Radium 226 by SM 7500

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 79885	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB12-039	02/13/12 15:35	Radium 226 (Suspended)	ND	0.2					

Sample ID	RunNo: 79887	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB12-037	02/15/12 14:03	Radium 226 (Dissolved)	ND	0.2					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 79885	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS12-039	02/13/12 15:35	Radium 226 (Suspended)	4.9	0.2	5.06		95.9	56.4 - 132	

Sample ID	RunNo: 79887	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS12-037	02/15/12 14:03	Radium 226 (Dissolved)	5.3	0.2	5.06		105	62.5 - 125	

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 79885	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1202020-001FMS	02/13/12 15:35	Radium 226 (Suspended)	4.6	0.2	5.03	ND	92.1	51.4 - 135	

Sample ID	RunNo: 79887	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1202008-002CMS	02/15/12 14:03	Radium 226 (Dissolved)	5.3	0.2	5.06	0.6	91.8	41.6 - 146	

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 79885	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202019-001FDUP	02/13/12 15:35	Radium 226 (Suspended)	ND	0.2	ND			20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1202019
Project: Reno Creek Project

Date: 2/23/2012
Report ID: S1202019001

Radium 226 by SM 7500

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 79887	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202008-001CDUP	02/15/12 14:03	Radium 226 (Dissolved)	601	0.2	585	2.72		20	

Thorium 230 in Water

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 79772	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB12-039	02/14/12 8:13	Thorium 230 (Suspended)	ND	0.2					

Sample ID	RunNo: 80060	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB12-046	02/22/12 9:14	Thorium 230 (Dissolved)	ND	0.2					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 79772	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS12-039	02/10/12 13:25	Thorium 230 (Dissolved)	11	0.2	12		91.5	67.1 - 134	
		Thorium 230 (Suspended)	11	0.2	12		91.5	69.2 - 127	

Sample ID	RunNo: 80060	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS12-046	02/22/12 9:14	Thorium 230 (Dissolved)	12	0.2	12		95.7	67.1 - 134	

Radon 222 by SM7500-RN

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 79683	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB-R79683	02/02/12 0:00	Radon-222	ND	50					L

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 79683	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS-R79683	02/02/12 0:00	Radon-222	23200	100	22500		103	75 - 125	L

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 79683	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202019-001A	02/02/12 0:00	Radon-222	70	50	80	6.41		20	L

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1202019
Project: Reno Creek Project

Date: 2/23/2012
Report ID: S1202019001

Silica as SiO2

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 79578	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MBLK DISS/CAT	02/06/12 12:12	Silica as SiO2	ND	1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 79578	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DISS LCS Q	02/06/12 12:14	Silica as SiO2	2	1	2.14		98.9	85 - 115	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 79578	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202042-001DD	02/06/12 12:49	Silica as SiO2	4	1	4	0.141		20	

Solids By SM 2540

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 79638	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DI	02/03/12 14:42	Total Dissolved Solids (180)	ND	10					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 79638	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
CONTROL	02/03/12 14:43	Total Dissolved Solids (180)	220	10	226		95.6	90 - 110	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 79638	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202064-009A	02/03/12 14:53	Total Dissolved Solids (180)	280	10	290	2.76		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1202019
Project: Reno Creek Project

Date: 2/23/2012
Report ID: S1202019001

Total (200.2) Metals by EPA 200.7 ICP - Water

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 79626	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB-5826	02/07/12 16:45	Iron	ND	0.05					
		Manganese	ND	0.02					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 79626	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS-5826	02/07/12 16:47	Iron	0.50	0.05	0.5		99.7	85 - 115	
		Manganese	0.20	0.02	0.2		97.8	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 79626	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1202020-001ES	02/07/12 16:57	Iron	0.51	0.05	0.5	ND	98.1	70 - 130	
		Manganese	0.19	0.02	0.2	ND	96.8	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 79626	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1202020-001EMSD	02/07/12 16:59	Iron	0.48	0.05	0.51	5.77		20	
		Manganese	0.18	0.02	0.19	5.57		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 79626	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202019-001ED	02/07/12 16:52	Iron	ND	0.05	ND			20	
		Manganese	ND	0.02	ND			20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1202019
Project: Reno Creek Project

Date: 2/23/2012
Report ID: S1202019001

Suspended Natural Uranium

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 79645	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLK	02/07/12 12:36	Uranium Suspended	ND	0.0003					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 79645	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS	02/07/12 12:32	Uranium Suspended	0.0979	0.0003	0.1		97.9	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 79645	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1201414-001FS	02/07/12 13:45	Uranium Suspended	1.11	0.0003	1.1	0.0015	101	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 79645	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1201414-001FMSD	02/07/12 13:49	Uranium Suspended	1.09	0.0003	1.11	2.02		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 79645	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1201414-001FD	02/07/12 13:40	Uranium Suspended	0.0008	0.0003	0.0015	47.5		20	R

pH Water

Sample Type **LCS** Units: s.u.

Sample ID	RunNo: 79579	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	02/06/12 15:42	pH	8.7	0.1	8.48		103	90 - 110	

Sample Type **DUP** Units: s.u.

Sample ID	RunNo: 79579	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1202043-001AD	02/06/12 20:16	pH	9.2	0.1	9.1	0.505		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



Inter-Mountain Labs
 Sheridan, WY and Gillette, WY
 1673 Terra Ave.
 Sheridan, WY 82801
 (307) 672-8845

- CHAIN OF CUSTODY RECORD -

This is a LEGAL DOCUMENT. All shaded fields must be completed. See reverse for instructions. #RC 08183

Client Name		Project Identification		Sampler (Signature/Printed)		Telephone #	
AUC LLC		Reno Creek Project		Tracey Nickerson			
Report Address		Contact Name		ANALYSES / PARAMETERS		REMARKS	
peavendor@auc-llc.com; trickerson@treecorp.com; johniver@treecorp.com		Tracey Nickerson		See Attached Table 1 (4, 14, Guide 8, Radon 222)		Groundwater	
Invoice Address AUC LLC		Purchase Order #		See Attached Table 2 (Air Particulate)			
909 Waenae Ct. Ste 150		307-265-0696		See Attached Table 3 (Vegetation)			
Casper WY 82601				See Attached Table 4 (Sediment)			
		SAMPLE IDENTIFICATION		See Attached Table 5 (Soil)			
ITEM	LAB ID (Lab Use Only)	DATE SAMPLED	TIME SAMPLED	Matrix	# of Containers	See Attached Table 6 (WYPDES Discharge)	See Attached Table 7 (Blanks)
1							
2	251202019-01	4/12	9:30	WT 13			
3							
4							
5							
6							
7							
8							
9							
10							
11							
12							
13							
14							

LAB COMMENTS	RELINQUISHED BY (Signature/Printed)	DATE	TIME	RECEIVED BY (Signature/Printed)	DATE	TIME
3.1	Tracey Nickerson	4/12	1530	Tracey Nickerson	4/12	1608
				Kathy Bop	4/12	9:49

SHIPPING INFO	MATRIX CODES	TURNAROUND TIMES	COMPLIANCE INFORMATION	ADDITIONAL REMARKS
<input type="checkbox"/> UPS <input type="checkbox"/> Fed Express <input type="checkbox"/> US Mail <input type="checkbox"/> Hand Carried <input checked="" type="checkbox"/> Other	water <input checked="" type="checkbox"/> WT Soil <input type="checkbox"/> SL Solid <input type="checkbox"/> SD Trip Blank <input type="checkbox"/> TB Other <input type="checkbox"/> OT	<input checked="" type="checkbox"/> Check desired service <input type="checkbox"/> Standard turnaround <input type="checkbox"/> RUSH - 5 Working Days <input type="checkbox"/> URGENT - < 2 Working Days Rush & Urgent Surcharges will be applied	Compliance Monitoring? <input checked="" type="checkbox"/> Y / <input type="checkbox"/> N Program (SDWA, NPDES,...) PWSID / Permit # Chlorinated? <input type="checkbox"/> Y / <input checked="" type="checkbox"/> N Sample Disposal: Lab <input checked="" type="checkbox"/> Client	