

U.S. NRC Region III
2443 Warrenville Road
Suite 210
Lisle, Illinois 60532-4352

Attn: Mr. Frank Tran

License # 12-13568-02 Diagnostic Health Services

Please accept this letter, as our response to your e-mail dated July 8, 2013.

1. Isotopes stored at the facilities where Tc-99m, Tl-201, I-123 allowed under 10 CFR 35.100/200 of our Materials License.
2. All Licensed materials have been removed from these annex facilities prior to the closeout surveys. Final surveys are documented with no detectable activity found above natural background.
3. Resubmitting closeout surveys with wipe surveys noted in DPM. These areas remain in-active and are secured.

Should you have any questions, or need additional information, please contact me at (618)530-0290.

Sincerely,

WMS Gooch 7/9/13

William S. Gooch
Radiation Safety Officer
Diagnostic Health Services
9785 Mackenzie Rd. Suite 101
St Louis, Mo 63123
www.dhsinc.com

DHS Storage Facility
 Outpatient Specialty
 Physician Clinic
 700 S. Main Street
 Clinton, Indiana

CLOSE OUT SURVEYS

Location: Union - Clinton Annex Building
 700 South Main St
 825 South Main St Clinton, IN 47842

DHS Storage Facility
 Annex Building
 825 South Main Street
 Clinton, Indiana

Date: 4-26-13

	Survey Meter	Leak/Wipe Instrument
Make	Ludlum	Ludlum
Model	14C	3-6
Serial #	214857	68644
Cal. Date	11-26-12	7-31-12
Check Source	0.60x1 mR/hr	8K cpm
Background	0.02	250 cpm dpm

	*Area	Survey (mR/hr)	Leak/Wipe (dpm/100 sq. cm.)
	700 S MAIN ANNEX		
A	Prod. Syringe Decay	0.02	250 dpm
B	Prod. Waste Storage	0.02	200 dpm
C	Work Area	0.02	250 dpm
F	Ammo Box	0.02	250 dpm
E	Lab Degr	0.02	250 dpm
	825 SOUTH MAIN ANNEX	0.02	250 dpm
A	RAD PAPER Receipt Area	0.02	250 dpm
B	RAD mat waste storage	0.02	250 dpm
C	Work Area	0.02	250 dpm
S	Secured door	0.02	250 dpm

All 10 Areas surveyed at
 Background ~~dpm~~

* Please refer to the attached sketch for the location of the areas surveyed
 and wipe tested

Performed By: Rodney Hille
 Rodney Hille
 RSH

Date: 4-26-13

CLOSE OUT SURVEYS

Location:

St. Mary's Warrick Hospital

Date: 4-30-13

DHS Storage Facility
Maintenance Building
1118 Millis Avenue
Boonville, Indiana

	Survey Meter	Leak/Wipe Instrument
Make	Ludlum	Ludlum
Model	14C	14C
Serial #	132123	# 132123
Cal. Date	12-Feb-2013	12-Feb-2013
Check Source	3.5 mR/hr	3.5 mR/hr
Background	0.02	200cpm 0.02 \emptyset dpm

*Area	Survey (mr/hr)	Leak/Wipe (dpm/100 sq. cm.)
A. Radioactive Material Storage	0.02	200cpm \emptyset dpm
B. Radioactive Waste Storage	0.02	200cpm \emptyset dpm
C. Door to Shop	0.02	200cpm \emptyset dpm
D. Outside Wall	0.02	200cpm \emptyset dpm
E. Outside Wall	0.02	200cpm \emptyset dpm

All 5 Area's Surveyed At
Background \emptyset dpm

* Please refer to the attached sketch for the location of the areas surveyed and wipe tested

Performed By: Laura Bennett
✓ B. David RSO

Date: 4-30-13

CLOSE OUT SURVEYS

Location: Green County General Hospital
 Annex building
 390 N. E. "A" Street Linton, IN 47441

Date: 4-26-13

	Survey Meter	Leak/Wipe Instrument
Make	Ludlum	Ludlum
Model	14C	16
Serial #	144722	57568
Cal. Date	11-28-2012	8-14-2012
Check Source	1.5mB/hr	135 cpm
Background	0.02	900 cpm

*Area	Survey (mr/hr)	Leak/Wipe (dpm/100 sq. cm.)
A. Radioactive material Storage	0.02	800 cpm < 0 dpm
B. Rad. Waste Storage	0.02	900 cpm 0 dpm
C. Exterior surface of doors	0.02	900 cpm 0 dpm
C. Exterior Surface of walls	0.02	900 cpm 0 dpm
D. Surface of cabinets	0.02	800 cpm 0 dpm
E. Radioactive Package Receipt Area	0.02	900 cpm 0 dpm
F. Rad Package Return Area	0.02	900 cpm 0 dpm
		All 7 Area's
		Are surveyed At
		Background 0 dpm

* Please refer to the attached sketch for the location of the areas surveyed and wipe tested

Performed By: Rodney Hille

Date: 4-26-13

Bill Goch 250
Rodney Hille

DHS Storage Facility
 Sullivan County Specialty Clinic
 Suite B
 2204 Mary Sherman Drive
 Sullivan, Indiana

CLOSE OUT SURVEYS

Location:
 Sullivan County Specialty Clinic
 Suite B
 2204 Mary Sherman Dr.
 Sullivan, IN 47882

Date: 4-26-13

	Survey Meter	Leak/Wipe Instrument
Make	Ludlum	Ludlum
Model	14C	3
Serial #	156770	164341
Cal. Date	2-22-2013	1-29-2013
Check Source	8.5 mc/hr	0.7Kcpm x 10
Background	0.02 mc/hr	6.5K dpm

*Area	Survey (mr/hr)	Leak/Wipe (dpm/100 sq. cm.)
D1 Radioactive Decay	0.02	0 dpm
D2 Syringe Waste	0.02	0 dpm
E Work Area	0.02	0 dpm
F Receipt Area	0.02	0 dpm
G Return Area	0.02	0 dpm
H Lab Door	0.02	0 dpm
I Floor Source	0.02	0 dpm

All Areas Surveyed At Background & dpm.

* Please refer to the attached sketch for the location of the areas surveyed and wipe tested

Performed By: Alice Stoly CNMT
 Bill Dovich RSO.

Date: 4-26-13

Tran, Frank

From: Gooch, Bill <goochb@dhsinc.com>
Sent: Tuesday, July 09, 2013 9:48 AM
To: Tran, Frank
Cc: Gooch, Bill
Subject: Mail Control No. 580650 (12-13568-02)
Attachments: annexcloseoutnrc.pdf

Dear Mr. Tran:
Please find attached the information you requested for our amendment.

Thank You,

Bill Gooch

Your message is ready to be sent with the following file or link attachments:

annexcloseoutnrc.pdf

Note: To protect against computer viruses, e-mail programs may prevent sending or receiving certain types of file attachments. Check your e-mail security settings to determine how attachments are handled.