



Thursday, July 07, 2011

Phil Cavendor
AUC, LLC
1536 Cole Blvd
Suite 330
Lakewood, CO 80401

RE: Reno Creek Project
Orde

r No.: S1103485

Dear Phil Cavendor:

Inter-Mountain Laboratories received samples PZM8-002-110329 on March 31, 2011.

There were no problems with the analyses and all data for associated QC met EPA or laboratory specifications except where noted in the Case Narrative.

If you have any questions regarding these tests results, please feel free to call.

Sincerely,

Lacey Ketron
Project Manager

xc: File
Encl.



Date: 7/7/2011

CLIENT: AUC LLC
Project: Reno Creek Project
Lab Order: S1103485

CASE NARRATIVE
Report ID: S1103485002
(Replaces S1103485001)

Sample PZM8-002-110329 was received on March 29, 2011.

All samples were received and analyzed within the EPA recommended holding times, except those noted in this case narrative. Samples were analyzed using the methods outlined in the following references:

- U.S.E.P.A. 600 "Methods for Chemical Analysis of Water and Wastes", 1993
- "Standard Methods For The Examination of Water and Wastewater", 20th ed., 1998
- Test Methods for Evaluating Solid Waste, Physical/Chemical Methods, SW846, 3rd Edition
- Methods indicated with the Monday, March 12, 2007 Federal Register, 40 CFR Part 122, 136 et al.

All Quality Control parameters met the acceptance criteria defined by EPA and Inter-Mountain Laboratories except as indicated in this case narrative.

For the purposes of this report, the Reporting Limit (RL) is the Practical Quantitation Limit (PQL).

Radon 222 was analyzed through ALS Laboratory Group Certification #: CO00078.

Qualifiers by sample

S1103485-001 - General Parameters/Radon-222 - Analyzed by a contract laboratory

Qualifiers by sample

- S1103482-008 - Cations/Calcium - Spike Recovery outside accepted recovery limits (1)
- S1103474-001 - Radiochemistry/Radium 226 (Dissolved) - RPD outside accepted recovery limits
- MB-R69021 - General Parameters/Radon-222 - Analyzed by a contract laboratory
- LCS-R69021 - General Parameters/Radon-222 - Analyzed by a contract laboratory
- S1103485-001 - General Parameters/Radon-222 - Analyzed by a contract laboratory
- S1103455-001 - General Parameters/Radon-222 - RPD outside accepted recovery limits
- S1103485-001 - General Parameters/Radon-222 - RPD outside accepted recovery limits

(1) Spike recovery outside acceptance criteria due to presence of analyte in the sample that greatly exceeded level of spike used. Data approved because all other QC were within acceptance limits.

Reviewed by:

Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
1536 Cole Blvd
Lakewood, CO 80401

Date Reported 7/7/2011
Report ID S1103485002
(Replaces S1103485001)

ProjectName: Reno Creek Project
Lab ID: S1103485-001
ClientSample ID: PZM8-002-110329
COC: RC 08265

WorkOrder: S1103485
CollectionDate: 3/29/2011 2:30:00 PM
DateReceived: 3/31/2011 10:17:00 AM
FieldSampler: TN
Matrix: Water

Table with 7 columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init. Rows include pH Reading, Conductivity Reading, Turbidity Reading, Dissolved Oxygen Reading, Temperature Reading, Oxygen Reduction Potential Reading, Purge Rate Reading, and Water Level Reading.

These results apply only to the samples tested.

RL - Reporting Limit

- Qualifiers: * Value exceeds Maximum Contaminant Level
C Calculated Value
H Holding times for preparation or analysis exceeded
L Analyzed by a contract laboratory
ND Not Detected at the Reporting Limit
S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
E Value above quantitation range
J Analyte detected below quantitation limits
M Value exceeds Monthly Ave or MCL
O Outside the Range of Dilutions

Reviewed by: [Signature]
Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
1536 Cole Blvd
Lakewood, CO 80401

ProjectName: Reno Creek Project
Lab ID: S1103485-001
ClientSample ID: PZM8-002-110329
COC: RC 08265

Date Reported: 7/7/2011
Report ID: S1103485002
(Replaces S1103485001)

WorkOrder: S1103485
CollectionDate: 3/29/2011 2:30:00 PM
DateReceived: 3/31/2011 10:17:00 AM
FieldSampler: TN
Matrix: Water

Analyses	Result	Units	Qual	RL	Method	Date Analyzed/Init	
Anions/Cations							
Alkalinity, Total (As CaCO3)	78	mg/L		5	SM 2320B	3/31/2011 7:06:32 PM	AMB
Alkalinity, Bicarbonate as HCO3	96	mg/L		5	SM 2320B	3/31/2011 7:06:32 PM	AMB
Alkalinity, Carbonate as CO3	ND	mg/L		5	SM 2320B	3/31/2011 7:06:32 PM	AMB
Chloride	7	mg/L		1	EPA 300.0	4/1/2011 6:22:00 PM	KO
Fluoride	ND	mg/L		0.1	SM 4500FC	3/31/2011 7:06:32 PM	AMB
Nitrogen, Nitrate-Nitrite (as N)	ND	mg/L		0.1	EPA 353.2	4/11/2011 4:24:00 PM	AS
Sulfate	846	mg/L		1	EPA 300.0	4/1/2011 6:22:00 PM	KO
Calcium	110	mg/L		1	EPA 200.7	4/5/2011 8:32:37 PM	DG
Magnesium	24	mg/L		1	EPA 200.7	3/31/2011 6:21:28 PM	MS
Potassium	9	mg/L		1	EPA 200.7	3/31/2011 6:21:28 PM	MS
Sodium	320	mg/L		1	EPA 200.7	3/31/2011 6:21:28 PM	MS
Nitrogen, Ammonia (As N)	ND	mg/L		0.1	EPA 350.1	4/1/2011 11:43:00 AM	AS
Silica as SiO2	11	mg/L		1	EPA 200.7	3/31/2011 6:21:28 PM	MS
General Parameters							
pH	8.3	s.u.		0.1	SM 4500 H B	3/31/2011 7:06:32 PM	AMB
Electrical Conductivity	1700	µmhos/cm		5	SM 2510B	3/31/2011 7:06:32 PM	AMB
Total Dissolved Solids (180)	1520	mg/L		10	SM 2540	4/1/2011 2:40:00 PM	JF
Data Quality							
Cation Sum	21.60	meq/L		0.00	SM 1030E	4/8/2011 4:34:37 PM	LJK
Anion Sum	19.55	meq/L		0.00	SM 1030E	4/8/2011 4:34:37 PM	LJK
Cation-Anion Balance (± 5%)	4.97	%		0.00	SM 1030E	4/8/2011 4:34:37 PM	LJK
Solids, Total Dissolved (Calc)	1360	mg/L		10	SM 1030E	4/8/2011 4:34:37 PM	LJK
Calculated TDS/TDS Ratio (0.80-1.20)	1.12	dec. %		0.01	Calculation	4/8/2011 4:35:55 PM	LJK

These results apply only to the samples tested.

RL - Reporting Limit

Qualifiers:

- * Value exceeds Maximum Contaminant Level
- C Calculated Value
- H Holding times for preparation or analysis exceeded
- L Analyzed by a contract laboratory
- ND Not Detected at the Reporting Limit
- S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
- E Value above quantitation range
- J Analyte detected below quantitation limits
- M Value exceeds Monthly Ave or MCL
- O Outside the Range of Dilutions

Reviewed by: 
Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
 1536 Cole Blvd
 Lakewood, CO 80401

ProjectName: Reno Creek Project
Lab ID: S1103485-001
ClientSample ID: PZM8-002-110329
COC: RC 08265

Date Reported: 7/7/2011
Report ID: S1103485002
 (Replaces S1103485001)

WorkOrder: S1103485
CollectionDate: 3/29/2011 2:30:00 PM
DateReceived: 3/31/2011 10:17:00 AM
FieldSampler: TN
Matrix: Water

Analyses	Result	Units	Qual	RL	Method	Date Analyzed/Init	
Metals - Dissolved							
Aluminum	ND	mg/L		0.1	EPA 200.7	3/31/2011 6:21:28 PM	MS
Arsenic	0.003	mg/L		0.001	EPA 200.8	4/1/2011 10:54:59 AM	MS
Barium	ND	mg/L		0.1	EPA 200.8	4/1/2011 10:54:59 AM	MS
Boron	ND	mg/L		0.1	EPA 200.7	3/31/2011 6:21:28 PM	MS
Cadmium	ND	mg/L		0.001	EPA 200.8	4/1/2011 10:54:59 AM	MS
Chromium	ND	mg/L		0.01	EPA 200.7	3/31/2011 6:21:28 PM	MS
Copper	ND	mg/L		0.01	EPA 200.8	4/1/2011 10:54:59 AM	MS
Iron	ND	mg/L		0.05	EPA 200.7	3/31/2011 6:21:28 PM	MS
Lead	ND	mg/L		0.01	EPA 200.8	4/1/2011 10:54:59 AM	MS
Manganese	0.09	mg/L		0.01	EPA 200.7	3/31/2011 6:21:28 PM	MS
Mercury	ND	mg/L		0.001	EPA 245.1	4/5/2011 12:26:26 PM	BK
Molybdenum	0.02	mg/L		0.01	EPA 200.8	4/1/2011 10:54:59 AM	MS
Nickel	ND	mg/L		0.05	EPA 200.7	3/31/2011 6:21:28 PM	MS
Selenium	ND	mg/L		0.005	EPA 200.8	4/1/2011 10:54:59 AM	MS
Uranium	0.0194	mg/L		0.0003	EPA 200.8	4/1/2011 10:54:59 AM	MS
Vanadium	ND	mg/L		0.1	EPA 200.8	4/1/2011 10:54:59 AM	MS
Zinc	ND	mg/L		0.01	EPA 200.7	3/31/2011 6:21:28 PM	MS
Metals - Suspended							
Uranium	ND	mg/L		0.0003	EPA 200.8	4/5/2011 10:42:59 AM	MS
Metals - Total							
Iron	0.23	mg/L		0.05	EPA 200.7	4/4/2011 5:45:36 PM	MS
Manganese	0.10	mg/L		0.01	EPA 200.7	4/4/2011 5:45:36 PM	MS

These results apply only to the samples tested.

RL - Reporting Limit

Qualifiers:

- * Value exceeds Maximum Contaminant Level
- C Calculated Value
- H Holding times for preparation or analysis exceeded
- L Analyzed by a contract laboratory
- ND Not Detected at the Reporting Limit
- S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
- E Value above quantitation range
- J Analyte detected below quantitation limits
- M Value exceeds Monthly Ave or MCL
- O Outside the Range of Dilutions

Reviewed by: 
 Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
 1536 Cole Blvd
 Lakewood, CO 80401

ProjectName: Reno Creek Project
Lab ID: S1103485-001
ClientSample ID: PZM8-002-110329
COC: RC 08265

Date Reported: 7/7/2011
Report ID: S1103485002
 (Replaces S1103485001)

WorkOrder: S1103485
CollectionDate: 3/29/2011 2:30:00 PM
DateReceived: 3/31/2011 10:17:00 AM
FieldSampler: TN
Matrix: Water

Analyses	Result	Units	Qual	RL	Method	Date Analyzed/Init	
Radionuclides - Dissolved							
Gross Alpha	133	pCi/L		4	SM 7110B	4/17/2011 9:40:04 PM	SH
Gross Alpha Precision (±)	8.3	pCi/L			SM 7110B	4/17/2011 9:40:04 PM	SH
Gross Beta	60.6	pCi/L		7	SM 7110B	4/17/2011 9:40:04 PM	SH
Gross Beta Precision (±)	5.3	pCi/L			SM 7110B	4/17/2011 9:40:04 PM	SH
Lead 210	2.7	pCi/L		1	OTW01	4/20/2011 12:38:11 PM	SH
Lead 210 Precision (±)	0.4	pCi/L			OTW01	4/20/2011 12:38:11 PM	SH
Polonium 210	ND	pCi/L		1	OTW01	4/21/2011 12:25:00 PM	SH
Polonium 210 Precision (±)	NA	pCi/L			OTW01	4/21/2011 12:25:00 PM	SH
Radium 226	46.9	pCi/L		0.2	SM 7500-Ra B	4/20/2011 6:59:28 PM	SH
Radium 226 Precision (±)	0.7	pCi/L			SM 7500-Ra B	4/20/2011 6:59:28 PM	SH
Radium 228	1.0	pCi/L		1	Ra-05	4/18/2011 4:25:00 PM	SH
Radium 228 Precision (±)	0.8	pCi/L			Ra-05	4/18/2011 4:25:00 PM	SH
Thorium 230	ND	pCi/L		0.2	ACW10	4/11/2011 5:53:00 PM	SH
Thorium 230 Precision (±)	NA	pCi/L			ACW10	4/11/2011 5:53:00 PM	SH
Radionuclides - Suspended							
Lead 210	17.7	pCi/L		1	OTW01	4/25/2011 10:14:35 AM	SH
Lead 210 Precision (±)	1.0	pCi/L			OTW01	4/25/2011 10:14:35 AM	SH
Polonium 210	ND	pCi/L		1	OTW01	4/28/2011 8:30:00 AM	SH
Polonium 210 Precision (±)	NA	pCi/L			OTW01	4/28/2011 8:30:00 AM	SH
Radium 226	0.7	pCi/L		0.2	SM 7500-Ra B	4/25/2011 8:23:34 AM	SH
Radium 226 Precision (±)	0.1	pCi/L			SM 7500-Ra B	4/25/2011 8:23:34 AM	SH
Thorium 230	0.2	pCi/L		0.2	ACW10	4/18/2011 5:46:00 PM	SH
Thorium 230 Precision (±)	0.1	pCi/L			ACW10	4/18/2011 5:46:00 PM	SH
Radionuclides - Total							
Radon 222	40900	pCi/L	L	50	SM7500-RN	3/31/2011	LJK
Radon 222 Precision (±)	4700	pCi/L	L		SM7500-RN	3/31/2011	LJK

These results apply only to the samples tested.

RL - Reporting Limit

Qualifiers:

- * Value exceeds Maximum Contaminant Level
- C Calculated Value
- H Holding times for preparation or analysis exceeded
- L Analyzed by a contract laboratory
- ND Not Detected at the Reporting Limit
- S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
- E Value above quantitation range
- J Analyte detected below quantitation limits
- M Value exceeds Monthly Ave or MCL
- O Outside the Range of Dilutions

Reviewed by: 
 Lacey Ketron, Water Lab Supervisor

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
 (Replaces S1103485001)

AlkalinitySample Type **MBLK** Units: mg/L

Sample ID	RunNo: 68778	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	03/31/11 14:45	Alkalinity, Total (As CaCO3)	ND	5					

Sample ID	RunNo: 68913	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	04/05/11 14:57	Alkalinity, Total (As CaCO3)	ND	5					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 68778	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	03/31/11 14:34	Alkalinity, Total (As CaCO3)	474	5	473		100	96.6 - 103	

Sample ID	RunNo: 68913	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	04/05/11 14:48	Alkalinity, Total (As CaCO3)	474	5	473		100	96.6 - 103	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 68778	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103445-009A	03/31/11 15:01	Alkalinity, Bicarbonate as HCO3	ND	5	ND	0		20	
		Alkalinity, Carbonate as CO3	ND	5	ND	0		20	
		Alkalinity, Total (As CaCO3)	ND	5	ND	0		20	

Sample ID	RunNo: 68778	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103482-008A	03/31/11 17:58	Alkalinity, Bicarbonate as HCO3	541	5	539	0.320		20	
		Alkalinity, Carbonate as CO3	ND	5	ND	0		20	
		Alkalinity, Total (As CaCO3)	443	5	442	0.320		20	

Sample ID	RunNo: 68778	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103492-006A	03/31/11 22:03	Alkalinity, Bicarbonate as HCO3	78	5	77	2.03		20	
		Alkalinity, Carbonate as CO3	55	5	57	2.93		20	
		Alkalinity, Total (As CaCO3)	156	5	158	0.956		20	

Sample ID	RunNo: 68913	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1104036-003A	04/05/11 17:01	Alkalinity, Bicarbonate as HCO3	ND	5	ND	0		20	
		Alkalinity, Total (As CaCO3)	ND	5	ND	0		20	

Sample ID	RunNo: 68913	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1104041-009A	04/05/11 20:37	Alkalinity, Bicarbonate as HCO3	12	5	10	11.7		20	
		Alkalinity, Total (As CaCO3)	10	5	9	11.7		20	

Qualifiers:

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
(Replaces S1103485001)

Alkalinity

Sample Type **DUP**

Units: mg/L

Sample ID	RunNo: 68913	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1104041-019A	04/05/11 22:25	Alkalinity, Bicarbonate as HCO3	49	5	47	4.57		20	
		Alkalinity, Total (As CaCO3)	40	5	38	4.57		20	

Sample ID	RunNo: 68913	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1103243-005B	04/06/11 2:14	Alkalinity, Bicarbonate as HCO3	46	5	45	1.78		20	
		Alkalinity, Carbonate as CO3	ND	5	ND	0		20	
		Alkalinity, Total (As CaCO3)	37	5	37	1.77		20	

Sample ID	RunNo: 68913	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1103491-001B	04/06/11 3:10	Alkalinity, Bicarbonate as HCO3	49	5	50	1.70		20	
		Alkalinity, Carbonate as CO3	5	5	ND	0		20	
		Alkalinity, Total (As CaCO3)	49	5	48	1.34		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
 (Replaces S1103485001)

ConductivitySample Type **MBLK**

Units: µmhos/cm

Sample ID	RunNo: 68778	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	03/31/11 14:45	Electrical Conductivity	ND	5					

Sample ID	RunNo: 68913	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	04/05/11 14:57	Electrical Conductivity	ND	5					

Sample Type **LCS**

Units: µmhos/cm

Sample ID	RunNo: 68778	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	03/31/11 14:34	Electrical Conductivity	985	5	1000		98.5	90 - 110	

Sample ID	RunNo: 68913	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	04/05/11 14:48	Electrical Conductivity	975	5	1000		97.5	90 - 110	

Sample Type **DUP**

Units: µmhos/cm

Sample ID	RunNo: 68778	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103445-009A	03/31/11 15:01	Electrical Conductivity	38400	5	41800	8.13		20	

Sample ID	RunNo: 68778	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103482-008A	03/31/11 17:58	Electrical Conductivity	1380	5	1390	0.288		20	

Sample ID	RunNo: 68778	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103492-006A	03/31/11 22:03	Electrical Conductivity	278	5	279	0.358		20	

Sample ID	RunNo: 68913	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1104036-003A	04/05/11 17:01	Electrical Conductivity	850	5	836	1.67		20	

Sample ID	RunNo: 68913	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1104041-009A	04/05/11 20:37	Electrical Conductivity	53	5	54	2.57		20	

Sample ID	RunNo: 68913	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1104041-019A	04/05/11 22:25	Electrical Conductivity	190	5	190	0.263		20	

Qualifiers:

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
(Replaces S1103485001)

Dissolved Mercury by EPA 245.1 - Water

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 68885	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LRB	04/05/11 9:18	Mercury	ND	0.001					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 68885	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS	04/05/11 9:16	Mercury	0.002	0.001	0.002		93.5	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 68885	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1103473-014C	04/05/11 12:16	Mercury	0.002	0.001	0.00244	ND	100	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 68885	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1103473-014C	04/05/11 12:17	Mercury	0.002	0.001	0.002	2.00	98.5	0	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 68885	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103473-014C	04/05/11 12:14	Mercury	ND	0.001	ND	0		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
 (Replaces S1103485001)

Dissolved Metals by ICP - EPA 200.7 - Water

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 68791	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MBLK DISS/CAT	03/31/11 15:18	Aluminum	ND	0.1					
		Boron	ND	0.1					
		Chromium	ND	0.01					
		Iron	ND	0.05					
		Manganese	ND	0.02					
		Nickel	ND	0.01					
		Zinc	ND	0.01					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 68791	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DISS LCS Q	03/31/11 15:20	Aluminum	1.0	0.1	1		96.8	85 - 115	
		Boron	1.0	0.1	1		99.6	85 - 115	
		Chromium	0.98	0.01	1		98.4	85 - 115	
		Iron	0.98	0.05	1		97.7	85 - 115	
		Manganese	0.98	0.02	1		97.8	85 - 115	
		Nickel	0.96	0.01	1		95.8	85 - 115	
		Zinc	0.97	0.01	1		97.1	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 68791	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1103474-001ES	03/31/11 17:19	Aluminum	0.5	0.1	0.5	ND	99.5	70 - 130	
		Beryllium	0.51	0.01	0.5	ND	101	70 - 130	
		Boron	0.5	0.1	0.5	ND	96.3	70 - 130	
		Chromium	0.47	0.01	0.5	ND	94.7	70 - 130	
		Iron	0.48	0.05	0.5	ND	96.1	70 - 130	
		Manganese	0.94	0.02	0.5	0.48	91.0	70 - 130	
		Nickel	0.46	0.01	0.5	ND	92.8	70 - 130	
		Zinc	0.48	0.01	0.5	ND	94.1	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 68791	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1103474-001ESD	03/31/11 17:21	Aluminum	0.5	0.1	0.5	0.281	99.8	20	
		Beryllium	0.51	0.01	0.51	0.198	101	20	
		Boron	0.5	0.1	0.5	2.39	98.8	20	
		Chromium	0.47	0.01	0.47	0.106	94.8	20	
		Iron	0.48	0.05	0.48	0.808	95.4	20	
		Manganese	0.94	0.02	0.94	0.288	91.5	20	
		Nickel	0.47	0.01	0.46	0.345	93.1	20	
		Zinc	0.48	0.01	0.48	0.168	94.2	20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
(Replaces S1103485001)

Dissolved Metals by ICP - EPA 200.7 - Water

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 68791	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103474-001ED	03/31/11 17:16	Aluminum	ND	0.1	ND	0		20	
		Boron	ND	0.1	ND	0		20	
		Chromium	ND	0.01	ND	0		20	
		Iron	ND	0.05	ND	0		20	
		Manganese	0.48	0.02	0.48	1.10		20	
		Nickel	ND	0.01	ND	0		20	
		Zinc	ND	0.01	ND	0		20	

Dissolved Metals by ICPMS EPA 200.8 - Water

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 68812	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLK	04/01/11 10:27	Arsenic	ND	0.005					
		Barium	ND	0.5					
		Cadmium	ND	0.002					
		Copper	ND	0.01					
		Lead	ND	0.02					
		Molybdenum	ND	0.02					
		Selenium	ND	0.005					
		Uranium	ND	0.001					
		Vanadium	ND	0.02					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 68812	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS	04/01/11 10:23	Arsenic	0.103	0.005	0.1		103	85 - 115	
		Barium	ND	0.5	0.1		89.0	85 - 115	
		Cadmium	0.102	0.002	0.1		102	85 - 115	
		Copper	0.09	0.01	0.1		94.3	85 - 115	
		Lead	0.10	0.02	0.1		96.3	85 - 115	
		Molybdenum	0.10	0.02	0.1		98.9	85 - 115	
		Selenium	0.103	0.005	0.1		103	85 - 115	
		Uranium	0.100	0.001	0.1		99.6	85 - 115	
		Vanadium	0.09	0.02	0.1		93.2	85 - 115	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011

Report ID: S1103485002
 (Replaces S1103485001)

Dissolved Metals by ICPMS EPA 200.8 - WaterSample Type **MS** Units: mg/L

Sample ID	RunNo: 68812	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1103455-001ES	04/01/11 10:43	Arsenic	1.15	0.005	1.1	ND	104	70 - 130	
		Barium	1.0	0.5	1.1	ND	89.2	70 - 130	
		Cadmium	1.12	0.002	1.1	ND	102	70 - 130	
		Copper	1.02	0.01	1.1	ND	92.6	70 - 130	
		Lead	1.05	0.02	1.1	ND	95.5	70 - 130	
		Molybdenum	1.13	0.02	1.1	ND	101	70 - 130	
		Selenium	1.15	0.005	1.1	ND	105	70 - 130	
		Uranium	1.11	0.001	1.1	0.019	98.8	70 - 130	
		Vanadium	1.05	0.02	1.1	ND	95.4	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 68812	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1103455-001EMSD	04/01/11 10:47	Arsenic	1.14	0.005	1.15	0.591	104	20	
		Barium	1.1	0.5	1.0	2.25	91.3	20	
		Cadmium	1.14	0.002	1.12	2.38	104	20	
		Copper	1.00	0.01	1.02	1.75	91.0	20	
		Lead	1.09	0.02	1.05	3.49	98.8	20	
		Molybdenum	1.15	0.02	1.13	2.24	104	20	
		Selenium	1.15	0.005	1.15	0.652	104	20	
		Uranium	1.13	0.001	1.11	1.89	101	20	
		Vanadium	1.02	0.02	1.05	3.06	92.5	20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 68812	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103455-001ED	04/01/11 10:39	Arsenic	ND	0.005	ND	0		20	
		Barium	ND	0.5	ND	0		20	
		Cadmium	ND	0.002	ND	0		20	
		Copper	ND	0.01	ND	0		20	
		Lead	ND	0.02	ND	0		20	
		Molybdenum	ND	0.02	ND	0		20	
		Selenium	ND	0.005	ND	0		20	
		Uranium	0.020	0.001	0.019	2.15		20	
		Vanadium	ND	0.02	ND	0		20	

Qualifiers:

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
 (Replaces S1103485001)

Fluoride by SM 4500Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 68778	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	03/31/11 14:45	Fluoride	ND	0.1					

Sample ID	RunNo: 68913	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	04/05/11 14:57	Fluoride	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 68778	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	03/31/11 14:34	Fluoride	2.6	0.1	2.82		91.6	86.1 - 114	

Sample ID	RunNo: 68913	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	04/05/11 14:48	Fluoride	2.8	0.1	2.82		99.9	86.1 - 114	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 68778	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1103482-008A	03/31/11 18:15	Fluoride	2.6	0.1	2.5	0.2	95.9	80 - 120	

Sample ID	RunNo: 68778	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1103492-015ASPK	03/31/11 23:56	Fluoride	2.4	0.1	2.5	ND	96.1	80 - 120	

Sample ID	RunNo: 68913	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1104041-2SPK	04/05/11 19:12	Fluoride	2.4	0.1	2.5	ND	94.3	80 - 120	

Sample ID	RunNo: 68913	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1104041-23SPK	04/05/11 23:08	Fluoride	2.1	0.1	2.5	ND	84.8	80 - 120	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 68778	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1103482-008A	03/31/11 18:27	Fluoride	2.7	0.1	2.6	2.48	98.5	20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 68778	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103482-008A	03/31/11 17:58	Fluoride	0.2	0.1	0.2	8.00		20	

Sample ID	RunNo: 68778	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103492-006A	03/31/11 22:03	Fluoride	0.4	0.1	0.5	10.4		20	

Sample ID	RunNo: 68778	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103493-001A	04/01/11 0:09	Fluoride	ND	0.1	ND	0		20	

Qualifiers:

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
 (Replaces S1103485001)

Fluoride by SM 4500Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 68913	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1104036-003A	04/05/11 17:01	Fluoride	ND	0.1	ND	0		20	
Sample ID	RunNo: 68913	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1104041-009A	04/05/11 20:37	Fluoride	0.2	0.1	0.2	2.17		20	
Sample ID	RunNo: 68913	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1104041-019A	04/05/11 22:25	Fluoride	0.6	0.1	0.6	2.43		20	

Gross Alpha, Beta by SM 7110BSample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 69264	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-104	04/17/11 21:40	Gross Alpha (Dissolved)	ND	2					
		Gross Beta (Dissolved)	ND	3					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 69264	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-104	04/17/11 21:40	Gross Alpha (Dissolved)	71	2	69		103	72.6 - 137	
		Gross Beta (Dissolved)	129	3	130		100	72.6 - 137	

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 69264	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1103491-001GMS	04/17/11 21:40	Gross Alpha (Dissolved)	58	2	69	ND	84.1	44.4 - 137	
		Gross Beta (Dissolved)	138	3	130	4	104	44.4 - 137	
Sample ID	RunNo: 69264	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1103495-001FMS	04/17/11 21:40	Gross Alpha (Dissolved)	86	2	69	12	108	44.4 - 137	
		Gross Beta (Dissolved)	140	3	130	5	104	44.4 - 137	

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 69264	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103494-001FDUP	04/17/11 21:40	Gross Alpha (Dissolved)	110 ± 3.7	2	120	8.66		20	
		Gross Beta (Dissolved)	31.0 ± 1.6	3	34	7.57		20	
Sample ID	RunNo: 69264	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103495-002FDUP	04/17/11 21:40	Gross Alpha (Dissolved)	17.3 ± 1.5	2	15	12.4		20	
		Gross Beta (Dissolved)	6.3 ± 1.2	3	7	11.2		20	

Qualifiers:

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
(Replaces S1103485001)

Anions by ION Chromatography

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 68828	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLK	04/01/11 11:01	Chloride	ND	1					
		Sulfate	ND	1					

Sample ID	RunNo: 68964	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLK	04/06/11 16:25	Chloride	ND	1					
		Sulfate	ND	1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 68828	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DIONEX	04/01/11 8:50	Chloride	30	1	30		99.7	90 - 110	
		Sulfate	142	1	150		94.4	90 - 110	

Sample ID	RunNo: 68964	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DIONEX	04/06/11 13:03	Chloride	30	1	30		98.7	90 - 110	
		Sulfate	139	1	150		92.8	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 68828	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1103483-001BSPK	04/01/11 19:10	Chloride	1320	1	505	770	108	80 - 120	
		Sulfate	4230	1	4040	ND	105	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 68828	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103483-001B	04/01/11 18:58	Chloride	687	1	770	10.7		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011

Report ID: S1103485002
 (Replaces S1103485001)

Cations by ICP (Method 200.7)Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 68790	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MBLK DISS/CAT	03/31/11 15:18	Calcium	ND	1					
		Magnesium	ND	1					
		Potassium	ND	1					
		Sodium	ND	1					

Sample ID	RunNo: 68912	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MBLK DISS/CAT	04/05/11 16:31	Calcium	ND	1					
		Magnesium	ND	1					
		Potassium	ND	1					
		Sodium	ND	1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 68790	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
CAT LCS IML3	03/31/11 15:23	Calcium	40	1	40		99.6	85 - 115	
		Magnesium	39	1	40		98.4	85 - 115	
		Potassium	39	1	40		98.6	85 - 115	
		Sodium	38	1	40		94.9	85 - 115	

Sample ID	RunNo: 68912	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
CAT LCS IML3	04/05/11 16:36	Calcium	38	1	40		96.0	85 - 115	
		Magnesium	39	1	40		96.7	85 - 115	
		Potassium	39	1	40		96.2	85 - 115	
		Sodium	38	1	40		94.5	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 68790	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1103482-008AS	03/31/11 18:12	Calcium	204	1	25	187	67.0	70 - 130	S
		Magnesium	125	1	25	103	87.0	70 - 130	
		Potassium	38	1	25	9	117	70 - 130	
		Sodium	60	1	25	35	99.4	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 68790	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1103482-008ASD	03/31/11 18:14	Calcium	206	1	204	0.937	74.6	20	
		Magnesium	126	1	125	0.888	91.5	20	
		Potassium	37	1	38	2.19	114	20	
		Sodium	59	1	60	2.23	94.0	20	

Qualifiers:

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
(Replaces S1103485001)

Cations by ICP (Method 200.7)

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 68790	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103482-008AD	03/31/11 18:09	Calcium	185	1	187	1.03		20	
		Magnesium	103	1	103	0.619		20	
		Potassium	9	1	9	1.77		20	
		Sodium	36	1	35	2.13		20	

Nitrogen, Ammonia (as N)

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 68802	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	04/01/11 9:57	Nitrogen, Ammonia (As N)	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 68802	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
QC	04/01/11 9:59	Nitrogen, Ammonia (As N)	9.2	0.1	8.93		103	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 68802	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1103482-001B	04/01/11 11:50	Nitrogen, Ammonia (As N)	9.6	0.1	10	ND	96.1	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 68802	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103482-001B	04/01/11 11:49	Nitrogen, Ammonia (As N)	ND	0.1	ND	0		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
(Replaces S1103485001)

Nitrogen, Nitrate-Nitrite (as N)

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 69088	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLK	04/11/11 16:20	Nitrogen, Nitrate-Nitrite (as N)	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 69088	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
QC	04/11/11 16:17	Nitrogen, Nitrate-Nitrite (as N)	7.7	0.1	7.69		100	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 69088	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1104001-003B	04/11/11 17:05	Nitrogen, Nitrate-Nitrite (as N)	8.3	0.1	10	ND	82.5	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 69088	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1104001-003B	04/11/11 17:04	Nitrogen, Nitrate-Nitrite (as N)	ND	0.1	ND	0		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
(Replaces S1103485001)

Lead 210 and Polonium 210

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 69362	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB11-094 04/19/11 19:25 Lead 210 (Dissolved) ND 1

Sample ID	RunNo: 69436	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB11-104 04/25/11 10:14 Lead 210 (Suspended) ND 1

Sample ID	RunNo: 69436	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB11-104B 04/25/11 12:24 Lead 210 (Suspended) ND 1

Sample ID	RunNo: 69473	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB11-094 04/21/11 9:24 Polonium 210 (Dissolved) ND 1

Sample ID	RunNo: 69550	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB11-104 04/28/11 8:30 Polonium 210 (Suspended) ND 1

Sample ID	RunNo: 69550	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB11-104B 04/28/11 9:44 Polonium 210 (Suspended) ND 1

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 69362	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

LCS11-094 04/19/11 19:25 Lead 210 (Dissolved) 15 1 13.1 116 67.6 - 131

Sample ID	RunNo: 69436	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

LCS11-104 04/25/11 10:14 Lead 210 (Suspended) 14 1 13.1 107 67.6 - 131

Sample ID	RunNo: 69473	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

LCS11-094 04/21/11 9:24 Polonium 210 (Dissolved) 12 1 13.1 93.6 67.6 - 131

Sample ID	RunNo: 69550	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

LCS11-104 04/28/11 8:30 Polonium 210 (Suspended) 13 1 13.1 102 67.6 - 131

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 69362	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

S1103348-001GMS 04/19/11 19:25 Lead 210 (Dissolved) 14 1 13.1 1 92.5 66 - 133

Sample ID	RunNo: 69362	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

S1103348-002GMS 04/19/11 19:25 Lead 210 (Dissolved) 19 1 13.1 5 107 66 - 133

Sample ID	RunNo: 69436	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

S1103455-001GMS 04/25/11 10:14 Lead 210 (Suspended) 18 1 13.1 5 96.7 66 - 133

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
 (Replaces S1103485001)

Lead 210 and Polonium 210Sample Type **MS** Units: pCi/L

Sample ID	RunNo:	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------	---------	--------	----	-------	----------	------	--------------	------

S1103485-001GMS	04/25/11 10:14	Lead 210 (Suspended)	31	1	13.1	18	103	66 - 133	
-----------------	----------------	----------------------	----	---	------	----	-----	----------	--

Sample ID	RunNo:	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------	---------	--------	----	-------	----------	------	--------------	------

S1103348-002GMS	04/21/11 9:24	Polonium 210 (Dissolved)	3	1	13.1	ND	23.7	66 - 133	
-----------------	---------------	--------------------------	---	---	------	----	------	----------	--

Sample ID	RunNo:	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------	---------	--------	----	-------	----------	------	--------------	------

S1103455-001GMS	04/28/11 8:30	Polonium 210 (Suspended)	7	1	13.1	ND	52.4	66 - 133	
-----------------	---------------	--------------------------	---	---	------	----	------	----------	--

Sample ID	RunNo:	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------	---------	--------	----	-------	----------	------	--------------	------

S1103485-001GMS	04/28/11 8:30	Polonium 210 (Suspended)	5	1	13.1	ND	40.4	66 - 133	
-----------------	---------------	--------------------------	---	---	------	----	------	----------	--

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo:	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------	---------	--------	----	----------	------	------	--------------	------

S1103410-001DDUP	04/19/11 19:25	Lead 210 (Dissolved)	1.1 ± 0.5	1	1	8.36		20	
------------------	----------------	----------------------	-----------	---	---	------	--	----	--

Sample ID	RunNo:	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------	---------	--------	----	----------	------	------	--------------	------

S1103410-001DDUP	04/21/11 9:24	Polonium 210 (Dissolved)	ND	1	ND	0		20	
------------------	---------------	--------------------------	----	---	----	---	--	----	--

Sample ID	RunNo:	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------	---------	--------	----	----------	------	------	--------------	------

S1103474-001GDUP	04/28/11 8:30	Polonium 210 (Suspended)	ND	1	ND	0		20	
------------------	---------------	--------------------------	----	---	----	---	--	----	--

Radium 228Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo:	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------	---------	--------	----	-------	----------	------	--------------	------

MB11-103	04/18/11 16:25	Radium 228 (Dissolved)	ND	1					
----------	----------------	------------------------	----	---	--	--	--	--	--

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo:	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------	---------	--------	----	-------	----------	------	--------------	------

LCS11-103	04/18/11 16:25	Radium 228 (Dissolved)	15	1	13.9		111	74.9 - 122	
-----------	----------------	------------------------	----	---	------	--	-----	------------	--

Sample Type **MS** Units: pCi/L

Sample ID	RunNo:	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------	---------	--------	----	-------	----------	------	--------------	------

S1103474-001GMS	04/18/11 16:25	Radium 228 (Dissolved)	13	1	13.9	ND	91.6	48.4 - 167	
-----------------	----------------	------------------------	----	---	------	----	------	------------	--

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo:	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------	---------	--------	----	----------	------	------	--------------	------

S1103491-001GDUP	04/18/11 16:25	Radium 228 (Dissolved)	ND	1	ND	0		20	
------------------	----------------	------------------------	----	---	----	---	--	----	--

Qualifiers:

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
 (Replaces S1103485001)

Radium 226 by SM 7500Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 69370	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-096	04/20/11 16:50	Radium 226 (Dissolved)	ND	0.2					

Sample ID	RunNo: 69475	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-103	04/25/11 8:23	Radium 226 (Suspended)	ND	0.2					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 69370	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-096	04/20/11 16:50	Radium 226 (Dissolved)	5.7	0.2	5		113	62.5 - 125	

Sample ID	RunNo: 69475	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-103	04/25/11 8:23	Radium 226 (Suspended)	4.8	0.2	5		95.5	62.5 - 125	

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 69370	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1103411-001DMS	04/20/11 16:50	Radium 226 (Dissolved)	5.7	0.2	5	0.7	101	41.6 - 146	

Sample ID	RunNo: 69370	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1103485-001GMS	04/20/11 18:59	Radium 226 (Dissolved)	53.0	0.2	5	46.9	123	41.6 - 146	

Sample ID	RunNo: 69475	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1103494-002FMS	04/26/11 11:46	Radium 226 (Suspended)	6.5	0.2	5	ND	129	41.6 - 146	

Sample ID	RunNo: 69475	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1103485-001GMS	04/26/11 11:46	Radium 226 (Suspended)	5.5	0.2	5	ND	110	41.6 - 146	

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 69370	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103411-002DDUP	04/20/11 16:50	Radium 226 (Dissolved)	ND	0.2	ND	0		20	

Sample ID	RunNo: 69370	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103474-001GDUP	04/20/11 18:59	Radium 226 (Dissolved)	0.5	0.2	0.6	25.1		20	R

Sample ID	RunNo: 69475	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103494-001FDUP	04/26/11 11:46	Radium 226 (Suspended)	0.3	0.2	0.3	10.7		20	

Sample ID	RunNo: 69475	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1104083-001GDUP	04/25/11 9:28	Radium 226 (Suspended)	ND	0.2	ND	0		20	

Qualifiers:

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
 (Replaces S1103485001)

Thorium 230 in WaterSample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 69252	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-095	04/11/11 9:40	Thorium 230 (Dissolved)	ND	0.2					

Sample ID	RunNo: 69452	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-102B	04/20/11 14:08	Thorium 230 (Dissolved)	ND	0.2					

Sample ID	RunNo: 69452	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DMB11-102	04/21/11 12:32	Thorium 230 (Suspended)	ND	0.2					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 69252	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-095	04/11/11 9:40	Thorium 230 (Dissolved)	11	0.2	12		91.9	50 - 150	

Sample ID	RunNo: 69452	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-102B	04/20/11 14:08	Thorium 230 (Dissolved)	12	0.2	12		98.4	50 - 150	
		Thorium 230 (Suspended)	12	0.2	12		98.4	50 - 150	

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 69252	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103485-001GDUP	04/11/11 17:53	Thorium 230 (Dissolved)	ND	0.2	ND	0		20	

Sample ID	RunNo: 69452	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103494-001FDUP	04/20/11 14:09	Thorium 230 (Dissolved)	ND	0.2	ND	0		20	

Sample ID	RunNo: 69452	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103455-001GDUP	04/18/11 17:46	Thorium 230 (Suspended)	ND	0.2	ND	0		20	

Radon 222 by SM7500-RNSample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 69021	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB-R69021	03/31/11 0:00	Radon-222	ND	50					L

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 69021	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS-R69021	03/31/11 0:00	Radon-222	22200	100	22500		98.6	75 - 125	L

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 69021	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103485-001B	03/31/11 0:00	Radon-222	32100	40	40900	21.5		20	RL

Qualifiers:

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
(Replaces S1103485001)

Silica as SiO2

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 68792	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MBLK DISS/CAT	03/31/11 15:18	Silica as SiO2	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 68792	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DISS LCS Q	03/31/11 15:20	Silica as SiO2	2.2	0.1	2.14		100	85 - 115	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 68792	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103474-001ED	03/31/11 17:16	Silica as SiO2	9.8	0.1	9.9	1.10		20	

Solids By SM 2540

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 68854	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DI	04/01/11 10:10	Total Dissolved Solids (180)	10.0	10					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 68854	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
CONTROL	04/01/11 10:15	Total Dissolved Solids (180)	230	10	226		103	90 - 110	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 68854	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103472-001A	04/01/11 11:30	Total Dissolved Solids (180)	1690	10	1700	0.939		20	

Sample ID	RunNo: 68854	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103487-001A	04/01/11 14:55	Total Dissolved Solids (180)	4660	10	4590	1.53		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
(Replaces S1103485001)

Total (200.2) Metals by EPA 200.7 ICP - Water

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 68869	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLK 4/4 TOT -5008	04/04/11 17:22	Iron	ND	0.05					
		Manganese	ND	0.02					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 68869	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS 4/4 TOT -5008	04/04/11 17:24	Iron	0.47	0.05	0.5		94.8	85 - 115	
		Manganese	0.19	0.02	0.2		95.5	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 68869	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1104006-001DS	04/04/11 17:52	Iron	1.09	0.05	0.5	0.58	102	70 - 130	
		Manganese	0.40	0.02	0.2	0.20	103	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 68869	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1104006-001DSD	04/04/11 17:54	Iron	1.07	0.05	1.09	1.60	98.9	20	
		Manganese	0.40	0.02	0.40	0.742	101	20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 68869	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103485-001FD	04/04/11 17:47	Iron	0.23	0.05	0.23	0.597		20	
		Manganese	0.10	0.02	0.10	2.12		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
(Replaces S1103485001)

Suspended Natural Uranium

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 68908	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLK	04/05/11 9:39	Uranium Suspended	ND	0.0003					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 68908	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS	04/05/11 9:34	Uranium Suspended	0.0918	0.0003	0.1		91.8	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 68908	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1103374-001HS	04/05/11 10:25	Uranium Suspended	1.00	0.0003	1.1	ND	91.0	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 68908	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1103374-001HMSD	04/05/11 10:28	Uranium Suspended	0.987	0.0003	1.00	1.41	89.7	20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 68908	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103374-001HD	04/05/11 10:22	Uranium Suspended	ND	0.0003	ND	0		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1103485
Project: Reno Creek Project

Date: 7/13/2011
Report ID: S1103485002
(Replaces S1103485001)

pH Water

Sample Type **LCS** Units: s.u.

Sample ID	RunNo: 68778	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	03/31/11 14:34	pH	10.7	0.1	10.7		100	94.8 - 105	

Sample ID	RunNo: 68913	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	04/05/11 14:48	pH	10.8	0.1	10.7		101	94.8 - 105	

Sample Type **DUP** Units: s.u.

Sample ID	RunNo: 68778	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103445-009A	03/31/11 15:01	pH	ND	0.1	ND	0		20	

Sample ID	RunNo: 68778	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103482-008A	03/31/11 17:58	pH	7.7	0.1	7.8	0.186		20	

Sample ID	RunNo: 68778	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103488-001A	03/31/11 20:15	pH	8.0	0.1	7.7	3.14		20	

Sample ID	RunNo: 68778	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1103492-006A	03/31/11 22:03	pH	10.0	0.1	10.0	0.0726		20	

Sample ID	RunNo: 68913	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1104036-003A	04/05/11 17:01	pH	2.9	0.1	2.9	1.18		20	

Sample ID	RunNo: 68913	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1104037-001B	04/05/11 18:38	pH	8.5	0.1	8.5	0.139		20	

Sample ID	RunNo: 68913	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1104041-009A	04/05/11 20:37	pH	7.1	0.1	7.0	1.42		20	

Sample ID	RunNo: 68913	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1104041-019A	04/05/11 22:25	pH	8.1	0.1	8.1	0.602		20	

Sample ID	RunNo: 68913	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1104046-004B	04/05/11 23:56	pH	7.9	0.1	7.9	0.0517		20	

Sample ID	RunNo: 68913	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1104052-001C	04/06/11 0:31	pH	7.8	0.1	7.8	0.360		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits

Client Name: **AUC LLC** Report Address: **peavendor@auc-llc.com; mickerson@treccorp.com; jshriver@treccorp.com**
 Invoice Address: **AUC LLC** Invoice Address: **981 W. 1st St. Casper, WY 82401**
 Project Identification: **Reno Creek Project** Contact Name: **Traci Nickerson** Sampler (Signature/Printed): **Traci Nickerson 265-0696**
 Email: **tracynickerson@treccorp.com** Phone: **307-265-0696** Purchase Order #: **265-0696** Quote #: **265-0696**
 Telephone #: **307-265-0696**

ITEM	LAB ID (Lab Use Only)	DATE SAMPLED	TIME	SAMPLE IDENTIFICATION	Matrix	# of Containers	ANALYSES / PARAMETERS	REMARKS
1							See Attached Table (4.14, Guide 8, Radon 222)	
2	51103485	3/29/11	1430	D2M8-002410389	W	13	See Attached Table 2 (Air Particulate)	Groundwater
3							See Attached Table 3 (Vegetation)	
4							See Attached Table 4 (Sediment)	
5							See Attached Table 5 (Soil)	
6							See Attached Table 8 WYPDES Discharge	
7							See Attached Table 9 (Blanks)	
8								
9								
10								
11								
12								
13								
14								

LAB COMMENTS: **Relinquished By (Signature/Printed): Traci Nickerson 3/29/11 1430**
 RECEIVED BY (Signature/Printed): **Alvin Q. McLeod 3/29/11 1718**
 DATE TIME: **3/29/11 1718**
 RECEIVED BY (Signature/Printed): **Rodney Pope 3/29/11 10:17**
 DATE TIME: **3/29/11 10:17**

SHIPPING INFO: UPS Fed Express US Mail Hand Carried Other _____

MATRIX CODES: Water Soil Solid Tip Blank Other _____

TURNAROUND TIMES: Check desired service Standard turnaround RUSH - 5 Working Days URGENT - < 2 Working Days

COMPLIANCE INFORMATION: Compliance Monitoring? Program (SDWA, NPDES, ...) PWSID / Permit # _____ Chlorinated? _____ Sample Disposal: Lab Client

ADDITIONAL REMARKS: **37**