



Tuesday, June 05, 2012

Phil Cavendor  
AUC, LLC  
1536 Cole Blvd  
Suite 330  
Lakewood, CO 80401

RE: Reno Creek Project  
Orde

r No.: S1205139

Dear Phil Cavendor:

Inter-Mountain Laboratories received sample SM7-004-120508 were received on May 8, 2012.

There were no problems with the analyses and all data for associated QC met EPA or laboratory specifications except where noted in the Case Narrative.

If you have any questions regarding these tests results, please feel free to call.

Sincerely,

Lacey Ketron  
Project Manager

xc: File  
Encl.



Date: 6/5/2012

**CLIENT:** AUC LLC  
**Project:** Reno Creek Project  
**Lab Order:** S1205139

**CASE NARRATIVE**  
**Report ID:** S1205139001

Sample SM7-004-120508 was received on May 8, 2012.

All samples were received and analyzed within the EPA recommended holding times, except those noted in this case narrative. Samples were analyzed using the methods outlined in the following references:

U.S.E.P.A. 600 "Methods for Chemical Analysis of Water and Wastes", 1993  
"Standard Methods For The Examination of Water and Wastewater", 20th ed., 1998  
Test Methods for Evaluating Solid Waste, Physical/Chemical Methods, SW846, 3rd Edition  
Methods indicated with the Monday, March 12, 2007 Federal Register, 40 CFR Part 122, 136 et al.

All Quality Control parameters met the acceptance criteria defined by EPA and Inter-Mountain Laboratories except as indicated in this case narrative.

For the purposes of this report, the Reporting Limit (RL) is the Practical Quantitation Limit (PQL).

Radon 222 was analyzed through ALS Laboratory Group Certification #: CO00078.

Qualifiers by sample

S1205136-001 - Radiochemistry/Gross Alpha (Dissolved) - RPD outside accepted recovery limits  
S1205139-001 - General Parameters/Radon-222 - Analyzed by a contract laboratory  
LCS-R84003 - General Parameters/Radon-222 - Analyzed by a contract laboratory  
MB-R84003 - General Parameters/Radon-222 - Analyzed by a contract laboratory

Reviewed by:

Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
1536 Cole Blvd
Lakewood, CO 80401

Date Reported 6/5/2012
Report ID S1205139001

ProjectName: Reno Creek Project
Lab ID: S1205139-001
ClientSample ID: SM7-004-120508
COC: RC 08471

WorkOrder: S1205139
CollectionDate: 5/8/2012 10:00:00 AM
DateReceived: 5/8/2012 3:47:00 PM
FieldSampler: TN
Matrix: Water

Comments

Table with 7 columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init. Rows include Field parameters (pH, Conductivity, etc.), Anions/Cations (Alkalinity, Chloride, etc.), General Parameters (pH, Electrical Conductivity, etc.), and Data Quality (Cation Sum, Anion Sum, etc.).

These results apply only to the samples tested.

RL - Reporting Limit

- Qualifiers: \* Value exceeds Maximum Contaminant Level
C Calculated Value
H Holding times for preparation or analysis exceeded
L Analyzed by a contract laboratory
ND Not Detected at the Reporting Limit
S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
E Value above quantitation range
J Analyte detected below quantitation limits
M Value exceeds Monthly Ave or MCL
O Outside the Range of Dilutions

Reviewed by: [Signature]
Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
1536 Cole Blvd
Lakewood, CO 80401
ProjectName: Reno Creek Project
Lab ID: S1205139-001
ClientSample ID: SM7-004-120508
COC: RC 08471

Date Reported 6/5/2012
Report ID S1205139001
WorkOrder: S1205139
CollectionDate: 5/8/2012 10:00:00 AM
DateReceived: 5/8/2012 3:47:00 PM
FieldSampler: TN
Matrix: Water

Comments

Table with 7 columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init. Rows include Metals - Dissolved (Aluminum, Arsenic, Barium, Boron, Cadmium, Chromium, Copper, Iron, Lead, Manganese, Mercury, Molybdenum, Nickel, Selenium, Uranium, Vanadium, Zinc) and Metals - Suspended (Uranium) and Metals - Total (Iron, Manganese).

These results apply only to the samples tested.

RL - Reporting Limit

- Qualifiers: \* Value exceeds Maximum Contaminant Level
C Calculated Value
H Holding times for preparation or analysis exceeded
L Analyzed by a contract laboratory
ND Not Detected at the Reporting Limit
S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
E Value above quantitation range
J Analyte detected below quantitation limits
M Value exceeds Monthly Ave or MCL
O Outside the Range of Dilutions

Reviewed by: [Signature]
Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
1536 Cole Blvd
Lakewood, CO 80401
ProjectName: Reno Creek Project
Lab ID: S1205139-001
ClientSample ID: SM7-004-120508
COC: RC 08471

Date Reported 6/5/2012
Report ID S1205139001
WorkOrder: S1205139
CollectionDate: 5/8/2012 10:00:00 AM
DateReceived: 5/8/2012 3:47:00 PM
FieldSampler: TN
Matrix: Water

Comments

Table with 7 columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init. Rows include Radionuclides - Dissolved (Gross Alpha, Lead 210, Polonium 210, Radium 226, Radium 228, Thorium 230) and Radionuclides - Suspended (Lead 210, Polonium 210, Radium 226, Thorium 230). Total radionuclides include Radon 222.

These results apply only to the samples tested.

RL - Reporting Limit

- Qualifiers: \* Value exceeds Maximum Contaminant Level
C Calculated Value
H Holding times for preparation or analysis exceeded
L Analyzed by a contract laboratory
ND Not Detected at the Reporting Limit
S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
E Value above quantitation range
J Analyte detected below quantitation limits
M Value exceeds Monthly Ave or MCL
O Outside the Range of Dilutions

Reviewed by: [Signature]
Lacey Ketron, Water Lab Supervisor



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1205139  
**Project:** Reno Creek Project

**Date:** 6/5/2012  
**Report ID:** S1205139001

**Alkalinity**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 82721	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	05/11/12 11:32	Alkalinity, Total (As CaCO3)	38	5					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 82721	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	05/11/12 11:19	Alkalinity, Total (As CaCO3)	604	5	601		100	90 - 110	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 82721	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205135-008AD	05/11/12 16:18	Alkalinity, Bicarbonate as HCO3	2250	5	2220	0.958		20	
		Alkalinity, Carbonate as CO3	ND	5	ND			20	
		Alkalinity, Total (As CaCO3)	1840	5	1820	0.958		20	

**Conductivity**

Sample Type **MBLK** Units: µmhos/cm

Sample ID	RunNo: 82721	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	05/11/12 11:32	Electrical Conductivity	ND	5					

Sample Type **LCS** Units: µmhos/cm

Sample ID	RunNo: 82721	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	05/11/12 11:19	Electrical Conductivity	997	5	1030		96.8	90 - 110	

Sample Type **DUP** Units: µmhos/cm

Sample ID	RunNo: 82721	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205135-008AD	05/11/12 16:18	Electrical Conductivity	7130	5	7160	0.419		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1205139  
**Project:** Reno Creek Project

**Date:** 6/5/2012  
**Report ID:** S1205139001

**Dissolved Mercury by EPA 245.1 - Water**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 82789	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LRB	05/15/12 9:18	Mercury	ND	0.001					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 82789	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS	05/15/12 9:16	Mercury	0.002	0.001	0.002		100	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 82789	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1205139-001D	05/15/12 11:09	Mercury	0.002	0.001	0.00244	ND	92.1	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 82789	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1205139-001D	05/15/12 11:10	Mercury	0.002	0.001	0.002	2.15		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 82789	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205139-001D	05/15/12 11:07	Mercury	ND	0.001	ND			20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits

**ANALYTICAL QC SUMMARY REPORT**

**CLIENT:** AUC LLC  
**Work Order:** S1205139  
**Project:** Reno Creek Project

**Date:** 6/5/2012  
**Report ID:** S1205139001

**Dissolved Metals by ICP - EPA 200.7 - Water**Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 82657	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MBLK DISS/CAT	05/10/12 11:39	Aluminum	ND	0.1					
		Boron	ND	0.1					
		Chromium	ND	0.01					
		Iron	ND	0.05					
		Manganese	ND	0.02					
		Nickel	ND	0.01					
		Zinc	ND	0.01					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 82657	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DISS LCS Q	05/10/12 11:42	Aluminum	1.1	0.1	1		105	85 - 115	
		Boron	1.0	0.1	1		101	85 - 115	
		Chromium	1.05	0.01	1		105	85 - 115	
		Iron	1.03	0.05	1		103	85 - 115	
		Manganese	1.07	0.02	1		107	85 - 115	
		Nickel	1.06	0.01	1		106	85 - 115	
		Zinc	1.04	0.01	1		104	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 82657	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1205140-001DS	05/10/12 16:31	Aluminum	0.5	0.1	0.5	ND	109	70 - 130	
		Boron	0.5	0.1	0.5	ND	99.1	70 - 130	
		Chromium	0.52	0.01	0.5	ND	104	70 - 130	
		Iron	0.53	0.05	0.5	ND	105	70 - 130	
		Manganese	0.56	0.02	0.5	0.03	107	70 - 130	
		Nickel	0.53	0.01	0.5	ND	106	70 - 130	
		Zinc	0.55	0.01	0.5	ND	109	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 82657	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1205140-001DSD	05/10/12 16:33	Aluminum	0.5	0.1	0.5	1.33		20	
		Boron	0.5	0.1	0.5	1.40		20	
		Chromium	0.50	0.01	0.52	2.99		20	
		Iron	0.51	0.05	0.53	3.02		20	
		Manganese	0.54	0.02	0.56	3.35		20	
		Nickel	0.52	0.01	0.53	2.74		20	
		Zinc	0.53	0.01	0.55	2.38		20	

**Qualifiers:**

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		





### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1205139  
**Project:** Reno Creek Project

**Date:** 6/5/2012  
**Report ID:** S1205139001

**Dissolved Metals by ICP - EPA 200.7 - Water**

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 82657	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205139-001DD	05/10/12 16:26	Aluminum	ND	0.1	ND			20	
		Boron	ND	0.1	ND			20	
		Chromium	ND	0.01	ND			20	
		Iron	ND	0.05	ND			20	
		Manganese	0.87	0.02	0.86	1.25		20	
		Nickel	ND	0.01	ND			20	
		Zinc	ND	0.01	ND			20	

**Dissolved Metals by ICPMS EPA 200.8 - Water**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 82591	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLK	05/09/12 14:13	Arsenic	ND	0.005					
		Barium	ND	0.1					
		Cadmium	ND	0.002					
		Copper	ND	0.01					
		Lead	ND	0.02					
		Molybdenum	ND	0.02					
		Selenium	ND	0.005					
		Uranium	ND	0.001					
		Vanadium	ND	0.02					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 82591	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS	05/09/12 14:08	Arsenic	0.100	0.005	0.1		99.5	85 - 115	
		Barium	ND	0.1	0.1		95.7	85 - 115	
		Cadmium	0.098	0.002	0.1		97.5	85 - 115	
		Copper	0.10	0.01	0.1		101	85 - 115	
		Lead	0.10	0.02	0.1		101	85 - 115	
		Molybdenum	0.10	0.02	0.1		101	85 - 115	
		Selenium	0.100	0.005	0.1		99.8	85 - 115	
		Uranium	0.098	0.001	0.1		98.2	85 - 115	
		Vanadium	0.10	0.02	0.1		99.4	85 - 115	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits

**ANALYTICAL QC SUMMARY REPORT**

**CLIENT:** AUC LLC  
**Work Order:** S1205139  
**Project:** Reno Creek Project

**Date:** 6/5/2012  
**Report ID:** S1205139001

**Dissolved Metals by ICPMS EPA 200.8 - Water**Sample Type **MS** Units: mg/L

Sample ID	RunNo: 82591	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1205159-004CS	05/09/12 15:47	Arsenic	1.11	0.005	1.1	ND	101	70 - 130	
		Barium	1.1	0.1	1.1	ND	96.9	70 - 130	
		Cadmium	1.05	0.002	1.1	ND	95.4	70 - 130	
		Copper	1.11	0.01	1.1	ND	101	70 - 130	
		Lead	1.08	0.02	1.1	ND	98.2	70 - 130	
		Molybdenum	1.09	0.02	1.1	ND	98.9	70 - 130	
		Selenium	1.09	0.005	1.1	ND	99.4	70 - 130	
		Uranium	1.12	0.001	1.1	0.021	99.6	70 - 130	
		Vanadium	1.14	0.02	1.1	ND	103	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 82591	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1205159-004CMSD	05/09/12 16:02	Arsenic	1.08	0.005	1.11	2.57		20	
		Barium	1.1	0.1	1.1	0.423		20	
		Cadmium	1.06	0.002	1.05	1.32		20	
		Copper	1.13	0.01	1.11	1.06		20	
		Lead	1.06	0.02	1.08	1.53		20	
		Molybdenum	1.09	0.02	1.09	0.0578		20	
		Selenium	1.06	0.005	1.09	3.33		20	
		Uranium	1.13	0.001	1.12	0.834		20	
		Vanadium	1.10	0.02	1.14	3.29		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 82591	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205159-004CD	05/09/12 15:42	Arsenic	ND	0.005	ND			20	
		Barium	ND	0.1	ND			20	
		Cadmium	ND	0.002	ND			20	
		Copper	ND	0.01	ND			20	
		Lead	ND	0.02	ND			20	
		Molybdenum	ND	0.02	ND			20	
		Selenium	ND	0.005	ND			20	
		Uranium	0.020	0.001	0.021	3.94		20	
		Vanadium	ND	0.02	ND			20	

**Qualifiers:**

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1205139  
**Project:** Reno Creek Project

**Date:** 6/5/2012  
**Report ID:** S1205139001

**Fluoride by SM 4500**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 82721	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	05/11/12 11:32	Fluoride	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 82721	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	05/11/12 11:19	Fluoride	3.2	0.1	3.18		102	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 82721	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1205145-001ASPK	05/11/12 19:29	Fluoride	2.6	0.1	2.5	0.4	89.0	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 82721	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205135-008AD	05/11/12 16:18	Fluoride	0.5	0.1	0.5	0.514		20	

**Gross Alpha, Beta by SM 7110B**

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 83405	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB12-152	05/31/12 19:58	Gross Alpha (Dissolved)	ND	2					
		Gross Beta (Dissolved)	ND	3					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 83405	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS12-152	05/31/12 19:58	Gross Alpha (Dissolved)	76	2	69		110	83 - 133	
		Gross Beta (Dissolved)	140	3	126		111	88 - 115	

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 83405	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1205142-001FMS	05/31/12 19:58	Gross Alpha (Dissolved)	69	2	69	ND	99.3	48 - 153	
		Gross Beta (Dissolved)	141	3	126	ND	112	60 - 140	

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 83405	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205136-001FDUP	05/31/12 19:58	Gross Alpha (Dissolved)	2	2	ND			20	
		Gross Beta (Dissolved)	10	4	11	10.2		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1205139  
**Project:** Reno Creek Project

**Date:** 6/5/2012  
**Report ID:** S1205139001

**Anions by ION Chromatography**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 82640	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLK	05/10/12 9:33	Chloride	ND	1					
		Sulfate	ND	1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 82640	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DIONEX	05/10/12 9:03	Sulfate	149	1	150		99.3	90 - 110	

Sample ID	RunNo: 82640	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DIONEX	05/10/12 9:13	Chloride	28	1	30		94.3	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 82640	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1205162-001BSPK	05/10/12 21:24	Chloride	788	1	765	33	98.7	80 - 120	
		Sulfate	4260	1	2040	1820	119	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 82640	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205162-001B	05/10/12 21:14	Chloride	33	1	33	1.27		20	
		Sulfate	1810	1	1820	0.748		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits

**ANALYTICAL QC SUMMARY REPORT**

**CLIENT:** AUC LLC  
**Work Order:** S1205139  
**Project:** Reno Creek Project

**Date:** 6/5/2012  
**Report ID:** S1205139001

**Cations by ICP (Method 200.7)**Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 82641	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MBLK DISS/CAT	05/10/12 11:39	Calcium	ND	1					
		Magnesium	ND	1					
		Potassium	ND	1					
		Sodium	ND	1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 82641	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
CAT LCS IML3	05/10/12 11:44	Calcium	41	1	40		102	85 - 115	
		Magnesium	41	1	40		103	85 - 115	
		Potassium	41	1	40		103	85 - 115	
		Sodium	42	1	40		104	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 82641	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1205140-001AS	05/10/12 16:31	Calcium	127	1	50	81	91.8	70 - 130	
		Magnesium	70	1	50	17	106	70 - 130	
		Potassium	65	1	50	7	115	70 - 130	
		Sodium	177	1	50	129	97.0	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 82641	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1205140-001ASD	05/10/12 16:33	Calcium	125	1	127	1.41		20	
		Magnesium	69	1	70	1.77		20	
		Potassium	64	1	65	1.61		20	
		Sodium	176	1	177	0.538		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 82641	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205139-001AD	05/10/12 16:26	Calcium	363	1	359	1.10		20	
		Magnesium	63	1	61	3.52		20	
		Potassium	13	1	13	0.260		20	
		Sodium	430	1	434	0.970		20	

**Qualifiers:**

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1205139  
**Project:** Reno Creek Project

**Date:** 6/5/2012  
**Report ID:** S1205139001

**Nitrogen, Ammonia (as N)**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 82736	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	05/14/12 10:17	Nitrogen, Ammonia (As N)	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 82736	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
QC	05/14/12 10:19	Nitrogen, Ammonia (As N)	10.8	0.1	10		108	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 82736	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1205141-001C	05/14/12 10:49	Nitrogen, Ammonia (As N)	10.6	0.1	10	ND	106	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 82736	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205141-001C	05/14/12 10:48	Nitrogen, Ammonia (As N)	ND	0.1	ND			20	

**Nitrogen, Nitrate-Nitrite (as N)**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 82633	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	05/10/12 14:17	Nitrogen, Nitrate-Nitrite (as N)	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 82633	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
QC	05/10/12 14:19	Nitrogen, Nitrate-Nitrite (as N)	4.4	0.1	4.35		101	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 82633	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1205141-001C	05/10/12 14:56	Nitrogen, Nitrate-Nitrite (as N)	6.2	0.1	5	1.3	99.1	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 82633	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205141-001C	05/10/12 14:55	Nitrogen, Nitrate-Nitrite (as N)	1.3	0.1	1.3	0.797		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1205139  
**Project:** Reno Creek Project

**Date:** 6/5/2012  
**Report ID:** S1205139001

**Lead 210 and Polonium 210**

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 83407	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB12-132 05/25/12 9:42 Lead 210 (Dissolved) ND 1

Sample ID	RunNo: 83408	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB12-132 05/24/12 14:05 Polonium 210 (Dissolved) ND 1

Sample ID	RunNo: 83415	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB12-137 05/29/12 15:32 Lead 210 (Suspended) ND 1

Sample ID	RunNo: 83416	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB12-137 05/29/12 12:51 Polonium 210 (Suspended) ND 1

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 83407	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

LCS12-132 05/25/12 9:42 Lead 210 (Dissolved) 15 1 12.7 120 67.6 - 131

Sample ID	RunNo: 83408	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

LCS12-132 05/24/12 14:05 Polonium 210 (Dissolved) 5 1 5.21 92.5 35.1 - 175

Sample ID	RunNo: 83415	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

LCS12-137 05/29/12 15:32 Lead 210 (Suspended) 14 1 12.7 112 60.3 - 141

Sample ID	RunNo: 83416	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

LCS12-137 05/29/12 12:51 Polonium 210 (Suspended) 7 1 6.83 100 48.6 - 156

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 83407	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

S1205136-001FMS 05/25/12 9:42 Lead 210 (Dissolved) 13 1 12.5 ND 105 66 - 133

Sample ID	RunNo: 83407	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

S1205137-001FMS 05/25/12 11:53 Lead 210 (Dissolved) 13 1 12.5 ND 104 66 - 133

Sample ID	RunNo: 83408	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

S1205136-001FMS 05/24/12 14:05 Polonium 210 (Dissolved) 8 1 12.5 ND 63.9 20.9 - 160

Sample ID	RunNo: 83408	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

S1205137-001FMS 05/24/12 15:35 Polonium 210 (Dissolved) 17 1 12.5 ND 136 20.9 - 160

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1205139  
**Project:** Reno Creek Project

**Date:** 6/5/2012  
**Report ID:** S1205139001

**Lead 210 and Polonium 210**

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 83407	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1205140-001FDUP 05/25/12 11:53 Lead 210 (Dissolved) 2 1 2 5.02 20

Sample ID	RunNo: 83408	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1205140-001FDUP 05/24/12 15:35 Polonium 210 (Dissolved) ND 1 ND 20

Sample ID	RunNo: 83415	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1205136-001FDUP 05/29/12 15:32 Lead 210 (Suspended) 1 1 1 1.65 20

Sample ID	RunNo: 83416	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1205136-001FDUP 05/29/12 12:51 Polonium 210 (Suspended) ND 1 ND 20

**Radium 228**

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 83255	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB12-135 05/18/12 20:13 Radium 228 (Dissolved) ND 1

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 83255	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

LCS12-135 05/18/12 16:42 Radium 228 (Dissolved) 9 1 9.52 95.6 74.9 - 122

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 83255	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

S1205140-001FMS 05/18/12 16:42 Radium 228 (Dissolved) 11 1 9.52 1 108 48.4 - 167

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 83255	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1205141-001FDUP 05/18/12 16:42 Radium 228 (Dissolved) 4 1 4 1.76 20

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits



**ANALYTICAL QC SUMMARY REPORT**

**CLIENT:** AUC LLC  
**Work Order:** S1205139  
**Project:** Reno Creek Project

**Date:** 6/5/2012  
**Report ID:** S1205139001

**Radium 226 by SM 7500**Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 82998	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB12-135	05/21/12 10:02	Radium 226 (Dissolved)	ND	0.2					

Sample ID	RunNo: 83261	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DMB12-137	05/21/12 15:45	Radium 226 (Dissolved)	ND	0.2					
		Radium 226 (Suspended)	ND	0.2					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 82998	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS12-135	05/21/12 10:02	Radium 226 (Dissolved)	5.5	0.2	5.06		108	62.5 - 125	

Sample ID	RunNo: 83261	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DLCS12-137	05/21/12 15:45	Radium 226 (Dissolved)	5.1	0.2	5.06		101	62.5 - 125	
		Radium 226 (Suspended)	5.1	0.2	5.06		101	56.4 - 132	

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 82998	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
TAP WATER MS	05/21/12 10:02	Radium 226 (Dissolved)	5.6	0.2	5.06	0.4	104	41.6 - 146	

Sample ID	RunNo: 83261	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1205139-001FMS	05/21/12 15:45	Radium 226 (Suspended)	5.6	0.2	5.06	0.3	104	51.4 - 135	

Sample Type **MSD** Units: pCi/L

Sample ID	RunNo: 82998	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
TAP WATER MSD	05/21/12 10:02	Radium 226 (Dissolved)	5.5	0.2	5.6	2.86	101	20	

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 83261	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205137-001FDUP	05/21/12 15:45	Radium 226 (Suspended)	ND	0.2	ND			20	

**Qualifiers:**

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		

**ANALYTICAL QC SUMMARY REPORT**

**CLIENT:** AUC LLC  
**Work Order:** S1205139  
**Project:** Reno Creek Project

**Date:** 6/5/2012  
**Report ID:** S1205139001

**Thorium 230 in Water - Dissolved**Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 82986	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB12-131	05/17/12 9:14	Thorium 230 (Dissolved)	ND	0.2					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 82986	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS12-131	05/17/12 9:14	Thorium 230 (Dissolved)	13	0.2	12		105	67.1 - 134	

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 82986	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205142-001F DUP	05/17/12 13:19	Thorium 230 (Dissolved)	ND	0.2	ND			20	

**Thorium 230 in Water - Suspended**Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 83113	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB12-135	05/22/12 9:10	Thorium 230 (Suspended)	ND	0.2					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 83113	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS12-135	05/21/12 9:10	Thorium 230 (Suspended)	15	0.2	12		117	67.1 - 134	

Sample Type **LCS D** Units: pCi/L

Sample ID	RunNo: 83113	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
MDL-2	05/21/12 0:00	Thorium 230 (Suspended)	0.5	0.2	ND	ND	104	30	

**Radon 222 in Water - Total**Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 84003	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB-R84003	05/09/12 0:00	Radon-222	ND	50					L

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 84003	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS-R84003	05/09/12 0:00	Radon-222	21200	100	22500		94.2	75 - 125	L

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 84003	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205139-001A	05/10/12 0:00	Radon-222	ND	50	ND			20	L

**Qualifiers:**

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1205139  
**Project:** Reno Creek Project

**Date:** 6/5/2012  
**Report ID:** S1205139001

**Silica as SiO2**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 82660	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MBLK DISS/CAT	05/10/12 11:39	Silica as SiO2	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 82660	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DISS LCS Q	05/10/12 11:42	Silica as SiO2	2.2	0.1	2.14		101	85 - 115	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 82660	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205139-001DD	05/10/12 16:26	Silica as SiO2	9.4	0.1	9.4	0.513		20	

**Solids By SM 2540**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 82727	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DI	05/09/12 9:00	Total Dissolved Solids (180)	ND	10					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 82727	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
CONTROL	05/09/12 9:00	Total Dissolved Solids (180)	220	10	226		95.6	90 - 110	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 82727	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205094-002B	05/07/12 9:00	Total Dissolved Solids (180)	1180	10	1170	1.37		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1205139  
**Project:** Reno Creek Project

**Date:** 6/5/2012  
**Report ID:** S1205139001

**Total (200.2) Metals by EPA 200.7 ICP - Water**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 82724	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB-6052	05/11/12 17:37	Iron	ND	0.05					
		Manganese	ND	0.02					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 82724	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS-6052	05/11/12 17:39	Iron	0.49	0.05	0.5		98.6	85 - 115	
		Manganese	0.20	0.02	0.2		101	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 82724	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1205137-001ES	05/11/12 17:49	Iron	0.55	0.05	0.5	0.06	97.7	70 - 130	
		Manganese	0.92	0.02	0.2	0.71	107	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 82724	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1205137-001EMSD	05/11/12 17:51	Iron	0.55	0.05	0.55	0.0181		20	
		Manganese	0.92	0.02	0.92	0.108		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 82724	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205136-001ED	05/11/12 17:44	Iron	6.07	0.05	6.11	0.630		20	
		Manganese	0.47	0.02	0.48	0.652		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits



### ANALYTICAL QC SUMMARY REPORT

**CLIENT:** AUC LLC  
**Work Order:** S1205139  
**Project:** Reno Creek Project

**Date:** 6/5/2012  
**Report ID:** S1205139001

**Suspended Natural Uranium**

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 82698	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLK	05/11/12 13:14	Uranium Suspended	ND	0.0003					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 82698	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS	05/11/12 13:09	Uranium Suspended	0.103	0.0003	0.1		103	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 82698	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1205136-001FS	05/11/12 20:27	Uranium Suspended	1.11	0.0003	1.1	ND	101	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 82698	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1205136-001FMSD	05/11/12 20:32	Uranium Suspended	1.13	0.0003	1.11	1.82		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 82698	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205136-001FD	05/11/12 20:22	Uranium Suspended	ND	0.0003	ND			20	

**pH Water**

Sample Type **LCS** Units: s.u.

Sample ID	RunNo: 82721	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	05/11/12 11:19	pH	9.1	0.1	8.69		104	90 - 110	

Sample Type **DUP** Units: s.u.

Sample ID	RunNo: 82721	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1205135-008AD	05/11/12 16:18	pH	8.0	0.1	8.0	0.125		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
  - H Holding times for preparation or analysis exceeded
  - L Analyzed by a contract laboratory
  - O Outside the Range of Dilutions
  - S Spike Recovery outside accepted recovery limits
  - E Value above quantitation range
  - J Analyte detected below quantitation limits
  - ND Not Detected at the Reporting Limit
  - R RPD outside accepted recovery limits



**Inter-Mountain Labs**  
 Sheridan, WY and Gillette, WY  
 1673 Terra Ave.  
 Sheridan, WY 82801  
 (307) 672-8945

**- CHAIN OF CUSTODY RECORD -**

This is a **LEGAL DOCUMENT**. All shaded fields must be completed. See reverse for instructions. #RC 08471

Client Name		Project Identification		Sampler (Signature/Printed)		Telephone #	
AUC LLC		Reno Creek Project		<i>Juan Alvarez</i>			
Report Address		Contact Name		ANALYSES / PARAMETERS		REMARKS	
trickerson@treccorp.com; jshriver@treccorp.com		Tara Dickerson		See Attached Table 1 (4, 14, Guide 8, Radon 222)			
Invoice Address AUC LLC		Email trickerson@treccorp.com		See Attached Table 2 (Air Particulate)			
900 W. Hwy 82, 150		Phone 307-265-0196		See Attached Table 3 (Vegetation)			
Casper, WY 82401		Purchase Order #		See Attached Table 4 (Sediment)			
LAB ID	DATE	TIME	SAMPLE IDENTIFICATION	Matrix	# of Containers		
(Lab Use Only)	SAMPLED						
1							
2	5/12/05	13:00	SM7-004-120508 WT		16	X	Groundwater
3							
4							
5							
6							
7							
8							
9							
10							
11							
12							
13							
14							

  

LAB COMMENTS	Relinquished By (Signature/Printed)	DATE	TIME	Received By (Signature/Printed)	DATE	TIME
4.0°C	<i>Juan Alvarez</i>	5/8/05	10:00	<i>Tara Dickerson</i>	5/8/05	15:17
				<i>Alan J. Miller</i>	5/8/05	15:47
				<i>Kathy Bob</i>	5/9/05	9:36

  

SHIPPING INFO	MATRIX CODFS	TURNAROUND TIMES	COMPLIANCE INFORMATION	ADDITIONAL REMARKS
<input type="checkbox"/> UPS <input type="checkbox"/> Fed Express <input type="checkbox"/> US Mail <input type="checkbox"/> Hand Carried <input checked="" type="checkbox"/> Other	water <input checked="" type="checkbox"/> WT Soil <input type="checkbox"/> SL Solid <input type="checkbox"/> SD Trip Blank <input type="checkbox"/> TB Other <input type="checkbox"/> OT	<input checked="" type="checkbox"/> Check desired service <input type="checkbox"/> Standard turnaround <input type="checkbox"/> RUSH - 5 Working Days <input type="checkbox"/> URGENT - < 2 Working Days Rush & Urgent Surcharges will be applied	Compliance Monitoring? <input checked="" type="checkbox"/> Y <input type="checkbox"/> N Program (SDWA, NPDES,...) PWSID / Permit # Chlorinated? <input type="checkbox"/> Y <input checked="" type="checkbox"/> N Sample Disposal: Lab <input checked="" type="checkbox"/> Client	3 Radon logs to Jim Swider