



Thursday, January 12, 2012

Phil Cavendor
AUC, LLC
1536 Cole Blvd
Suite 330
Lakewood, CO 80401

RE: Reno Creek Project
Orde

r No.: S1111459

Dear Phil Cavendor:

Inter-Mountain Laboratories received sample UM3R-001-111129 on November 29, 2011.

There were no problems with the analyses and all data for associated QC met EPA or laboratory specifications except where noted in the Case Narrative.

If you have any questions regarding these tests results, please feel free to call.

Sincerely,

Lacey Ketron
Project Manager

xc: File
Encl.



Date: 1/12/2012

CLIENT: AUC LLC
Project: Reno Creek Project
Lab Order: S1111459

CASE NARRATIVE
Report ID: S1111459001

Sample UM3R-001-111129 was received on November 29, 2011.

All samples were received and analyzed within the EPA recommended holding times, except those noted in this case narrative. Samples were analyzed using the methods outlined in the following references:

- U.S.E.P.A. 600 "Methods for Chemical Analysis of Water and Wastes", 1993
- "Standard Methods For The Examination of Water and Wastewater", 20th ed., 1998
- Test Methods for Evaluating Solid Waste, Physical/Chemical Methods, SW846, 3rd Edition
- Methods indicated with the Monday, March 12, 2007 Federal Register, 40 CFR Part 122, 136 et al.

All Quality Control parameters met the acceptance criteria defined by EPA and Inter-Mountain Laboratories except as indicated in this case narrative.

For the purposes of this report, the Reporting Limit (RL) is the Practical Quantitation Limit (PQL).

Radon 222 was analyzed through ALS Laboratory Group Certification #: CO00078.

Qualifiers by sample

- MB-R78191 - General Parameters/Radon-222 - Analyzed by a contract laboratory
- LCS-R78191 - General Parameters/Radon-222 - Analyzed by a contract laboratory
- S1111459-001 - General Parameters/Radon-222 - Analyzed by a contract laboratory
- S1111463-004 - Cations/Calcium - Spike Recovery outside accepted recovery limits (2)
- S1111235-024 - Anions/Chloride - Spike Recovery outside accepted recovery limits (3)
- S1111235-024 - Anions/Sulfate - Spike Recovery outside accepted recovery limits (3)
- S1111459-001 - Anions/Alkalinity, Bicarbonate as HCO₃ - Holding times for preparation or analysis exceeded (1)
- S1111459-001 - Anions/Alkalinity, Carbonate as CO₃ - Holding times for preparation or analysis exceeded (1)
- S1111459-001 - General Parameters/Alkalinity, Total (As CaCO₃) - Holding times for preparation or analysis exceeded (1)

- (1) Alkalinity was received and analyzed within holding. Reanalyzed for balance purposes at a later date.
- (2) Spike recovery outside acceptance criteria due to presence of analyte in the sample that greatly exceeded level of spike used. Data approved because all other QC were within acceptance limits.
- (3) Spike failed due to possible matrix interference from the sample.

Reviewed by:

Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
1536 Cole Blvd
Lakewood, CO 80401
ProjectName: Reno Creek Project
Lab ID: S1111459-001
ClientSample ID: UM3R-001-111129
COC: RC 08148

Date Reported 1/12/2012
Report ID S1111459001
WorkOrder: S1111459
CollectionDate: 11/29/2011 9:45:00 AM
DateReceived: 11/29/2011 4:39:00 PM
FieldSampler: TN
Matrix: Water

Comments

Table with 7 columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init

Field data table with columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init

Anions/Cations

Table of anions and cations with columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init

General Parameters

Table of general parameters with columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init

These results apply only to the samples tested.

RL - Reporting Limit

- Qualifiers: * Value exceeds Maximum Contaminant Level
C Calculated Value
H Holding times for preparation or analysis exceeded
L Analyzed by a contract laboratory
ND Not Detected at the Reporting Limit
S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
E Value above quantitation range
J Analyte detected below quantitation limits
M Value exceeds Monthly Ave or MCL
O Outside the Range of Dilutions

Reviewed by: Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
1536 Cole Blvd
Lakewood, CO 80401
ProjectName: Reno Creek Project
Lab ID: S1111459-001
ClientSample ID: UM3R-001-111129
COC: RC 08148

Date Reported: 1/12/2012
Report ID: S1111459001
WorkOrder: S1111459
CollectionDate: 11/29/2011 9:45:00 AM
DateReceived: 11/29/2011 4:39:00 PM
FieldSampler: TN
Matrix: Water

Comments

Table with 7 columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init. Rows include Data Quality (Cation Sum, Anion Sum, etc.), Metals - Dissolved (Aluminum, Arsenic, etc.), Metals - Suspended (Uranium), and Metals - Total (Iron, Manganese).

These results apply only to the samples tested.

RL - Reporting Limit

- Qualifiers: * Value exceeds Maximum Contaminant Level
C Calculated Value
H Holding times for preparation or analysis exceeded
L Analyzed by a contract laboratory
ND Not Detected at the Reporting Limit
S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
E Value above quantitation range
J Analyte detected below quantitation limits
M Value exceeds Monthly Ave or MCL
O Outside the Range of Dilutions

Reviewed by: [Signature]
Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
1536 Cole Blvd
Lakewood, CO 80401
ProjectName: Reno Creek Project
Lab ID: S1111459-001
ClientSample ID: UM3R-001-111129
COC: RC 08148

Date Reported 1/12/2012
Report ID S1111459001
WorkOrder: S1111459
CollectionDate: 11/29/2011 9:45:00 AM
DateReceived: 11/29/2011 4:39:00 PM
FieldSampler: TN
Matrix: Water

Comments

Table with 7 columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init. Rows include Radionuclides - Dissolved (Gross Alpha, Lead 210, Polonium 210, Radium 226, Thorium 230) and Radionuclides - Suspended (Lead 210, Polonium 210, Radium 226, Thorium 230). Total radionuclides section includes Radon 222.

These results apply only to the samples tested.

RL - Reporting Limit

- Qualifiers: * Value exceeds Maximum Contaminant Level
C Calculated Value
H Holding times for preparation or analysis exceeded
L Analyzed by a contract laboratory
ND Not Detected at the Reporting Limit
S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
E Value above quantitation range
J Analyte detected below quantitation limits
M Value exceeds Monthly Ave or MCL
O Outside the Range of Dilutions

Reviewed by: [Signature]
Lacey Ketron, Water Lab Supervisor



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1111459
Project: Reno Creek Project

Date: 1/12/2012
Report ID: S1111459001

Alkalinity

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 78460	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	12/30/11 10:50	Alkalinity, Total (As CaCO3)	ND	5					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 78460	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	12/30/11 10:39	Alkalinity, Total (As CaCO3)	616	5	601		102	90 - 110	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 78460	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1112402-005AD	12/30/11 13:51	Alkalinity, Bicarbonate as HCO3	414	5	413	0.416		20	
		Alkalinity, Carbonate as CO3	ND	5	ND			20	
		Alkalinity, Total (As CaCO3)	340	5	338	0.416		20	

Conductivity

Sample Type **MBLK** Units: µmhos/cm

Sample ID	RunNo: 77558	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	12/01/11 16:25	Electrical Conductivity	ND	5					

Sample Type **LCS** Units: µmhos/cm

Sample ID	RunNo: 77558	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	12/01/11 16:11	Electrical Conductivity	1090	5	1120		97.0	90 - 110	

Sample Type **DUP** Units: µmhos/cm

Sample ID	RunNo: 77558	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111483-002AD	12/01/11 20:44	Electrical Conductivity	2230	5	2250	0.889		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1111459
Project: Reno Creek Project

Date: 1/12/2012
Report ID: S1111459001

Dissolved Mercury by EPA 245.1 - Water

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 77659	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LRB	12/06/11 8:04	Mercury	ND	0.001					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 77659	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS	12/06/11 8:02	Mercury	0.002	0.001	0.002		100	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 77659	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111459-001D	12/06/11 9:28	Mercury	0.002	0.001	0.00244	ND	95.4	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 77659	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1111459-001D	12/06/11 9:30	Mercury	0.002	0.001	0.002	3.18		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 77659	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111459-001D	12/06/11 9:26	Mercury	ND	0.001	ND			20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1111459
Project: Reno Creek Project

Date: 1/12/2012
Report ID: S1111459001

Dissolved Metals by ICP - EPA 200.7 - WaterSample Type **MBLK** Units: mg/L

Sample ID	RunNo: 77535	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MBLK DISS/CAT	12/01/11 11:45	Aluminum	ND	0.1					
		Boron	ND	0.1					
		Chromium	ND	0.01					
		Iron	ND	0.05					
		Manganese	ND	0.02					
		Nickel	ND	0.01					
		Zinc	ND	0.01					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 77535	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DISS LCS Q	12/01/11 11:48	Aluminum	1.0	0.1	1		101	85 - 115	
		Boron	1.0	0.1	1		100	85 - 115	
		Chromium	1.01	0.01	1		101	85 - 115	
		Iron	1.01	0.05	1		101	85 - 115	
		Manganese	1.04	0.02	1		104	85 - 115	
		Nickel	1.04	0.01	1		104	85 - 115	
		Zinc	1.02	0.01	1		102	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 77535	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111485-001DS	12/01/11 14:48	Aluminum	0.5	0.1	0.5	ND	91.2	70 - 130	
		Boron	1.0	0.1	0.5	0.7	73.5	70 - 130	
		Chromium	0.46	0.01	0.5	ND	92.5	70 - 130	
		Iron	0.94	0.05	0.5	0.58	73.7	70 - 130	
		Manganese	0.51	0.02	0.5	0.07	88.1	70 - 130	
		Nickel	0.45	0.01	0.5	ND	89.5	70 - 130	
		Zinc	0.47	0.01	0.5	ND	92.9	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 77535	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1111485-001DSD	12/01/11 14:51	Aluminum	0.5	0.1	0.5	3.09		20	
		Boron	1.1	0.1	1.0	2.91		20	
		Chromium	0.48	0.01	0.46	2.70		20	
		Iron	0.96	0.05	0.94	1.95		20	
		Manganese	0.52	0.02	0.51	2.48		20	
		Nickel	0.46	0.01	0.45	2.93		20	
		Zinc	0.48	0.01	0.47	2.73		20	

Qualifiers:

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1111459
Project: Reno Creek Project

Date: 1/12/2012
Report ID: S1111459001

Dissolved Metals by ICP - EPA 200.7 - Water

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 77535	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111485-007DD	12/01/11 15:14	Aluminum	ND	0.1	ND			20	
		Boron	0.9	0.1	0.8	7.24		20	
		Chromium	ND	0.01	ND			20	
		Iron	ND	0.05	ND			20	
		Manganese	ND	0.02	ND			20	
		Nickel	ND	0.01	ND			20	
		Zinc	ND	0.01	ND			20	

Dissolved Metals by ICPMS EPA 200.8 - Water

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 77467	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLK	11/30/11 16:09	Arsenic	ND	0.005					
		Barium	ND	0.5					
		Cadmium	ND	0.002					
		Copper	ND	0.01					
		Lead	ND	0.02					
		Molybdenum	ND	0.02					
		Selenium	ND	0.005					
		Uranium	ND	0.001					
		Vanadium	ND	0.02					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 77467	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS	11/30/11 16:05	Arsenic	0.100	0.005	0.1		100	85 - 115	
		Barium	ND	0.5	0.1		106	85 - 115	
		Cadmium	0.101	0.002	0.1		101	85 - 115	
		Copper	0.10	0.01	0.1		98.1	85 - 115	
		Lead	0.10	0.02	0.1		97.6	85 - 115	
		Molybdenum	0.10	0.02	0.1		99.0	85 - 115	
		Selenium	0.101	0.005	0.1		101	85 - 115	
		Uranium	0.096	0.001	0.1		95.5	85 - 115	
		Vanadium	0.10	0.02	0.1		95.8	85 - 115	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1111459
Project: Reno Creek Project

Date: 1/12/2012
Report ID: S1111459001

Dissolved Metals by ICPMS EPA 200.8 - Water

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 77467	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111485-002DS	11/30/11 18:36	Arsenic	1.12	0.005	1.1	ND	102	70 - 130	
		Barium	1.2	0.5	1.1	ND	101	70 - 130	
		Cadmium	1.11	0.002	1.1	ND	100	70 - 130	
		Copper	1.06	0.01	1.1	ND	96.6	70 - 130	
		Lead	1.08	0.02	1.1	ND	98.6	70 - 130	
		Molybdenum	1.17	0.02	1.1	ND	106	70 - 130	
		Selenium	1.11	0.005	1.1	ND	100	70 - 130	
		Uranium	1.04	0.001	1.1	ND	94.3	70 - 130	
		Vanadium	1.08	0.02	1.1	ND	97.9	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 77467	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1111485-002DMSD	11/30/11 18:40	Arsenic	1.14	0.005	1.12	1.20		20	
		Barium	1.2	0.5	1.2	0.102		20	
		Cadmium	1.11	0.002	1.11	0.257		20	
		Copper	1.07	0.01	1.06	0.493		20	
		Lead	1.09	0.02	1.08	0.0624		20	
		Molybdenum	1.17	0.02	1.17	0.248		20	
		Selenium	1.11	0.005	1.11	0.577		20	
		Uranium	1.04	0.001	1.04	0.377		20	
		Vanadium	1.09	0.02	1.08	1.23		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 77467	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111485-002DD	11/30/11 18:32	Arsenic	ND	0.005	ND			20	
		Barium	ND	0.5	ND			20	
		Cadmium	ND	0.002	ND			20	
		Copper	ND	0.01	ND			20	
		Lead	ND	0.02	ND			20	
		Molybdenum	ND	0.02	ND			20	
		Selenium	ND	0.005	ND			20	
		Uranium	ND	0.001	ND			20	
		Vanadium	ND	0.02	ND			20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1111459
Project: Reno Creek Project

Date: 1/12/2012
Report ID: S1111459001

Fluoride by SM 4500

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 77558	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	12/01/11 16:25	Fluoride	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 77558	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	12/01/11 16:15	Fluoride	3.1	0.1	2.96		104	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 78460	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1112443-001ASPK	12/31/11 1:06	Fluoride	6.5	0.1	2.5	4.0	101	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 77558	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111485-003AD	12/01/11 22:30	Fluoride	5.6	0.1	5.5	2.39		20	

Gross Alpha, Beta by SM 7110B

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 78253	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-350	12/18/11 19:56	Gross Alpha (Dissolved)	ND	2					
		Gross Beta (Dissolved)	ND	3					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 78253	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-350	12/18/11 19:56	Gross Alpha (Dissolved)	83	2	73.6		113	83 - 133	
		Gross Beta (Dissolved)	139	3	126		110	88 - 115	

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 78253	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111397-001EMS	12/18/11 19:56	Gross Alpha (Dissolved)	73	2	73.6	ND	99.7	48 - 153	
		Gross Beta (Dissolved)	135	3	126	ND	107	60 - 140	

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 78253	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1112147-001EDUP	12/21/11 20:42	Gross Alpha (Dissolved)	31	3	28	11.8		20	
		Gross Beta (Dissolved)	14	4	15	4.98		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1111459
Project: Reno Creek Project

Date: 1/12/2012
Report ID: S1111459001

Anions by ION ChromatographySample Type **MBLK** Units: mg/L

Sample ID	RunNo: 77583	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLANK	12/02/11 8:46	Chloride	ND	1					
		Sulfate	ND	1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 77583	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DIONEX	12/02/11 10:18	Chloride	32	1	30		107	90 - 110	
		Sulfate	139	1	150		92.9	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 77583	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111235-024ASPK	12/02/11 12:31	Chloride	40	1	55	ND	73.1	80 - 120	S
		Sulfate	709	1	440	379	74.9	80 - 120	S

Cations by ICP (Method 200.7)Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 78413	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MBLK DISS/CAT	12/29/11 12:47	Calcium	ND	1					
		Magnesium	ND	1					
		Potassium	ND	1					
		Sodium	ND	1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 78413	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
CAT LCS IML3	12/29/11 12:51	Calcium	41	1	40		103	85 - 115	
		Magnesium	40	1	40		101	85 - 115	
		Potassium	41	1	40		102	85 - 115	
		Sodium	40	1	40		99.4	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 77533	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111463-004AS	12/01/11 13:23	Calcium	603	1	50	646	-85.6	70 - 130	S
		Magnesium	145	1	50	105	80.1	70 - 130	
		Potassium	70	1	50	13	115	70 - 130	
		Sodium	98	1	50	42	112	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 77533	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1111463-004ASD	12/01/11 13:25	Calcium	598	1	603	0.737		20	
		Magnesium	144	1	145	0.990		20	
		Potassium	71	1	70	1.17		20	
		Sodium	98	1	98	0.551		20	

Qualifiers:

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1111459
Project: Reno Creek Project

Date: 1/12/2012
Report ID: S1111459001

Cations by ICP (Method 200.7)

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 77533	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111463-003AD	12/01/11 13:18	Calcium	490	1	483	1.43		20	
		Magnesium	75	1	74	2.40		20	
		Potassium	5	1	5	1.99		20	
		Sodium	41	1	41	1.34		20	

Nitrogen, Ammonia (as N)

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 77623	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
CCB	12/05/11 14:31	Nitrogen, Ammonia (As N)	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 77623	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
QC	12/05/11 13:44	Nitrogen, Ammonia (As N)	10.1	0.1	10		101	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 77623	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111485-001C	12/05/11 14:59	Nitrogen, Ammonia (As N)	11.6	0.1	10	1.2	104	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 77623	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111485-001C	12/05/11 14:58	Nitrogen, Ammonia (As N)	1.1	0.1	1.2	2.96		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1111459
Project: Reno Creek Project

Date: 1/12/2012
Report ID: S1111459001

Nitrogen, Nitrate-Nitrite (as N)

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 77456	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
BLK	11/30/11 15:18	Nitrogen, Nitrate-Nitrite (as N)	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 77456	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
QC	11/30/11 15:15	Nitrogen, Nitrate-Nitrite (as N)	2.4	0.1	2.31		104	90 - 110	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 77456	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111463-004B	11/30/11 15:46	Nitrogen, Nitrate-Nitrite (as N)	9.52	0.05	10	ND	95.2	80 - 120	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 77456	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111463-004B	11/30/11 15:45	Nitrogen, Nitrate-Nitrite (as N)	ND	0.05	ND			20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1111459
Project: Reno Creek Project

Date: 1/12/2012
Report ID: S1111459001

Lead 210 and Polonium 210Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 78153	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB11-341	12/16/11 12:52	Polonium 210 (Dissolved)	ND	1					
----------	----------------	--------------------------	----	---	--	--	--	--	--

Sample ID	RunNo: 78153	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB11-341	12/16/11 12:52	Polonium 210 (Suspended)	ND	1					
----------	----------------	--------------------------	----	---	--	--	--	--	--

Sample ID	RunNo: 78154	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB11-341	12/19/11 15:38	Lead 210 (Dissolved)	ND	1					
----------	----------------	----------------------	----	---	--	--	--	--	--

Sample ID	RunNo: 78154	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB11-341	12/19/11 15:38	Lead 210 (Suspended)	ND	1					
----------	----------------	----------------------	----	---	--	--	--	--	--

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 78153	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

LCS11-341 PO	12/16/11 12:52	Polonium 210 (Dissolved)	4	1	3.48		115	35.1 - 175	
		Polonium 210 (Suspended)	4	1	3.48		115	48.6 - 156	

Sample ID	RunNo: 78154	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

LCS11-341	12/19/11 15:38	Lead 210 (Dissolved)	14	1	12.8		106	67.6 - 131	
		Lead 210 (Suspended)	14	1	12.8		106	60.3 - 141	

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 78153	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

TAP WATER MS	12/16/11 12:52	Polonium 210 (Dissolved)	13	1	12.8	ND	97.7	20.9 - 160	
		Polonium 210 (Suspended)	13	1	12.8	ND	97.7	28 - 173	

Sample ID	RunNo: 78154	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

TAP WATER MS	12/19/11 15:38	Lead 210 (Dissolved)	13	1	12.8	ND	103	66 - 133	
		Lead 210 (Suspended)	13	1	12.8	ND	103	62.4 - 132	

Sample Type **MSD** Units: pCi/L

Sample ID	RunNo: 78153	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	------	------	------	--------------	------

TAP WATER MSD	12/16/11 12:52	Polonium 210 (Dissolved)	13	1	13	4.60		30	
		Polonium 210 (Suspended)	13	1	13	4.60		30	

Sample ID	RunNo: 78154	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	------	------	------	--------------	------

TAP WATER MSD	12/19/11 15:38	Lead 210 (Dissolved)	14	1	13	4.42		30	
		Lead 210 (Suspended)	14	1	13	4.42		30	

Qualifiers:

B	Analyte detected in the associated Method Blank	E	Value above quantitation range
H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits
L	Analyzed by a contract laboratory	ND	Not Detected at the Reporting Limit
O	Outside the Range of Dilutions	R	RPD outside accepted recovery limits
S	Spike Recovery outside accepted recovery limits		



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1111459
Project: Reno Creek Project

Date: 1/12/2012
Report ID: S1111459001

Lead 210 and Polonium 210

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 78153	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1111459-001FDUP 12/16/11 12:52 Polonium 210 (Dissolved) ND 1 ND 20

Sample ID	RunNo: 78154	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1111459-001FDUP 12/19/11 15:38 Lead 210 (Dissolved) ND 1 ND 20

Radium 228

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 78628	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

MB11-361 01/04/12 17:51 Radium 228 (Dissolved) ND 1

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 78628	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

LCS11-361 01/04/12 17:51 Radium 228 (Dissolved) 9 1 9.96 92.7 74.9 - 122

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 78628	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
-----------	--------------	---------	--------	----	-------	----------	------	--------------	------

S1111448-001FMS 01/04/12 17:51 Radium 228 (Dissolved) 11 1 9.96 ND 107 48.4 - 167

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 78628	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
-----------	--------------	---------	--------	----	----------	------	------	--------------	------

S1112147-002EDUP 01/04/12 20:07 Radium 228 (Dissolved) ND 1 ND 20

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1111459
Project: Reno Creek Project

Date: 1/12/2012
Report ID: S1111459001

Radium 226 by SM 7500

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 78093	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-346	12/15/11 17:28	Radium 226 (Dissolved)	ND	0.2					

Sample ID	RunNo: 78142	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-346	12/15/11 14:52	Radium 226 (Suspended)	ND	0.2					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 78093	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-346	12/15/11 17:28	Radium 226 (Dissolved)	9.6	0.2	10.1		94.6	62.5 - 125	

Sample ID	RunNo: 78142	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-346	12/15/11 14:52	Radium 226 (Suspended)	10.1	0.2	10.1		99.4	56.4 - 132	

Sample Type **MS** Units: pCi/L

Sample ID	RunNo: 78093	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111449-007DMS	12/15/11 17:28	Radium 226 (Dissolved)	10.9	0.2	10.1	0.2	105	41.6 - 146	

Sample ID	RunNo: 78093	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111449-010DMS	12/16/11 9:44	Radium 226 (Dissolved)	11.3	0.2	10.1	0.7	105	41.6 - 146	

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 78093	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111449-005DDUP	12/15/11 17:28	Radium 226 (Dissolved)	0.3	0.2	0.3	9.38		20	

Sample ID	RunNo: 78093	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111449-008DDUP	12/15/11 17:28	Radium 226 (Dissolved)	ND	0.2	ND			20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1111459
Project: Reno Creek Project

Date: 1/12/2012
Report ID: S1111459001

Thorium 230 in Water

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 77876	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-341	12/13/11 9:13	Thorium 230 (Suspended)	ND	0.2					

Sample ID	RunNo: 77876	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB11-341	12/13/11 9:13	Thorium 230 (Dissolved)	ND	0.2					

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 77876	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS11-341	12/13/11 9:13	Thorium 230 (Dissolved)	14	0.2	12		109	67.1 - 134	
		Thorium 230 (Suspended)	14	0.2	12		109	69.2 - 127	

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 77876	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111459-001FDUP	12/13/11 9:13	Thorium 230 (Suspended)	ND	0.2	ND				

Radon 222 by SM7500-RN

Sample Type **MBLK** Units: pCi/L

Sample ID	RunNo: 78191	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB-R78191	12/01/11 0:00	Radon-222	ND	50					L

Sample Type **LCS** Units: pCi/L

Sample ID	RunNo: 78191	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS-R78191	12/01/11 0:00	Radon-222	20700	100	22500		92.0	75 - 125	L

Sample Type **DUP** Units: pCi/L

Sample ID	RunNo: 78191	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111459-001A	12/01/11 0:00	Radon-222	ND	50	ND			20	L

Silica as SiO2

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 77536	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MBLK DISS/CAT	12/01/11 11:45	Silica as SiO2	ND	0.1					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 77536	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DISS LCS Q	12/01/11 11:48	Silica as SiO2	2.1	0.1	2.14		98.0	85 - 115	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1111459
Project: Reno Creek Project

Date: 1/12/2012
Report ID: S1111459001

Solids By SM 2540

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 77576	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
DI	11/30/11 15:32	Total Dissolved Solids (180)	ND	10					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 77576	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
CONTROL	11/30/11 15:33	Total Dissolved Solids (180)	210	10	226		92.0	90 - 110	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 77576	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111449-005A	11/30/11 15:44	Total Dissolved Solids (180)	420	10	420	0.472		20	

Total (200.2) Metals by EPA 200.7 ICP - Water

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 77594	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
MB-5661	12/02/11 17:14	Iron	ND	0.05					
		Manganese	ND	0.02					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 77594	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS-5661	12/02/11 17:23	Iron	0.48	0.05	0.5		95.9	85 - 115	
		Manganese	0.19	0.02	0.2		93.9	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 77594	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111487-001CS	12/02/11 17:33	Iron	0.60	0.05	0.5	0.06	109	70 - 130	
		Manganese	0.24	0.02	0.2	ND	111	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 77594	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1111487-001CMSD	12/02/11 17:35	Iron	0.60	0.05	0.60	0.681		20	
		Manganese	0.24	0.02	0.24	0.508		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 77594	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111459-001ED	12/02/11 17:28	Iron	ND	0.05	ND			20	
		Manganese	ND	0.02	ND			20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1111459
Project: Reno Creek Project

Date: 1/12/2012
Report ID: S1111459001

Suspended Natural Uranium

Sample Type **MBLK** Units: mg/L

Sample ID	RunNo: 77686	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ICB	12/06/11 15:05	Uranium Suspended	ND	0.0003					

Sample Type **LCS** Units: mg/L

Sample ID	RunNo: 77686	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
LCS	12/06/11 17:27	Uranium Suspended	0.109	0.0003	0.1		109	85 - 115	

Sample Type **MS** Units: mg/L

Sample ID	RunNo: 77686	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
S1111459-001FS	12/06/11 17:40	Uranium Suspended	1.08	0.0003	1.1	ND	98.5	70 - 130	

Sample Type **MSD** Units: mg/L

Sample ID	RunNo: 77686	Analyte	Result	RL	Conc	%RPD	%REC	% RPD Limits	Qual
S1111459-001FMSD	12/06/11 17:53	Uranium Suspended	1.22	0.0003	1.08	12.3		20	

Sample Type **DUP** Units: mg/L

Sample ID	RunNo: 77686	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111459-001FD	12/06/11 17:31	Uranium Suspended	0.0003	0.0003	ND			20	

pH Water

Sample Type **LCS** Units: s.u.

Sample ID	RunNo: 77558	Analyte	Result	RL	Spike	Ref Samp	%REC	% Rec Limits	Qual
ATQC	12/01/11 16:11	pH	8.8	0.1	8.48		104	90 - 110	

Sample Type **DUP** Units: s.u.

Sample ID	RunNo: 77558	Analyte	Result	RL	Ref Samp	%RPD	%REC	% RPD Limits	Qual
S1111462-001AD	12/01/11 19:02	pH	7.4	0.1	7.1	4.89		20	

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



Inter-Mountain Labs
 Sheridan, WY and Gillette, WY
 1673 Terra Ave. (307) 682-8845
 Sheridan, WY 82801 (307) 672-8845

- CHAIN OF CUSTODY RECORD -

This is a **LEGAL DOCUMENT**. All shaded fields must be completed. See reverse for instructions. #RC 08148

Client Name AUC LLC		Project Identification Reno Creek Project		Sampler (Signature/Printed) <i>Jeani Wickerson</i>		Telephone #	
Report Address pcavendor@auc-llc.com; tnickerson@treccorp.com; jshriver@treccorp.com		Contact Name <i>Trec: Wickerson</i>		Analyses / Parameters See Attached Table 1 (4,14, Guide 8, Radon 222)		REMARKS	
Invoice Address AUC LLC 900 Wrenn Ct. Ste 100 Casper WY 82401		Purchase Order #		See Attached Table 2 (Air Particulate)		See Attached Table 3 (Vegetation)	
DATE SAMPLED		SAMPLE IDENTIFICATION		See Attached Table 4 (Sediment)		See Attached Table 5 (Soil)	
ITEM	LAB ID (Lab Use Only)	DATE	TIME	MATRIX	# of Containers	See Attached Table 8 WYDES Discharge	See Attached Table 9 (Blanks)
1							
2	5111457-01	11/29/14	9:45	W	13		
3							
4							
5							
6							
7							
8							
9							
10							
11							
12							
13							
14							

LAB COMMENTS 7.0°C		Relinquished By (Signature/Printed) <i>Jeani Wickerson</i>		DATE 11/29/14		TIME 15:30			
RECEIVED BY (Signature/Printed) <i>Kathy Boop</i>		DATE 11/30/14		TIME 10:23		ADDITIONAL REMARKS			
SHIPPING INFO <input type="checkbox"/> UPS <input type="checkbox"/> Fed Express <input type="checkbox"/> US Mail <input type="checkbox"/> Hand Carried <input checked="" type="checkbox"/> Other		MATRIX CODES Water <input checked="" type="checkbox"/> WT Soil <input type="checkbox"/> SL Solid <input type="checkbox"/> SD Trip Blank <input type="checkbox"/> TB Other <input type="checkbox"/> OT		TURNAROUND TIMES <input checked="" type="checkbox"/> Check desired service <input type="checkbox"/> Standard turnaround <input type="checkbox"/> RUSH - 5 Working Days <input type="checkbox"/> URGENT - < 2 Working Days		COMPLIANCE INFORMATION Compliance Monitoring? <input checked="" type="checkbox"/> Y <input type="checkbox"/> N Program (SDWA, NPDES,...) PWSID / Permit # Chlorinated? <input checked="" type="checkbox"/> Y <input type="checkbox"/> N		Sample Disposal: Lab <input checked="" type="checkbox"/> Client	
Rush & Urgent Surcharges will be applied									