



Tuesday, April 10, 2012

Phil Cavendor
AUC, LLC
1536 Cole Blvd
Suite 330
Lakewood, CO 80401

RE: Reno Creek Project
Orde

r No.: S1203095

Dear Phil Cavendor:

Inter-Mountain Laboratories received sample UM3R-003-120307 on March 7, 2012.

There were no problems with the analyses and all data for associated QC met EPA or laboratory specifications except where noted in the Case Narrative.

If you have any questions regarding these tests results, please feel free to call.

Sincerely,

Lacey Ketron
Project Manager

xc: File
Encl.



Date: 4/10/2012

CLIENT: AUC LLC
Project: Reno Creek Project
Lab Order: S1203095

CASE NARRATIVE
Report ID: S1203095001

Sample UM3R-003-120307 was received on March 7, 2012.

All samples were received and analyzed within the EPA recommended holding times, except those noted in this case narrative. Samples were analyzed using the methods outlined in the following references:

U.S.E.P.A. 600 "Methods for Chemical Analysis of Water and Wastes", 1993
"Standard Methods For The Examination of Water and Wastewater", 20th ed., 1998
Test Methods for Evaluating Solid Waste, Physical/Chemical Methods, SW846, 3rd Edition
Methods indicated with the Monday, March 12, 2007 Federal Register, 40 CFR Part 122, 136 et al.

All Quality Control parameters met the acceptance criteria defined by EPA and Inter-Mountain Laboratories except as indicated in this case narrative.

For the purposes of this report, the Reporting Limit (RL) is the Practical Quantitation Limit (PQL).

Radon 222 was analyzed through ALS Laboratory Group Certification #: CO00078.

Qualifiers by sample

S1203095-001 - General Parameters/Radon-222 - Analyzed by a contract laboratory
LCS-R81054 - General Parameters/Radon-222 - Analyzed by a contract laboratory
MB-R81054 - General Parameters/Radon-222 - Analyzed by a contract laboratory
S1203094-003 - Anions/Chloride - Spike Recovery outside accepted recovery limits (1)

(1) Spike failed due to possible matrix interference from the sample.

Reviewed by:

Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
1536 Cole Blvd
Lakewood, CO 80401
ProjectName: Reno Creek Project
Lab ID: S1203095-001
ClientSample ID: UM3R-003-120307
COC: RC 08268

Date Reported: 4/10/2012
Report ID: S1203095001
WorkOrder: S1203095
CollectionDate: 3/7/2012 9:30:00 AM
DateReceived: 3/7/2012 3:52:00 PM
FieldSampler: TN
Matrix: Water

Comments

Table with 7 columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init. Rows include pH Reading, Conductivity Reading, Turbidity Reading, Dissolved Oxygen Reading, Temperature Reading, and Oxygen Reduction Potential Reading.

These results apply only to the samples tested.

RL - Reporting Limit

- Qualifiers: * Value exceeds Maximum Contaminant Level
C Calculated Value
H Holding times for preparation or analysis exceeded
L Analyzed by a contract laboratory
ND Not Detected at the Reporting Limit
S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
E Value above quantitation range
J Analyte detected below quantitation limits
M Value exceeds Monthly Ave or MCL
O Outside the Range of Dilutions

Reviewed by: [Signature]
Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
1536 Cole Blvd
Lakewood, CO 80401
ProjectName: Reno Creek Project
Lab ID: S1203095-001
ClientSample ID: UM3R-003-120307
COC: RC 08268

Date Reported 4/10/2012
Report ID S1203095001
WorkOrder: S1203095
CollectionDate: 3/7/2012 9:30:00 AM
DateReceived: 3/7/2012 3:52:00 PM
FieldSampler: TN
Matrix: Water

Comments

Table with 7 columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init. Rows include Anions/Cations (Alkalinity, Chloride, etc.), General Parameters (pH, Conductivity), and Data Quality (Cation Sum, Anion Sum, etc.).

These results apply only to the samples tested.

RL - Reporting Limit

- Qualifiers: * Value exceeds Maximum Contaminant Level, C Calculated Value, H Holding times for preparation or analysis exceeded, L Analyzed by a contract laboratory, ND Not Detected at the Reporting Limit, S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank, E Value above quantitation range, J Analyte detected below quantitation limits, M Value exceeds Monthly Ave or MCL, O Outside the Range of Dilutions

Reviewed by: [Signature]
Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
1536 Cole Blvd
Lakewood, CO 80401
ProjectName: Reno Creek Project
Lab ID: S1203095-001
ClientSample ID: UM3R-003-120307
COC: RC 08268

Date Reported 4/10/2012
Report ID S1203095001
WorkOrder: S1203095
CollectionDate: 3/7/2012 9:30:00 AM
DateReceived: 3/7/2012 3:52:00 PM
FieldSampler: TN
Matrix: Water

Comments

Table with 7 columns: Analyses, Result, Units, Qual, RL, Method, Date Analyzed/Init. Rows include Metals - Dissolved (Aluminum, Arsenic, Barium, Boron, Cadmium, Chromium, Copper, Iron, Lead, Manganese, Mercury, Molybdenum, Nickel, Selenium, Uranium, Vanadium, Zinc) and Metals - Suspended (Uranium) and Metals - Total (Iron, Manganese).

These results apply only to the samples tested.

RL - Reporting Limit

- Qualifiers: * Value exceeds Maximum Contaminant Level
C Calculated Value
H Holding times for preparation or analysis exceeded
L Analyzed by a contract laboratory
ND Not Detected at the Reporting Limit
S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
E Value above quantitation range
J Analyte detected below quantitation limits
M Value exceeds Monthly Ave or MCL
O Outside the Range of Dilutions

Reviewed by: [Signature]
Lacey Ketron, Water Lab Supervisor



Sample Analysis Report

Company: AUC LLC
 1536 Cole Blvd
 Lakewood, CO 80401

ProjectName: Reno Creek Project
Lab ID: S1203095-001
ClientSample ID: UM3R-003-120307
COC: RC 08268

Date Reported 4/10/2012
Report ID S1203095001

WorkOrder: S1203095
CollectionDate: 3/7/2012 9:30:00 AM
DateReceived: 3/7/2012 3:52:00 PM
FieldSampler: TN
Matrix: Water

Comments

| Analyses | Result | Units | Qual | RL | Method | Date Analyzed/Init | |
|----------------------------------|--------|-------|------|-----|--------------|--------------------|-----|
| Radionuclides - Dissolved | | | | | | | |
| Gross Alpha | ND | pCi/L | | 2 | SM 7110B | 03/27/2012 2205 | SH |
| Gross Alpha Precision (±) | NA | pCi/L | | | SM 7110B | 03/27/2012 2205 | SH |
| Gross Beta | 22.1 | pCi/L | | 4 | SM 7110B | 03/27/2012 2205 | SH |
| Gross Beta Precision (±) | 2.5 | pCi/L | | | SM 7110B | 03/27/2012 2205 | SH |
| Lead 210 | ND | pCi/L | | 1 | OTW01 | 03/23/2012 1247 | SH |
| Lead 210 Precision (±) | NA | pCi/L | | | OTW01 | 03/23/2012 1247 | SH |
| Polonium 210 | ND | pCi/L | | 1 | OTW01 | 03/20/2012 1325 | SH |
| Polonium 210 Precision (±) | NA | pCi/L | | | OTW01 | 03/20/2012 1325 | SH |
| Radium 226 | 0.4 | pCi/L | | 0.2 | SM 7500-Ra B | 03/19/2012 1607 | SH |
| Radium 226 Precision (±) | 0.1 | pCi/L | | | SM 7500-Ra B | 03/19/2012 1607 | SH |
| Radium 228 | ND | pCi/L | | 1 | Ra-05 | 03/15/2012 1647 | SH |
| Radium 228 Precision (±) | NA | pCi/L | | | Ra-05 | 03/15/2012 1647 | SH |
| Thorium 230 | ND | pCi/L | | 0.2 | ACW10 | 03/21/2012 917 | WL |
| Thorium 230 Precision (±) | NA | pCi/L | | | ACW10 | 03/21/2012 917 | WL |
| Radionuclides - Suspended | | | | | | | |
| Lead 210 | 1.5 | pCi/L | | 1 | OTW01 | 03/28/2012 1323 | SH |
| Lead 210 Precision (±) | 0.5 | pCi/L | | | OTW01 | 03/28/2012 1323 | SH |
| Polonium 210 | ND | pCi/L | | 1 | OTW01 | 03/27/2012 1514 | SH |
| Polonium 210 Precision (±) | NA | pCi/L | | | OTW01 | 03/27/2012 1514 | SH |
| Radium 226 | ND | pCi/L | | 0.2 | SM 7500-Ra B | 03/20/2012 1123 | SH |
| Radium 226 Precision (±) | NA | pCi/L | | | SM 7500-Ra B | 03/20/2012 1123 | SH |
| Thorium 230 | ND | pCi/L | | 0.2 | ACW10 | 03/22/2012 1412 | WL |
| Thorium 230 Precision (±) | NA | pCi/L | | | ACW10 | 03/22/2012 1412 | WL |
| Radionuclides - Total | | | | | | | |
| Radon 222 | ND | pCi/L | L | 50 | SM7500-RN | 03/09/2012 000 | LJK |
| Radon 222 Precision (±) | NA | pCi/L | L | | SM7500-RN | 03/09/2012 000 | LJK |


These results apply only to the samples tested.

RL - Reporting Limit

Qualifiers:

- * Value exceeds Maximum Contaminant Level
- C Calculated Value
- H Holding times for preparation or analysis exceeded
- L Analyzed by a contract laboratory
- ND Not Detected at the Reporting Limit
- S Spike Recovery outside accepted recovery limits

- B Analyte detected in the associated Method Blank
- E Value above quantitation range
- J Analyte detected below quantitation limits
- M Value exceeds Monthly Ave or MCL
- O Outside the Range of Dilutions

Reviewed by: 
 Lacey Ketron, Water Lab Supervisor



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1203095
Project: Reno Creek Project

Date: 4/10/2012
Report ID: S1203095001

Alkalinity

Sample Type **MBLK** Units: mg/L

| Sample ID | RunNo: 80610 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|------------------------------|--------|----|-------|----------|------|--------------|------|
| BLANK | 03/13/12 13:37 | Alkalinity, Total (As CaCO3) | ND | 5 | | | | | |

Sample Type **LCS** Units: mg/L

| Sample ID | RunNo: 80610 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|------------------------------|--------|----|-------|----------|------|--------------|------|
| ATQC | 03/13/12 13:25 | Alkalinity, Total (As CaCO3) | 593 | 5 | 601 | | 98.7 | 90 - 110 | |

Sample Type **DUP** Units: mg/L

| Sample ID | RunNo: 80610 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|----------------|----------------|---------------------------------|--------|----|----------|------|------|--------------|------|
| S1203179-022AD | 03/13/12 19:51 | Alkalinity, Bicarbonate as HCO3 | 11 | 5 | 10 | 8.79 | | 20 | |
| | | Alkalinity, Carbonate as CO3 | ND | 5 | ND | | | 20 | |
| | | Alkalinity, Total (As CaCO3) | 9 | 5 | 8 | 8.79 | | 20 | |

Conductivity

Sample Type **MBLK** Units: µmhos/cm

| Sample ID | RunNo: 80610 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|-------------------------|--------|----|-------|----------|------|--------------|------|
| BLANK | 03/13/12 13:37 | Electrical Conductivity | ND | 5 | | | | | |

Sample Type **LCS** Units: µmhos/cm

| Sample ID | RunNo: 80610 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|-------------------------|--------|----|-------|----------|------|--------------|------|
| ATQC | 03/13/12 13:25 | Electrical Conductivity | 1100 | 5 | 1120 | | 97.5 | 90 - 110 | |

Sample Type **DUP** Units: µmhos/cm

| Sample ID | RunNo: 80610 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|----------------|----------------|-------------------------|--------|----|----------|------|------|--------------|------|
| S1203179-022AD | 03/13/12 19:51 | Electrical Conductivity | 24 | 5 | 23 | 1.72 | | 20 | |

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1203095
Project: Reno Creek Project

Date: 4/10/2012
Report ID: S1203095001

Dissolved Mercury by EPA 245.1 - Water

Sample Type **MBLK** Units: mg/L

| Sample ID | RunNo: 80578 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|---------------|---------|--------|-------|-------|----------|------|--------------|------|
| LRB | 03/13/12 8:11 | Mercury | ND | 0.001 | | | | | |

Sample Type **LCS** Units: mg/L

| Sample ID | RunNo: 80578 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|---------------|---------|--------|-------|-------|----------|------|--------------|------|
| LCS | 03/13/12 8:09 | Mercury | 0.002 | 0.001 | 0.002 | | 100 | 85 - 115 | |

Sample Type **MS** Units: mg/L

| Sample ID | RunNo: 80578 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|---------------|---------------|---------|--------|--------|---------|----------|------|--------------|------|
| S1203094-006D | 03/13/12 9:30 | Mercury | 0.0024 | 0.0001 | 0.00244 | ND | 98.2 | 70 - 130 | |

Sample Type **MSD** Units: mg/L

| Sample ID | RunNo: 80578 | Analyte | Result | RL | Conc | %RPD | %REC | % RPD Limits | Qual |
|---------------|---------------|---------|--------|--------|--------|-------|------|--------------|------|
| S1203094-006D | 03/13/12 9:32 | Mercury | 0.0024 | 0.0001 | 0.0024 | 0.113 | | 20 | |

Sample Type **DUP** Units: mg/L

| Sample ID | RunNo: 80578 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|---------------|---------------|---------|--------|--------|----------|------|------|--------------|------|
| S1203094-006D | 03/13/12 9:28 | Mercury | ND | 0.0001 | ND | | | 20 | |

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1203095
Project: Reno Creek Project

Date: 4/10/2012
Report ID: S1203095001

Dissolved Metals by ICP - EPA 200.7 - WaterSample Type **MBLK** Units: mg/L

| Sample ID | RunNo: 80504 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|---------------|----------------|-----------|--------|------|-------|----------|------|--------------|------|
| MBLK DISS/CAT | 03/09/12 13:13 | Aluminum | ND | 0.1 | | | | | |
| | | Boron | ND | 0.1 | | | | | |
| | | Chromium | ND | 0.01 | | | | | |
| | | Iron | ND | 0.05 | | | | | |
| | | Manganese | ND | 0.02 | | | | | |
| | | Nickel | ND | 0.01 | | | | | |
| | | Zinc | ND | 0.01 | | | | | |

Sample Type **LCS** Units: mg/L

| Sample ID | RunNo: 80504 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|------------|----------------|-----------|--------|------|-------|----------|------|--------------|------|
| DISS LCS Q | 03/09/12 13:16 | Aluminum | 1.0 | 0.1 | 1 | | 99.4 | 85 - 115 | |
| | | Boron | 1.0 | 0.1 | 1 | | 99.4 | 85 - 115 | |
| | | Chromium | 0.98 | 0.01 | 1 | | 98.0 | 85 - 115 | |
| | | Iron | 1.00 | 0.05 | 1 | | 100 | 85 - 115 | |
| | | Manganese | 1.00 | 0.02 | 1 | | 99.8 | 85 - 115 | |
| | | Nickel | 0.99 | 0.01 | 1 | | 99.4 | 85 - 115 | |
| | | Zinc | 1.00 | 0.01 | 1 | | 100 | 85 - 115 | |

Sample Type **MS** Units: mg/L

| Sample ID | RunNo: 80504 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|----------------|----------------|----------|--------|------|-------|----------|------|--------------|------|
| S1203104-003CS | 03/09/12 13:56 | Aluminum | 0.7 | 0.1 | 0.5 | 0.2 | 104 | 70 - 130 | |
| | | Boron | 1.0 | 0.1 | 0.5 | 0.6 | 90.0 | 70 - 130 | |
| | | Chromium | 0.47 | 0.01 | 0.5 | ND | 93.7 | 70 - 130 | |
| | | Iron | 0.61 | 0.05 | 0.5 | 0.12 | 98.4 | 70 - 130 | |
| | | Nickel | 0.48 | 0.01 | 0.5 | ND | 96.4 | 70 - 130 | |
| | | Zinc | 0.47 | 0.01 | 0.5 | ND | 93.5 | 70 - 130 | |

Sample Type **MSD** Units: mg/L

| Sample ID | RunNo: 80504 | Analyte | Result | RL | Conc | %RPD | %REC | % RPD Limits | Qual |
|-----------------|----------------|----------|--------|------|------|--------|------|--------------|------|
| S1203104-003CSD | 03/09/12 13:58 | Aluminum | 0.7 | 0.1 | 0.7 | 0.627 | | 20 | |
| | | Boron | 1.0 | 0.1 | 1.0 | 0.0957 | | 20 | |
| | | Chromium | 0.47 | 0.01 | 0.47 | 0.956 | | 20 | |
| | | Iron | 0.60 | 0.05 | 0.61 | 1.93 | | 20 | |
| | | Nickel | 0.47 | 0.01 | 0.48 | 1.91 | | 20 | |
| | | Zinc | 0.47 | 0.01 | 0.47 | 1.55 | | 20 | |

Qualifiers:

| | | | |
|---|--|----|--|
| B | Analyte detected in the associated Method Blank | E | Value above quantitation range |
| H | Holding times for preparation or analysis exceeded | J | Analyte detected below quantitation limits |
| L | Analyzed by a contract laboratory | ND | Not Detected at the Reporting Limit |
| O | Outside the Range of Dilutions | R | RPD outside accepted recovery limits |
| S | Spike Recovery outside accepted recovery limits | | |

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1203095
Project: Reno Creek Project

Date: 4/10/2012
Report ID: S1203095001

Dissolved Metals by ICP - EPA 200.7 - WaterSample Type **DUP** Units: mg/L

| Sample ID | RunNo: 80504 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|----------------|----------------|----------|--------|------|----------|-------|------|--------------|------|
| S1203104-002CD | 03/09/12 13:51 | Aluminum | 0.2 | 0.1 | 0.2 | 11.5 | | 20 | |
| | | Boron | 1.0 | 0.1 | 1.0 | 0.187 | | 20 | |
| | | Chromium | ND | 0.01 | ND | | | 20 | |
| | | Iron | 0.14 | 0.05 | 0.14 | 0.661 | | 20 | |
| | | Nickel | ND | 0.01 | ND | | | 20 | |
| | | Zinc | ND | 0.01 | ND | | | 20 | |

Dissolved Metals by ICPMS EPA 200.8 - WaterSample Type **MBLK** Units: mg/L

| Sample ID | RunNo: 80498 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|------------|--------|-------|-------|----------|------|--------------|------|
| BLK | 03/09/12 10:54 | Arsenic | ND | 0.005 | | | | | |
| | | Barium | ND | 0.5 | | | | | |
| | | Cadmium | ND | 0.002 | | | | | |
| | | Copper | ND | 0.01 | | | | | |
| | | Lead | ND | 0.02 | | | | | |
| | | Molybdenum | ND | 0.02 | | | | | |
| | | Selenium | ND | 0.005 | | | | | |
| | | Uranium | ND | 0.001 | | | | | |
| | | Vanadium | ND | 0.02 | | | | | |

Sample Type **LCS** Units: mg/L

| Sample ID | RunNo: 80498 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|------------|--------|-------|-------|----------|------|--------------|------|
| LCS | 03/09/12 10:50 | Arsenic | 0.100 | 0.005 | 0.1 | | 100 | 85 - 115 | |
| | | Barium | ND | 0.5 | 0.1 | | 94.5 | 85 - 115 | |
| | | Cadmium | 0.094 | 0.002 | 0.1 | | 94.1 | 85 - 115 | |
| | | Copper | 0.10 | 0.01 | 0.1 | | 97.1 | 85 - 115 | |
| | | Lead | 0.10 | 0.02 | 0.1 | | 95.0 | 85 - 115 | |
| | | Molybdenum | 0.10 | 0.02 | 0.1 | | 99.1 | 85 - 115 | |
| | | Selenium | 0.099 | 0.005 | 0.1 | | 99.1 | 85 - 115 | |
| | | Uranium | 0.096 | 0.001 | 0.1 | | 96.0 | 85 - 115 | |
| | | Vanadium | 0.10 | 0.02 | 0.1 | | 95.7 | 85 - 115 | |

Qualifiers:

| | | | |
|---|--|----|--|
| B | Analyte detected in the associated Method Blank | E | Value above quantitation range |
| H | Holding times for preparation or analysis exceeded | J | Analyte detected below quantitation limits |
| L | Analyzed by a contract laboratory | ND | Not Detected at the Reporting Limit |
| O | Outside the Range of Dilutions | R | RPD outside accepted recovery limits |
| S | Spike Recovery outside accepted recovery limits | | |



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1203095
Project: Reno Creek Project

Date: 4/10/2012
Report ID: S1203095001

Dissolved Metals by ICPMS EPA 200.8 - Water

Sample Type **MS** Units: mg/L

| Sample ID | RunNo: 80498 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|----------------|----------------|------------|--------|-------|-------|----------|------|--------------|------|
| S1203095-001DS | 03/09/12 18:15 | Arsenic | 1.02 | 0.005 | 1.1 | ND | 92.7 | 70 - 130 | |
| | | Barium | 1.1 | 0.5 | 1.1 | ND | 98.3 | 70 - 130 | |
| | | Cadmium | 1.12 | 0.002 | 1.1 | ND | 101 | 70 - 130 | |
| | | Copper | 1.16 | 0.01 | 1.1 | ND | 105 | 70 - 130 | |
| | | Lead | 1.12 | 0.02 | 1.1 | ND | 102 | 70 - 130 | |
| | | Molybdenum | 1.22 | 0.02 | 1.1 | 0.03 | 109 | 70 - 130 | |
| | | Selenium | 1.05 | 0.005 | 1.1 | ND | 95.5 | 70 - 130 | |
| | | Uranium | 1.17 | 0.001 | 1.1 | ND | 107 | 70 - 130 | |
| | | Vanadium | 1.15 | 0.02 | 1.1 | ND | 104 | 70 - 130 | |

Sample Type **MSD** Units: mg/L

| Sample ID | RunNo: 80498 | Analyte | Result | RL | Conc | %RPD | %REC | % RPD Limits | Qual |
|------------------|----------------|------------|--------|-------|------|-------|------|--------------|------|
| S1203095-001DMSD | 03/09/12 18:19 | Arsenic | 1.04 | 0.005 | 1.02 | 1.92 | | 20 | |
| | | Barium | 1.1 | 0.5 | 1.1 | 0.893 | | 20 | |
| | | Cadmium | 1.11 | 0.002 | 1.12 | 0.966 | | 20 | |
| | | Copper | 1.15 | 0.01 | 1.16 | 0.342 | | 20 | |
| | | Lead | 1.11 | 0.02 | 1.12 | 1.39 | | 20 | |
| | | Molybdenum | 1.18 | 0.02 | 1.22 | 3.31 | | 20 | |
| | | Selenium | 1.04 | 0.005 | 1.05 | 1.01 | | 20 | |
| | | Uranium | 1.16 | 0.001 | 1.17 | 0.979 | | 20 | |
| | | Vanadium | 1.11 | 0.02 | 1.15 | 3.24 | | 20 | |

Sample Type **DUP** Units: mg/L

| Sample ID | RunNo: 80498 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|----------------|----------------|------------|--------|-------|----------|------|------|--------------|------|
| S1203095-001DD | 03/09/12 18:10 | Arsenic | ND | 0.005 | ND | | | 20 | |
| | | Barium | ND | 0.5 | ND | | | 20 | |
| | | Cadmium | ND | 0.002 | ND | | | 20 | |
| | | Copper | ND | 0.01 | ND | | | 20 | |
| | | Lead | ND | 0.02 | ND | | | 20 | |
| | | Molybdenum | 0.03 | 0.02 | 0.03 | 3.12 | | 20 | |
| | | Selenium | ND | 0.005 | ND | | | 20 | |
| | | Uranium | ND | 0.001 | ND | | | 20 | |
| | | Vanadium | ND | 0.02 | ND | | | 20 | |

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1203095
Project: Reno Creek Project

Date: 4/10/2012
Report ID: S1203095001

Fluoride by SM 4500

Sample Type **MBLK** Units: mg/L

| Sample ID | RunNo: 80610 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|----------|--------|-----|-------|----------|------|--------------|------|
| BLANK | 03/13/12 13:37 | Fluoride | ND | 0.1 | | | | | |

Sample Type **LCS** Units: mg/L

| Sample ID | RunNo: 80610 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|----------|--------|-----|-------|----------|------|--------------|------|
| ATQC | 03/13/12 13:25 | Fluoride | 2.9 | 0.1 | 2.96 | | 96.9 | 90 - 110 | |

Sample Type **MS** Units: mg/L

| Sample ID | RunNo: 80610 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|------------------|----------------|----------|--------|-----|-------|----------|------|--------------|------|
| S1203161-001BSPK | 03/13/12 14:56 | Fluoride | 2.5 | 0.1 | 2.5 | 0.2 | 92.8 | 80 - 120 | |

Gross Alpha, Beta by SM 7110B

Sample Type **MBLK** Units: pCi/L

| Sample ID | RunNo: 81158 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|-------------------------|--------|----|-------|----------|------|--------------|------|
| MB12-081 | 03/27/12 22:05 | Gross Alpha (Dissolved) | ND | 2 | | | | | |
| | | Gross Beta (Dissolved) | ND | 3 | | | | | |

Sample Type **LCS** Units: pCi/L

| Sample ID | RunNo: 81158 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|-------------------------|--------|----|-------|----------|------|--------------|------|
| LCS12-081 | 03/27/12 22:05 | Gross Alpha (Dissolved) | 58 | 2 | 69 | | 83.8 | 83 - 133 | |
| | | Gross Beta (Dissolved) | 122 | 3 | 126 | | 96.8 | 88 - 115 | |

Sample Type **MS** Units: pCi/L

| Sample ID | RunNo: 81158 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------------|----------------|-------------------------|--------|----|-------|----------|------|--------------|------|
| S1203245-006EMS | 03/27/12 22:05 | Gross Alpha (Dissolved) | 57 | 2 | 73.6 | 7 | 68.3 | 48 - 153 | |
| | | Gross Beta (Dissolved) | 144 | 3 | 126 | 7 | 109 | 60 - 140 | |

Sample Type **DUP** Units: pCi/L

| Sample ID | RunNo: 81158 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|------------------|----------------|-------------------------|--------|----|----------|------|------|--------------|------|
| S1203246-002FDUP | 03/28/12 17:44 | Gross Alpha (Dissolved) | 1400 | 2 | 1340 | 4.47 | | 20 | |
| | | Gross Beta (Dissolved) | 491 | 3 | 458 | 7.25 | | 20 | |

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1203095
Project: Reno Creek Project

Date: 4/10/2012
Report ID: S1203095001

Anions by ION Chromatography

Sample Type **MBLK** Units: mg/L

| Sample ID | RunNo: 80525 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|----------|--------|----|-------|----------|------|--------------|------|
| BLK | 03/09/12 11:00 | Chloride | ND | 1 | | | | | |
| | | Sulfate | ND | 1 | | | | | |

Sample Type **LCS** Units: mg/L

| Sample ID | RunNo: 80525 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|----------|--------|----|-------|----------|------|--------------|------|
| DIONEX | 03/09/12 10:36 | Chloride | 30 | 1 | 30 | | 101 | 90 - 110 | |
| | | Sulfate | 156 | 1 | 150 | | 104 | 90 - 110 | |

Sample Type **MS** Units: mg/L

| Sample ID | RunNo: 80525 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------------|---------------|----------|--------|----|-------|----------|------|--------------|------|
| S1203094-003BSP | 03/10/12 0:38 | Chloride | 6 | 1 | 5 | 3 | 75.8 | 80 - 120 | S |
| | | Sulfate | 48 | 1 | 40 | 8 | 98.2 | 80 - 120 | |

Sample Type **DUP** Units: mg/L

| Sample ID | RunNo: 80525 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|---------------|---------------|----------|--------|----|----------|-------|------|--------------|------|
| S1203094-003B | 03/10/12 0:27 | Chloride | 3 | 1 | 3 | 0.296 | | 20 | |
| | | Sulfate | 8 | 1 | 8 | 0.807 | | 20 | |

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1203095
Project: Reno Creek Project

Date: 4/10/2012
Report ID: S1203095001

Cations by ICP (Method 200.7)Sample Type **MBLK** Units: mg/L

| Sample ID | RunNo: 80503 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|---------------|----------------|-----------|--------|----|-------|----------|------|--------------|------|
| MBLK DISS/CAT | 03/09/12 13:13 | Calcium | ND | 1 | | | | | |
| | | Magnesium | ND | 1 | | | | | |
| | | Potassium | ND | 1 | | | | | |
| | | Sodium | ND | 1 | | | | | |

| Sample ID | RunNo: 81453 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|---------------|----------------|-----------|--------|----|-------|----------|------|--------------|------|
| MBLK DISS/CAT | 04/04/12 11:42 | Calcium | ND | 1 | | | | | |
| | | Magnesium | ND | 1 | | | | | |
| | | Potassium | ND | 1 | | | | | |
| | | Sodium | ND | 1 | | | | | |

Sample Type **LCS** Units: mg/L

| Sample ID | RunNo: 80503 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|--------------|----------------|-----------|--------|----|-------|----------|------|--------------|------|
| CAT LCS IML3 | 03/09/12 13:18 | Calcium | 40 | 1 | 40 | | 99.9 | 85 - 115 | |
| | | Magnesium | 39 | 1 | 40 | | 97.8 | 85 - 115 | |
| | | Potassium | 41 | 1 | 40 | | 103 | 85 - 115 | |
| | | Sodium | 39 | 1 | 40 | | 97.5 | 85 - 115 | |

| Sample ID | RunNo: 81453 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|--------------|----------------|-----------|--------|----|-------|----------|------|--------------|------|
| CAT LCS IML3 | 04/04/12 11:47 | Calcium | 41 | 1 | 40 | | 103 | 85 - 115 | |
| | | Magnesium | 41 | 1 | 40 | | 103 | 85 - 115 | |
| | | Potassium | 41 | 1 | 40 | | 102 | 85 - 115 | |
| | | Sodium | 40 | 1 | 40 | | 98.8 | 85 - 115 | |

Sample Type **MS** Units: mg/L

| Sample ID | RunNo: 80503 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|----------------|----------------|-----------|--------|----|-------|----------|------|--------------|------|
| S1203113-003AS | 03/09/12 14:26 | Calcium | 62 | 1 | 50 | 13 | 96.3 | 70 - 130 | |
| | | Magnesium | 54 | 1 | 50 | 5 | 98.6 | 70 - 130 | |
| | | Potassium | 58 | 1 | 50 | 8 | 101 | 70 - 130 | |
| | | Sodium | 68 | 1 | 50 | 20 | 97.5 | 70 - 130 | |

| Sample ID | RunNo: 81453 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|----------------|----------------|-----------|--------|----|-------|----------|------|--------------|------|
| S1204032-012AS | 04/04/12 15:53 | Calcium | 53 | 1 | 50 | ND | 105 | 70 - 130 | |
| | | Magnesium | 54 | 1 | 50 | ND | 109 | 70 - 130 | |
| | | Potassium | 55 | 1 | 50 | ND | 110 | 70 - 130 | |
| | | Sodium | 53 | 1 | 50 | ND | 105 | 70 - 130 | |

Qualifiers:

| | | | |
|---|--|----|--|
| B | Analyte detected in the associated Method Blank | E | Value above quantitation range |
| H | Holding times for preparation or analysis exceeded | J | Analyte detected below quantitation limits |
| L | Analyzed by a contract laboratory | ND | Not Detected at the Reporting Limit |
| O | Outside the Range of Dilutions | R | RPD outside accepted recovery limits |
| S | Spike Recovery outside accepted recovery limits | | |

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1203095
Project: Reno Creek Project

Date: 4/10/2012
Report ID: S1203095001

Cations by ICP (Method 200.7)Sample Type **MSD** Units: mg/L

| Sample ID | RunNo: 80503 | Analyte | Result | RL | Conc | %RPD | %REC | % RPD Limits | Qual |
|-----------------|----------------|-----------|--------|----|------|-------|------|--------------|------|
| S1203113-003ASD | 03/09/12 14:28 | Calcium | 61 | 1 | 62 | 0.851 | | 20 | |
| | | Magnesium | 53 | 1 | 54 | 2.27 | | 20 | |
| | | Potassium | 58 | 1 | 58 | 1.02 | | 20 | |
| | | Sodium | 68 | 1 | 68 | 0.532 | | 20 | |

| Sample ID | RunNo: 81453 | Analyte | Result | RL | Conc | %RPD | %REC | % RPD Limits | Qual |
|-----------------|----------------|-----------|--------|----|------|------|------|--------------|------|
| S1204032-012ASD | 04/04/12 15:55 | Calcium | 54 | 1 | 53 | 1.82 | | 20 | |
| | | Magnesium | 55 | 1 | 54 | 1.82 | | 20 | |
| | | Potassium | 57 | 1 | 55 | 2.64 | | 20 | |
| | | Sodium | 54 | 1 | 53 | 2.37 | | 20 | |

Sample Type **DUP** Units: mg/L

| Sample ID | RunNo: 80503 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|----------------|----------------|-----------|--------|----|----------|-------|------|--------------|------|
| S1203113-002AD | 03/09/12 14:21 | Calcium | 10 | 1 | 10 | 0.632 | | 20 | |
| | | Magnesium | 4 | 1 | 4 | 0.424 | | 20 | |
| | | Potassium | 8 | 1 | 8 | 0.623 | | 20 | |
| | | Sodium | 22 | 1 | 22 | 0.593 | | 20 | |

| Sample ID | RunNo: 81453 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|----------------|----------------|-----------|--------|----|----------|-------|------|--------------|------|
| S1204032-011AD | 04/04/12 15:48 | Calcium | 8 | 1 | 8 | 0.139 | | 20 | |
| | | Magnesium | ND | 1 | ND | | | 20 | |
| | | Potassium | ND | 1 | ND | | | 20 | |
| | | Sodium | ND | 1 | ND | | | 20 | |

Nitrogen, Ammonia (as N)Sample Type **MBLK** Units: mg/L

| Sample ID | RunNo: 80649 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|---------------|--------------------------|--------|-----|-------|----------|------|--------------|------|
| BLANK | 03/14/12 9:17 | Nitrogen, Ammonia (As N) | ND | 0.1 | | | | | |

Sample Type **LCS** Units: mg/L

| Sample ID | RunNo: 80649 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|---------------|--------------------------|--------|-----|-------|----------|------|--------------|------|
| QC | 03/14/12 9:19 | Nitrogen, Ammonia (As N) | 9.9 | 0.1 | 10 | | 98.8 | 90 - 110 | |

Sample Type **MS** Units: mg/L

| Sample ID | RunNo: 80649 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|---------------|---------------|--------------------------|--------|-----|-------|----------|------|--------------|------|
| S1203096-001C | 03/14/12 9:34 | Nitrogen, Ammonia (As N) | 9.8 | 0.1 | 10 | 0.1 | 97.1 | 80 - 120 | |

Sample Type **DUP** Units: mg/L

| Sample ID | RunNo: 80649 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|---------------|---------------|--------------------------|--------|-----|----------|------|------|--------------|------|
| S1203096-001C | 03/14/12 9:33 | Nitrogen, Ammonia (As N) | 0.1 | 0.1 | 0.1 | 1.24 | | 20 | |

Qualifiers:

| | | | |
|---|--|----|--|
| B | Analyte detected in the associated Method Blank | E | Value above quantitation range |
| H | Holding times for preparation or analysis exceeded | J | Analyte detected below quantitation limits |
| L | Analyzed by a contract laboratory | ND | Not Detected at the Reporting Limit |
| O | Outside the Range of Dilutions | R | RPD outside accepted recovery limits |
| S | Spike Recovery outside accepted recovery limits | | |



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1203095
Project: Reno Creek Project

Date: 4/10/2012
Report ID: S1203095001

Nitrogen, Nitrate-Nitrite (as N)

Sample Type **MBLK** Units: mg/L

| Sample ID | RunNo: 80582 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|----------------------------------|--------|-----|-------|----------|------|--------------|------|
| BLK | 03/13/12 10:31 | Nitrogen, Nitrate-Nitrite (as N) | ND | 0.1 | | | | | |

Sample Type **LCS** Units: mg/L

| Sample ID | RunNo: 80582 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|----------------------------------|--------|-----|-------|----------|------|--------------|------|
| QC | 03/13/12 10:33 | Nitrogen, Nitrate-Nitrite (as N) | 4.6 | 0.1 | 4.35 | | 106 | 90 - 110 | |

Sample Type **MS** Units: mg/L

| Sample ID | RunNo: 80582 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|---------------|----------------|----------------------------------|--------|-----|-------|----------|------|--------------|------|
| S1203099-001C | 03/13/12 11:19 | Nitrogen, Nitrate-Nitrite (as N) | 10.8 | 0.1 | 10 | ND | 108 | 80 - 120 | |

Sample Type **DUP** Units: mg/L

| Sample ID | RunNo: 80582 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|---------------|----------------|----------------------------------|--------|-----|----------|------|------|--------------|------|
| S1203099-001C | 03/13/12 11:18 | Nitrogen, Nitrate-Nitrite (as N) | ND | 0.1 | ND | | | 20 | |

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1203095
Project: Reno Creek Project

Date: 4/10/2012
Report ID: S1203095001

Lead 210 and Polonium 210Sample Type **MBLK** Units: pCi/L

| Sample ID | RunNo: | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|

| | | | | | | | | | |
|----------|---------------|----------------------|----|---|--|--|--|--|--|
| MB12-072 | 03/23/12 9:43 | Lead 210 (Dissolved) | ND | 1 | | | | | |
|----------|---------------|----------------------|----|---|--|--|--|--|--|

| Sample ID | RunNo: | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|

| | | | | | | | | | |
|----------|----------------|--------------------------|----|---|--|--|--|--|--|
| MB12-072 | 03/20/12 12:17 | Polonium 210 (Dissolved) | ND | 1 | | | | | |
|----------|----------------|--------------------------|----|---|--|--|--|--|--|

| Sample ID | RunNo: | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|

| | | | | | | | | | |
|----------|----------------|----------------------|----|---|--|--|--|--|--|
| MB12-079 | 03/27/12 16:21 | Lead 210 (Dissolved) | ND | 1 | | | | | |
| | | Lead 210 (Suspended) | ND | 1 | | | | | |

| Sample ID | RunNo: | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|

| | | | | | | | | | |
|-----------|----------------|--------------------------|----|---|--|--|--|--|--|
| DMB12-079 | 03/27/12 15:14 | Polonium 210 (Suspended) | ND | 1 | | | | | |
|-----------|----------------|--------------------------|----|---|--|--|--|--|--|

Sample Type **LCS** Units: pCi/L

| Sample ID | RunNo: | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|

| | | | | | | | | | |
|-----------|---------------|----------------------|----|---|------|--|-----|------------|--|
| LCS12-072 | 03/23/12 9:43 | Lead 210 (Dissolved) | 14 | 1 | 12.8 | | 110 | 67.6 - 131 | |
| | | Lead 210 (Suspended) | 14 | 1 | 12.8 | | 110 | 60.3 - 141 | |

| Sample ID | RunNo: | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|

| | | | | | | | | | |
|-----------|----------------|--------------------------|---|---|------|--|------|------------|--|
| LCS12-072 | 03/20/12 12:17 | Polonium 210 (Dissolved) | 5 | 1 | 7.03 | | 72.9 | 35.1 - 175 | |
| | | Polonium 210 (Suspended) | 5 | 1 | 7.03 | | 72.9 | 48.6 - 156 | |

| Sample ID | RunNo: | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|

| | | | | | | | | | |
|-----------|----------------|----------------------|----|---|------|--|-----|------------|--|
| LCS12-079 | 03/27/12 16:21 | Lead 210 (Dissolved) | 15 | 1 | 12.8 | | 118 | 67.6 - 131 | |
| | | Lead 210 (Suspended) | 15 | 1 | 12.8 | | 118 | 60.3 - 141 | |

| Sample ID | RunNo: | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|

| | | | | | | | | | |
|-----------|----------------|--------------------------|----|---|------|--|-----|------------|--|
| LCS12-079 | 03/27/12 13:02 | Polonium 210 (Dissolved) | 7 | 1 | 6.83 | | 107 | 35.1 - 175 | |
| | | Polonium 210 (Suspended) | ND | 1 | | | | | |

Sample Type **MS** Units: pCi/L

| Sample ID | RunNo: | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|

| | | | | | | | | | |
|-----------------|----------------|----------------------|----|---|------|---|------|------------|--|
| S1203064-001FMS | 03/23/12 12:47 | Lead 210 (Suspended) | 14 | 1 | 12.8 | 1 | 96.9 | 62.4 - 132 | |
|-----------------|----------------|----------------------|----|---|------|---|------|------------|--|

| Sample ID | RunNo: | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|

| | | | | | | | | | |
|-------------------|----------------|--------------------------|---|---|------|----|------|----------|--|
| S1203065-001FMSPO | 03/20/12 13:25 | Polonium 210 (Suspended) | 7 | 1 | 7.03 | ND | 98.3 | 28 - 173 | |
|-------------------|----------------|--------------------------|---|---|------|----|------|----------|--|

| Sample ID | RunNo: | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|
|-----------|--------|---------|--------|----|-------|----------|------|--------------|------|

| | | | | | | | | | |
|-----------------|----------------|----------------------|----|---|------|---|------|----------|--|
| S1203217-005DMS | 03/27/12 16:21 | Lead 210 (Dissolved) | 14 | 1 | 12.8 | 2 | 92.0 | 66 - 133 | |
|-----------------|----------------|----------------------|----|---|------|---|------|----------|--|

Qualifiers:

| | | | |
|---|--|----|--|
| B | Analyte detected in the associated Method Blank | E | Value above quantitation range |
| H | Holding times for preparation or analysis exceeded | J | Analyte detected below quantitation limits |
| L | Analyzed by a contract laboratory | ND | Not Detected at the Reporting Limit |
| O | Outside the Range of Dilutions | R | RPD outside accepted recovery limits |
| S | Spike Recovery outside accepted recovery limits | | |



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1203095
Project: Reno Creek Project

Date: 4/10/2012
Report ID: S1203095001

Lead 210 and Polonium 210

Sample Type **MS** Units: pCi/L

| Sample ID | RunNo: 81340 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------------|----------------|--------------------------|--------|----|-------|----------|------|--------------|------|
| S1203217-006DMS | 03/27/12 13:02 | Polonium 210 (Dissolved) | 5 | 1 | 6.83 | ND | 75.5 | 20.9 - 160 | |

Sample Type **DUP** Units: pCi/L

| Sample ID | RunNo: 81108 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|------------------|---------------|----------------------|--------|----|----------|------|------|--------------|------|
| S1203067-001FDUP | 03/23/12 9:43 | Lead 210 (Dissolved) | ND | 1 | 2 | | | 20 | |

| Sample ID | RunNo: 81108 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|------------------|----------------|----------------------|--------|----|----------|------|------|--------------|------|
| S1203065-001FDUP | 03/23/12 12:47 | Lead 210 (Suspended) | ND | 1 | 1 | | | 20 | |

| Sample ID | RunNo: 81109 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|------------------|----------------|--------------------------|--------|----|----------|------|------|--------------|------|
| S1203067-001FDUP | 03/20/12 12:17 | Polonium 210 (Dissolved) | ND | 1 | ND | | | 20 | |

| Sample ID | RunNo: 81340 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|------------------|----------------|--------------------------|--------|----|----------|------|------|--------------|------|
| S1203095-001FDUP | 03/27/12 15:14 | Polonium 210 (Suspended) | ND | 1 | ND | | | 20 | |

Radium 228

Sample Type **MBLK** Units: pCi/L

| Sample ID | RunNo: 80754 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|------------------------|--------|----|-------|----------|------|--------------|------|
| MB12-072 | 03/15/12 16:47 | Radium 228 (Dissolved) | ND | 1 | | | | | |

Sample Type **LCS** Units: pCi/L

| Sample ID | RunNo: 80754 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|------------------------|--------|----|-------|----------|------|--------------|------|
| LCS12-072 | 03/15/12 16:47 | Radium 228 (Dissolved) | 11 | 1 | 9.72 | | 112 | 74.9 - 122 | |

Sample Type **MS** Units: pCi/L

| Sample ID | RunNo: 80754 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------------|----------------|------------------------|--------|----|-------|----------|------|--------------|------|
| S1203064-001FMS | 03/15/12 16:47 | Radium 228 (Dissolved) | 10 | 1 | 9.72 | ND | 98.4 | 48.4 - 167 | |

Sample Type **DUP** Units: pCi/L

| Sample ID | RunNo: 80754 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|------------------|----------------|------------------------|--------|----|----------|------|------|--------------|------|
| S1203065-001FDUP | 03/15/12 14:46 | Radium 228 (Dissolved) | ND | 1 | ND | | | 20 | |

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1203095
Project: Reno Creek Project

Date: 4/10/2012
Report ID: S1203095001

Radium 226 by SM 7500

Sample Type **MBLK** Units: pCi/L

| Sample ID | RunNo: 80815 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|------------------------|--------|-----|-------|----------|------|--------------|------|
| MB12-072 | 03/19/12 13:10 | Radium 226 (Dissolved) | ND | 0.2 | | | | | |

| Sample ID | RunNo: 80817 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|------------------------|--------|-----|-------|----------|------|--------------|------|
| MB12-074 | 03/20/12 11:23 | Radium 226 (Suspended) | ND | 0.2 | | | | | |

Sample Type **LCS** Units: pCi/L

| Sample ID | RunNo: 80815 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|------------------------|--------|-----|-------|----------|------|--------------|------|
| LCS12-072 | 03/19/12 13:10 | Radium 226 (Dissolved) | 5.1 | 0.2 | 5.06 | | 100 | 62.5 - 125 | |

| Sample ID | RunNo: 80817 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|------------------------|--------|-----|-------|----------|------|--------------|------|
| LCS12-074 | 03/20/12 11:23 | Radium 226 (Suspended) | 5.3 | 0.2 | 5.06 | | 104 | 56.4 - 132 | |

Sample Type **MS** Units: pCi/L

| Sample ID | RunNo: 80815 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------------|----------------|------------------------|--------|-----|-------|----------|------|--------------|------|
| S1203089-002DMS | 03/19/12 13:10 | Radium 226 (Dissolved) | 8.9 | 0.2 | 10.1 | 0.3 | 85.0 | 41.6 - 146 | |

| Sample ID | RunNo: 80817 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|--------------|----------------|------------------------|--------|-----|-------|----------|------|--------------|------|
| TAP WATER MS | 03/20/12 11:23 | Radium 226 (Suspended) | 5.1 | 0.2 | 5.06 | ND | 101 | 51.4 - 135 | |

Sample Type **DUP** Units: pCi/L

| Sample ID | RunNo: 80815 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|------------------|----------------|------------------------|--------|-----|----------|------|------|--------------|------|
| S1203104-004EDUP | 03/19/12 16:07 | Radium 226 (Dissolved) | 1.2 | 0.2 | 1.2 | 6.76 | | 20 | |

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits

**ANALYTICAL QC SUMMARY REPORT**

CLIENT: AUC LLC
Work Order: S1203095
Project: Reno Creek Project

Date: 4/10/2012
Report ID: S1203095001

Thorium 230 in WaterSample Type **MBLK** Units: pCi/L

| Sample ID | RunNo: 81037 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|---------------|-------------------------|--------|-----|-------|----------|------|--------------|------|
| MB12-073 | 03/20/12 9:06 | Thorium 230 (Dissolved) | ND | 0.2 | | | | | |

| Sample ID | RunNo: 81037 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|---------------|-------------------------|--------|-----|-------|----------|------|--------------|------|
| MB12-073 | 03/23/12 9:19 | Thorium 230 (Suspended) | ND | 0.2 | | | | | |

Sample Type **LCS** Units: pCi/L

| Sample ID | RunNo: 81037 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|---------------|-------------------------|--------|-----|-------|----------|------|--------------|------|
| LCS12-073 | 03/20/12 9:07 | Thorium 230 (Dissolved) | 13 | 0.2 | 12 | | 102 | 67.1 - 134 | |
| | | Thorium 230 (Suspended) | 13 | 0.2 | 12 | | 102 | 69.2 - 127 | |

Sample Type **DUP** Units: pCi/L

| Sample ID | RunNo: 81037 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|-------------------|----------------|-------------------------|--------|-----|----------|------|------|--------------|------|
| S1203095-001F DUP | 03/22/12 14:12 | Thorium 230 (Dissolved) | ND | 0.2 | ND | | | 20 | |

| Sample ID | RunNo: 81037 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|-------------------|---------------|-------------------------|--------|-----|----------|------|------|--------------|------|
| S1203096-001F DUP | 03/23/12 9:19 | Thorium 230 (Suspended) | ND | 0.2 | ND | | | 20 | |

Radon 222 by SM7500-RNSample Type **MBLK** Units: pCi/L

| Sample ID | RunNo: 81054 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|---------------|-----------|--------|----|-------|----------|------|--------------|------|
| MB-R81054 | 03/09/12 0:00 | Radon-222 | ND | 50 | | | | | L |

Sample Type **LCS** Units: pCi/L

| Sample ID | RunNo: 81054 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|------------|---------------|-----------|--------|-----|-------|----------|------|--------------|------|
| LCS-R81054 | 03/09/12 0:00 | Radon-222 | 22200 | 100 | 22500 | | 98.7 | 75 - 125 | L |

Sample Type **DUP** Units: pCi/L

| Sample ID | RunNo: 81054 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|---------------|---------------|-----------|--------|----|----------|------|------|--------------|------|
| S1203095-001G | 03/09/12 0:00 | Radon-222 | ND | 50 | ND | | | 20 | L |

Silica as SiO2Sample Type **MBLK** Units: mg/L

| Sample ID | RunNo: 80505 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|---------------|----------------|----------------|--------|-----|-------|----------|------|--------------|------|
| MBLK DISS/CAT | 03/09/12 13:13 | Silica as SiO2 | ND | 0.1 | | | | | |

Sample Type **LCS** Units: mg/L

| Sample ID | RunNo: 80505 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|------------|----------------|----------------|--------|-----|-------|----------|------|--------------|------|
| DISS LCS Q | 03/09/12 13:16 | Silica as SiO2 | 2.1 | 0.1 | 2.14 | | 97.2 | 85 - 115 | |

Qualifiers:

| | | | |
|---|--|----|--|
| B | Analyte detected in the associated Method Blank | E | Value above quantitation range |
| H | Holding times for preparation or analysis exceeded | J | Analyte detected below quantitation limits |
| L | Analyzed by a contract laboratory | ND | Not Detected at the Reporting Limit |
| O | Outside the Range of Dilutions | R | RPD outside accepted recovery limits |
| S | Spike Recovery outside accepted recovery limits | | |



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1203095
Project: Reno Creek Project

Date: 4/10/2012
Report ID: S1203095001

Solids By SM 2540

Sample Type **MBLK** Units: mg/L

| Sample ID | RunNo: 80665 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|------------------------------|--------|----|-------|----------|------|--------------|------|
| DI | 03/08/12 11:32 | Total Dissolved Solids (180) | ND | 10 | | | | | |

Sample Type **LCS** Units: mg/L

| Sample ID | RunNo: 80665 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|------------------------------|--------|----|-------|----------|------|--------------|------|
| CONTROL | 03/08/12 11:33 | Total Dissolved Solids (180) | 220 | 10 | 226 | | 95.6 | 90 - 110 | |

Sample Type **DUP** Units: mg/L

| Sample ID | RunNo: 80665 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|---------------|---------------|------------------------------|--------|----|----------|-------|------|--------------|------|
| S1203159-001A | 03/13/12 9:03 | Total Dissolved Solids (180) | 1700 | 10 | 1700 | 0.235 | | 20 | |

Total (200.2) Metals by EPA 200.7 ICP - Water

Sample Type **MBLK** Units: mg/L

| Sample ID | RunNo: 80615 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|-----------|--------|------|-------|----------|------|--------------|------|
| MB-5907 | 03/13/12 16:21 | Iron | ND | 0.05 | | | | | |
| | | Manganese | ND | 0.02 | | | | | |

Sample Type **LCS** Units: mg/L

| Sample ID | RunNo: 80615 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|-----------|--------|------|-------|----------|------|--------------|------|
| LCS-5907 | 03/13/12 16:24 | Iron | 0.50 | 0.05 | 0.5 | | 99.3 | 85 - 115 | |
| | | Manganese | 0.20 | 0.02 | 0.2 | | 98.0 | 85 - 115 | |

Sample Type **MS** Units: mg/L

| Sample ID | RunNo: 80615 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|----------------|----------------|-----------|--------|------|-------|----------|------|--------------|------|
| S1203096-001ES | 03/13/12 16:33 | Iron | 0.57 | 0.05 | 0.5 | 0.11 | 92.6 | 70 - 130 | |
| | | Manganese | 0.20 | 0.02 | 0.2 | ND | 97.4 | 70 - 130 | |

Sample Type **MSD** Units: mg/L

| Sample ID | RunNo: 80615 | Analyte | Result | RL | Conc | %RPD | %REC | % RPD Limits | Qual |
|------------------|----------------|-----------|--------|------|------|------|------|--------------|------|
| S1203096-001EMSD | 03/13/12 16:35 | Iron | 0.58 | 0.05 | 0.57 | 1.81 | | 20 | |
| | | Manganese | 0.20 | 0.02 | 0.20 | 1.46 | | 20 | |

Sample Type **DUP** Units: mg/L

| Sample ID | RunNo: 80615 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|----------------|----------------|-----------|--------|------|----------|------|------|--------------|------|
| S1203095-001ED | 03/13/12 16:28 | Iron | ND | 0.05 | ND | | | 20 | |
| | | Manganese | ND | 0.02 | ND | | | 20 | |

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



ANALYTICAL QC SUMMARY REPORT

CLIENT: AUC LLC
Work Order: S1203095
Project: Reno Creek Project

Date: 4/10/2012
Report ID: S1203095001

Suspended Natural Uranium

Sample Type **MBLK** Units: mg/L

| Sample ID | RunNo: 80587 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|-------------------|--------|--------|-------|----------|------|--------------|------|
| BLK | 03/13/12 12:40 | Uranium Suspended | ND | 0.0003 | | | | | |

Sample Type **LCS** Units: mg/L

| Sample ID | RunNo: 80587 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|-------------------|--------|--------|-------|----------|------|--------------|------|
| LCS | 03/13/12 12:35 | Uranium Suspended | 0.104 | 0.0003 | 0.1 | | 104 | 85 - 115 | |

Sample Type **MS** Units: mg/L

| Sample ID | RunNo: 80587 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|----------------|----------------|-------------------|--------|--------|-------|----------|------|--------------|------|
| S1203095-001FS | 03/13/12 13:10 | Uranium Suspended | 1.13 | 0.0003 | 1.1 | ND | 103 | 70 - 130 | |

Sample Type **MSD** Units: mg/L

| Sample ID | RunNo: 80587 | Analyte | Result | RL | Conc | %RPD | %REC | % RPD Limits | Qual |
|------------------|----------------|-------------------|--------|--------|------|-------|------|--------------|------|
| S1203095-001FMSD | 03/13/12 13:14 | Uranium Suspended | 1.14 | 0.0003 | 1.13 | 0.534 | | 20 | |

Sample Type **DUP** Units: mg/L

| Sample ID | RunNo: 80587 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|----------------|----------------|-------------------|--------|--------|----------|------|------|--------------|------|
| S1203095-001FD | 03/13/12 13:05 | Uranium Suspended | ND | 0.0003 | ND | | | 20 | |

pH Water

Sample Type **LCS** Units: s.u.

| Sample ID | RunNo: 80610 | Analyte | Result | RL | Spike | Ref Samp | %REC | % Rec Limits | Qual |
|-----------|----------------|---------|--------|-----|-------|----------|------|--------------|------|
| ATQC | 03/13/12 13:25 | pH | 8.7 | 0.1 | 8.48 | | 102 | 90 - 110 | |

Sample Type **DUP** Units: s.u.

| Sample ID | RunNo: 80610 | Analyte | Result | RL | Ref Samp | %RPD | %REC | % RPD Limits | Qual |
|----------------|----------------|---------|--------|-----|----------|-------|------|--------------|------|
| S1203179-022AD | 03/13/12 19:51 | pH | 7.2 | 0.1 | 7.2 | 0.854 | | 20 | |

- Qualifiers:**
- B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - L Analyzed by a contract laboratory
 - O Outside the Range of Dilutions
 - S Spike Recovery outside accepted recovery limits
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - ND Not Detected at the Reporting Limit
 - R RPD outside accepted recovery limits



Inter-Mountain Labs
 Sheridan, WY and Gillette, WY
 1673 Terra Ave (307) 682-8945
 Sheridan, WY 82801 (307) 672-8946

- CHAIN OF CUSTODY RECORD -

This is a LEGAL DOCUMENT. All shaded fields must be completed. See reverse for instructions. #RC 08268

| | | | | | | | |
|---|--|--|--|---|--|----------------------|--|
| Client Name AUC LLC | | Project Identification Reno Creek Project | | Sampler (Signature/Printed) <i>Traci Mickelson</i> | | Telephone # | |
| Report Address pcavendor@auc-llc.com; tmickerson@treccorp.com; johnw@treccorp.com | | Contact Name <i>Traci Mickelson</i> | | Analyses / Parameters | | | |
| Invoice Address AUC LLC 900 Federal Ct. Ste 150 Casper, WY 82601 | | Purchase Order # | | See Attached Table (14, Guide 8, Radon 22) | | | |
| LAB ID (Lab Use Only) | | DATE SAMPLED | | TIME SAMPLED | | REMARKS | |
| 1 | | | | | | | |
| 2 <i>51203095</i> | | <i>3/7/12 9:30</i> | | <i>UM3R-003-120307 w 10 X</i> | | <i>Stourderwater</i> | |
| 3 | | | | | | | |
| 4 | | | | | | | |
| 5 | | | | | | | |
| 6 | | | | | | | |
| 7 | | | | | | | |
| 8 | | | | | | | |
| 9 | | | | | | | |
| 10 | | | | | | | |
| 11 | | | | | | | |
| 12 | | | | | | | |
| 13 | | | | | | | |
| 14 | | | | | | | |

| | | | | | | |
|----------------------------|---|------|------|---|---------------|-------------|
| LAB COMMENTS <i>7.0</i> | Relinquished By (Signature/Printed) <i>Traci Mickelson</i> | DATE | TIME | Received By (Signature/Printed) <i>Alice L. McNeil</i> | DATE | TIME |
| | | | | | <i>3/7/12</i> | <i>1552</i> |

| | | | | | | | | | |
|---|---------------------------------------|--------------------------------|-------------------------------------|---|---|--------------------------|--|--------------------|--|
| SHIPPING INFO | | MATRIX CODES | | TURNAROUND TIMES | | COMPLIANCE INFORMATION | | ADDITIONAL REMARKS | |
| <input type="checkbox"/> UPS | <input type="checkbox"/> Fed Express | <input type="checkbox"/> Water | <input type="checkbox"/> Soil | <input checked="" type="checkbox"/> Check desired service | <input checked="" type="checkbox"/> Standard turnaround | Compliance Monitoring? | Y / <input checked="" type="radio"/> N | | |
| <input type="checkbox"/> US Mail | <input type="checkbox"/> Hand Carried | <input type="checkbox"/> Solid | <input type="checkbox"/> Trip Blank | <input type="checkbox"/> RUSH - 5 Working Days | <input type="checkbox"/> URGENT - < 2 Working Days | Program (SDWA, NPDES...) | | | |
| <input checked="" type="checkbox"/> Other | | <input type="checkbox"/> Other | | Rush & Urgent Surcharges will be applied | | PWSID / Permit # | | | |
| | | | | | | Chlorinated? | Y / <input checked="" type="radio"/> N | | |
| | | | | | | Sample Disposal: Lab | <input checked="" type="checkbox"/> Client | | |