

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

<p>1. LICENSEE/LOCATION INSPECTED:</p> <p>ProMedica Northwest Ohio Cardiology Consultants 2940 North McCord Road Toledo, OH 43614</p> <p>REPORT NUMBER(S) 2013-003</p>	<p>2. NRC/REGIONAL OFFICE</p> <p>Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352</p>	
<p>3. DOCKET NUMBER(S)</p> <p>150-00034</p>	<p>4. LICENSE NUMBER(S)</p> <p>OH-02220490001</p>	<p>5. DATE(S) OF INSPECTION</p> <p>8/16/2013</p>

**LICENSEE:**

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.  
(Violations and Corrective Actions)

**Statement of Corrective Actions**

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Ryan Craffey		8/19/13
BRANCH CHIEF	Aaron McCraw		8/19/13

**Docket File Information**

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

<p>1. LICENSEE/LOCATION INSPECTED:</p> <p>ProMedica Northwest Ohio Cardiology Consultants 2940 North McCord Road Toledo, OH 43614</p> <p>REPORT NUMBER(S) 2013-003</p>	<p>2. NRC/REGIONAL OFFICE</p> <p>Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352</p>
--	---

<p>3. DOCKET NUMBER(S)</p> <p>150-00034</p>	<p>4. LICENSE NUMBER(S)</p> <p>OH-02220490001</p>	<p>5. DATE(S) OF INSPECTION</p> <p>8/16/2013</p>
---	---	--

<p>6. INSPECTION PROCEDURES USED</p> <p>87130</p>	<p>7. INSPECTION FOCUS AREAS</p> <p>All</p>
---	---

**SUPPLEMENTAL INSPECTION INFORMATION**

<p>1. PROGRAM CODE(S)</p> <p>02220</p>	<p>2. PRIORITY</p> <p>3</p>	<p>3. LICENSEE CONTACT</p> <p>Norm Szakovits, Nuclear Testing Mgr.</p>	<p>4. TELEPHONE NUMBER</p> <p>(419) 842-3006</p>
--	-----------------------------	--	--

Main Office Inspection                      Next Inspection Date:                      TBD

Field Office Inspection

Temporary Job Site Inspection    219 North Front Street, Monroe, MI

**PROGRAM SCOPE**

This licensee is a mobile nuclear medicine service authorized by its State of Ohio license to use byproduct material for diagnostic imaging and localization procedures under rule 3701:1-58-34 of the Ohio Administrative Code at its facility in Toledo, Ohio, and at temporary client sites. The licensee currently conducts these licensed activities at client sites within NRC jurisdiction under an initial reciprocity report filed in December 2012 for calendar year 2013. The licensee receives unit doses for cardiac stress tests at its facility in Ohio, and after assaying the doses there transports them to client locations for use. Immediately prior to administration, the nuclear medicine technologist performs decay calculations based on the assay results to determine the dose to the patient. All waste is returned to the licensee's facility for decay in storage.

**PERFORMANCE OBSERVATIONS**

The inspector observed the mobile nuclear medicine technologist perform one cardiac stress test using technetium-99m, noting satisfactory use of dosimetry and ALARA practices. The technologist, through interviews, demonstrated adequate knowledge of radiation protection principles and NRC requirements specific to mobile nuclear medicine services. The inspector observed the technologist perform survey meter quality control, package receipt and return procedures and area surveys. The inspector also observed the technologist prepare equipment, including all remaining radioactive material, for transport upon completion of licensed activities at the client's location.

The inspector reviewed available licensee records for package receipt and return, area surveys, and the provision of mobile nuclear medicine service signed by the client's management. The inspector also reviewed the licensee's shipping papers, and verified that the emergency contact numbers were valid and provided adequate guidance in the event of an emergency.

No violations of NRC requirements were identified as a result of this inspection.