

Docket File Information
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

| | | | |
|--|---|---|--|
| 1. LICENSEE/LOCATION INSPECTED: InnerVision Advanced Medical Imaging 3801 Amelia Avenue Lafayette, IN 47905 REPORT NUMBER(S) 13-01 | | 2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352 | |
| 3. DOCKET NUMBER(S) 030-35552 | 4. LICENSE NUMBER(S) 13-32273-01 | 5. DATE(S) OF INSPECTION August 15, 2013 | |
| 6. INSPECTION PROCEDURES USED 87131 | | 7. INSPECTION FOCUS AREAS 3.01-3.07 | |

SUPPLEMENTAL INSPECTION INFORMATION

| | | | |
|-----------------------------|------------------|--|---------------------------------------|
| 1. PROGRAM CODE(S) 02200 | 2. PRIORITY 3 | 3. LICENSEE CONTACT Kent Lancaster, M.D., RSO | 4. TELEPHONE NUMBER (765) 447-7447 |
|-----------------------------|------------------|--|---------------------------------------|

Main Office Inspection Next Inspection Date: August 15, 2016

Field Office Inspection _____

Temporary Job Site Inspection _____

PROGRAM SCOPE

The licensee was a diagnostic clinic authorized for materials in Section 35.100, 35.200, and 35.300. The nuclear medicine department was staffed with two full-time nuclear medicine technologists (NMT) who performed approximately 45 diagnostic nuclear medicine procedures per week. These studies included a full spectrum of diagnostic imaging studies (primarily PET), and the clinic received unit doses from a local radiopharmacy. The licensee rarely performed therapies utilizing I-131 authorized under Section 35.300; on average, only one I-131 treatment was performed per year at the clinic. Additionally, the license employed an outside health physics consultant to audit the radiation safety program on a quarterly basis.

Performance Observations

During the time of the inspection, the inspector observed three patient administrations. The inspector observed an NMT assay a unit dose and use a syringe shield while administering the dose to the patient. The licensee had recently purchased a PET infusion machine; the inspector observed the NMTs use this machine to deliver doses to patients as well. The NMTs demonstrated: (1) package receipt procedures; (2) daily calibrator checks; (3) daily surveys; (4) security of byproduct material; (5) spill response procedures; and (6) an inventory check with not identified issues. The inspection also consisted of a review of select licensee records including: (1) linearity test records; (2) weekly wipe test results; (3) quarterly audit reports; (4) disposal logs; (5) Radiation Safety Committee (RSC) meeting minutes; (6) ALARA notification reports; (7) dosimetry reports; and (8) written directives. The inspector reviewed all of the written directives since the time of the last inspection (last treatments on 3/27/13 and 11/02/12); the directives contained appropriate information including the patient name, prescribed dose, administered dose, and had been signed and dated by an appropriate Authorized User. Of the written directives reviewed, the inspector observed administered doses that ranged from approximately 15-70 mCi of I-131.