



Designer and Manufacturer  
of  
Scientific and Industrial  
Instruments

# CERTIFICATE OF CALIBRATION

## LUDLUM MEASUREMENTS, INC.

501 Oak Street  
325-235-5494  
Sweetwater, TX 79556, U.S.A.

231 Sam Rayburn Parkway  
865-270-8962  
Lenoir City, TN 37771, U.S.A.

CUSTOMER CABRERA SERVICES ORDER NO. 20215937/387745

Mfg. Ludlum Measurements, Inc. Model 3 Serial No. 79511

Mfg. Ludlum Measurements, Inc. Model 44-9 Serial No. PR 137499

Cal. Date 5-Feb-13 Cal Due Date 5-Feb-14 Cal. Interval 1 Year Meterface 202-002

Check mark  applies to applicable instr. and/or detector IAW mfg. spec. T. 73 °F RH 20 % Alt 699.8 mm Hg

New Instrument  Instrument Received  Within Toler. +-10%  10-20%  Out of Tol.  Requiring Repair  Other-See comments

Mechanical ck.  Meter Zeroed  Background Subtract  Input Sens. Linearity

F/S Resp. ck.  Reset ck.  Window Operation  Geotropism

Audio ck.  Alarm Setting ck.  Batt. ck. (Min. Volt) 2.2 VDC

Calibrated in accordance with LMI SOP 14.8 rev 12/05/89.  Calibrated in accordance with LMI SOP 14.9 rev 02/07/97.

Instrument Volt Set 900 V Input Sens. 26 mV Det. Oper. 900 V at 26 mV Threshold Dial Ratio          =          mV

HV Readout (2 points) Ref./Inst.          /          V Ref./Inst.          /          V

### COMMENTS:

4 Pi EFF. FOR Tc99: 24.06 %  
Source S/N: 5296  
Background: 41 cpm.  
Source Count: 8,000 cpm.  
Source Activity: 33,077 dpm.

Gamma Calibration: GM detectors positioned perpendicular to source except for M 44-9 in which the front of probe faces source.

RANGE/MULTIPLIER	REFERENCE CAL. POINT	INSTRUMENT REC'D "AS FOUND READING"	INSTRUMENT METER READING*
x100	400 K cpm	<u>4K</u>	<u>4K</u>
x100	100 K cpm	<u>1K</u>	<u>1K</u>
x10	40 K cpm	<u>4K</u>	<u>4K</u>
x10	10 K cpm	<u>1K</u>	<u>1K</u>
x1	4 K cpm	<u>4K</u>	<u>4K</u>
x1	1 K cpm	<u>1K</u>	<u>1K</u>
x0.1	400 cpm	<u>3.9 K</u>	<u>4K</u>
x0.1	100 cpm	<u>1K</u>	<u>1K</u>

\*Uncertainty within ± 10% C.F. within ± 20%

ALL Range(s) Calibrated Electronically

REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*	REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*
Digital Readout					

Ludlum Measurements, Inc. certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology, or to the calibration facilities of other International Standards Organization members, or have been derived from accepted values of natural physical constants or have been derived by the ratio type of calibration techniques.  
The calibration system conforms to the requirements of ANSI/NCSL Z540-1-1994 and ANSI N323-1978  
State of Texas Calibration License No. LO-1963

Reference Instruments and/or Sources:  059  280  720  734  781  1131  1616  1696  5105  5717CO  5719CO  
 60646  70897  73410  E551  E552  G112  M565  S-394  S-1054  T-304  T879  T10081  T10082  Y982  
 Alpha S/N           Beta S/N           Other           
 m 500 S/N 189506  Oscilloscope S/N           Multimeter S/N 3810039

Calibrated By: William Tinsley Date 5 February - 2013  
Reviewed By: Donnie Miexos Date 7-Feb-13

AC Inst.  Passed Dielectric (Hi-Pot) and Continuity Test  
Only  Failed:



Designer and Manufacturer  
of  
Scientific and Industrial  
Instruments

**LUDLUM MEASUREMENTS, INC.**

501 Oak Street  
325-235-5494  
Sweetwater, TX 79556, U.S.A.

231 Sam Rayburn Parkway  
865-270-8962  
Lenoir City, TN 37771, U.S.A.

CONVERSION CHART

Customer CABRERA SERVICES Date 5-Feb-13 Order #. 20215937/387745

Model 3 Serial No. 79511 Detector Model 44-9 Serial No. PR 137499

Source Cs<sup>137</sup>: 194.6 mCi, Cs<sup>137</sup>: 21 mCi High Voltage 900 V  
Input Sensitivity 26 mV

Reference Point	"As Found" Readings (CPM):		After Adjustment Readings (CPM):	
	Meter Reading	Range/Scale	Meter Reading	Range/Scale
150 mR/hr	1.6 K	X 100	1.6 K	X 100
50 mR/hr	1 K	X 100	1 K	X 100
15 mR/hr	4.1 K	X 10	4.1 K	X 10
5 mR/hr	1.6 K	X 10	1.6 K	X 10
1.5 mR/hr	0.5 K	X 10	0.5 K	X 10
1.0 mR/hr	3.2 K	X 1	3.2 K	X 1

Signature: William Tensley Date 5-February 2013