


Radiological Survey Sheet

Job Location Pfizer GEORON CT Page 1 of 23
 Survey Purpose CAB DECON 118 211 DECOMMISSION WHC Date 12/11/05
 Performed by P. KEEFE (Print)  (Sign)

Instrument (Model/S/N)	Packard Tri-Carb S/N 431521	Packard Tri-Carb S/N 431520	Ludlum 2241-2 163603	Wallac Model 1409 S/N 4061042
Det. (Model/SN)	Internal	Internal	Ludlum 43-68 PR 149613	Internal
Type Rad.	β	β	β	β
Cal. Due:	10/12/06	10/12/06	04/14/06	12/01/06
Check Instruments Used			+	+

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)	
1.	1110	BACKGROUND		
2.		CAB ENTRANCE DOOR	- OUTSIDE	
3.			- INSIDE	
4.		SM WALL	- UPPER	
5.			- LOWER	
6.		SM WALL	- UPPER	
7.			- LOWER	
8.		BENCH TOP		
9.		SPLASH GUARD		
10.		DRAWERS	- TOP - OUT IN	
11.			- UPPER MID - ↓ ↓	
12.			- LOWER MID - ↓ ↓	
13.			- BOTTOM - OUT IN	
14.		HOOD 118E HOOD 047	- OUTSIDE - LEFT - UPPER	
15.			- LOWER	
16.			- FRONT - TOP - LEFT	
17.		↓		- RIGHT
18.		1140		VENT AREA

Radiological Survey Continuation Sheet

Page 2 of 23

Date 12/11/05

LAB DECON 118E 211

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
19	1140	HOOD - OUTSIDE - CONTROLS	- LEFT
20			- RIGHT
21			- GLASS SLASH - LEFT
22			- RIGHT
23			- SPLASH GUARD - LEFT
24			- RIGHT
25		(CABINET - LEFT - DOORS (OUTSIDE)	- LEFT
26			- RIGHT
27			- CENTER - DOORS (OUTSIDE) - LEFT
28			- RIGHT
29			- RIGHT - DOORS (OUTSIDE) - LEFT
30			- RIGHT
31		HOOD - INSIDE - LEFT	- UPPER
32			- MID
33			- BOTTOM
34			- REAR (LEFT) - UPPER
35			- MID
36			- BOTTOM
37			(RIGHT) - UPPER
38			- MID
39			- BOTTOM
40			- RIGHT - UPPER
41			- MID
42			- BOTTOM
43			- BENCHTOP (LEFT) - LEFT
44	▽		- MID
45	1200		- RIGHT

- See Note.

Radiological Survey Continuation Sheet

Page 3 of 23

Date 12/11/05

LAB DECON 118 211#

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
46	1250	HOOD - INSIDE	- BENCHTOP (RIGHT) - LEFT
47			- MID
48			- RIGHT
49			- BASIN - LEFT / RIGHT
50			- FRONT / REAR
51			- BOTTOM / DRAIN
52		CABINET (LEFT) INSIDE	- LEFT / RIGHT
53			- REAR / BOTTOM
54			- DOORS
55		(CENTER) INSIDE	- LEFT / RIGHT
56			- REAR / BOTTOM
57			- DOORS
58		(RIGHT) INSIDE	- LEFT / RIGHT
59			- REAR / BOTTOM
60			- DOORS
61		GLASS RACK	
62		WALL SWIPE	
63		PAPER TOWEL DISPENSER	
64		FACET	
65		BENCHTOP AROUND BASIN	
66		BACKSPASH	
67		BASIN	- LEFT / RIGHT
68			- FRONT / REAR
69			- BOTTOM / DRAIN
70		CABINET - OUTSIDE DOORS	- LEFT
71	▽		- RIGHT
72	1310		- INSIDE - LEFT / RIGHT

Radiological Survey Continuation Sheet

Page 4 of 23

Date 12/11/05

LAB DECOW 118 211

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
73	1310	CABINET - INSIDE - REAR / BOTTOM	
74		- DOORS	
75		EYEWASH	
76		OVERHEAD CABINET - OUTSIDE - LEFT	
77		- BOTTOM	
78		- DOORS - LEFT	
79		- RIGHT	
80		BENCHTOP	
81		DRAWS - TOP - OUT / IN (LEFT)	
82		- UPPER MID ↓ ↓ ↓	
83		- LOWER MID ↓ ↓ ↓	
84		- BOTTOM - OUT / IN ↓	
85		DRAWS - UPPER - OUT / IN (RIGHT)	
86		- LOWER - OUT / IN ↓	
87		CABINET - DOOR - OUT / IN	
88		- INSIDE - LEFT / RIGHT	
89		- REAR / BOTTOM	
90		OVERHEAD CABINET - OUTSIDE - RIGHT	
91		- BOTTOM	
92		- DOORS - LEFT	
93		- RIGHT	
94		- INSIDE - LEFT / RIGHT	
95		- REAR / BOTTOM	
96		- DOORS	
897		BENCHTOP	
98	▽	- DRAWS (LEFT) - TOP - OUT / IN	
99	1330	- UPPER MID - OUT / IN	

Radiological Survey Continuation Sheet

Page 5 of 23

Date 12/11/05

(AB DECON 118 211

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
100	1330	DRAWS (LEFT) - LOWER MID	- OUT/IN
101			- BOTTOM - OUT/IN
102		DRAWS (RIGHT) - UPPER	- OUT/IN
103			- LOWER - OUT/IN
104		CABINET - DOOR	- OUT/IN
105			- INSIDE - LEFT / REAR
106			- RIGHT / BOTTOM
107		WALL Shelves - TOP	
108			- MID
109			- BOTTOM
110		SUPPORTS	
111		Wall under shelves - upper	
112			- lower
113		GLASS DOOR - OUTSIDE	
114			- INSIDE
115		OFFICE shelves	
116		DESKTOP SWIPE	
117		DRAWS (LEFT) - OUT/IN	
118		(RIGHT) - OUT/IN	
119		OFFICE shelves	
120		DESKTOP SWIPE	
121		DRAWS (LEFT) - OUT/IN	
122		(RIGHT) - OUT/IN	
123		Wall swiipe - upper	
124			- mid
125	▽		- lower
126	1350	GLASS DOOR - OUTSIDE	

Radiological Survey Continuation Sheet

Page 6 of 23

Date 12/11/05

(LAB DECON) 118 211

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
127	1350	Glass door - INSIDE	
128		White board	
129		File CABINET (LEFT) -OUT	
130			-IN
131			(RIGHT) -OUT
132			-IN
133		Shelves	
134		SUPPORTS	
135		DESKTOP	
136		DRAWERS (LEFT) -OUT	
137			-IN
138			(RIGHT) -OUT
139			-IN
140		Wall swive - UPPER	
141			- LOWER
142		Overhead CABINET - OUTSIDE - LEFT	
143			- BOTTOM
144			- DOORS
145			- INSIDE - LEFT / RIGHT
146			- REAR / BOTTOM
147			- DOORS
148		BENCHTOP	
149		DRAWERS - TOP - OUT / IN	
150			- UPPER MID ↓ ↓
151			- MID ↓ ↓
152	▽		- LOWER MID ↓ ↓
153	1420		- BOTTOM - OUT / IN

Radiological Survey Continuation Sheet

Page 7 of 23

Date 12/11/05

LAB DECON 118 211

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
154	1420	Accore under BENCH	- LEFT / RIGHT
155			- REAR / FLOOR
156		Overhead CABINET	- OUTSIDE - BOTTOM
157			- DOORS
158			- INSIDE - LEFT / RIGHT
159			- REAR / BOTTOM
160			- shelves
161		BENCH TOP	
162		DRAWERS - LEFT	- OUT / IN
163			- RIGHT - OUT / IN
164		CABINET - DOORS	- OUT / IN
165			- INSIDE - LEFT / RIGHT
166			- REAR / BOTTOM
167			- shelf
168		Overhead CABINET - OUTSIDE	- DOORS
169			- BOTTOM
170			- INSIDE - LEFT / RIGHT
171			- REAR / BOTTOM
172			- shelves
173		BENCH TOP	
174		DRAWERS - TOP	- OUT / IN
175			- upper mid ↓ ↓
176			- lower mid ↓ ↓
177			- BOTTOM - OUT / IN
178		CABINET - DOOR	- OUT / IN
179			- INSIDE - LEFT / RIGHT
180	1446		- REAR / BOTTOM

Radiological Survey Continuation Sheet

Page 8 of 23

Date 12/01/05

LAB DECON 118211

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
181	1440	BENCHTOP SWIPE	
182		DRAWS - LEFT - TOP	- OUT/IN
183			- UPPER MID ↓ ↓
184			- LOWER MID ↓ ↓
185			- BOTTOM - OUT/IN
186			- RIGHT - UPPER - OUT/IN
187			- LOWER ↓ ↓
188		CABINET - OUTSIDE	- DOOR
189			- RIGHT
190			- INSIDE - LEFT / RIGHT
191			- REAR / BOTTOM
192			- Shelves
193		BENCHTOP Shelf	
194		SUPPORTS	
195		BENCHTOP	
196		DRAWS - LEFT	- OUT/IN
197			- RIGHT - OUT/IN
198		ALCOVE - LEFT / RIGHT	
199			- REAR / FLOOR
200		BENCHTOP Shelf.	
201		SUPPORTS	
202		BENCHTOP	
203		DRAWS - LEFT - TOP	- OUT/IN
204			- UPPER MID ↓ ↓
205			- LOWER MID ↓ ↓
206	▽		- BOTTOM - OUT/IN
207	1530		- RIGHT - TOP - OUT/IN


Radiological Survey Continuation Sheet

Page 9 of 23

Date 12/11/05

(LAB DECON 118 211

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
208	1510	DRAWS - RIGHT - LOWER	
209		CABINET - DOOR - OUTSIDE	
210		- INSIDE - DOOR	
211		- LEFT / RIGHT	
212		- REAR / BOTTOM	
213		- shelves	
214		BENCHTOP	
215		DRAWS - LEFT - TOP - OUT/IN	
216		- UPPER MID ↓ ↓	
217		- lower mid ↓ ↓	
218		- Bottom - OUT/IN	
219		DRAWS - RIGHT - UPPER - OUT/IN	
220		- lower ↓ ↓	
221		BENCHTOP	
222		ALCOVE - LEFT / RIGHT	
223		- REAR / FLOOR	
224		BENCHTOP	
225		DRAWS - LEFT - TOP - OUT/IN	
226		- UPPER MID ↓ ↓	
227		- lower mid ↓ ↓	
228		- Bottom - OUT/IN	
229		- RIGHT - UPPER ↓ ↓	
230		- lower ↓ ↓	
231		CABINET - OUTSIDE - DOOR	
232		- LEFT	
233	↓	- INSIDE - LEFT / RIGHT	
234	1530	- REAR / BOTTOM	



ALL SURFACES WERE SURVEYED WITH Ludlum model 2241-C AND FOUND TO BE WITHIN LIMITS.

○ = wipe location

PROTOCOL : 97 118 211
 DATE : 2005/12/12
 TIME : 11:39
 ID : P97AS007

Lab Decon

Wallac 1400 DSA ver 2.50 S/N 4091042

Counting mode : CPM
 Isotope(s) : H3,C14,K42
 H3 = 5- 350,12.43 y
 C14 = 5- 660,5730.00 y
 K42 = 5-1024,12.36 h
 Protocol name : 118 211
 Counting time : 60
 Repeats : 1
 Cycles : 1
 Replicates : 1
 2 sigma % : 0.01
 Minimum cpm : 0.00 Checking time: 10
 Advanced modes : Chemilum
 Output to Display :
 POS,DATE,TIME,CPM1,CPM2,CPM3,RACKPOS,CPM,CTIME,SQPI
 Additions to Display : Header,Listing,Spectrum
 Header : Lab Decon
 Spectrum : Beta
 Window 1 : 1-1024 /Beta
 Window 2 : 1-1024 /Beta
 Window 3 : 1-1024 /Beta
 Window 4 : 1-1024 /Beta
 Window 5 : 1-1024 /Beta
 Window 6 : 1-1024 /Beta
 FNCT1 = FNCT1 :
 FNCT2 = FNCT2 :
 FNCT3 = FNCT3 :
 FNCT4 = FNCT4 :

POS	DATE	TIME	CPM1	CPM2	CPM3
1	12/12/2005	11:39 AM	18.80	29.20	43.80
2	12/12/2005	11:41 AM	22.90	28.10	33.30
3	12/12/2005	11:42 AM	22.90	30.10	42.60
4	12/12/2005	11:44 AM	17.40	23.60	30.70
5	12/12/2005	11:46 AM	24.90	35.30	42.60

BACKGROUND

POS	DATE	TIME	CPM1	CPM2	CPM3
6	12/12/2005	11:47 AM	29.20	34.40	38.50
7	12/12/2005	11:49 AM	14.50	21.80	28.00
8	12/12/2005	11:50 AM	19.80	26.10	30.20
9	12/12/2005	11:52 AM	17.70	24.90	27.00
10	12/12/2005	11:53 AM	18.70	27.00	35.30
11	12/12/2005	11:55 AM	21.80	28.00	35.30
12	12/12/2005	11:56 AM	18.70	31.20	39.50
13	12/12/2005	11:58 AM	18.70	27.00	30.10
14	12/12/2005	12:00 PM	16.60	20.80	26.00
15	12/12/2005	12:01 PM	20.80	30.10	36.30
16	12/12/2005	12:03 PM	17.70	28.10	31.20
17	12/12/2005	12:04 PM	12.50	16.60	22.80
18	12/12/2005	12:06 PM	14.50	21.80	30.10
19	12/12/2005	12:07 PM	21.80	27.00	29.10
20	12/12/2005	12:09 PM	18.80	25.00	33.30
21	12/12/2005	12:11 PM	22.90	28.10	35.30
22	12/12/2005	12:12 PM	21.80	24.90	24.90
23	12/12/2005	12:14 PM	17.70	26.00	34.30
24	12/12/2005	12:15 PM	26.00	33.20	45.70
25	12/12/2005	12:17 PM	17.70	23.90	30.10
26	12/12/2005	12:18 PM	10.40	12.50	16.60
27	12/12/2005	12:20 PM	19.80	25.00	32.30
28	12/12/2005	12:21 PM	26.00	35.30	40.50
29	12/12/2005	12:23 PM	20.80	27.10	34.40
30	12/12/2005	12:25 PM	8.30	16.70	20.80
31	12/12/2005	12:26 PM	18.70	24.90	28.10
32	12/12/2005	12:28 PM	29.10	35.30	38.40
33	12/12/2005	12:29 PM	22.90	30.10	34.30
34	12/12/2005	12:31 PM	9.40	17.70	28.10
35	12/12/2005	12:32 PM	14.60	25.00	30.20
36	12/12/2005	12:34 PM	32.30	38.50	44.80

POS	DATE	TIME	CPM1	CPM2	CPM3
37	12/12/2005	12:36 PM	21.80	30.10	36.30
38	12/12/2005	12:37 PM	18.70	22.90	28.10
39	12/12/2005	12:39 PM	19.70	32.20	38.40
40	12/12/2005	12:40 PM	22.80	28.00	36.30
41	12/12/2005	12:42 PM	34.30	45.70	51.90
42	12/12/2005	12:43 PM	13.50	14.50	14.50
43	12/12/2005	12:45 PM	69.60	85.20	88.40
44	12/12/2005	12:46 PM	19.70	27.00	32.20
45	12/12/2005	12:48 PM	16.60	22.80	26.00
46	12/12/2005	12:50 PM	23.90	32.20	38.50
47	12/12/2005	12:51 PM	35.30	42.60	53.00
48	12/12/2005	12:53 PM	29.10	38.50	41.60
49	12/12/2005	12:54 PM	18.70	27.00	32.20
50	12/12/2005	12:56 PM	19.70	27.00	33.20
51	12/12/2005	12:57 PM	16.40	23.60	31.70
52	12/12/2005	12:59 PM	22.80	28.00	36.30
53	12/12/2005	1:01 PM	10.40	15.60	20.80
54	12/12/2005	1:02 PM	20.80	26.00	31.20
55	12/12/2005	1:04 PM	15.60	20.80	28.10
56	12/12/2005	1:05 PM	18.70	21.80	24.90
57	12/12/2005	1:07 PM	19.80	31.20	36.50
58	12/12/2005	1:08 PM	17.70	28.00	36.30
59	12/12/2005	1:10 PM	17.60	20.80	31.10
60	12/12/2005	1:12 PM	20.80	24.90	28.00
61	12/12/2005	1:13 PM	14.60	19.80	28.10
62	12/12/2005	1:15 PM	19.70	21.80	21.80
63	12/12/2005	1:16 PM	13.50	18.70	26.00
64	12/12/2005	1:18 PM	18.70	23.90	32.20
65	12/12/2005	1:19 PM	13.50	26.00	34.40
66	12/12/2005	1:21 PM	21.80	28.00	37.40
67	12/12/2005	1:22 PM	12.50	18.70	26.00

POS	DATE	TIME	CPM1	CPM2	CPM3
68	12/12/2005	1:24 PM	18.70	21.80	26.00
69	12/12/2005	1:26 PM	13.50	21.80	27.00
70	12/12/2005	1:27 PM	9.30	19.70	32.20
71	12/12/2005	1:29 PM	21.80	30.10	40.50
72	12/12/2005	1:30 PM	14.60	24.90	30.10
73	12/12/2005	1:32 PM	26.00	29.10	33.20
74	12/12/2005	1:33 PM	12.50	20.80	31.20
75	12/12/2005	1:35 PM	5.20	12.50	22.90
76	12/12/2005	1:37 PM	14.50	18.70	28.00
77	12/12/2005	1:38 PM	21.80	31.20	34.30
78	12/12/2005	1:40 PM	24.90	29.10	35.30
79	12/12/2005	1:41 PM	24.00	29.20	31.20
80	12/12/2005	1:43 PM	13.50	19.80	26.00
81	12/12/2005	1:44 PM	26.00	32.20	38.40
82	12/12/2005	1:46 PM	27.00	30.10	30.10
83	12/12/2005	1:47 PM	11.40	18.70	19.70
84	12/12/2005	1:49 PM	22.90	32.30	42.70
85	12/12/2005	1:51 PM	21.90	25.00	32.30
86	12/12/2005	1:52 PM	29.10	33.20	40.50
87	12/12/2005	1:54 PM	18.70	24.90	30.10
88	12/12/2005	1:55 PM	17.70	24.00	29.20
89	12/12/2005	1:57 PM	24.00	30.20	36.50
90	12/12/2005	1:58 PM	22.90	33.30	39.50
91	12/12/2005	2:00 PM	11.50	19.80	27.10
92	12/12/2005	2:02 PM	26.00	33.20	44.70
93	12/12/2005	2:03 PM	28.00	37.40	44.70
94	12/12/2005	2:05 PM	18.70	22.90	30.10
95	12/12/2005	2:06 PM	10.40	15.60	19.80
96	12/12/2005	2:08 PM	23.60	28.70	34.80
97	12/12/2005	2:09 PM	19.70	24.90	30.10
98	12/12/2005	2:11 PM	16.60	28.10	38.40

POS	DATE	TIME	CPM1	CPM2	CPM3
99	12/12/2005	2:13 PM	12.50	26.10	35.40
100	12/12/2005	2:14 PM	21.80	27.00	35.30
101	12/12/2005	2:16 PM	22.90	31.30	37.50
102	12/12/2005	2:17 PM	15.60	24.90	33.30
103	12/12/2005	2:19 PM	15.60	21.80	29.10
104	12/12/2005	2:20 PM	11.50	17.70	24.00
105	12/12/2005	2:22 PM	22.90	30.10	33.20
106	12/12/2005	2:23 PM	18.70	20.80	24.90
107	12/12/2005	2:25 PM	15.60	27.10	33.40
108	12/12/2005	2:27 PM	25.00	35.40	41.70
109	12/12/2005	2:28 PM	20.80	28.00	35.30
110	12/12/2005	2:30 PM	16.60	27.00	34.30
111	12/12/2005	2:31 PM	22.90	30.10	34.30
112	12/12/2005	2:33 PM	17.70	26.00	30.10
113	12/12/2005	2:34 PM	31.30	36.50	39.60
114	12/12/2005	2:36 PM	27.10	32.30	35.40
115	12/12/2005	2:37 PM	21.80	30.10	36.40
116	12/12/2005	2:39 PM	19.80	26.00	31.30
117	12/12/2005	2:41 PM	29.20	36.50	44.80
118	12/12/2005	2:42 PM	15.60	22.90	28.10
119	12/12/2005	2:44 PM	19.70	23.90	29.10
120	12/12/2005	2:45 PM	14.60	22.90	29.10
121	12/12/2005	2:47 PM	18.80	21.90	27.10
122	12/12/2005	2:48 PM	28.10	32.20	37.40
123	12/12/2005	2:50 PM	27.00	33.20	35.30
124	12/12/2005	2:52 PM	27.10	34.40	37.50
125	12/12/2005	2:53 PM	20.80	23.90	27.00
126	12/12/2005	2:55 PM	17.70	26.00	29.10
127	12/12/2005	2:56 PM	9.40	15.60	20.80
128	12/12/2005	2:58 PM	16.60	19.70	29.10
129	12/12/2005	2:59 PM	22.90	28.10	32.30

POS	DATE	TIME	CPM1	CPM2	CPM3
130	12/12/2005	3:01 PM	29.10	36.40	40.50
131	12/12/2005	3:03 PM	19.70	28.00	33.20
132	12/12/2005	3:04 PM	26.10	33.40	39.60
133	12/12/2005	3:06 PM	32.20	37.40	48.80
134	12/12/2005	3:07 PM	13.50	15.60	24.00
135	12/12/2005	3:09 PM	17.70	24.90	29.10
136	12/12/2005	3:10 PM	16.60	23.90	27.00
137	12/12/2005	3:12 PM	25.00	32.30	35.40
138	12/12/2005	3:13 PM	16.60	22.90	24.90
139	12/12/2005	3:15 PM	14.50	23.90	30.10
140	12/12/2005	3:17 PM	20.80	26.00	34.30
141	12/12/2005	3:18 PM	24.90	29.10	40.50
142	12/12/2005	3:20 PM	14.60	18.70	22.90
143	12/12/2005	3:21 PM	19.70	26.00	31.20
144	12/12/2005	3:23 PM	24.90	36.40	40.50
145	12/12/2005	3:24 PM	12.50	16.70	24.00
146	12/12/2005	3:26 PM	27.00	32.20	34.30
147	12/12/2005	3:28 PM	22.80	33.20	37.40
148	12/12/2005	3:29 PM	20.80	29.10	35.30
149	12/12/2005	3:31 PM	17.70	21.80	30.10
150	12/12/2005	3:32 PM	28.00	36.40	37.40
151	12/12/2005	3:34 PM	18.70	29.10	33.30
152	12/12/2005	3:35 PM	30.10	38.40	43.60
153	12/12/2005	3:37 PM	17.70	22.80	26.00
154	12/12/2005	3:39 PM	16.60	22.90	27.00
155	12/12/2005	3:40 PM	30.20	36.50	39.60
156	12/12/2005	3:42 PM	19.70	24.90	30.10
157	12/12/2005	3:43 PM	19.70	27.00	32.20
158	12/12/2005	3:45 PM	20.80	22.90	28.00
159	12/12/2005	3:46 PM	19.70	27.00	31.20
160	12/12/2005	3:48 PM	21.80	27.00	34.30

POS	DATE	TIME	CPM1	CPM2	CPM3
161	12/12/2005	3:49 PM	15.60	23.90	29.10
162	12/12/2005	3:51 PM	11.40	15.60	18.70
163	12/12/2005	3:53 PM	19.70	27.00	33.20
164	12/12/2005	3:54 PM	17.60	19.70	22.80
165	12/12/2005	3:56 PM	16.60	20.80	28.00
166	12/12/2005	3:57 PM	18.70	24.00	27.10
167	12/12/2005	3:59 PM	18.70	19.70	27.00
168	12/12/2005	4:00 PM	15.60	23.90	31.20
169	12/12/2005	4:02 PM	19.70	26.00	33.20
170	12/12/2005	4:04 PM	22.80	30.10	34.30
171	12/12/2005	4:05 PM	20.80	26.00	30.10
172	12/12/2005	4:07 PM	13.50	18.70	21.80
173	12/12/2005	4:08 PM	14.60	22.90	32.30
174	12/12/2005	4:10 PM	30.10	37.40	40.50
175	12/12/2005	4:11 PM	18.80	26.00	33.30
176	12/12/2005	4:13 PM	20.80	27.00	31.20
177	12/12/2005	4:14 PM	22.90	27.00	38.40
178	12/12/2005	4:16 PM	21.90	25.00	31.30
179	12/12/2005	4:18 PM	13.50	18.70	28.00
180	12/12/2005	4:19 PM	26.00	33.20	42.60
181	12/12/2005	4:21 PM	21.80	30.10	39.50
182	12/12/2005	4:22 PM	21.80	28.10	34.30
183	12/12/2005	4:24 PM	24.90	35.30	43.70
184	12/12/2005	4:26 PM	23.90	28.00	37.40
185	12/12/2005	4:27 PM	13.50	17.70	27.00
186	12/12/2005	4:29 PM	18.70	28.00	34.30
187	12/12/2005	4:30 PM	16.60	27.00	36.40
188	12/12/2005	4:32 PM	19.70	26.00	34.30
189	12/12/2005	4:33 PM	9.40	13.60	21.90
190	12/12/2005	4:35 PM	16.60	20.80	24.90
191	12/12/2005	4:36 PM	27.10	33.40	40.70

POS	DATE	TIME	CPM1	CPM2	CPM3
192	12/12/2005	4:38 PM	19.70	24.90	30.10
193	12/12/2005	4:40 PM	11.40	18.70	22.90
194	12/12/2005	4:41 PM	15.60	20.80	27.00
195	12/12/2005	4:43 PM	26.00	33.20	35.30
196	12/12/2005	4:44 PM	27.00	35.30	38.40
197	12/12/2005	4:46 PM	16.60	21.80	28.10
198	12/12/2005	4:47 PM	23.90	34.30	39.50
199	12/12/2005	4:49 PM	11.40	21.80	28.00
200	12/12/2005	4:51 PM	22.90	31.20	38.40
201	12/12/2005	4:52 PM	25.00	33.40	35.40
202	12/12/2005	4:54 PM	24.90	34.30	37.40
203	12/12/2005	4:55 PM	31.20	40.50	45.70
204	12/12/2005	4:57 PM	9.40	13.50	17.70
205	12/12/2005	4:58 PM	13.50	20.80	25.00
206	12/12/2005	5:00 PM	20.80	28.10	36.40
207	12/12/2005	5:01 PM	15.60	23.90	30.10
208	12/12/2005	5:03 PM	14.60	17.70	20.80
209	12/12/2005	5:05 PM	19.80	26.10	36.50
210	12/12/2005	5:06 PM	19.70	26.00	28.10
211	12/12/2005	5:08 PM	21.80	31.20	35.30
212	12/12/2005	5:09 PM	19.70	27.00	31.20
213	12/12/2005	5:11 PM	33.80	38.90	44.10
214	12/12/2005	5:12 PM	15.60	26.00	33.30
215	12/12/2005	5:14 PM	24.00	27.10	31.30
216	12/12/2005	5:15 PM	22.90	25.00	36.50
217	12/12/2005	5:17 PM	16.70	21.90	27.10
218	12/12/2005	5:19 PM	19.80	26.10	31.30
219	12/12/2005	5:20 PM	20.80	29.10	33.20
220	12/12/2005	5:22 PM	16.60	27.00	31.20
221	12/12/2005	5:23 PM	26.00	37.40	43.60
222	12/12/2005	5:25 PM	23.90	31.20	36.40

POS	DATE	TIME	CPM1	CPM2	CPM3
223	12/12/2005	5:26 PM	18.70	27.00	34.30
224	12/12/2005	5:28 PM	19.70	28.10	36.40
225	12/12/2005	5:30 PM	23.90	28.00	31.20
226	12/12/2005	5:31 PM	26.10	30.20	36.50
227	12/12/2005	5:33 PM	13.50	19.70	22.90
228	12/12/2005	5:34 PM	19.70	26.00	33.20
229	12/12/2005	5:36 PM	24.00	34.40	43.80
230	12/12/2005	5:37 PM	17.70	22.90	28.00
231	12/12/2005	5:39 PM	26.00	30.10	36.30
232	12/12/2005	5:40 PM	18.70	29.10	32.20
233	12/12/2005	5:42 PM	23.90	32.20	39.50
234	12/12/2005	5:44 PM	13.50	17.70	24.90
235	12/12/2005	5:45 PM	20.80	31.30	36.50
236	12/12/2005	5:47 PM	26.10	36.50	44.80
237	12/12/2005	5:48 PM	27.00	34.30	37.40
238	12/12/2005	5:50 PM	27.00	32.20	40.50
239	12/12/2005	5:51 PM	18.70	23.90	27.00
240	12/12/2005	5:53 PM	28.10	30.20	40.60
241	12/12/2005	5:55 PM	15.60	22.90	28.10
242	12/12/2005	5:56 PM	20.80	26.00	30.10
243	12/12/2005	5:58 PM	28.10	37.50	44.80
244	12/12/2005	5:59 PM	12.50	26.10	33.40
245	12/12/2005	6:01 PM	19.80	27.10	30.20
246	12/12/2005	6:02 PM	20.80	29.20	36.50
247	12/12/2005	6:04 PM	22.90	31.20	39.50
248	12/12/2005	6:05 PM	14.50	20.80	30.10
249	12/12/2005	6:07 PM	21.80	34.30	36.40
250	12/12/2005	6:09 PM	18.70	23.90	29.10
251	12/12/2005	6:10 PM	23.90	33.30	40.50
252	12/12/2005	6:12 PM	26.00	33.30	41.60
253	12/12/2005	6:13 PM	27.10	34.40	44.80

POS	DATE	TIME	CPM1	CPM2	CPM3
254	12/12/2005	6:15 PM	24.00	28.10	38.60
255	12/12/2005	6:16 PM	12.50	20.80	23.90
256	12/12/2005	6:18 PM	25.00	33.30	39.60
257	12/12/2005	6:20 PM	24.90	33.20	36.40
258	12/12/2005	6:21 PM	19.80	27.10	30.20
259	12/12/2005	6:23 PM	23.90	36.30	43.60
260	12/12/2005	6:24 PM	21.80	29.10	37.40

PROTOCOL : 97 118 211
 DATE : 2005/12/13
 TIME : 08:19
 ID : P97AS008

Lab Decon

Wallac 1400 DSA ver 2.50 S/N 4091042

Counting mode : CPM
 Isotope(s) : H3,C14,K42
 H3 = 5- 350,12.43 y
 C14 = 5- 660,5730.00 y
 K42 = 5-1024,12.36 h
 Protocol name : 118 211
 Counting time : 60
 Repeats : 1
 Cycles : 1
 Replicates : 1
 2 sigma % : 0.01
 Minimum cpm : 0.00 Checking time: 10
 Advanced modes : Chemilum
 Output to Display :
 POS,DATE,TIME,CPM1,CPM2,CPM3,RACKPOS,CPM,CTIME,SQPI
 Additions to Display : Header,Listing,Spectrum
 Header : Lab Decon
 Spectrum : Beta
 Window 1 : 1-1024 /Beta
 Window 2 : 1-1024 /Beta
 Window 3 : 1-1024 /Beta
 Window 4 : 1-1024 /Beta
 Window 5 : 1-1024 /Beta
 Window 6 : 1-1024 /Beta
 FNCT1 = FNCT1 :
 FNCT2 = FNCT2 :
 FNCT3 = FNCT3 :
 FNCT4 = FNCT4 :

Total count rate:

H3 40.5 CPM
 C14 48.8 CPM
 K42 63.4 CPM

POS	DATE	TIME	CPM1	CPM2	CPM3
1	12/13/2005	8:19 AM	12.50	17.70	24.90

Background

POS	DATE	TIME	CPM1	CPM2	CPM3	
* 2	12/13/2005	8:21 AM	28.10	31.20	38.50	RESWIPE # 43

NOTE:

* AREA WAS DECONED AND RESWIPE.