

Radiological Survey Sheet

Job Location Pfizer, Inc. Groton, Ct. 06340 118 / 214 Page 1 5
 Survey Purpose Radiologic Survey of Equipment LAB, DECOM, Date 9-27-12
 Performed by E. SAILOR E. Sailor
 (Print) (Sign)

Instrument (Model/S/N)	Packard Tri-Carb S/N 431521	Packard Tri-Carb S/N 431520	#2 Ludlum 2241-2 163603	Wallac Model 1409 S/N 4061042
Det. (Model/SN)	Internal	Internal	Ludlum 43-68 PR 149613	Internal
Type Rad.	β	β	β	β
Cal. Due:	03/05/2013	03/05/2013	04/17/2013	09/15/2012
Check Instruments Used	✓		✓	

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
1.	0800	BACKGROUND	
2.		door	
3.		FLOOR	
4.		UPPER WALL	
5.		LOWER "	
6.		UPPER "	
7.		LOWER "	
8.		UPPER "	
9.		LOWER "	
10.		UPPER "	
11.		LOWER "	
12.		UPPER "	
13.		LOWER "	
14.		UPPER "	
15.		LOWER "	
16.		UPPER "	
17.		LOWER "	
18.	✓	UPPER	

Radiological Survey Continuation Sheet

Page 2 of 5


Date 9-27-12

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
19	080	lower wall	
20		upper ✓	
21		lower "	
22		RACKS	
23		"	
24		"	
25		"	
26		"	
27		"	
28		"	
29		"	
30		"	
31		FLOOR	
32		"	
33		"	
34		"	
35		"	
36		"	
37		"	
38		"	
39		"	
40		"	
41		"	
42	✓	"	

118 / 214

pg 3 of 5

9-27-12



Surveyed 100% of lower surfaces with instrument #2.

no problems noted.

○ = wipe location.

Protocol# 4 - Eds Monthly Swipes.lsa

User: Ed Gailor

Assay Definition-

Assay Description:
Equipment Swipes

Assay Type: CPM
Report Name: Report1
Output Data Path: C:\Packard\Tricarb\Results\Ed Gailor\Eds Monthly Swipes
Raw Results Path: C:\Packard\Tricarb\Results\Ed Gailor\Eds Monthly Swipes\20120927_0848
\20120927_0848.results
RTF File Name: C:\Packard\Tricarb\Results\Ed Gailor\Eds Monthly Swipes\Ed's Monthly Swipes.rtf
Assay File Name: C:\Packard\TriCarb\Assays\Eds Monthly Swipes.lsa

Count Conditions-

Nuclide: Triple
Quench Indicator: tSIE/AEC
External Std Terminator (sec): 5 sec
Pre-Count Delay (min): 0.00
Quench Set: n/a
Count Time (min): 1.00
Count Mode: Normal
Assay Count Cycles: 1 Repeat Sample Count: 1
#Vials/Sample: 1 Calculate % Reference: Off

Background Subtract: On - Manual
Low CPM Threshold: Off
2 Sigma % Terminator: Off

Regions	LL	UL	Bkg Subtract
A	0.0	18.6	0.00
B	18.6	156.0	0.00
C	0.0	2000.0	0.00

Count Corrections-

Static Controller: On Luminescence Correction: On
Colored Samples: n/a Heterogeneity Monitor: n/a
Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

S#	CPMA	CPMB	CPMC	MESSAGES
1 <i>BE</i>	18	8	32	
2	12	10	25	
3	18	9	32	
4	6	4	24	
5	20	7	34	
6	1	13	23	
7	11	8	24	
8	11	8	23	
9	6	3	16	
10	16	7	25	

Protocol# 4 - Eds Monthly Swipes.lsa

User: Ed ^{Garlor} ~~Garlor~~

11	11	8	27
12	9	5	17
13	6	4	15
14	17	13	38
15	20	10	38
16	10	4	23
17	14	6	23
18	17	6	31
19	5	12	21
20	14	6	27
21	8	7	23
22	3	6	11
23	1	8	14
24	8	6	16
25	9	11	27
26	8	8	23
27	16	8	27
28	11	4	21
29	11	11	28
30	6	10	22
31	6	10	23
32	0	8	13
33	11	14	32
34	10	10	27
35	15	7	27
36	9	3	19
37	16	6	25
38	11	9	30
39	10	4	18
40	6	7	18
41	11	10	24
42	3	2	11