

Radiological Survey Sheet

118/229

Job Location PFIZER INC, GASTON, CT 06340 ^{118 2nd Floor} Page 1 of 5

Survey Purpose WALKIN COOLER DECOMMISSIONING Date 3-23-10

Performed by E. GAILOR (Print) E. Gailor (Sign)

Instrument (Model/S/N)	Packard Tri-Carb S/N 431521	Packard Tri-Carb S/N 431520	Ludlum 2241-2 163603	Wallac Model 1409 S/N 4061042
Det. (Model/SN)	Internal	Internal	Ludlum 43-68 PR 149613	Internal
Type Rad.	β	β	β	β
Cal. Due:	04/03/2010	04/03/2010	04/03/2010 ²¹	04/03/2010
Check Instruments Used	✓		✓	

No.	Time	Locations
1.	1230	BACKGROUND
2.		OUTSIDE DOOR & HNDL
3.		INSIDE SHR " "
4.		FLOOR
5.		UPPER WALL
6.		LOWER "
7.		UPPER "
8.		LOWER "
9.		UPPER "
10.		LOWER "
11.		UPPER "
12.		LOWER "
13.		UPPER "
14.		LOWER "
15.		UPPER "
16.		LOW ER "
17.		UPPER "
18.		LOW ER "


See attached LSC printout for Total Activity (cpm/100 cm²)

WALK IN COOLER

Pg 3 of 5
3-23-10

O₂ WIFE LOCATION

Surveyed 100% OF LOWER SURFACES USING INSTRUMENT 43-68. All READINGS NOTED TO BE AT NORMAL BACKGROUND LEVELS, WITH THE FOLLOWING 2 EXCEPTIONS:



△ 100 cm² SPOT OF FLOOR CONTAMINATED TO A LEVEL OF 5,767 dpm. DISCONTAMINATED TO NORMAL BACKGROUND LEVEL.

△ 100 cm² SPOT OF FLOOR CONTAMINATED TO A LEVEL OF 17,300 dpm. DISCONTAMINATED TO NORMAL BACKGROUND LEVEL.

Protocol# 2 - Room Decon.lsa

User: Bill Calver

Equipment Swipes 220 236 Area

Assay Definition-

Assay Description:
Decon D408-410

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\Bill Calver\Room Decon\20100323_1603

Raw Results Path: C:\Packard\Tricarb\Results\Bill Calver\Room Decon\20100323_1603\20100323_1603.results

RTF File Name: C:\Packard\Tricarb\Results\Bill Calver\Room Decon\20100323_1603\Lab Decon 118 237.rtf

Assay File Name: C:\Packard\TriCarb\Assays\Room Decon.lsa

Count Conditions-

Nuclide: Triple

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 1.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	0.0	18.6
B	18.6	156.0
C	0.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: On

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

S#	CPMA	CPMB	CPMC	MESSAGES
1	9	3	17	
2	18	10	35	
3	8	5	20	
4	3	5	17	
5	16	12	33	
6	15	8	28	
7	10	10	29	
8	11	11	33	
9	8	7	21	
10	9	7	21	

Equipment Swipes 220 236 Area

11	13	6	23
12	7	9	21
13	2	8	18
14	5	8	16
15	5	8	19
16	11	4	21
17	12	9	26
18	18	11	36
19	25	6	38
20	9	5	19
21	13	5	27
22	13	5	22
23	10	10	27
24	12	4	22
25	11	9	26
26	5	13	27
27	18	10	33
28	0	9	14
29	8	8	20
30	10	7	23
31	23	6	32
32	7	11	26
33	12	8	23
34	4	4	13

Note: All of the results are SATISFACTORY, including #s 18 & 19.
 w/ Calver