

Radiological Survey Continuation Sheet

Page 2 of

Date 1-23-13

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
19 20	1135	LOWER WALL	
21		UPPER "	
22		LOWER "	
23		FLOOR	
24		"	
25		"	
26		HOOD LEFT SIDE	
27		" BACK	
28		" VENTS	
29		" RIGHT SIDE	
30		" LEFT BOTTOM	
31		" RIGHT "	
32		" SASH	
33		" LP	
34		CAB DOORS	
35		INSIDE CAB	
36		SIDE OF HOOD	
37		WALL	
38		CAB DOORS	
39		" TOP SHELF	
40		" 2ND "	
41		" BOTTOM "	
42		" DOORS	
43		" TOP SHELF	
44		" 2ND "	
45		" BOTTOM "	
46	✓	CAB DOORS	

Radiological Survey Continuation Sheet


Page 3 of

Date 1-23-12

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
46	1135	CAD TOP SHEET	
47		" 2ND "	
48		" BOTTOM "	
49		FLOOR	
50		"	
51		"	
52		UPPER WALL	
53		LOWER "	
54		UPPER "	
55		LOWER "	
56		door	
57		UPPER WALL	
58		LOWER "	
59		UPPER "	
60		LOWER "	
61		FLOOR	
62		"	
63		"	
64		"	
65		"	
66		"	
67		"	
68		"	
69		"	
70		"	
71		"	
72	✓	"	

pg 5 of 20

118/252 A pg of 1-23-13



~~SPAT checked~~ Surveyed 100% of lower surfaces with instrument #7.
All readings found to be at normal background levels.

O = wipe test location

Assay Definition-

Assay Description:
Equipment Swipes

Assay Type: CPM
Report Name: Report1
Output Data Path: C:\Packard\Tricarb\Results\Ed Gailor\Eds Monthly Swipes
Raw Results Path: C:\Packard\Tricarb\Results\Ed Gailor\Eds Monthly Swipes\20130123_1211
\20130123_1211.results
RTF File Name: C:\Packard\Tricarb\Results\Ed Gailor\Eds Monthly Swipes\Ed's Monthly
Swipes.rtf
Assay File Name: C:\Packard\TriCarb\Assays\Eds Monthly Swipes.lsa

Count Conditions-

Nuclide: Triple
Quench Indicator: tSIE/AEC
External Std Terminator (sec): 5 sec
Pre-Count Delay (min): 0.00
Quench Set: n/a
Count Time (min): 1.00
Count Mode: Normal
Assay Count Cycles: 1 Repeat Sample Count: 1
#Vials/Sample: 1 Calculate % Reference: Off

Background Subtract: On - Manual
Low CPM Threshold: Off
2 Sigma % Terminator: Off

Regions	LL	UL	Bkg Subtract
A	0.0	18.6	0.00
B	18.6	156.0	0.00
C	0.0	2000.0	0.00

Count Corrections-

Static Controller: On Luminescence Correction: On
Colored Samples: n/a Heterogeneity Monitor: n/a
Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

S#	CPMA	CPMB	CPMC	MESSAGES
1	10	2	17	
2	15	6	21	
3	12	7	24	
4	14	8	30	
5	10	8	24	
6	4	3	11	
7	5	5	14	
8	7	5	15	
9	10	8	24	
10	11	4	22	

11	11	8	23
12	7	9	17
13	9	1	17
14	10	8	21
15	4	6	15
16	8	9	23
17	15	3	23
18	10	6	20
19	8	6	20
20	14	8	25
21	11	6	24
22	6	10	20
23	7	8	20
24	9	12	25
25	11	3	17
26	9	10	27
27	11	10	23
28	8	8	19
29	5	9	19
30	10	4	19
31	10	7	22
32	12	2	17
33	15	2	26
34	5	3	15
35	10	6	18
36	18	7	25
37	6	3	14
38	9	6	19
39	14	6	28
40	7	6	17
41	3	5	12
42	10	6	18
43	1	4	9
44	4	16	20
45	3	10	16
46	9	7	19
47	17	5	29
48	11	6	21
49	17	8	29
50	14	6	22
51	12	8	22
52	10	11	26
53	5	6	14
54	10	8	24
55	7	6	18
56	15	2	21
57	8	4	16
58	3	8	18
59	11	5	20
60	8	10	20
61	3	5	12
62	9	4	15
63	8	5	17
64	9	5	19
65	9	5	17
66	13	8	25
67	7	8	20
68	5	9	17
69	11	4	21
70	5	1	9
71	7	7	19
72	7	6	14
73	10	5	16