

Radiological Survey Sheet

Job Location Pfizer, Inc. Groton, Ct. 06340 Page 1 of 2

Survey Purpose RADIATION ASSESSMENT 11E/252B Date 2/14/13

Performed by DAVID DAVAZZ (Print)  (Sign)

Instrument (Model/S/N)	Packard Tri-Carb S/N 431521	Packard Tri-Carb S/N 431520	Ludlum 2241-2 163603	
Det. (Model/SN)	Internal	Internal	Ludlum 43-68 PR 149613	
Type Rad.	β	β	β	
Cal. Due:	03/05/2013	03/05/2013	04/17/2013	
Check Instruments Used	✓		✓	

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
1.	1155	BACKGROUND	
2.		F1 Floor	
3.		F2 Floor	✓ (SEE #157)
4.		F3 Floor	
5.		F4 Floor	
6.		W1 WALL	
7.		W2 WALL	
8.		W3 WALL	
9.		W4 WALL	
10.		W5 WALL	
11.		W6 WALL	
12.		W7 WALL	
13.		W8 WALL	
14.		W9 WALL	
15.		W10 WALL	
16.		W11 WALL	
17.		W12 WALL	
18.	✓	W13 WALL	

Radiological Survey Continuation Sheet

Page 2 of 21

Date 2/14/13

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
19		W14 W11	✓ (SEE #158)
20		W15 W11	
21		AMP STAND	✓ (SEE #159)
22		VD1 IN DUCT	
23		VD2 IN DUCT	✓ 2 (SEE #160)
24		VD3 IN DUCT	
25		VD4 DUCT	
26		VD5 DUCT	✓ (SEE #160)
27		VD6 DUCT	
28		VD7 DUCT	
29		VD8 DUCT	✓ (SEE #161)
30		VD9 DUCT	✓ (SEE #162)
31		VD10 DUCT	✓ (SEE #163)
32		VD11 DUCT	✓ (SEE #164)
33		VD12 IN DUCT	
34		VD13 DUCT	✓ (SEE #165)
35		VD14 DUCT	
36		VD15 DUCT	
37		VD16 DUCT	✓ (SEE #166)
38		L1 TOP LIGHT	✓ (SEE #167)
39		L1 BOTTOM LIGHT	
40		VD17 IN DUCT	
41		F5 Floor	
42		F6 Floor	
43		F7 Floor	
44		F8 Floor	
45		✓ F9 Floor	

Radiological Survey Continuation Sheet

Page 3 of 21

Date 2/14/13

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
46		F10 Floor	✓ (SEE #168)
47		F11 Floor	✓ (SEE #169)
48		F12 Floor	✓ (SEE #170)
49		F13 Floor	
50		F14 Floor	✓ (SEE #171)
51		F15 Floor	
52		F16 Floor	✓ (SEE #172)
53		F17 Floor	
54		F18 Floor	
55		F19 Floor	✓ (SEE #173)
56		F20 Floor	✓ (SEE #174)
57		F21 Floor	
58		F22 Floor	
59		F23 Floor	✓ (SEE #175)
60		F24 Floor	✓ (SEE #176)
61		W16 WALL	✓ (SEE #177)
62		W17 WALL	
63		W18 WALL	
64		W19 WALL	
65		W20 WALL	
66		W21 WALL	
67		W22 WALL	
68		W23 WALL	✓ (SEE #178)
69		W24 WALL	✓ (SEE #179)
70		W25 WALL	
71		W26 WALL	
72	↓	W27 WALL	✓ (SEE #180)

Radiological Survey Continuation Sheet

Page 4 of 21

Date 2/14/13

No.	Time	Locations		See attached LSC printout for Total Activity (cpm/100 cm ²)
73		W28 WALL	✓	(SEE # 181)
74		W29 WALL	✓	(SEE # 182)
75		W30 WALL		
76		V18 IN DUCT	✓	(reductions)
77		V19 DUCT	✓	↓ (SEE # 202)
78		V20 IN DUCT	✓	
79		V21 DUCT	✓	
80		V22 IN DUCT	✓	
81		V23 DUCT	✓	
82		V24 IN DUCT	✓	
83		V25 DUCT	✓	
84		V26 DUCT	✓	
85		V27 DUCT	✓	
86		V28 DUCT	✓	
87		VD29 DUCT	✓	
88		VD30 IN DUCT		
89		VD31 DUCT		
90		L2 TOP LIGHT		
91		L2 BOTTOM LIGHT	✓	
92		VD32 DUCT	✓	(SEE # 184)
93		P1 PIPE	✓	(SEE # 185)
94		P2 PIPE	✓	(SEE # 186)
95		P3 PIPE	✓	(SEE # 187)
96		P4 PIPE	✓	(SEE # 188)
97		VD33 DUCT		
98		VD34 IN DUCT		
99	↓	VD35 IN DUCT		

Radiological Survey Continuation Sheet

Page 5 of 21

Date 2/14/13

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
100		VD36 DUCT	✓ (SEE #195)
101		VD37 DUCT	
102		VD38 DUCT	
103		VD39 IN DUCT	
104		VD40 DUCT	
105		VD41 DUCT	
106		W31 WALL	
107		W32 WALL	✓ (SEE #190)
108		W33 WALL	✓ (SEE #191)
109		W34 WALL	
110		W35 WALL	
111		W36 WALL	✓ (SEE #192)
112		W37 WALL	
113		W38 WALL	
114		W39 WALL	✓ (SEE #193)
115		W40 WALL	
116		W41 WALL	
117		W42 WALL	
118		S1 TOP SHELF	✓ (SEE #194)
119		S1 BOTTOM SHELF	
120		S2 TOP SHELF	
121		S2 BOTTOM SHELF	
122		S3 TOP SHELF	
123		S3 BOTTOM SHELF	
124		S4 TOP SHELF	
125		S4 BOTTOM SHELF	
126	↓	S5 TOP SHELF	✓ (SEE #195)

Radiological Survey Continuation Sheet

Page 6 of 21

Date 2/14/13

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
127		S5 Bottom Shelf	✓ (SEE #196)
128		S6 Top Shelf	
129		S6 Bottom Shelf	✓ (SEE #197)
130		W43 WALL	
131		W44 WALL	
132		W45 WALL	
133		W46 WALL	
134		W47 WALL	✓ (SEE #198)
135		W48 WALL	
136		W49 WALL	
137		W50 WALL	
138		W51 WALL	
139		W52 WALL	
140		W53 WALL	
141		W54 WALL	
142		W55 WALL	
143		W56 WALL	
144		W57 WALL	
145		W58 WALL	✓ (SEE #199)
146		W59 WALL	
147		W60 WALL	
148		L3 Top LIGHT	
149		L3 Bottom LIGHT	
150		VD42 DUCT	
151		VD43 in DUCT	
152		Floor Pit	
153	↓	Floor Pit DRAIN	

Radiological Survey Continuation Sheet

Page 7 of 21

Date 2/14/13


No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
154		Flare PIT GATE	
155		P5 Pipe	✓ (see # 200)
156		P6 Pipe	✓ (see # 201)
4/7/13 157	1358	Post Decm #3	
158		Post Decm #19	
159		Post Decm #21	
160		Post Decm #26	
161		Post Decm #29	
162		Post Decm #30	
163		Post Decm #31	
164		Post Decm #32	
165		Post Decm #34	
166		Post Decm #37	
167		Post Decm #38	
168		Post Decm #46	
169		Post Decm #47	
170		Post Decm #49	
171		Post Decm #50	
172		Post Decm #52	
173		Post Decm #55	
174		Post Decm #56	
175		Post Decm #59	
176		Post Decm #60	
177		Post Decm #61	
178		Post Decm #68	
179		Post Decm #69	
180		Post Decm #72	


Radiological Survey Continuation Sheet

Page 8 of 21

Date 4/7/13


No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
181		POST DECON # 73	
182		POST DECON # 74	
183		POST DECON # 91	
184		POST DECON # 92	
185		POST DECON # 93	
186		POST DECON # 94	
187		POST DECON # 95	
188		POST DECON # 96	
189		POST DECON # 100	
190		POST DECON # 107	
191		POST DECON # 108	
192		POST DECON # 111	
193		POST DECON # 114	
194		POST DECON # 118	
195		POST DECON # 126	
196		POST DECON # 127	
197		POST DECON # 129	
198		POST DECON # 134	
199		POST DECON # 145	
200		POST DECON # 155	
201		POST DECON # 156	
202		POST DECON # 87	
203	✓	INSIDE V029 (radioster)	
204		BACKGROUND	
205		POST DECON Δ 1	
206		POST DECON Δ 2	
207		POST DECON Δ 3	






118 / 252 B

page 1 of 21
DATE: 2/14/13



EXHAUST VENT DUCTING



 Removed
TO RAO WASTE

118/252B

Surveyed approximately 100% of surfaces using the 43-68. All areas were found to be at normal background levels, with the following exceptions:

1. A small area on the floor by the put was found to be contaminated to a maximum level of 29,039 dpm/100cm². The area was decontaminated to normal background levels.
2. An area of the floor was found to be contaminated to a maximum level of 25,971 dpm/100 cm². The area was decontaminated to normal background levels.
3. An area of the floor was found to be contaminated to a maximum level of 1,528,016 dpm/100 cm². The area was decontaminated to normal background levels.
4. An area of the floor was found to be contaminated to a maximum level of 18,814 dpm/100 cm². The area was decontaminated to normal background levels.
5. An area of the floor was found to be contaminated to a maximum level of 127,198 dpm/100 cm². The area was decontaminated to normal background levels.
6. An area of the floor was found to be contaminated to a maximum level of 9,816 dpm/100 cm². The area was decontaminated to normal background levels.
7. An area of the floor was found to be contaminated to a maximum level of 17,791 dpm/100 cm². The area was decontaminated to normal background levels.
8. An area of the wall was found to be contaminated to a maximum level of 101,636 dpm/100 cm². The area was decontaminated to normal background levels.
9. An area of the wall was found to be contaminated to a maximum level of 174,233 dpm/100 cm². The area was decontaminated to normal background levels.
10. Exhaust ducting was found to be contaminated to a maximum level 116,973 dpm/100 cm². This ducting was removed and disposed of as radioactive waste.

Equipment Swipes 220 236 Area

Assay Definition-

Assay Description:
Decon D408-410

Assay Type: CPM
Report Name: Report1
Output Data Path: C:\Packard\Tricarb\Results\Bill Calver\Room Decon\20130214_1309
Raw Results Path: C:\Packard\Tricarb\Results\Bill Calver\Room Decon\20130214_1309\20130214_1309.results
RTF File Name: C:\Packard\Tricarb\Results\Bill Calver\Room Decon\20130214_1309\Lab Decon 118 237.rtf
Assay File Name: C:\Packard\TriCarb\Assays\Room Decon.lsa

Count Conditions-

Nuclide: Triple
Quench Indicator: tSIE/AEC
External Std Terminator (sec): 0.5 2s%
Pre-Count Delay (min): 0.00
Quench Set: n/a
Count Time (min): 1.00
Count Mode: Normal
Assay Count Cycles: 1 Repeat Sample Count: 1
#Vials/Sample: 1 Calculate % Reference: Off

Background Subtract: Off
Low CPM Threshold: Off
2 Sigma % Terminator: Off

Regions	LL	UL
A	0.0	18.6
B	18.6	156.0
C	0.0	2000.0

Count Corrections-

Static Controller: On Luminescence Correction: On
Colored Samples: n/a Heterogeneity Monitor: n/a
Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

S#	CPMA	CPMB	CPMC	MESSAGES
1	10	7	22	
2	74	10	91	
3	92	18	122	
4	24	9	36	
5	46	28	81	
6	16	11	32	
7	27	12	42	
8	37	23	67	
9	10	12	25	
10	28	19	51	

5.6-2

Protocol# 2 - Room Decon.lsa

User: Bill Calver

Equipment Swipes 220 236 Area

11	23	15	41
12	11	7	25
13	36	27	66
14	20	10	40
15	19	18	39
16	46	25	75
17	28	17	51
18	47	34	85
19	76	38	120 ✓
20	30	27	60
21	73	43	123 ✓
22	43	34	88
23	89	49	143 ✓
24	34	27	61
25	41	28	71
26	49	45	101 ✓
27	36	14	59
28	42	50	96
29	93	37	135 ✓
30	238	139	383 ✓
31	110	64	176 ✓
32	61	42	109 ✓
33	9	5	19
34	95	60	158 ✓
35	45	21	70
36	47	43	99 ✓
37	63	35	104 ✓
38	74	48	125 ✓
39	15	10	31
40	38	18	58
41	19	12	34
42	18	21	44
43	22	19	48
44	28	15	48
45	29	25	60
46	71	54	128 ✓
47	96	14	117 ✓
48	92	44	141 ✓
49	29	18	54
50	202	110	318 ✓
51	46	16	65
52	185	91	288 ✓
53	51	36	93
54	29	19	54
55	75	37	121 ✓
56	206	119	347 ✓
57	28	21	56
58	51	29	82
59	324	175	503 ✓
60	862	449	1322 ✓
61	54	43	105
62	26	26	56
63	27	16	45
64	41	48	94
65	25	12	44
66	24	19	51
67	22	30	55
68	96	55	154 ✓
69	100	98	202 ✓
70	32	23	62
71	46	25	80
72	83	77	164 ✓
73	149	136	292 ✓

Equipment Swipes 220 236 Area

74	151	134	293 ✓
75	29	33	67
76	185	112	306 ✓
77	123	83	212 ✓
78	262	169	444 ✓
79	140	123	273 ✓
80	284	182	524 ✓
81	64	55	129 ✓
82	731	531	1279 ✓
83	166	268	436 ✓
84	95	72	172 ✓
85	224	163	391 ✓
86	159	150	319 ✓
87	66	52	128 ✓
88	8	5	17
89	39	16	60
90	28	9	40
91	220	157	388 ✓
92	111	73	187 ✓
93	79	49	134 ✓
94	144	72	218 ✓
95	97	68	170 ✓
96	132	64	200 ✓
97	25	16	46
98	18	7	30
99	41	24	72
100	76	49	129 ✓
101	35	20	59
102	22	9	34
103	7	13	23
104	19	18	46
105	27	15	46
106	27	24	54
107	106	87	198 ✓
108	54	49	105 ✓
109	12	7	23
110	24	34	59
111	71	60	138 ✓
112	15	6	24
113	18	14	40
114	67	75	149 ✓
115	49	39	95
116	15	10	27
117	10	12	26
118	78	36	120 ✓
119	50	30	90
120	58	36	97
121	23	35	64
122	22	29	58
123	56	35	96
124	33	37	76
125	30	26	62
126	83	52	139 ✓
127	62	42	108 ✓
128	20	30	54
129	77	61	147 ✓
130	13	17	35
131	24	22	50
132	27	28	57
133	24	24	55
134	70	72	145 ✓
135	28	29	61
136	16	15	35

Equipment Swipes 220 236 Area

137	31	25	61
138	26	23	54
139	22	16	42
140	23	20	45
141	39	28	70
142	33	21	60
143	33	29	73
144	39	20	63
145	53	48	107 ✓
146	25	11	41
147	32	52	91
148	31	11	46
149	19	10	33
150	16	3	24
151	8	7	19
152	41	41	92
153	34	21	61
154	23	13	39
155	137	104	247 ✓
156	467	191	666 ✓

Assay Definition-

Assay Description:
Background check for PM Lab (Mikey)

Assay Type: CPM
Report Name: Report1
Output Data Path: C:\Packard\Tricarb\Results\bc\bills swipes
Raw Results Path: C:\Packard\Tricarb\Results\bc\bills swipes\20130407_1452\20130407_1452.results
RTF File Name: C:\Packard\Tricarb\Results\bc\bills swipes\Report1.rtf
Assay File Name: C:\Packard\TriCarb\Assays\bills swipes.lsa

Count Conditions-

Nuclide: TRIPLE
Quench Indicator: tSIE/AEC
External Std Terminator (sec): 0.5 2s%
Pre-Count Delay (min): 0.00
Quench Set: n/a
Count Time (min): 1.00
Count Mode: Normal
Assay Count Cycles: 1 Repeat Sample Count: 1
#Vials/Sample: 1 Calculate % Reference: Off

Background Subtract: Off
Low CPM Threshold: Off
2 Sigma % Terminator: Off

Regions	LL	UL
A	0.0	18.6
B	18.6	156.0
C	0.0	2000.0

Count Corrections-

Static Controller: On Luminescence Correction: On
Colored Samples: n/a Heterogeneity Monitor: n/a
Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

S#	CPMA	CPMB	CPMC	MESSAGES	PID	Custom1
1	15	4	21		26	
2	6	11	21		26	
3	9	6	16		26	
4	28	5	38		26	
5	14	21	43		26	
6	29	32	64		26	
7	21	22	48		26	
8	30	29	61		26	
9	23	18	44		26	
10	18	14	39		26	
11	34	20	57		26	

19 of 21

167	12	30	25	58	26
	13	11	5	16	26
	14	17	6	27	26
170	15	10	12	22	26
	16	12	10	26	26
	17	19	6	30	26
	18	12	8	26	26
	19	20	17	40	6
175	20	15	13	33	6
	21	14	18	39	6
	22	19	8	31	6
	23	5	4	10	6
	24	71	4	79	6
180	25	20	13	38	6
	26	42	19	69	6
	27	28	17	53	6
	28	20	9	31	6
	29	25	15	44	6
185	30	19	11	36	6
	31	51	16	73	6
	32	16	17	42	6
	33	20	11	35	6
	34	32	39	78	6
190	35	22	24	48	6
	36	10	12	27	6
	37	11	12	29	5
	38	30	12	48	5
	39	11	2	17	5
195	40	13	3	21	5
	41	20	5	30	5
	42	17	9	32	5
	43	20	7	29	5
	44	16	12	30	5
200	45	14	11	27	5
201	46	13	17	31	5
202	47	11	16	29	5
203	48	104	125	239	5

19

Assay Definition-

Assay Description:
Background check for PM Lab (Mikey)

Assay Type: CPM
Report Name: Report1
Output Data Path: C:\Packard\Tricarb\Results\bc\bills swipes
Raw Results Path: C:\Packard\Tricarb\Results\bc\bills swipes\20130621_1150\20130621_1150.results
RTF File Name: C:\Packard\Tricarb\Results\bc\bills swipes\Report1.rtf
Assay File Name: C:\Packard\TriCarb\Assays\bills swipes.lsa

Count Conditions-

Nuclide: TRIPLE
Quench Indicator: tSIE/AEC
External Std Terminator (sec): 0.5 2s%
Pre-Count Delay (min): 0.00
Quench Set: n/a
Count Time (min): 1.00
Count Mode: Normal
Assay Count Cycles: 1 Repeat Sample Count: 1
#Vials/Sample: 1 Calculate % Reference: Off

Background Subtract: Off
Low CPM Threshold: Off
2 Sigma % Terminator: Off

Regions	LL	UL
A	0.0	18.6
B	18.6	156.0
C	0.0	2000.0

Count Corrections-

Static Controller: On Luminescence Correction: On
Colored Samples: n/a Heterogeneity Monitor: n/a
Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

S#	CPMA	CPMB	CPMC	MESSAGES	PID	Custom1
1 204	9	5	17		10	
2 205	20	2	27		10	
3 206	15	9	32		10	
4 207	10	8	21		10	
5 208	21	10	31		10	
6 209	12	11	31		10	
7 210	42	24	69		10	
8 211	14	10	28		10	
9 212	16	13	31		10	
10 213	22	13	37		10	
11 214	19	10	36		10	

21510p 13 12 29 10