


Radiological Survey Sheet

Job Location Pfizer GROTON CT 06340 Page 1 of 30

Survey Purpose LAB DECON 118B B401 Date 12/22/05

Performed by P. Keefe (Print)  (Sign)

Instrument (Model/S/N)	Packard Tri-Carb S/N 431521	Packard Tri-Carb S/N 431520	Ludlum 2241-2 163603	Wallac Model 1409 S/N 4061042
Det. (Model/SN)	Internal	Internal	Ludlum 43-68 PR 149613	Internal
Type Rad.	β	β	β	β
Cal. Due:	10/12/06	10/12/06	04/14/06	12/01/06
Check Instruments Used			X	X

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
1.	0830	BACKGROUND	
2.		LAB DOOR - OUTSIDE - UPPER	
3.		- LOWER	
4.		- INSIDE - UPPER	
5.		- LOWER	
6.		5m Wall - Upper	
7.		- MID	
8.		- LOWER	
9.		Alcove Walls - LEFT - UPPER	
10.		- MID	
11.		- Bottom	
12.		- REAR - UPPER	
13.		- WINDOW	
14.		- LOWER	
15.		- RIGHT - UPPER	
16.		- MID	
17.	▽		- LOWER
18.	0900	Network Connection	

Radiological Survey Continuation Sheet

Page 2 of 38

Date 12/22/05

LMB DECON 118B B401

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
19	0900	Hood - OUTSIDE	SECTION #1 - TOP
20			- GLASS SASH OUT IN
21			- CONTROLS
22			- CABINET - DOORS (OUTSIDE)
23			SECTION #2 - TOP
24			- GLASS SASH - OUT IN
25			- CONTROLS
26			- DRAW (OUTSIDE)
27			- CABINET - DOORS (OUTSIDE)
28			SECTION #3 - TOP
29			- GLASS SASH - OUT IN
30			- CONTROLS
31			- ALCOVE - LEFT RIGHT
32			- REAR FLOOR
33			- DRAW - OUTSIDE
34			- CABINET DOOR - LEFT (OUT)
35			SECTION #4 - TOP
36			- GLASS SASH - OUT IN
37			- CONTROLS
38			- DRAW - OUTSIDE
39			- CABINET DOOR - RIGHT (OUT)
40			- ALCOVE WALL - LEFT
41			SECTION #5 - TOP
42			- GLASS SASH - OUT IN
43			- CONTROLS
44	▽		- DRAW (OUTSIDE)
45	0920		- CABINET DOOR - LEFT (OUT)

Radiological Survey Continuation Sheet

Page 3 of 30

Date 12/22/05

LAB DECON 118 B B401

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
46	0130	HOOD - OUTSIDE - SECTION #6 - TOP	
47			- GLASS SLASH - OUT/IN
48			- CONTROLS
49			- DOORS - CABINET
50			- OUTSIDE (CABINET (RIGHT))
51		HOOD - INSIDE - SECTION #6 - RIGHT - UPPER	
52			- MID
53			- BOTTOM
54			- REAR - UPPER
55			- MID
56			- LOWER
57			- WORK AREA - LEFT
58			- MID
59			- BASIN
60			CABINET - LEFT / RIGHT
61			- REAR / BOTTOM
62			- TRAY
63			- DOORS
64			
65			- SECTION #5 - REAR - UPPER
66			- MID
67			- LOWER
68			- METAL RACK
69			- WORK AREA - LEFT
70			- MID
71	▽		- BASIN
72	1200		CABINET - LEFT / RIGHT

Radiological Survey Continuation Sheet

Page 4 of 28

Date 12/22/05

CAB DECON 118 B 3401

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
73	1200	HOOD - INSIDE - SECTION #5 - CABINET - REAR BOTTOM	
74			- TRAY
75			- DOORS ✓ (see #478)
76			- DRAW (INSIDE)
77			- SECTION #4 - REAR - UPPER
78			- MID
79			- LOWER
80			- METAL RACK
81			- WORK AREA - BASIN
82			- CENTER
83			- RIGHT
84			- SECTION #3 - REAR - UPPER
85			- MID
86			- LOWER
87			- WORK AREA - LEFT
88			- MID
89			- BASIN
90			- METAL RACK
91			- DRAW
92			- CABINET - LEFT RIGHT
93			- REAR BOTTOM
94			- TRAY
95			- DOORS
96			- SECTION #2 - REAR - UPPER
97			- MID
98	▽		- LOWER
99	1230		- WORK AREA - LEFT

✓ (see #47)

✓ (see #478)

Radiological Survey Continuation Sheet

Page 5 of 38

Date 12/22/05

LAB DECON 118B B401

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
100	1230	HOOD INSIDE	-SECTION #2 - WORK AREA - CENTER
101			- BASIN
102			- DRAW
103			- CABINET - LEFT RIGHT
104			- REAR BOTTOM
105			- TRAY
106			- DOORS
107			-SECTION #1 - REAR - UPPER
108			- MID
109			- LOWER
110			- LEFT - UPPER
111			- MID
112			- LOWER
113			- CABINET - LEFT RIGHT
114			- REAR BOTTOM
115			- TRAY
116			- DOORS
117			- WORK AREA - LEFT
118			- MID
119			- Bottom
120		CAB DOOR	- INSIDE
121			- OUTSIDE
122		Wall SWIPE	- UPPER
123			- MID
124			- Bottom
125	▽	Window SWIPE	
126	1300	HEATER	

Radiological Survey Continuation Sheet

Page 6 of 38

Date 12/22/05

CAO DECON 1188 8401

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
127	1300	Wall SWIPES	- UPPER
128			- LOWER
129			- MID
130		Window SWIPE	
131		HEATER	
132		Window SWIPE	
133		HEATER	
134		Window SWIPE	
135		HEATER	
136		Window SWIPE	
137		HEATER	
138		Wall SWIPES	- UPPER
139			- MID
140			- LOWER
141		Door SWIPE	
142		HOOD - OUTSIDE - SECTION #1	- LEFT - UPPER
143			- MID
144			- BOTTOM
145			- TOP
146			- GLASS SASH - OUT / IN
147			- CONTROLS
148			- CABINET - DOORS (OUTSIDE)
149			- LEFT
150			- SECTION #2 - TOP
151			- GLASS SASH - OUT / IN
152			- CONTROLS
153	1400		- DRAW (OUTSIDE)

Radiological Survey Continuation Sheet

Page 7 of 38

Date 12/22/05

LAB DECON 118B B401

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
154	1400	HOOD - OUTSIDE - SECTION #2 - CABINET - DOORS	
155		- SECTION #3 - TOP	
156		- GLASS SASH - OUT IN	
157		- CONTROLS	
158		- CABINET DOOR (LG)	
159		- DRAW - OUT	
160		- CABINET - DOOR (LEFT)	
161		- SECTION #4 - TOP	
162		- GLASS SASH - OUT IN	
163		- CONTROLS	
164		- DRAW (OUT)	
165		- CABINET DOOR (OUT)	
166		- ALCOVE - LEFT REAR	
167		- SECTION #5 - TOP	
168		- GLASS SASH - OUT IN	
169		- CONTROLS	
170		- DRAW (OUT)	
171		- CABINET - DOORS	
172		- ALCOVE - RIGHT REAR	
173		- SECTION #6 - TOP	
6174		- GLASS SASH	
175		- CONTROLS	
1776		- CABINET - DOORS (OUT)	
1787		HOOD - INSIDE - SECTION #6 - RIGHT - UPPER	
178		- MID	
180179		- LOWER	
180	1430	- REAR - UPPER	

Radiological Survey Continuation Sheet

Page 8 of 38

Date 12/22/05

LAB DECON 118B B401

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
181	1430	HOOD - INSIDE	- SECTION # 6 - REAR - MID
182			- LOWER
183			- WORK AREA - BASIN
184			- CENTER
185			- RIGHT
186			- CABINET - LEFT / RIGHT
187			- REAR / BOTTOM
188			- TRAY
189			- DOORS
190			- SECTION # 35 - REAR - UPPER
191			- MID
192			- LOWER
193			- WORK AREA - LEFT
194			- BASIN
195			- RIGHT
196			- DRAW - INSIDE
197			- CABINET - LEFT / RIGHT
198			- REAR / BOT
199			- TRAY
200			- DOORS
201			- SECTION # 4 - REAR - UPPER
202			- MID
203			- LOWER
204			- WORK AREA - BASIN
205			- CENTER
206	▽		- RIGHT
207	1300 1500		- ABOVE - LEFT / RIGHT

Radiological Survey Continuation Sheet

Page 9 of 38

Date 12/22/05

(LAB DECON) 1183 B401

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
208	1350 ⁰⁰	HOOD - INSIDE	- SECTION #4 - QUONIA - REAR / FLOOR
209			- SECTION #3 - REAR - UPPER
210			- MID
211			- Bottom
212			- WORK AREA - BASIN
213			- CENTER
214			- RIGHT
215			- DRAW - INSIDE
216			- CABINET - LEFT / RIGHT
217			- REAR / Bottom
218			- TRAY
219			- DOORS
220			- SECTION #2 - REAR - UPPER
221			- MID
222			- LOWER
223			- WORK AREA - BASIN
224			- CENTER
225			- RIGHT
226			- DRAW - INSIDE
227			- CABINET - LEFT / RIGHT
228			- REAR / Bottom
229			- TRAY
230			- DOORS
231			- CABINET (LG) - LEFT / RT
232		- REAR / Bot	
233	▽		- DOOR
234	1510		RADIATOR / cooler

Radiological Survey Continuation Sheet

Page 10 of 38

Date 12/22/05

LAB DECON 118B B461

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
235	1520	HOOD - INSIDE - SECTION #1 - LEFT - UPPER	
236			- MID
237			- LOWER
238			- REAR - UPPER
239			- MID
240			- LOWER
241			- WORK AREA - BASIN
242			- CENTER
243			- RIGHT
244			- DRAIN HOSE
245			- CABINET - LEFT / RIGHT
246			- REAR / BOT
247			- TRAY
248			- DOORS
249		BENCHTOP shelf	- TOP
250			- MID
251			- BOTTOM
252		SUPPORTS	
253		BENCHTOP - LEFT	
254			- RIGHT
255		DRAW - OUTLINE	
256		CABINET - OUTSIDE - DOORS	
257			- LEFT
258			- INSIDE - LEFT / RIGHT
259			- REAR BOTTOM
260	▽		- shelf
261	1600		- DOORS

Radiological Survey Continuation Sheet

Page 11 of 38

Date 12/23/05

LAB DECON 118B B401

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
262	1130	BENCHTOP shelf	-TOP
263			-MID
264			-Bottom
265		SUPPORTS	
266		BENCHTOP	-LEFT
267			-RIGHT
268		CABINET	-OUTSIDE -DOOR
269			-INSIDE -LEFT /RIGHT
270			-REAR /Bottom
271			-DOOR
272			-shelf
273		BENCHTOP shelf	-TOP
274			-MID
275			-Bottom
276		SUPPORTS	
277		BENCHTOP	-LEFT
278			-RIGHT
279		DRAWS (LEFT)	-TOP (two) -OUT/IN
280			-lower -OUT/IN
281		(RIGHT)	-TOP (two) -OUT/IN
282			-lower -OUT/IN
283		(LARGE)	-upper -OUT/IN
284			-lower ↓ ↓
285		BENCHTOP shelves	-TOP
286			-MID
287			-Bottom
288	1215	SUPPORTS	

Radiological Survey Continuation Sheet

Page 12 of 38

Date 12/23/05

LAB DECON 118B B401

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
289	1215	BENCHTOP - LEFT	
290		- RIGHT	
291		CABINET - OUTSIDE - DOOR	
292		- INSIDE - LEFT RIGHT	
293		- REAR BOTTOM	
294		- DOOR	
295		Water Purify Machine - FRONT	
296		- SIDES	
297		- BOTTOM	
298		- HOSES	
299		GLASS RACK	
300		FAUCET	
301		EYEWASH	
302		BASIN - LEFT RIGHT	
303		- REAR FRONT	
304		- BOTTOM DRAIN	
305		CABINET - OUTSIDE - DOORS	
306		- INSIDE - LEFT RIGHT	
307		- REAR BOTTOM	
308		- DOORS	
309		HOOD - OUTSIDE - FRONT - TOP	
310		- VENTED AREA	
311		- SASH - LEFT	
312		- MID	
313		- RIGHT	
314	▽		- SPLASH GUARD - LEFT
315	1235		- RIGHT

Radiological Survey Continuation Sheet

Page 13 of 38

Date 12/23/05

(LAB DECON 118B B401

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
316	1235	HOOD - OUTSIDE - CABINET - DOORS	
317		- Shelf	
318		- DRAWS - TOP - OUT/IN	
319		- UPPER MID - OUT/IN	
320		- LOWER MID - OUT/IN	
321		- BOTTOM - OUT/IN	
322		- CONTROLS - LEFT	
323		- RIGHT	
324		- LEFT - UPPER	
325		- LOWER	
326		- RIGHT - UPPER	
327		- MID	
328		- LOWER	
329		HOOD - INSIDE - LEFT - UPPER	
330		- MID	
331		- LOWER	
332		- REAR - LEFT - UPPER	
333		- MID	
334		- LOWER	
335		- RIGHT - UPPER	
336		- MID	
337		- LOWER	
338		- RIGHT - UPPER	
339		- MID	
340		- LOWER	
341	▽	- WORK AREA - LEFT	
342	1300	- RIGHT	

Radiological Survey Continuation Sheet

Page 14 of 38

Date 12/23/05

LAB DECON 118B B401

No.	
343	1300 HOOD - INSIDE - CABINET - LEFT RIGHT
353	
	- VENTED AREA
259	
360	

Radiological Survey Continuation Sheet

Page 15 of 30

Date 12/23/05

LAB DECON 118B B401

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
370	1330	HOOD - OUTSIDE - DRAWS - LOWER MID - OUT/IN	
371		- BOTTOM - OUT/IN	
372		HOOD - INSIDE - LEFT - UPPER	
373		- MID	
374		- BOTTOM	
375		- REAR - LEFT - UPPER	
376		- MID	
377		- BOTTOM	
378		- RIGHT - UPPER	
379		- MID	
380		- BOTTOM	
381		- RIGHT - UPPER	
382		- MID	
383		- BOTTOM	
384		- WORK AREA - LEFT - FRONT	
385		- REAR	
386		- RIGHT - FRONT	
387		- REAR	
388		- SPLASH GUARD	
389		- CABINET - LEFT / RIGHT	
390		- REAR / BOTTOM	
391		- TRAY	
392		- DOORS	
393		- DRAWS - TOP - OUT/IN	
394		- UPPER MID ↓ ↓	
395	▽	- LOWER MID ↓ ↓	
396	1400	- BOTTOM - OUT/IN	

Radiological Survey Continuation Sheet

Page 16 of 38

Date 12/23/05

LAB DECON 118B B401

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
397	1400	GLASS RACK	
398		FAUCET	
399		EYEWASH	
400		BASIN - LEFT / RIGHT	
401		- REAR / BOTTOM	
402		- FRONT / DRAIN	
403		CABINET - OUTSIDE - DOORS	
404		- INSIDE - LEFT / RIGHT	
405		- REAR / BOTTOM	
406		- DOORS	
407		LAB BENCH shelves - TOP	
408		- MID	
409		- BOTTOM	
410		SUPPORTS	
411		BENCHTOP - LEFT	
412		- RIGHT	
413		CABINET - OUTSIDE - DOOR	
414		- INSIDE - LEFT / RIGHT	
415		- REAR / BOTTOM	
416		- DOOR	
417		- shelves	
418		BENCHTOP shelf - TOP	
419		- MID	
420		- BOTTOM	
421		SUPPORTS	
422	∇	BENCHTOP - LEFT	
423	1430	- RIGHT	

Radiological Survey Continuation Sheet

Page 17 of 38

Date 12/23/05

LAB DECON 118B B401

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
424	1430	DRAWS -TOP - RIGHT (2) -OUT/IN	
425		-LEFT (2) -OUT/IN	
426		-MID -LEFT -OUT/IN	
427		-RIGHT -OUT/IN	
428		-LG -UPPER -OUT/IN	
429		-LOWER -OUT/IN	
430		BENCHTOP shelves -TOP	
431		-MID	
432		-Bottom	
433		SUPPORTS	
434		BENCHTOP -LEFT	
435		-RIGHT	
436		CABINET -OUTSIDE -DOOR	
437		-INSIDE -LEFT /RIGHT	
438		-REAR /Bottom	
439		-Shelves	
440		-DOOR	
441		BENCHTOP shelves -TOP	
442		-MID	
443		-Bottom	
445		Shelves SUPPORTS	
446		BENCHTOP -LEFT	
447		-RIGHT	
448		DRAW -OUT /IN	
449		CABINET -OUTSIDE -DOORS	
450	▽	-INSIDE -LEFT /RIGHT	
451	1500	-REAR /Bottom	

Radiological Survey Continuation Sheet


Page 18 of 38

Date 12/23/05

LAB DECON 118B B401

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm ²)
452	1500	CABINET - INSIDE - shelf	
453		- DOORS	
454		Waste CABINET - OUTSIDE - DOORS	
455		- INSIDE - LEFT / RIGHT	
456		- REAR / BOTTOM	
457		- DOORS	
458		FLOOR SWIPES	
↓	↓	↓ ↓	
476	1530	FLOOR SWIPES	
477	0800	POST DECON # 73	
478	↓	POST DECON # 75	

13/06



ALL SURFACES SURVEYED WITH LUDLUM MODEL 2241-2
 AND FOUND TO BE WITHIN LIMITS

PROTOCOL : 94 118B B401
 DATE : 2005/12/27
 TIME : 12:00
 ID : P94AS002

Lab Decon

Wallac 1400 DSA ver 2.50 S/N 4091042

Counting mode : CPM
 Isotope(s) : H3,C14
 H3 = 5- 350,12.43 y
 C14 = 5- 660,5730.00 y
 Protocol name : 118B B401
 Counting time : 60
 Repeats : 1
 Cycles : 1
 Replicates : 1
 2 sigma % : 0.01
 Minimum cpm : 0.00 Checking time: 10
 Advanced modes : Chemilum
 Output to Display :
 POS,DATE,TIME,CPM1,CPM2,CPMw3,RACKPOS,CPM,CTIME,SQPI
 Additions to Display : Spectrum,Header,Listing
 Header : Lab Decon
 Spectrum : Beta
 Window 1 : 1-1024 /Beta
 Window 2 : 1-1024 /Beta
 Window 3 : 1-1024 /Beta
 Window 4 : 1-1024 /Beta
 Window 5 : 1-1024 /Beta
 Window 6 : 1-1024 /Beta
 FNCT1 = FNCT1 :
 FNCT2 = FNCT2 :
 FNCT3 = FNCT3 :
 FNCT4 = FNCT4 :

Total count rate:

H3 9840.2 CPM
 C14 13423.2 CPM

POS	DATE	TIME	CPM1	CPM2	CPMW3
1	12/27/2005	12:00 PM	18.70	30.20	37.50
2	12/27/2005	12:02 PM	14.50	22.80	29.10

Background

POS	DATE	TIME	CPM1	CPM2	CPMW3
3	12/27/2005	12:04 PM	25.00	34.30	39.60
4	12/27/2005	12:05 PM	30.10	36.30	43.60
5	12/27/2005	12:07 PM	22.90	32.30	39.60
6	12/27/2005	12:08 PM	36.30	46.70	49.80
7	12/27/2005	12:10 PM	16.60	29.00	35.30
8	12/27/2005	12:11 PM	22.80	25.90	28.00
9	12/27/2005	12:13 PM	25.90	35.30	43.60
10	12/27/2005	12:15 PM	15.60	21.90	29.10
11	12/27/2005	12:16 PM	27.10	35.40	39.60
12	12/27/2005	12:18 PM	21.80	24.90	31.10
13	12/27/2005	12:19 PM	21.90	36.50	46.90
14	12/27/2005	12:21 PM	16.60	23.90	33.30
15	12/27/2005	12:22 PM	21.80	30.10	37.40
16	12/27/2005	12:24 PM	25.00	29.20	35.40
17	12/27/2005	12:25 PM	25.90	32.20	40.50
18	12/27/2005	12:27 PM	22.50	30.70	37.90
19	12/27/2005	12:29 PM	19.70	24.90	31.10
20	12/27/2005	12:30 PM	10.40	15.60	22.80
21	12/27/2005	12:32 PM	27.10	34.30	38.50
22	12/27/2005	12:33 PM	19.70	27.00	32.20
23	12/27/2005	12:35 PM	19.70	28.00	35.30
24	12/27/2005	12:36 PM	17.60	21.80	28.00
25	12/27/2005	12:38 PM	19.70	25.90	31.10
26	12/27/2005	12:40 PM	17.70	23.90	36.40
27	12/27/2005	12:41 PM	19.70	31.10	35.30
28	12/27/2005	12:43 PM	19.70	27.00	34.20
29	12/27/2005	12:44 PM	27.00	35.30	42.50
30	12/27/2005	12:46 PM	25.90	33.20	34.20
31	12/27/2005	12:47 PM	20.80	27.00	34.30
32	12/27/2005	12:49 PM	17.60	23.90	34.20
33	12/27/2005	12:50 PM	27.00	37.40	43.60

POS	DATE	TIME	CPM1	CPM2	CPMW3
34	12/27/2005	12:52 PM	20.80	27.00	31.10
35	12/27/2005	12:54 PM	16.60	21.80	25.90
36	12/27/2005	12:55 PM	19.70	32.20	38.40
37	12/27/2005	12:57 PM	27.10	36.40	42.70
38	12/27/2005	12:58 PM	21.80	24.90	33.20
39	12/27/2005	1:00 PM	17.60	24.90	29.00
40	12/27/2005	1:01 PM	12.50	18.70	28.00
41	12/27/2005	1:03 PM	21.80	28.00	32.20
42	12/27/2005	1:05 PM	14.50	20.80	28.00
43	12/27/2005	1:06 PM	15.60	24.90	30.10
44	12/27/2005	1:08 PM	17.70	28.10	38.50
45	12/27/2005	1:09 PM	17.70	27.10	35.40
46	12/27/2005	1:11 PM	21.80	28.00	30.10
47	12/27/2005	1:12 PM	19.70	27.00	36.30
48	12/27/2005	1:14 PM	23.90	32.30	40.60
49	12/27/2005	1:16 PM	20.80	29.10	36.30
50	12/27/2005	1:17 PM	14.60	14.60	19.80
51	12/27/2005	1:19 PM	11.40	19.70	22.80
52	12/27/2005	1:20 PM	15.60	25.00	32.30
53	12/27/2005	1:22 PM	33.20	42.60	46.70
54	12/27/2005	1:23 PM	19.70	24.90	31.10
55	12/27/2005	1:25 PM	15.60	22.80	25.90
56	12/27/2005	1:26 PM	18.70	30.10	36.30
57	12/27/2005	1:28 PM	22.90	32.30	40.60
58	12/27/2005	1:30 PM	21.80	33.20	35.30
59	12/27/2005	1:31 PM	15.60	18.70	27.10
60	12/27/2005	1:33 PM	18.70	25.90	29.10
61	12/27/2005	1:34 PM	17.60	23.90	33.20
62	12/27/2005	1:36 PM	17.60	21.80	29.10
63	12/27/2005	1:37 PM	28.10	33.30	41.60
64	12/27/2005	1:39 PM	22.90	26.00	31.20

POS	DATE	TIME	CPM1	CPM2	CPMW3
65	12/27/2005	1:40 PM	21.80	27.00	30.10
66	12/27/2005	1:42 PM	18.70	30.10	38.40
67	12/27/2005	1:44 PM	18.70	28.10	36.40
68	12/27/2005	1:45 PM	33.20	40.50	49.80
69	12/27/2005	1:47 PM	18.70	21.80	31.10
70	12/27/2005	1:48 PM	31.10	40.50	43.60
71	12/27/2005	1:50 PM	27.10	37.50	41.60
72	12/27/2005	1:51 PM	22.80	27.00	30.10
* 73	12/27/2005	1:53 PM	208.70	344.70	350.90
74	12/27/2005	1:55 PM	39.60	54.10	62.50
* 75	12/27/2005	1:56 PM	102.80	137.00	144.30
76	12/27/2005	1:58 PM	17.70	26.00	37.50
77	12/27/2005	1:59 PM	36.30	49.80	55.00
78	12/27/2005	2:01 PM	37.40	55.00	62.30
79	12/27/2005	2:02 PM	22.90	27.10	33.30
80	12/27/2005	2:04 PM	19.80	29.20	39.60
81	12/27/2005	2:05 PM	17.70	29.10	33.30
82	12/27/2005	2:07 PM	26.00	35.30	43.60
83	12/27/2005	2:09 PM	15.60	24.90	32.20
84	12/27/2005	2:10 PM	16.60	20.80	32.20
85	12/27/2005	2:12 PM	13.50	21.80	27.00
86	12/27/2005	2:13 PM	13.50	19.70	26.00
87	12/27/2005	2:15 PM	13.50	20.80	29.10
88	12/27/2005	2:16 PM	23.90	34.30	38.40
89	12/27/2005	2:18 PM	17.70	22.90	31.20
90	12/27/2005	2:19 PM	26.00	35.30	45.70
91	12/27/2005	2:21 PM	33.20	42.50	48.80
92	12/27/2005	2:23 PM	22.90	26.00	35.40
93	12/27/2005	2:24 PM	23.90	32.20	37.40
94	12/27/2005	2:26 PM	16.60	23.90	31.10
95	12/27/2005	2:27 PM	30.10	40.50	52.90

POS	DATE	TIME	CPM1	CPM2	CPMW3
96	12/27/2005	2:29 PM	14.60	21.90	28.10
97	12/27/2005	2:30 PM	20.80	32.20	38.40
98	12/27/2005	2:32 PM	15.60	28.00	34.20
99	12/27/2005	2:34 PM	20.80	27.10	29.10
100	12/27/2005	2:35 PM	14.50	24.90	31.10
101	12/27/2005	2:37 PM	19.80	30.20	38.50
102	12/27/2005	2:38 PM	21.80	25.90	37.40
103	12/27/2005	2:40 PM	22.90	30.20	32.30
104	12/27/2005	2:41 PM	27.00	37.40	45.70
105	12/27/2005	2:43 PM	19.70	32.20	39.40
106	12/27/2005	2:44 PM	27.00	34.20	47.70
107	12/27/2005	2:46 PM	19.70	26.00	30.10
108	12/27/2005	2:48 PM	24.90	36.30	39.50
109	12/27/2005	2:49 PM	24.00	33.30	39.60
110	12/27/2005	2:51 PM	20.80	27.00	36.30
111	12/27/2005	2:52 PM	21.80	27.00	30.10
112	12/27/2005	2:54 PM	14.50	22.80	31.10
113	12/27/2005	2:55 PM	25.00	37.50	43.70
114	12/27/2005	2:57 PM	19.80	27.10	35.40
115	12/27/2005	2:59 PM	20.80	27.00	42.60
116	12/27/2005	3:00 PM	12.50	20.80	28.10
117	12/27/2005	3:02 PM	17.60	22.80	29.10
118	12/27/2005	3:03 PM	11.40	20.80	23.90
119	12/27/2005	3:05 PM	12.50	22.90	29.10
120	12/27/2005	3:06 PM	18.70	33.20	41.50
121	12/27/2005	3:08 PM	18.70	23.90	30.10
122	12/27/2005	3:09 PM	23.90	34.30	38.40
123	12/27/2005	3:11 PM	23.90	29.10	34.20
124	12/27/2005	3:13 PM	19.80	28.10	35.40
125	12/27/2005	3:14 PM	20.80	28.00	32.20
126	12/27/2005	3:16 PM	24.90	36.30	46.70

POS	DATE	TIME	CPM1	CPM2	CPMW3
127	12/27/2005	3:17 PM	18.70	28.00	37.40
128	12/27/2005	3:19 PM	19.70	26.00	29.10
129	12/27/2005	3:20 PM	18.70	27.00	31.10
130	12/27/2005	3:22 PM	13.50	21.80	24.90
131	12/27/2005	3:24 PM	18.70	30.10	34.30
132	12/27/2005	3:25 PM	15.60	21.80	30.10
133	12/27/2005	3:27 PM	20.80	27.00	38.40
134	12/27/2005	3:28 PM	18.70	24.90	30.10
135	12/27/2005	3:30 PM	21.90	28.10	37.50
136	12/27/2005	3:31 PM	16.70	22.90	28.10
137	12/27/2005	3:33 PM	19.70	27.00	31.10
138	12/27/2005	3:34 PM	23.90	33.20	37.40
139	12/27/2005	3:36 PM	24.90	31.10	40.50
140	12/27/2005	3:38 PM	17.60	23.90	28.00
141	12/27/2005	3:39 PM	17.60	21.80	29.10
142	12/27/2005	3:41 PM	18.70	25.90	32.20
143	12/27/2005	3:42 PM	21.80	24.90	31.10
144	12/27/2005	3:44 PM	18.70	22.80	33.20
145	12/27/2005	3:45 PM	18.70	25.00	29.10
146	12/27/2005	3:47 PM	22.80	32.20	37.40
147	12/27/2005	3:49 PM	20.80	26.00	31.10
148	12/27/2005	3:50 PM	18.70	25.90	32.20
149	12/27/2005	3:52 PM	16.70	21.90	33.30
150	12/27/2005	3:53 PM	17.60	24.90	32.20
151	12/27/2005	3:55 PM	16.60	20.80	26.00
152	12/27/2005	3:56 PM	9.30	16.60	24.90
153	12/27/2005	3:58 PM	18.70	21.90	26.00
154	12/27/2005	4:00 PM	16.70	25.00	35.40
155	12/27/2005	4:01 PM	8.30	12.50	14.50
156	12/27/2005	4:03 PM	22.80	27.00	31.10
157	12/27/2005	4:04 PM	17.60	20.80	27.00

POS	DATE	TIME	CPM1	CPM2	CPMW3
158	12/27/2005	4:06 PM	14.50	22.80	26.00
159	12/27/2005	4:07 PM	17.70	21.80	30.10
160	12/27/2005	4:09 PM	28.00	32.20	37.40
161	12/27/2005	4:10 PM	19.70	27.00	36.30
162	12/27/2005	4:12 PM	20.80	34.30	46.70
163	12/27/2005	4:14 PM	17.70	21.80	23.90
164	12/27/2005	4:15 PM	20.80	26.00	38.40
165	12/27/2005	4:17 PM	9.30	17.70	22.90
166	12/27/2005	4:18 PM	16.60	16.60	23.90
167	12/27/2005	4:20 PM	18.70	27.00	35.30
168	12/27/2005	4:21 PM	14.50	19.70	32.20
169	12/27/2005	4:23 PM	16.70	24.00	32.30
170	12/27/2005	4:25 PM	17.70	30.10	34.30
171	12/27/2005	4:26 PM	22.90	28.10	39.60
172	12/27/2005	4:28 PM	12.50	19.80	27.10
173	12/27/2005	4:29 PM	19.80	26.00	30.20
174	12/27/2005	4:31 PM	20.80	29.10	33.20
175	12/27/2005	4:32 PM	29.10	35.30	42.60
176	12/27/2005	4:34 PM	23.90	32.20	37.40
177	12/27/2005	4:35 PM	15.60	21.80	30.10
178	12/27/2005	4:37 PM	16.60	20.80	32.20
179	12/27/2005	4:39 PM	19.70	27.00	32.20
180	12/27/2005	4:40 PM	20.80	27.00	36.40
181	12/27/2005	4:42 PM	19.80	22.90	30.20
182	12/27/2005	4:43 PM	20.80	29.10	39.50
183	12/27/2005	4:45 PM	20.80	24.90	33.30
184	12/27/2005	4:46 PM	12.50	17.70	28.10
185	12/27/2005	4:48 PM	19.80	26.00	32.30
186	12/27/2005	4:49 PM	18.70	22.80	31.20
187	12/27/2005	4:51 PM	16.60	25.90	31.10
188	12/27/2005	4:53 PM	29.10	36.40	44.70

POS	DATE	TIME	CPM1	CPM2	CPMW3
189	12/27/2005	4:54 PM	20.80	28.10	34.40
190	12/27/2005	4:56 PM	16.60	26.00	35.30
191	12/27/2005	4:57 PM	26.00	32.30	39.60
192	12/27/2005	4:59 PM	31.20	37.40	44.70
193	12/27/2005	5:00 PM	23.90	27.00	35.30
194	12/27/2005	5:02 PM	9.30	15.60	23.90
195	12/27/2005	5:04 PM	13.30	19.50	27.70
196	12/27/2005	5:05 PM	15.60	21.80	28.00
197	12/27/2005	5:07 PM	14.50	20.80	30.10
198	12/27/2005	5:08 PM	15.60	20.80	33.30
199	12/27/2005	5:10 PM	11.40	24.90	32.20
200	12/27/2005	5:11 PM	15.60	23.90	34.30
201	12/27/2005	5:13 PM	7.30	15.60	26.00
202	12/27/2005	5:15 PM	21.90	32.30	40.60
203	12/27/2005	5:16 PM	17.70	21.80	28.00
204	12/27/2005	5:18 PM	11.40	18.70	24.90
205	12/27/2005	5:19 PM	12.50	18.70	27.00
206	12/27/2005	5:21 PM	20.80	23.90	29.10
207	12/27/2005	5:22 PM	16.60	28.00	36.30
208	12/27/2005	5:24 PM	28.00	33.20	44.70
209	12/27/2005	5:25 PM	21.80	28.00	30.10
210	12/27/2005	5:27 PM	14.50	19.70	24.90
211	12/27/2005	5:29 PM	21.80	26.00	34.30
212	12/27/2005	5:30 PM	27.00	28.00	34.30
213	12/27/2005	5:32 PM	24.90	32.20	36.30
214	12/27/2005	5:33 PM	19.70	24.90	31.10
215	12/27/2005	5:35 PM	14.60	21.90	29.20
216	12/27/2005	5:36 PM	21.80	29.10	38.40
217	12/27/2005	5:38 PM	20.80	30.20	39.60
218	12/27/2005	5:40 PM	14.60	24.00	34.40
219	12/27/2005	5:41 PM	21.80	26.00	31.20

POS	DATE	TIME	CPM1	CPM2	CPMW3
220	12/27/2005	5:43 PM	15.60	19.70	29.10
221	12/27/2005	5:44 PM	16.60	20.80	22.80
222	12/27/2005	5:46 PM	21.80	24.90	32.20
223	12/27/2005	5:47 PM	24.90	30.10	40.50
224	12/27/2005	5:49 PM	18.70	22.90	31.20
225	12/27/2005	5:50 PM	23.90	31.20	40.50
226	12/27/2005	5:52 PM	18.70	24.90	29.10
227	12/27/2005	5:54 PM	18.70	24.90	29.10
228	12/27/2005	5:55 PM	16.70	25.00	33.30
229	12/27/2005	5:57 PM	18.70	30.10	41.50
230	12/27/2005	5:58 PM	15.60	19.70	29.10
231	12/27/2005	6:00 PM	18.70	23.90	35.30
232	12/27/2005	6:01 PM	18.70	22.80	31.20
233	12/27/2005	6:03 PM	13.50	19.70	28.00
234	12/27/2005	6:05 PM	12.50	19.70	28.00
235	12/27/2005	6:06 PM	18.70	25.00	35.40
236	12/27/2005	6:08 PM	17.70	21.80	26.00
237	12/27/2005	6:09 PM	18.70	24.90	36.30
238	12/27/2005	6:11 PM	16.60	23.90	28.00
239	12/27/2005	6:12 PM	20.80	28.00	33.20
240	12/27/2005	6:14 PM	12.50	19.80	27.10
241	12/27/2005	6:16 PM	29.10	36.30	39.50
242	12/27/2005	6:17 PM	12.30	19.50	25.60
243	12/27/2005	6:19 PM	23.90	31.10	38.40
244	12/27/2005	6:20 PM	14.50	21.80	28.00
245	12/27/2005	6:22 PM	16.70	21.90	28.10
246	12/27/2005	6:23 PM	13.50	21.80	26.00
247	12/27/2005	6:25 PM	24.90	31.10	37.40
248	12/27/2005	6:26 PM	16.60	29.10	36.40
249	12/27/2005	6:28 PM	16.70	18.80	26.00
250	12/27/2005	6:30 PM	22.90	34.40	39.60

POS	DATE	TIME	CPM1	CPM2	CPMW3
251	12/27/2005	6:31 PM	7.30	9.40	17.70
252	12/27/2005	6:33 PM	19.80	29.10	38.50
253	12/27/2005	6:34 PM	20.80	28.10	40.50
254	12/27/2005	6:36 PM	16.60	20.80	29.10
255	12/27/2005	6:38 PM	19.70	24.90	29.10
256	12/27/2005	6:39 PM	27.00	32.20	38.40
257	12/27/2005	6:41 PM	10.40	20.80	29.10
258	12/27/2005	6:42 PM	18.70	20.80	24.90
259	12/27/2005	6:44 PM	17.70	26.00	35.30
260	12/27/2005	6:45 PM	15.60	18.70	31.20
261	12/27/2005	6:47 PM	17.70	21.80	28.00
262	12/27/2005	6:48 PM	17.70	19.80	24.00
263	12/27/2005	6:50 PM	18.70	21.90	28.10
264	12/27/2005	6:51 PM	17.70	27.00	35.30
265	12/27/2005	6:53 PM	20.80	26.00	32.20
266	12/27/2005	6:55 PM	16.70	20.80	30.20
267	12/27/2005	6:56 PM	21.50	28.70	35.80
268	12/27/2005	6:58 PM	20.80	31.30	40.60
269	12/27/2005	6:59 PM	16.60	27.00	34.30
270	12/27/2005	7:01 PM	20.80	26.00	35.40
271	12/27/2005	7:02 PM	18.70	20.80	27.00
272	12/27/2005	7:04 PM	21.80	29.10	36.30
273	12/27/2005	7:06 PM	18.70	24.90	32.20
274	12/27/2005	7:07 PM	16.60	20.80	31.20
275	12/27/2005	7:09 PM	23.90	26.00	34.30
276	12/27/2005	7:10 PM	22.90	30.10	36.40
277	12/27/2005	7:12 PM	16.60	27.00	31.20
278	12/27/2005	7:13 PM	24.00	32.30	38.50
279	12/27/2005	7:15 PM	35.30	47.80	54.10
280	12/27/2005	7:17 PM	32.20	38.50	45.70
281	12/27/2005	7:18 PM	21.80	28.10	35.30

POS	DATE	TIME	CPM1	CPM2	CPMW3
282	12/27/2005	7:20 PM	14.50	22.90	32.20
283	12/27/2005	7:21 PM	26.00	29.20	35.40
284	12/27/2005	7:23 PM	15.60	22.80	32.20
285	12/27/2005	7:24 PM	20.80	26.00	34.30
286	12/27/2005	7:26 PM	20.80	28.00	35.30
287	12/27/2005	7:27 PM	19.80	27.10	35.40
288	12/27/2005	7:29 PM	13.50	18.70	28.10
289	12/27/2005	7:31 PM	22.80	29.10	33.20
290	12/27/2005	7:32 PM	21.90	29.20	35.40
291	12/27/2005	7:34 PM	26.00	34.30	40.50
292	12/27/2005	7:35 PM	32.20	41.50	47.80
293	12/27/2005	7:37 PM	20.80	27.00	35.30
294	12/27/2005	7:38 PM	20.80	31.10	37.40
295	12/27/2005	7:40 PM	26.00	33.20	43.60
296	12/27/2005	7:42 PM	21.80	30.10	35.30
297	12/27/2005	7:43 PM	24.90	33.20	47.80
298	12/27/2005	7:45 PM	15.60	20.80	29.10
299	12/27/2005	7:46 PM	28.10	32.30	40.60
300	12/27/2005	7:48 PM	10.40	15.60	19.80
301	12/27/2005	7:49 PM	17.70	22.90	31.20
302	12/27/2005	7:51 PM	17.60	26.00	30.10
303	12/27/2005	7:53 PM	12.50	24.90	29.10
304	12/27/2005	7:54 PM	16.60	27.00	35.30
305	12/27/2005	7:56 PM	19.70	29.10	33.20
306	12/27/2005	7:57 PM	14.50	21.80	31.20
307	12/27/2005	7:59 PM	18.70	23.90	28.00
308	12/27/2005	8:00 PM	27.00	35.30	40.50
309	12/27/2005	8:02 PM	18.70	29.10	32.20
310	12/27/2005	8:04 PM	19.70	30.10	37.40
311	12/27/2005	8:05 PM	22.90	28.00	34.30
312	12/27/2005	8:07 PM	17.60	23.90	31.10

POS	DATE	TIME	CPM1	CPM2	CPMW3
313	12/27/2005	8:08 PM	15.60	27.00	38.40
314	12/27/2005	8:10 PM	15.60	24.00	30.20
315	12/27/2005	8:11 PM	22.90	32.30	38.50
316	12/27/2005	8:13 PM	45.70	65.40	72.70
317	12/27/2005	8:14 PM	13.50	21.80	29.10
318	12/27/2005	8:16 PM	13.50	19.70	28.00
319	12/27/2005	8:18 PM	19.70	24.90	31.10
320	12/27/2005	8:19 PM	9.30	15.60	26.00
321	12/27/2005	8:21 PM	18.70	27.00	33.20
322	12/27/2005	8:22 PM	16.60	20.80	26.00
323	12/27/2005	8:24 PM	28.00	29.10	32.20
324	12/27/2005	8:25 PM	21.80	28.00	34.30
325	12/27/2005	8:27 PM	23.90	31.20	34.30
326	12/27/2005	8:29 PM	13.50	24.90	34.30
327	12/27/2005	8:30 PM	26.00	33.20	36.40
328	12/27/2005	8:32 PM	16.60	22.90	33.20
329	12/27/2005	8:33 PM	18.70	23.90	33.20
330	12/27/2005	8:35 PM	16.60	27.00	33.20
331	12/27/2005	8:36 PM	20.80	24.90	31.20
332	12/27/2005	8:38 PM	26.00	33.20	39.50
333	12/27/2005	8:39 PM	21.80	27.00	34.30
334	12/27/2005	8:41 PM	13.50	20.80	26.00
335	12/27/2005	8:43 PM	20.80	28.10	38.50
336	12/27/2005	8:44 PM	19.80	26.00	29.20
337	12/27/2005	8:46 PM	15.60	27.00	33.20
338	12/27/2005	8:47 PM	16.60	20.80	27.00
339	12/27/2005	8:49 PM	9.20	18.40	27.70
340	12/27/2005	8:50 PM	20.80	24.00	32.30
341	12/27/2005	8:52 PM	27.70	32.80	38.90
342	12/27/2005	8:53 PM	11.40	17.60	21.80
343	12/27/2005	8:55 PM	28.00	38.40	46.70

POS	DATE	TIME	CPM1	CPM2	CPMW3
344	12/27/2005	8:57 PM	13.50	19.70	30.10
345	12/27/2005	8:58 PM	20.80	29.10	41.50
346	12/27/2005	9:00 PM	36.50	44.80	57.30
347	12/27/2005	9:01 PM	19.80	25.00	32.30
348	12/27/2005	9:03 PM	15.60	21.90	26.00
349	12/27/2005	9:04 PM	16.70	26.00	30.20
350	12/27/2005	9:06 PM	22.80	29.10	35.30
351	12/27/2005	9:08 PM	18.70	27.00	33.30
352	12/27/2005	9:09 PM	21.90	31.30	36.50
353	12/27/2005	9:11 PM	15.60	22.90	32.20
354	12/27/2005	9:12 PM	18.70	24.90	32.20
355	12/27/2005	9:14 PM	16.60	28.00	35.30
356	12/27/2005	9:15 PM	15.60	24.90	37.40
357	12/27/2005	9:17 PM	22.90	28.10	35.30
358	12/27/2005	9:18 PM	11.50	15.60	22.90
359	12/27/2005	9:20 PM	26.00	32.20	37.40
360	12/27/2005	9:22 PM	21.80	27.00	36.30
361	12/27/2005	9:23 PM	27.60	35.80	43.00
362	12/27/2005	9:25 PM	18.70	23.90	32.20
363	12/27/2005	9:26 PM	14.60	20.80	27.10
364	12/27/2005	9:28 PM	16.60	19.70	25.90
365	12/27/2005	9:29 PM	17.70	27.00	33.20
366	12/27/2005	9:31 PM	26.00	32.20	43.60
367	12/27/2005	9:33 PM	18.70	26.00	32.20
368	12/27/2005	9:34 PM	26.00	35.30	39.50
369	12/27/2005	9:36 PM	19.70	26.00	34.30
370	12/27/2005	9:37 PM	13.50	26.00	34.40
371	12/27/2005	9:39 PM	22.80	27.00	34.30
372	12/27/2005	9:40 PM	21.80	27.00	35.30
373	12/27/2005	9:42 PM	21.80	29.10	36.30
374	12/27/2005	9:43 PM	19.80	25.00	32.30

POS	DATE	TIME	CPM1	CPM2	CPMW3
375	12/27/2005	9:45 PM	22.80	28.00	34.30
376	12/27/2005	9:47 PM	11.50	16.70	25.00
377	12/27/2005	9:48 PM	13.50	20.80	29.10
378	12/27/2005	9:50 PM	12.50	15.60	26.00
379	12/27/2005	9:51 PM	23.90	35.30	42.60
380	12/27/2005	9:53 PM	21.80	31.10	36.30
381	12/27/2005	9:54 PM	21.90	26.00	28.10
382	12/27/2005	9:56 PM	12.50	17.70	22.80
383	12/27/2005	9:58 PM	30.10	34.30	43.60
384	12/27/2005	9:59 PM	24.00	32.30	39.60
385	12/27/2005	10:01 PM	26.00	37.40	46.70
386	12/27/2005	10:02 PM	16.60	21.80	28.00
387	12/27/2005	10:04 PM	24.90	31.20	36.30
388	12/27/2005	10:05 PM	18.70	22.80	24.90
389	12/27/2005	10:07 PM	18.70	24.90	30.10
390	12/27/2005	10:08 PM	22.90	31.20	39.50
391	12/27/2005	10:10 PM	19.70	28.00	37.40
392	12/27/2005	10:12 PM	24.90	33.30	42.60
393	12/27/2005	10:13 PM	16.60	21.80	33.20
394	12/27/2005	10:15 PM	31.30	38.60	42.70
395	12/27/2005	10:16 PM	21.80	28.10	37.40
396	12/27/2005	10:18 PM	22.80	33.20	38.40
397	12/27/2005	10:19 PM	19.80	24.00	33.30
398	12/27/2005	10:21 PM	26.00	37.40	42.60
399	12/27/2005	10:23 PM	20.80	28.10	36.40
400	12/27/2005	10:24 PM	17.70	24.90	34.30
401	12/27/2005	10:26 PM	16.40	26.60	40.90
402	12/27/2005	10:27 PM	29.10	37.40	40.50
403	12/27/2005	10:29 PM	18.70	24.90	29.10
404	12/27/2005	10:30 PM	27.10	31.20	36.40
405	12/27/2005	10:32 PM	25.00	31.30	41.70

POS	DATE	TIME	CPM1	CPM2	CPMW3
406	12/27/2005	10:34 PM	27.00	40.50	49.80
407	12/27/2005	10:35 PM	11.50	16.70	24.00
408	12/27/2005	10:37 PM	31.20	45.70	58.20
409	12/27/2005	10:38 PM	22.90	28.10	36.40
410	12/27/2005	10:40 PM	21.80	26.00	31.10
411	12/27/2005	10:41 PM	19.80	30.20	34.40
412	12/27/2005	10:43 PM	22.80	28.00	38.40
413	12/27/2005	10:44 PM	20.80	26.00	32.20
414	12/27/2005	10:46 PM	24.90	33.20	40.50
415	12/27/2005	10:48 PM	28.10	36.40	42.70
416	12/27/2005	10:49 PM	23.90	28.00	37.40
417	12/27/2005	10:51 PM	17.70	24.00	31.30
418	12/27/2005	10:52 PM	27.00	30.10	36.30
419	12/27/2005	10:54 PM	19.80	30.20	34.40
420	12/27/2005	10:55 PM	20.80	31.20	40.50
421	12/27/2005	10:57 PM	20.80	30.20	38.50
422	12/27/2005	10:59 PM	18.80	26.00	37.50
423	12/27/2005	11:00 PM	20.80	27.00	31.20
424	12/27/2005	11:02 PM	22.80	28.00	33.20
425	12/27/2005	11:03 PM	25.00	33.30	39.60
426	12/27/2005	11:05 PM	22.80	32.20	38.40
427	12/27/2005	11:06 PM	22.80	34.30	37.40
428	12/27/2005	11:08 PM	18.70	24.90	33.20
429	12/27/2005	11:09 PM	28.00	35.30	42.60
430	12/27/2005	11:11 PM	22.80	30.10	36.30
431	12/27/2005	11:13 PM	29.10	42.60	49.90
432	12/27/2005	11:14 PM	16.70	23.90	29.20
433	12/27/2005	11:16 PM	28.10	33.20	41.60
434	12/27/2005	11:17 PM	13.50	20.80	31.20
435	12/27/2005	11:19 PM	9.40	15.60	21.80
436	12/27/2005	11:21 PM	27.10	36.50	49.00

POS	DATE	TIME	CPM1	CPM2	CPMW3
437	12/27/2005	11:22 PM	13.50	18.70	21.80
438	12/27/2005	11:24 PM	18.80	27.10	34.40
439	12/27/2005	11:25 PM	22.80	29.10	35.30
440	12/27/2005	11:27 PM	13.50	23.90	41.60
441	12/27/2005	11:28 PM	14.60	18.70	31.20
442	12/27/2005	11:30 PM	16.60	24.90	35.30
443	12/27/2005	11:31 PM	24.00	28.10	34.40
444	12/27/2005	11:33 PM	28.00	37.40	45.70
445	12/27/2005	11:35 PM	19.70	21.80	31.20
446	12/27/2005	11:36 PM	22.80	34.30	39.50
447	12/27/2005	11:38 PM	17.70	28.10	41.60
448	12/27/2005	11:39 PM	22.90	33.30	43.70
449	12/27/2005	11:41 PM	15.60	22.90	31.30
450	12/27/2005	11:42 PM	24.00	34.40	41.70
451	12/27/2005	11:44 PM	32.20	42.60	53.00
452	12/27/2005	11:46 PM	15.60	20.80	29.10
453	12/27/2005	11:47 PM	14.60	18.80	20.80
454	12/27/2005	11:49 PM	15.60	22.90	28.00
455	12/27/2005	11:50 PM	21.80	30.10	38.40
456	12/27/2005	11:52 PM	20.80	27.00	38.40
457	12/27/2005	11:53 PM	19.80	26.00	38.50
458	12/27/2005	11:55 PM	12.50	21.80	40.60
459	12/27/2005	11:56 PM	33.20	42.60	47.80
460	12/27/2005	11:58 PM	22.90	28.10	33.20
461	12/28/2005	0:00 AM	15.60	18.70	32.20
462	12/28/2005	0:01 AM	14.60	26.00	43.70
463	12/28/2005	0:03 AM	17.70	23.90	32.20
464	12/28/2005	0:04 AM	30.10	40.50	43.60
465	12/28/2005	0:06 AM	11.50	16.70	35.40
466	12/28/2005	0:07 AM	17.70	21.80	32.20
467	12/28/2005	0:09 AM	17.70	23.90	31.20

POS	DATE	TIME	CPM1	CPM2	CPMW3
468	12/28/2005	0:10 AM	19.70	28.10	36.40
469	12/28/2005	0:12 AM	19.80	33.30	39.60
470	12/28/2005	0:14 AM	20.80	28.10	35.40
471	12/28/2005	0:15 AM	20.80	28.00	40.50
472	12/28/2005	0:17 AM	12.50	15.60	19.80
473	12/28/2005	0:18 AM	27.00	32.20	37.40
474	12/28/2005	0:20 AM	20.80	27.00	29.10
475	12/28/2005	0:21 AM	19.70	24.90	32.20
476	12/28/2005	0:23 AM	16.70	25.00	29.20

* Samples were found to have scint fluid on outside of vials. I will clean and rerun.

Page 37 of 38

PROTOCOL : 94 118B B401
 DATE : 2006/01/03
 TIME : 08:20
 ID : P94AS003

Reswipes of Lab

Wallac 1400 DSA ver 2.50 S/N 4091042
 Counting mode : CPM
 Isotope(s) : H3,C14
 H3 = 5- 350,12.43 y
 C14 = 5- 660,5730.00 y
 Protocol name : 118B B401
 Counting time : 60
 Repeats : 1
 Cycles : 1
 Replicates : 1
 2 sigma % : 0.01
 Minimum cpm : 0.00 Checking time: 10
 Advanced modes : Chemilum
 Output to Display :
 POS,DATE,TIME,CPM1,CPM2,CPMw3,RACKPOS,CPM,CTIME,SQPI
 Additions to Display : Spectrum,Header,Listing
 Header : Reswipes of Lab
 Spectrum : Beta
 Window 1 : 1-1024 /Beta
 Window 2 : 1-1024 /Beta
 Window 3 : 1-1024 /Beta
 Window 4 : 1-1024 /Beta
 Window 5 : 1-1024 /Beta
 Window 6 : 1-1024 /Beta
 FNCT1 = FNCT1 :
 FNCT2 = FNCT2 :
 FNCT3 = FNCT3 :
 FNCT4 = FNCT4 :

Total count rate:

H3 50.9 CPM
 C14 70.7 CPM

POS	DATE	TIME	CPM1	CPM2	CPMW3	
1	1/3/2006	8:20 AM	27.00	32.20	38.40	BACKGROUND
2	1/3/2006	8:22 AM	8.30	16.70	30.20	PERON # 73

POS	DATE	TIME	CPM1	CPM2	CPMW3	
3	1/3/2006	8:24 AM	15.60	21.80	27.00	RERUN # 75

SAMPLES WERE CLEANED AND RERUN AND FOUND
TO BE WITHIN 2x BACKGROUND.