

# Radiological Survey Sheet

Job Location Pfizer, Inc. Groton, Ct. 06340 118 N / NO7 Page 1 of 6  
 Survey Purpose Radiologic Survey of Equipment LAB DECOMM Date 10-3-12  
 Performed by E. GAILOR E. Gailor  
 (Print) (Sign)

Instrument (Model/S/N)	Packard Tri-Carb S/N 431521	Packard Tri-Carb S/N 431520	Ludlum 2241-2 163603	Wallac Model 1409 S/N 4061042
Det. (Model/SN)	Internal	Internal	Ludlum 43-68 PR 149613	Internal
Type Rad.	$\beta$	$\beta$	$\beta$	$\beta$
Cal. Due:	03/05/2013	03/05/2013	04/17/2013	09/15/2012
Check Instruments Used	✓		✓	

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm <sup>2</sup> )
1.	0910	BACKGROUND	
2.		door	
3.		FLOOR	
4.		UPPER WALL	
5.		Lower "	
6.		UPPER "	
7.		Lower "	
8.		CAB. doors	
9.		" TOP SHELF	
10.		" 2ND "	
11.		" 3RD "	
12.		SINK AREA	
13.		UPPER wall	
14.		Lower "	
15.		UPPER "	
16.		Lower "	
17.		Hood LEFT SIDE	
18.	✓	" " BACK	

# Radiological Survey Continuation Sheet

Page 2 of 6

Date 10-3-12  
118N1N-07

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm <sup>2</sup> )
19	0910	HOOD RIGHT BACK	
20		" " BOTTOM	
21		" LEFT "	
22		" SASH	
23		" LIP	
24		CAB DOOR	
25		INSIDE CAB	
26		UPPER WALL	
27		LOWER "	
28		UPPER "	
29		LOWER "	
30		UPPER "	
31		LOWER "	
32		UPPER "	
33		LOWER "	
34		UPPER "	
35		LOWER "	
36		UPPER "	
37		LOWER "	
38		UPPER "	
39		LOWER "	
40		UPPER "	
41		LOWER "	
42		FLOOR	
43		"	
44		"	
45	✓	"	

## Radiological Survey Continuation Sheet

Page 3 of 6

Date 10-3-12


118N/NO7

No.	Time	Locations	See attached LSC printout for Total Activity (cpm/100 cm <sup>2</sup> )
46	0910	F100L	
47		"	
48		"	
49		"	
50		"	
51		"	
52		"	
53		"	
54		"	
55		"	
56		"	
57		"	
58		"	
59	✓	"	
60	✓	"	

118 N/N07

Pg 4 of 6

10-3-12



O = W.I.P.F. LOCATION

- Surveyed 100% OF LOWER SURFACES USING 43-68. NO REMAINS IN EXCESS  
OF NORMAL BACKGROUND LEVELS WERE NOTED.

Assay Definition-

Assay Description:  
Equipment Swipes

Assay Type: CPM  
Report Name: Report1  
Output Data Path: C:\Packard\Tricarb\Results\Ed Gailor\Eds Monthly Swipes  
Raw Results Path: C:\Packard\Tricarb\Results\Ed Gailor\Eds Monthly Swipes\20121003\_1249  
\20121003\_1249.results  
RTF File Name: C:\Packard\Tricarb\Results\Ed Gailor\Eds Monthly Swipes\Ed's Monthly  
Swipes.rtf  
Assay File Name: C:\Packard\TriCarb\Assays\Eds Monthly Swipes.lsa

Count Conditions-

Nuclide: Triple  
Quench Indicator: tSIE/AEC  
External Std Terminator (sec): 5 sec  
Pre-Count Delay (min): 0.00  
Quench Set: n/a  
Count Time (min): 1.00  
Count Mode: Normal  
Assay Count Cycles: 1  
#Vials/Sample: 1  
Repeat Sample Count: 1  
Calculate % Reference: Off

Background Subtract: On - Manual  
Low CPM Threshold: Off  
2 Sigma % Terminator: Off

Regions	LL	UL	Bkg Subtract
A	0.0	18.6	0.00
B	18.6	156.0	0.00
C	0.0	2000.0	0.00

Count Corrections-

Static Controller: On  
Colored Samples: n/a  
Coincidence Time (nsec): 18  
Luminescence Correction: On  
Heterogeneity Monitor: n/a  
Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

S#	CPMA	CPMB	CPMC	MESSAGES
1	3	11	17	
2	12	7	26	
3	11	9	26	
4	9	16	28	
5	6	8	24	
6	10	9	22	
7	6	12	26	
8	0	9	13	
9	2	11	18	
10	6	7	15	

Protocol# 3 - Eds Monthly Swipes.lsa

User: Ed Gailor

11	14	7	24
12	3	11	20
13	6	12	20
14	5	5	18
15	8	9	20
16	2	7	16
17	12	11	29
18	0	12	13
19	18	4	28
20	8	15	29
21	9	6	23
22	5	4	14
23	15	9	31
24	13	14	32
25	14	8	27
26	10	10	23
27	10	8	23
28	6	10	27
29	12	6	24
30	0	11	18
31	4	8	19
32	3	11	22
33	4	8	15
34	7	5	16
35	14	9	31
36	3	9	21
37	0	11	17
38	11	12	24
39	0	17	22
40	6	11	19
41	12	7	26
42	14	7	28
43	11	7	22
44	8	14	28
45	12	13	36
46	5	15	26
47	6	4	16
48	12	11	30
49	10	9	23
50	10	7	21
51	7	16	30
52	8	8	21
53	8	6	18
54	9	6	22
55	15	8	29
56	12	10	27
57	12	14	33
58	2	11	15
59	2	8	16
60	14	8	30