

Frank P.D. Tran
MATERIALS LICENSING BRANCH
UNITED STATES NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE ROAD SUITE 210
LISLE, ILLINOIS 60532-4352
OFFICE: (630)-829-9623 FAX: (630) 515-1078

CONVERSATION RECORD	TIME CALLED	DATE CALLED
Faxed to 417-347-9164 on 7/3/13	1:58 pm CST.	August 24, 2011

NAME OF PERSON(S) CONTACTED	ORGANIZATION	TELEPHONE NO.
Lesa Deardorff Director of Radiology	Freeman –Oak Hill Health System	O: (417) 347-4934

SUBJECT		
License No.: 24-17205-01	Control No.:	580262

SUMMARY

This refers to your renewal application dated March 18, 2013, and the phone conversation between Lesa Deardorff and Frank Tran on July 3, 2013.

During the conversation, the following information were requested.

- 1) The licensee's mailing address.
- 2) The total possession limit for 10 CFR 35.300 materials.
- 3) Confirm that the location of use located 932 East 34th Street, Joplin, Missouri, must be listed in the license.
- 4) Confirm that the licensee has not possessed any 3M Series 6500 Tube Sources.
- 5) Acknowledge that the authorization for the use of 35.400 for Dr. Michael Driver will be removed.
- 6) Acknowledge that License Condition 14 regarding the decommissioning financial assurance requirements will be removed because it did not apply to the scope of the licensee's program.
- 7) Provide names of authorized users that will be listed in the license.
- 8) Provide a copy of the shielding calculations (additional shieldings for the facilities and handling devices) regarding the use of the PET isotope(s).

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

ACTION REQUIRED

Please submit the requested information in writing with proper signature by July 22, 2013. Please referencing control number **580262** in the response to facilitate proper handling in our office. Your response could be faxed to 630-515-1078 or scan and email to frank.tran@nrc.gov (no need to mail the original).


PLEASE DIRECT ANY QUESTIONS YOU MAY HAVE REGARDING THIS REQUEST TO ME AT 630-829-9623 OR AT (800) 522-3025, EXT. 9623.

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

Frank P.D. Tran



July 3, 2013

TRANSMISSION VERIFICATION REPORT

TIME : 07/03/2013 15:58
 NAME : USNRC RIII
 FAX : 6308299782
 TEL :
 SER.# : 000A7J925774

DATE, TIME	07/03 15:57
FAX NO./NAME	84173479164
DURATION	00:00:48
PAGE(S)	03
RESULT	OK
MODE	STANDARD ECM

NRC FORM 386 (RIII)
 (4-2004)



UNITED STATES
 NUCLEAR REGULATORY COMMISSION
 REGION III
 2443 Warrenville Road, Suite 210
 Lisle, Illinois 60532-4352

TELEFAX TRANSMITTAL

DATE: 7/3/13 NUMBER OF PAGES: 3
 (including this page)

SEND TO: Lesia Deardorff, Director of Radiology

LOCATION: Freeman - Oak Hill Health System

FAX NUMBER: 417 - 347 - 9164 VERIFY BY CALLING SENDER

FROM: Frank Tran
 (SENDER)

TELEPHONE NUMBER: 630 - 829 - 9623 FAX NUMBER: 630 - 515 - 1078

If you do not receive the complete fax transmittal, please contact the sender as soon as possible at the telephone number provided above.

MESSAGE
As we discussed today, the additional information