

HOSPITAL SAN CRISTOBAL

6-8

September 10, 2013

HECTOR BERMUDEZ,

**LICENSING ASSISTANCE TEAM
DIVISION OF NUCLEAR MATERIALS SAFETY
U.S. NUCLEAR REGULATORY COMMISSION, REGION I
REGION I, 2100 RENAISSANCE BLVD, SUITE 100
KING OF PRUSSIA, PA 19406-2713**


SUBJ: Renewal of License NO. 52-25232-01 Docket Number 030-33026

Gentlemen:

Please find (Attachment 1) the additional information requested for our renewal license 52-25232-01.

Thanks in advance for your cooperation on this matter.

Cordially yours.


**Miguel A. Serpa, M.D.
RSO and Nuclear Medicine Director**

**Hospital San Cristobal
P.O. Box 800501
Coto Laurel
Ponce Puerto Rico 00780-0501**

Enclosures:

ATTACHMENT 1

580597
NMSS/RGN1 MATERIALS-002

Attachment 1

Radiation Safety Program

Item Number and Title	Suggested Response	Check Box if Included
Item 9: Dose Calibrator and other Dosage Measuring Equipment.	A statement that: "Equipment used to measure dosages will be calibrated in accordance with nationally recognized standards or the manufacturer's instructions."	X
Item 10: Safe Use of Unsealed Licensed Material	A statement that: "We have developed and will implement and maintain procedures for safe use of unsealed byproduct material that meet the requirements of 10 CFR 20.1101 and 10 CFR 20.1301	X
Item 10: Spill/Contamination Procedures	A statement that: "We have developed and will implement and maintain written procedures for safe response to spill of licensed material in accordance with 10 CFR 20.1101."	X