

QC Data for Surveys

Date	Room #	BKGD GM Counter	BKGD LB122	Standard LB122	MDA LB122	Bq	LSC avg BKGD	nch curve correction C-14	% Eff quench curve correction H-3
04/02/2013	1	< 100 cpm	0.72 kcpm	3.1 kcpm	2.8 kcpm		43 dpm	1.00	0.94
04/03/2013	2	<100	0.78	3.1	2.9		31	1.00	0.91
04/04/2013	4	<100	0.81	3.1	2.9		40	1.01	1.05
04/02/2013	6	<100	0.93	3.1	3.1		43	1.00	0.94
04/03/2013	10	<100	0.78	3.1	2.9		31	.100	0.92
04/02/2013	11	<100	0.76	3.1	2.8		27	1.00	0.96
03/17/2013	12	<100	0.79	3.1	2.9		26	1.00	0.95
03/26/2013	13	<100	0.77	3.1	2.9		31	1.00	0.95
04/01/2013	14	<100	0.8	3.1	2.9		34	1.00	0.95
03/29/2013	15	<100	0.87	3.1	3		25	1.00	0.95
04/01/2013	16	<100	0.8	3.1	2.9		34	1.00	0.95
03/29/2013	17	<100	0.8	3.1	2.9		31	100	0.95
04/01/2013	18	<100	0.8	3.1	2.9		34	1.00	0.95
03/29/2013	19	<100	0.77	3.1	2.9		25	1.00	0.95
03/29/2013	20	<100	0.85	3.1	3		44	1.00	0.95
04/02/2013	21	<100	0.8	3.1	2.9		27	1.00	0.94
04/03/2013	Custodial	<100	0.86	3.1	3		40	1.00	0.97
04/03/2013	Thin Closet	<100	0.86	3.1	3		40	1.00	0.97
04/01/2013	Men's	<100	0.89	3.1	3.1		34	1.00	0.96
04/01/2012	Women's	<100	0.9	3.1	3.1		34	1.00	0.96
04/01/2013	Hallway	<100	0.79	3.1	2.9		40	1.00	1.03
04/15/2013	QA/QC	<100	0.8	3.1	2.9		25	1.00	0.96