

Murnahan, Colleen

From: McKee, Bill [BMcKee@IPSAdvantage.com]
Sent: Thursday, September 19, 2013 7:57 AM
To: Hill, Carol
Cc: Rash, Marilyn K.; jacquelinecook@nrc.gov; Murnahan, Colleen
Subject: Termination of RA license
Attachments: Leak Tests.pdf; Termination Form 9-19-13.pdf

Attached is the information for termination of RA license # 42-29399-02. Thank you for your assistance in this matter.



Bill McKee
HSE Manager Gulf Coast Region
105 Bercegeay Rd.
Broussard, LA 70518
Office: (337) 714-2100
Fax: (337) 714-2112
Cell: (337) 288-3539
bmckee@ipsadvantage.com

Lack of SAFETY is a Problem
Be Part of the Solution

CERTIFICATE OF DISPOSITION OF MATERIALS

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Information Services Branch (T-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to Infocollects.Resource@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

LICENSEE NAME AND ADDRESS

Integrated Production Services, Inc.
16800 Greenspoint Park Drive
Suite 200S
Houston, TX. 77060

LICENSE NUMBER

42-29399-02

DOCKET NUMBER

LICENSE EXPIRATION DATE

09/30/2013

A. LICENSE STATUS (Check the appropriate box)

- ☐ This license has expired. ☒ This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

- ☐ 1. No radioactive materials have ever been procured or possessed by the licensee under this license.
- ☒ 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner.
- ☒ a. Transfer of radioactive materials to the licensee listed below:
Warrior Wireline Services
- ☐ b. Disposal of radioactive materials:
- ☐ 1. Directly by the licensee:
- ☐ 2. By licensed disposal site:
- ☐ 3. By waste contractor:
- ☒ c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

PUBLIC

- ☐ Immediate Release
☒ Normal Release

NON-PUBLIC

- ☐ A.3 Sensitive-Security Related
☐ A.7 Sensitive Internal
☐ Other: _____

Reviewer: MS Date: 10/3/13

C. SURVEYS PERFORMED AND REPORTED

- ☐ 1. A radiation survey was conducted by the licensee. The survey confirms:
- ☐ a. the absence of licensed radioactive materials
- ☐ b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
- ☐ 2. A copy of the radiation survey results:
- ☐ a. is attached; or ☐ b. is not attached (Provide explanation); or ☐ c. was forwarded to NRC on: _____ Date: _____
- ☒ 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
- ☒ a. The results of the latest leak test are attached; and/or ☐ b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:


NAME	TITLE	TELEPHONE (Include Area Code)	E-MAIL ADDRESS
Bill McKee	HSE Manager/RSO	(337) 288-3539	bmckee@ipsadvantage.com

Mail all future correspondence regarding this license to:

Marilyn Rash @ IPS 16800 Greenspoint Park Drive Suite 200S, Houston, TX 77060

C. CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE	SIGNATURE	DATE
Bill McKee		09/18/2013

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.



P.O. Box 8608
Midland, TX 79708
(432) 689-8043

CERTIFICATE OF LEAK TEST

INTEGRATED PRODUCTION SERVICES
4100 S. COUNCIL ROAD
OKLAHOMA CITY OK 73179

This is to certify that a swab identified as:

ISOTOPE: AmBe-241
ACTIVITY: 3.0 Ci
SERIAL NUMBER: NEE1-71-1

Analyzed By: Amanda Rosales

Date Leak Test Performed: 10/17/2012

Date to be Leak Tested Again: 4/17/2013

Analysis of the swab indicated less than 0.00025 microcuries of removable contamination.**

ISOTECH LABORATORY, INC.

BY: *Amanda Rosales*
Amanda Rosales
Laboratory Manager

** When analysis is approximately background, contamination level is recorded as Less Than 0.00025 microcuries of removable contamination.



P.O. Box 8608
Midland, TX 79708
(432) 689-8043

CERTIFICATE OF LEAK TEST

INTEGRATED PRODUCTION SERVICES
4100 S. COUNCIL ROAD
OKLAHOMA CITY OK 73179

This is to certify that a swab identified as:

ISOTOPE: AmBe-241
ACTIVITY: 100 mCi
SERIAL NUMBER: 539/08

Analyzed By: Amanda Rosales

Date Leak Test Performed: 10/17/2012

Date to be Leak Tested Again: 4/17/2013

Analysis of the swab indicated less than 0.00025 microcuries of removable contamination.**

ISOTECH LABORATORY, INC.

BY: *Amanda Rosales*
Amanda Rosales
Laboratory Manager

** When analysis is approximately background, contamination level is recorded as Less Than 0.00025 microcuries of removable contamination.



P.O. Box 8608
Midland, TX 79708
(432) 689-8043

CERTIFICATE OF LEAK TEST

INTEGRATED PRODUCTION SERVICES
4100 S. COUNCIL ROAD
OKLAHOMA CITY OK 73179

This is to certify that a swab identified as:

ISOTOPE: AmBe-241
ACTIVITY: 100 mCi
SERIAL NUMBER: 567/08

Analyzed By: Amanda Rosales

Date Leak Test Performed: 10/17/2012

Date to be Leak Tested Again: 4/17/2013

Analysis of the swab indicated less than 0.00025 microcuries of removable contamination.**

ISOTECH LABORATORY, INC.

BY: *Amanda Rosales*
Amanda Rosales
Laboratory Manager

** When analysis is approximately background, contamination level is recorded as Less Than 0.00025 microcuries of removable contamination.



P.O. Box 8608
Midland, TX 79708
(432) 689-8043

CERTIFICATE OF LEAK TEST

INTEGRATED PRODUCTION SERVICES
4100 S. COUNCIL ROAD
OKLAHOMA CITY OK 73179

This is to certify that a swab identified as:

ISOTOPE: AmBe-241
ACTIVITY: 100 mCi
SERIAL NUMBER: 566/08

Analyzed By: Amanda Rosales

Date Leak Test Performed: 10/17/2012

Date to be Leak Tested Again: 4/17/2013

Analysis of the swab indicated less than 0.00025 microcuries of removable contamination.**

ISOTECH LABORATORY, INC.

BY: *Amanda Rosales*
Amanda Rosales
Laboratory Manager

** When analysis is approximately background, contamination level is recorded as Less Than 0.00025 microcuries of removable contamination.



P.O. Box 8608
Midland, TX 79708
(432) 689-8043

CERTIFICATE OF LEAK TEST

INTEGRATED PRODUCTION SERVICES
4100 S. COUNCIL ROAD
OKLAHOMA CITY OK 73179

This is to certify that a swab identified as:

ISOTOPE: AmBe-241
ACTIVITY: 15.2 Ci
SERIAL NUMBER: N-523

Analyzed By: Amanda Rosales

Date Leak Test Performed: 10/17/2012

Date to be Leak Tested Again: 4/17/2013

Analysis of the swab indicated less than 0.00025 microcuries of removable contamination.**

ISOTECH LABORATORY, INC.

BY: *Amanda Rosales*
Amanda Rosales
Laboratory Manager

** When analysis is approximately background, contamination level is recorded as Less Than 0.00025 microcuries of removable contamination.



P.O. Box 8608
Midland, TX 79708
(432) 689-8043

CERTIFICATE OF LEAK TEST

INTEGRATED PRODUCTION SERVICES
4100 S. COUNCIL ROAD
OKLAHOMA CITY OK 73179

This is to certify that a swab identified as:

ISOTOPE: AmBe-241
ACTIVITY: 100 mCi
SERIAL NUMBER: 568/08

Analyzed By: Amanda Rosales

Date Leak Test Performed: 10/17/2012

Date to be Leak Tested Again: 4/17/2013

Analysis of the swab indicated less than 0.00025 microcuries of removable contamination.**

ISOTECH LABORATORY, INC.

BY: *Amanda Rosales*
Amanda Rosales
Laboratory Manager

** When analysis is approximately background, contamination level is recorded as Less Than 0.00025 microcuries of removable contamination.



P.O. Box 8608
Midland, TX 79708
(432) 689-8043

CERTIFICATE OF LEAK TEST

INTEGRATED PRODUCTION SERVICES
4100 S. COUNCIL ROAD
OKLAHOMA CITY OK 73179

This is to certify that a swab identified as:

ISOTOPE: AmBe-241
ACTIVITY: 3.0 Ci
SERIAL NUMBER: NEE1-71-1

Analyzed By: Amanda Rosales

Date Leak Test Performed: 10/17/2012

Date to be Leak Tested Again: 4/17/2013

Analysis of the swab indicated less than 0.00025 microcuries of removable contamination.**

ISOTECH LABORATORY, INC.

BY: *Amanda Rosales*
Amanda Rosales
Laboratory Manager

** When analysis is approximately background, contamination level is recorded as Less Than 0.00025 microcuries of removable contamination.



P.O. Box 8608
Midland, TX 79708
(432) 689-8043

CERTIFICATE OF LEAK TEST

INTEGRATED PRODUCTION SERVICES
4100 S. COUNCIL ROAD
OKLAHOMA CITY OK 73179

This is to certify that a swab identified as:

ISOTOPE: AmBe-241
ACTIVITY: 100 mCi
SERIAL NUMBER: 566/08

Analyzed By: Amanda Rosales

Date Leak Test Performed: 10/17/2012

Date to be Leak Tested Again: 4/17/2013

Analysis of the swab indicated less than 0.00025 microcuries of removable contamination.**

ISOTECH LABORATORY, INC.

BY: *Amanda Rosales*
Amanda Rosales
Laboratory Manager

** When analysis is approximately background, contamination level is recorded as Less Than 0.00025 microcuries of removable contamination.

582170

BETWEEN:

Accounts Receivable/Payable
and
Regional Licensing Branches

[FOR ARPB USE]
INFORMATION FROM WBL

Program Code: 03110
Status Code: Pending Termination
Fee Category: 5A
Exp. Date:
Fee Comments:
Decom Fin Assur Req: N

License Fee Worksheet - License Fee Transmittal

A. REGION

1. APPLICATION ATTACHED

Applicant/Licensee: Integrated Production Services, Inc.
Received Date: 09/19/2013
Docket Number: 3038322
Mail Control Number: 582170
License Number: 42-29399-02
Action Type: Termination

2. FEE ATTACHED

Amount: _____

Check No.: _____

3. COMMENTS

Signed: _____

Colleen Duranahan

Date: _____

9-25-13

B. LICENSE FEE MANAGEMENT BRANCH (Check when milestone 03 is entered / /)

1. Fee Category and Amount: _____

2. Correct Fee Paid. Application may be processed for:

Amendment: _____

Renewal: _____

License: _____

3. OTHER _____

Signed: _____

Date: _____



DATE

10/03/2013

NAME AND ADDRESS OF APPLICANT AND/OR LICENSEE

Integrated Production Services, Inc.
ATTN: Bill McKee, Radiation Safety Officer
c/o Marilyn Rash@ IPS
16800 Greenspoint Park Drive, Suite 200S
Houston, TX 77060

LICENSE NUMBER

42-29399-02

MAIL CONTROL NUMBER

582170

LICENSING AND/OR TECHNICAL REVIEWER

cmurnahan *CM*

This is to acknowledge the receipt of your:

☒ LETTER and/or ☐ APPLICATION

DATED: 09/18/2013

The initial processing, which included an administrative review, has been performed.

☐ AMENDMENT ☒ TERMINATION ☐ NEW LICENSE ☐ RENEWAL

- ☒ There were no administrative omissions identified during our initial review.
- ☐ This is to acknowledge receipt of your application for renewal of the material(s) license identified above. Your application is deemed timely filed, and accordingly, the license will not expire until final action has been taken by this office.
- ☐ Your application for a new NRC license did not include your taxpayer identification number. Please fill out NRC Form 531, located at the following link:

<http://www.nrc.gov/reading-rm/doc-collections/forms/nrc531.pdf>

Send the completed NRC Form 531, by facsimile, to the following number: (301) 415-5387

A copy of your action has been emailed to our License Fee and Accounts Receivable Branch, in our Headquarters office in Rockville, MD. You will be contacted separately if there is a fee issue involved.

Your application has been assigned the above listed **MAIL CONTROL NUMBER**. When calling to inquire about this action, please refer to this control number. Your application has been forwarded to a technical reviewer. Please note that the technical review, which is normally completed within 180 days for a renewal application (90 days for all other requests), may identify additional omissions or require additional information. If you have any questions concerning the processing of your application, our contact information is listed below:

Region IV
U. S. Nuclear Regulatory Commission
DNMS/NMSB - B
1600 E. Lamar Blvd.
Arlington, TX 76011-4511
(817) 200-1103 or (817) 200-1140