

Roldan, Lizette

From: Roldan, Lizette
Sent: Monday, October 21, 2013 3:10 PM
To: 'preynolds@rimrock.biz'
Subject: REQUEST FOR ADDITIONAL INFORMATION REGARDING RENEWAL CONTROL 580510
Attachments: apph-errata.pdf

License No.: 25-27752-01
Docket No.: 030-36242
Control No.: 580510

Dear Mr. Reynolds:

This is in reference to your letter dated April 22, 2013 requesting to renew Nuclear Regulatory Commission License Number 25-27752-01. In order to continue our review, we need the following additional information:

1. Establish a maximum possession limit for the cesium-137 and americium-241 you wish to have on the license, or state the amount of cameras you may possess at any one time. Please keep in mind you may want to anticipate future purchases when determining your maximum possession limit.
2. Provide a statement: "We will implement and maintain the operating, emergency, and security procedures in Errata: Appendix H of NUREG-1556, Vol. Rev. 1, Dated November 2001, and provide copies of these procedures to all gauge users and at each job site."

Appendix H of NUREG-1556 is attached for your convenience.

We will continue our review upon receipt of this information. Please reply to my attention by October 25, 2013, and refer to Mail Control No. 580510. You may send your response as an attachment in PDF format. If you have any technical questions regarding this deficiency letter, please call me at 817-200-1596.

Sincerely,

Lizette Roldán-Otero, Ph.D.

Health Physicist
U.S. Nuclear Regulatory Commission
1600 E. Lamar Blvd.
Arlington, TX 76011-4511
Office: 817-200-1596
Fax: 817-200-1188

PUBLIC

- Immediate Release
 Normal Release

NON-PUBLIC

- A.3 Sensitive-Security Related
 A.7 Sensitive Internal
 Other: _____

Reviewer: Jul Date: 10/22/13



RIMROCK ENGINEERING, INC.

Phone 406.294.8400

5440 Holiday Avenue • Billings, MT 59101
October 21, 2013

Fax 406.294.8405

U. S. Nuclear Regulatory Commission
1600 E. Lamar Blvd.
Arlington, TX 76011-4511

Attn: Lizette Roldan-Otero, Ph.D.

RE: Renewal of License Number 25-27752-01

Thank you for the phone call and information concerning your review of our license for renewal. Following is the additional information that you requested.

- 1) The maximum possession limit for cesium-137 and americium-241 should be sufficient for us to have 12 portable gauges. (Troxler 3400 series)
- 2) We will implement and maintain the operating, emergency and security procedures in Errata: Appendix H of NUREG-1556, Vol. Rev. 1, Dated November 2001, and provide copies of these procedures to all gauge users and at each job site.

Sincerely,
RIMROCK ENGINEERING, INC.

Paul R. Reynolds
Principal / RSO

Phone: 406-294-8400
Fax: 406-294-8405
Cell: 406-425-4104
Email: preynolds@rimrock.biz

PUBLIC

- Immediate Release
 Normal Release

NON-PUBLIC

- A.3 Sensitive-Security Related
 A.7 Sensitive Internal
 Other: _____

Reviewer: Jwo Date: 10/22/13