



UNITED STATES  
**NUCLEAR REGULATORY COMMISSION**  
REGION IV  
1600 EAST LAMAR BOULEVARD  
ARLINGTON, TEXAS 76011-4511

October 28, 2013

Rink International, LLC  
ATTN: Karl K. Rink, Ph.D.  
Radiation Safety Officer  
1277 Gold Hill Road  
Princeton, Idaho 83857

**SUBJECT: REQUEST FOR ADDITIONAL INFORMATION**

The Nuclear Regulatory Commission (NRC) has initiated the technical review of Rink International, LLC renewal application of NRC License 11-27756-01 dated September 20, 2013, and additional information is needed to continue the review process. Please provide the following information within 30 days of receipt of this letter. Make reference to mail control number 582160 when providing your response.

1. The renewal application did not contain decommissioning financial assurance documents. Provide decommissioning cost estimate and decommissioning funding plan for the amount of material authorized in license 11-27756-01 for NRC review and approval. Guidance on financial assurance is provided in NUREG-1757, Volume 3, Revision 1, "Financial Assurance, Recordkeeping, and Timeliness." This document can be found at: <http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1757/>.
2. The renewal application did not contain the commitment regarding radiation surveys. Provide the following commitment: "We will survey our facility and maintain contamination levels in accordance with the survey frequencies and contamination levels published in Appendix Q to NUREG-1556, Volume 7, 'Program-Specific Guidance About Academic, Research and Development, and Other Licenses of Limited Scope,' dated December 1999. Leak tests will be performed at the intervals approved by NRC or an Agreement State and specified in the SSD Registration Certificate. Leak tests will be performed by an organization authorized by NRC or an Agreement State to provide leak testing services to other licensees or using a leak test kits to other licensees and according to the sealed source or plated foil manufacturer's (distributor's) and kit supplier's instructions."
3. The renewal application did not contain the commitment regarding waste management. Provide the following commitment: "We will use the model waste procedures published in Appendix T to NUREG-1556, Volume 7, 'Program-Specific Guidance About Academic, Research and Development, and Other Licenses of Limited Scope,' dated December 1999."

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Thank you for your cooperation.

Sincerely,

**/RA/**

Roberto J. Torres, Senior Health Physicist  
Nuclear Materials Safety Branch B

Docket: 030-36268  
License: 11-27756-01  
Control: 582160