

TEST REPORT

REPORT NO. 54498-1
OUR JOB NO. ND 54498
YOUR P. O. NO. 7651
CONTRACT _____

WYLE LABORATORIES / Norco, California. 737-0871, 689-2104. TWX 910-332-1204. Cable WYLAB

JELCO, INC.
P. O. Box 2248
Pomona, California 91766

66 - Page Report

DATE 29 June 1976

Revision A
30 September 1976

SEISMIC TESTING

ON

CONTROL PANEL, SHIPPING SECTION NUMBER 3

FOR

JELCO, INC.

SUPERCEDES S#23-502-5-167

S#23-502-5-501-0 SCE#0396

STATE OF CALIFORNIA }
(COUNTY OF RIVERSIDE) ss.

Ray C. Myrick

, being duly sworn,
deposes and says: That the information contained in this report is the result of
complete and carefully conducted tests and is to the best of his knowledge true
and correct in all respects.

Ray C. Myrick

Subscribed and sworn to before me this 30th day of June, 1976


Notary Public in and for the County of Riverside, State of California



My Commission OFFICIAL SEAL JUNE 1979
CATHERINE C. KELTY
NOTARY PUBLIC - CALIFORNIA
RIVERSIDE COUNTY
My comm. expires JULY 14, 1979

135
W-867A

8103060 575



DEPARTMENT DYNAMICS

DEPT. MGR. *James J. Anderson*
James J. Anderson

TEST ENGINEER *Wayne K. Franz*
Wayne K. Franz

Registered
Professional
Engineer *George D. Shipway*
George D. Shipway

DCAS-QAR VERIFICATION

QUALITY CONTROL *Al Heeseman*

Revision A

REVISION SHEET

Revision Number	Date	Pages Affected	Par. No.	Description of Change
A	30 Sept. 1976	7	4.4.2.3	Furnishes supplemental information.
		12	Fig. 2	Corrects axes definitions

A

APPROVED BY:


W. Franz
Test Engineer
H. Heeseman
Quality Control

WYLE LABORATORIES Norco, California

1.0 REFERENCES

- 1.1 Jelco, Inc. Purchase Order No. 7651, dated 15 March 1976.
- 1.2 Bechtel Power Corporation Specification Number S023-502-5, Appendix 4F.
- 1.3 Bechtel Drawing Number 53018-C, entitled "Control Panel Layout Chemical and Volume Control, Reactor Coolant and Reactivity Systems Shipping Section 3".
- 1.4 Bechtel Drawing Number 53018-2, entitled "Control Panel Layout Chemical Control Shipping Section 3".
- 1.5 Wyle Laboratories Test Procedure No. 3570, Revision B.

2.0 GENERAL

Although Reference 1.1 above is applicable to the testing of two control panel specimens, Shipping Sections Number 7 and Number 3, only testing conducted on the latter is discussed in this report. Testing performed on Shipping Section Number 7 was described in an earlier report, Wyle Laboratories Test Report No. 54498, dated 31 March 1976.

3.0 PROCEDURES**3.1** Receiving Inspection

Prior to testing, the specimen, Shipping Section Number 3, was subjected to a visual examination for evidence of shipping damage. Specimen identification information was recorded on a receiving inspection data sheet included in the body of this report.

3.2 Test Fixture and Specimen Orientations

The specimen was fully supported on the test table by a rigid weldment of 12-inch structural steel I-beams. On this rigid base was welded a framework of four-inch square structural steel tubing. The specimen was placed on the tubing and

3.2

(continued)

welded in place employing a specified weld pattern of two-inch long welds on eight-inch centers across the entire front and rear edges of the console. The bottom angle on each open end was left unsupported underneath the spans for the first biaxial test plane (X-Y axes). For the second test plane (Z-Y axes) the end structure was supported underneath with four-inch square tubing. The open ends were not welded down, however. Additionally, in the Z-Y plane, the front and rear tubing lengths were braced (perpendicular to the longitudinal axis of the tubes) to prevent them from behaving as springs.

With the specimen in its normal upright position, its longitudinal axis was initially aligned parallel to the horizontal test machine driver axis. For the second test orientation the specimen was rotated ninety degrees about its vertical centerline such that its lateral axis was aligned with the horizontal driver. The specimen remained in its normal upright position throughout testing. Axis definitions are presented in Figures 1 and 2. The actual setups are shown in the attached photographs.

3.3

Instrumentation

3.3.1

Accelerometers

Twenty accelerometers were attached to the specimen near the mounting points for selected instruments in the panel assembly. The orientations were changed to suit each individual test run. The locations and orientations of each are shown in Figures 1 and 2 and Table I. These accelerometer data were recorded on a galvanometer recorder system for each test run.

3.3.2

Strain Gages

Eight strain gages were mounted at selected points on the specimen. Gages arbitrarily numbered 1 through 4 were mounted on the panel face at locations shown in the attached photographs. Gages 5 through 8 were mounted vertically one and one-eighth inches above the floor line on the rear structural members. Number 5 was located on the front face of the rearmost vertical strut located between the two center doors. Number 6 was on the left face of the same strut. Number 7 was on the rearmost outer

3.3.2 (continued)

surface of the leftmost structural angle. Number 8 was on the front face of the rearmost and leftmost vertical strut. These latter four gages were not photographed.

Strain gage data were recorded employing a signal conditioning/galvanometer recorder system.

3.4 Functional Testing

No electrical functional tests were conducted. The specimen was simply assembled with dummy instruments fabricated by Wyle Laboratories. For the middle left section of the control console, the dummies, composed of wood, masonite and steel, were designed to simulate the weight, center of gravity, and mounting method for each instrument at its proper location. For the remaining sections of the console face, no attempt was made to simulate the center of gravity or the standard instrument mounting method, rather only the total instrument weight for each general panel location was simulated. The dummy instruments are shown in the attached photographs.

3.5 Seismic Testing

3.5.1 Resonance Search

The specimen was subjected to sinusoidal sweep testing in the frequency range from 1 to 35 to 1 Hz. A logarithmic frequency sweep rate of one-half octave per minute was employed at an input level of 0.2g peak.

This type test was performed uniaxially, in the three principal axes, one at a time.

3.5.2 Random and Superimposed Sine Beat

Following iterative "bare table" motion calibrations the specimen was subjected to biaxially applied random motions with biaxial sine beat motions superimposed at specific frequencies.

The biaxial random motions were amplitude controlled with a series of adjustable attenuation one-third octave bandwidth filters whose center frequencies were tuned to frequencies in one-third octave



3.5.2 (continued)

increments from 1.25 to 35 Hz. Ten oscillation-per-beat sine beats were superimposed on the random excitation at frequencies of 1.6, 2.0, and 2.5 Hz. Twenty oscillations-per-beat sine beats were employed at 1.25 Hz.

One, three, four, and five beats per frequency were used for the 1.25, 1.6, 2.0, and the 2.5 Hz test conditions, respectively, with a two-second interbeat delay in each case.

Each test run consisted of thirty seconds of random excitation with the aforementioned appropriate sine beat excitations superimposed. A separate test run was made for each of two sine beat phasing conditions: i.e., the horizontal and vertical test machine drivers in phase and the two drivers 180° out of phase. The horizontal/vertical random waveform excitations were phase incoherent throughout the testing sequence.

The test response spectra were determined with the use of a shock spectra analyzer, tuned in one-third octave frequency increments from 1.25 to 100 Hz. The data were formatted in plots of peak acceleration versus the incremental frequency.

3.5.3 Test Sequence

The detailed sequence followed in the conduction of the test is given below.

- 3.5.3.1 Calibrated the biaxial seismic input motion so that an analysis of the random signal and the four sine beats enveloped the required response spectra.
- 3.5.3.2 Installed the specimen into the test setup as previously described.
- 3.5.3.3 Installed the instrumentation which is called out in Paragraph 3.3 and verified that it was being recorded on an oscilloscope.
- 3.5.3.4 Conducted a sine sweep resonance search in the longitudinal axis as detailed in Paragraph 3.5.1.
- 3.5.3.5 Conducted a sine sweep resonance search in the vertical axis.

- 3.5.3.6 Input the 30 seconds of biaxial seismic motion as detailed in Paragraph 3.5.2, with the 1.25 Hz sine beat superimposed; first with horizontal and vertical drivers in phase and then repeated the test with the drivers out of phase.
- 3.5.3.7 Repeated Paragraph 3.5.3.6 only input the sine beats at 1.6 Hz.
- 3.5.3.8 Repeated Paragraph 3.5.5.6 only input the sine beats at 2.0 Hz.
- 3.5.3.9 Repeated Paragraph 3.5.3.6 only input the sine beats at 2.5 Hz. Reoriented the specimen so that its lateral axis was parallel to the horizontal axis of excitation. Reoriented the appropriate accelerometers to coincide with the horizontal excitation axis.
- 3.5.3.10 Conducted a sine sweep as detailed in Paragraph 3.5.1 in the horizontal axis.
- 3.5.3.11 Repeated Paragraphs 3.5.3.6 through 3.5.3.9.

4.0 RESULTS

4.1 Receiving Inspection

Inspection of the specimen revealed no visible damage due to shipping. Receiving inspection data and specimen identification are shown on a following data sheet.

4.2 Test Fixture

No visible evidence of fixture or mounting method anomalies occurred.

4.3 Functional Tests

No visible anomalies occurred in the dummy instruments or in their mounting methods.

4.4 Seismic Tests

4.4.1 Resonance Searches

Resonance behavior, defined as an output/input acceleration ratio of at least 2.5 to 1, was evident in the lateral axis test data only. These frequency and response results are shown in Table II for each accelerometer displaying a resonance or resonances in its output.



4.4.2 Random with Sine Beats

4.4.2.1 Test Response Spectra (TRS)

The required response spectra (RRS) were enveloped by the TRS, for each sine beat condition, as shown in the attached plots.

4.4.2.2 Instrument Location Accelerations

The maximum instrument accelerations, as determined from galvanometer recordings of response accelerometer data, have been tabulated for each accelerometer. Table III is such a tabulation for the 2.5 Hz (out of phase) sine beat seismic test condition in the Z-Y biaxial test plane.

The data represent peak response accelerations for the peak table accelerations given. Since no response data have exceeded 2.5g, the requirement for less than 3.0g peak response is met, particularly when consideration is made that the inputs were significantly higher than those from the required spectra.

Only the 2.5 Hz sine beat seismic condition need be tabulated since it represents the worst case output/input amplification; i.e., it is the sine beat frequency closest to the first cabinet resonance frequency of 7.8 Hz. This rationale is valid since, from the galvanometer recordings, it is evident the peak input acceleration is derived from the sine beat input, not the random background excitation.

4.4.2.3 Strain Gages

No significant strains were measured throughout testing. The maximum strain recorded was on the order of 400 microinches per inch on gage Number 7 for the X-Y axes plane only. No other gages showed any measurable strain throughout testing. For the case of simple uniaxial strain in mild steel, 400 microinches per inch strain corresponds to 12,000 psi stress; far from its yield stress of approximately 45,000 psi (C1015, hot rolled 1-inch round).



WYLE LABORATORIES Norco, California

54498-1

REPORT NO.

8

PAGE NO.

TABLE I
ACCELEROMETER LOCATIONS

(See the attached photographs and Figures 1 and 2
for the locations)

Accelerometer Number	Orientations for Each Test					
	Resonance Search			Seismic Test		
	X	Y	Z	X-Y	Z-Y	
3	X	Y	Z	Y		Y
4	X	Y	Z	Y		Y
5	X	Y	Z	Y		Y
6	X	Y	Z	Y		Y
7	X	Y	Z	X		Z
8	X	Y	Z	X		Z
9	X	Y	Z	X		Z
10	X	Y	Z	X		Z
11	X	Y	Z	X		Z
12	X	Y	Z	X		Z
13	X	Y	Z	X		Z
14	X	Y	Z	X		Z
15	X	Y	Z	X		Z
16	X	Y	Z	X		Z
17	Y	Y	Y	Y		Y
18	X	X	Z	X		Z
19	Y	Y	Y	Y		Y
20	X	X	Z	X		Z
21	X	Y	Z	Y*		Y
22	X	Y	Z	Y*		Y

* Were in the X direction for both 1.25 Hz sine beat seismic test runs.



WYLE LABORATORIES Norco, California

REPORT NO.

54498-1

9

PAGE NO.

TABLE II
LATERAL AXIS RESONANCE SEARCH DATA (Z Axis)

<u>Direction</u>	<u>Accelerometer*</u>	<u>Frequency (Hz)</u>	<u>Maximum Response Acceleration (g peak)</u>
Sweep Down	8	7.8	1.8
	10	7.8	1.6
	10	9.0	1.2
	8	10.5	1.0
	10	22.5	0.9
	10	35.0	1.3
	8	35.0	0.7
Sweep Up	11	10.0	1.0
	10	10.0	1.2
	9	10.0	0.7
	8	10.0	0.9
	7	10.0	0.7
	6 (Vertical)	10.0	0.6
	5 (Vertical)	10.0	0.6
	4 (Vertical)	10.0	0.6
	3 (Vertical)	10.0	0.6
Sweep Down	20	9.0	0.9
	18	9.0	1.0
	16	9.0	0.9
	15	9.0	1.2
	14	9.0	1.0
	13	9.0	1.0
	20	10.5	1.0
	18	10.5	1.0
	16	10.5	0.9
	15	10.5	1.0
	14	10.5	0.8
	13	10.5	0.8
Sweep Up	22 (Vertical)	35.0	0.6
	21 (Vertical)	35.0	0.6
	20	20.0	0.5
	16	20.0	0.5
	22	10.0	0.8
	21	10.0	0.8
	20	10.0	1.5
	18	10.0	1.3
	16	10.0	1.2
	15	10.0	1.5
	14	10.0	1.2
	13	10.0	1.2
	12	10.0	0.7

* All mounted in the lateral direction except where noted.

144

TABLE III
PEAK SEISMIC RESPONSE

<u>Accelerometer</u>	<u>Acceleration Peak Response* (g)</u>
3 (Vertical)	1.9
4 (Vertical)	1.9
5 (Vertical)	1.9
6 (Vertical)	1.9
7	2.1
8	1.9
9	2.0
10	2.1
11	2.0
12	1.4
13	1.9
14	1.9
15	1.9
16	1.9
17 (Vertical)	1.4
18	2.1
19 (Vertical)	1.4
20	2.2
21 (Vertical)	1.9
22 (Vertical)	2.5

* At the 2.5 Hz sine beat condition (in phase) in the Z-Y axes plane. Lateral input was 1.7g peak; vertical input was 1.4g peak. All accelerometers were in the lateral direction except where noted.

DATA SHEET

Report No. 54498-1

Page No. 11

CUSTOMER

JELCO

Test Title:

SEISMIC RANDOM AND SINE BEAT

Specimen CONTROL CONSOLE


Job No. 54498

S/N

Date 6/18/76

Part No. 2CR-58, 50, 51 (#3)

FIGURE 1
ACCELEROMETER LOCATIONS AND
AXES DEFINITIONS



DATA SHEET

Report No. 54498-I

12

CUSTOMER

Test Title:

TELCO

SEISMIC RANDOM & SINE TEST

Revision

Specimen CONTROL CONSOLE

Job No. 54498

S/N

Date 10-18-76

Part No. 2CR-58,50,51 (#3)

FIGURE 2
ACCELEROMETER LOCATIONS
AND AXES DEFINITIONS

#19Y



#20X,Z

#21

#22



X



DATA SHEET

Customer JELCO Job No. 54498
Date 6-15-76

Specimen CONTROL CONSOLE

RECEIVING INSPECTION

No. of Specimens Received: (1) ONE

Record identification information exactly as it appears on the tag or specimen:

Manufacturer JELCO

Part Numbers ACR-58,50,51 (#3)

How does identification information appear: (name plate, tag, painted, imprinted, etc.)

BLUE PRINTS

Serial Numbers: _____

Examination: Visual, for evidence of damage, poor workmanship, or other defects, and completeness of identification.

Inspection Results: There was no visible evidence of damage to the specimens unless noted below.

NONE

- * If additional space is required for serial numbers, use an additional page, or reference first functional test data sheet (if applicable).

148

Inspected By St. Knoll
Sheet No. _____
Approved H. Brandy Date: 6/15/76

VIBRATION TEST DATA SHEET
RESONANCE SEARCH

Sheet _____

ID No. _____

b7I

Date	Time	Axis	Temp (°F)	SINUSOIDAL			Test Time (Min.)	Comments	Name
				Freq. (HZ)	Disp. ("DA)	Accel. (± G)			
76	NOTED	X-Y-Z	AMB	1-35-1	—	0.2	*	* ONE CYCLE 1-35-1 HZ. AT A SWEEP RATE OF APPROX. ONE HALF OCTAVE PER MINUTE.	
-16	1715	Y	AMB	1-4	—	0.2	4	START SWEEP	
	1719						4	SHUTDOWN SWITCH TO SERVO CONTROL.	
-16	1720	Y	AMB	4-35-4	—	0.2	12	RESUME SWEEP	
	1732						12	SHUTDOWN SWITCH TO MANUAL	
-16	1733	Y	AMB	4-1	—	0.2	4	RESUME SWEEP	
	1737						4	END OF SWEEP.	
.									
16	1740	X	AMB	1-4	—	0.2	4	START SWEEP.	
	1744						4	SHUTDOWN SWITCH TO SERVO CONTROL.	
-16	1745	X	AMB	4-35-4	—	0.2	12	RESUME SWEEP.	
	1757						12	SHUTDOWN SWITCH TO MANUAL.	
-16	1758	X	AMB	4-1	—	0.2	4	RESUME SWEEP.	
	1802						4	END OF SWEEP.	
.									
-18	1520	Z	AMB	1-4	—	0.2	4	START SWEEP.	
	1524						4	SHUTDOWN SWITCH TO SERVO CONTROL.	
-18	1525	Z	AMB	4-35-4	—	0.2	12	RESUME SWEEP.	
	1537						12	SHUTDOWN SWEEP. SWITCH TO MANUAL.	
-18	1538	Z	AMB	4-1	—		4	RESUME SWEEP	
	1542						4	END OF SWEEP.	

DATA SHEET

**CUSTOMER
Type:**

TELCO

SEISMIC RANDOM WITH SUPERIMPOSED SINE BEAT

Specimen CONTROL CONSOLIDATE

Job No. 5449

34

Date 6-18-16

wf

WYLE LABORATORIES

Customer JELCO Job No. 54498

Channel Identification: T/R 1 Trk. No. 1

Transducer S/N 1171 Control X.

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen CONTROL PANEL

Operator Kellogg


Date 6-16-76 Polarity + 0.5%

P/N 2CR-58,50,51 (43)

Axis of Test X-Y

1.25 Hz. id

HORIZONTAL RESPONSE SPECTRA



WYLE LABORATORIES

Customer JELCO Job No. 54498

Channel Identification: T/R 1 Trk. No. 1 Accel. No. 1

Transducer S/N 1171 Control (X), Response ()

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen CONTROL PANEL

Operator Kao "


P/N 2CR-58,50,51 (#3)

Date 6-16-76 Polarity + 0.5%

Axis of Test X-Y

1.25 Hz OUT

HORIZONTAL RESPONSE SPECTRA



WYLE LABORATORIES

Report No. 54498-1

Customer JELCO Job No. 54498

Page No. 19

Channel Identification: T/R 1 Trk. No. 2 Accel. No. 2

Transducer S/N 1034 Control (X), Response ()

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen CONTROL PANEL

Operator R.W. II


P/N 2CR-58, 50, 51 (4)

Date 6-16-76 Polarity + 0.5%

Axis of Test X-Y

1.25Hz. OUT Ø

VERTICAL RESPONSE SPECTRA



WYLE LABORATORIES

Customer JELCO

Job No. 54498

Report No. 54498-1

Page No. 20

Accel. No. 1

Channel Identification: T/R 1 Trk. No. 1

Transducer S/N 1171 Control (X).

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen CONTROL PANEL

Operator KD 011


Date 6-16-76 Polarity + 0.5%

P/N 2CR-58,50,51 (43)

Axis of Test X-Y

1.642. OUT Ø

HORIZONTAL RESPONSE SPECTRA



WYLE LABORATORIES

Customer JELCO Job No. 54498

Report No. 54498-1

Page No. 21

Channel Identification: T/R 1 Trk. No. 2

Accel. No. 2

Transducer S/N 103X Control (X),

Response ()

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen CONTROL PANEL

Operator Knoff


P/N ZCR-58, 50, 51

Date 6-16-76 Polarity + 0.5%

Axis of Test X-Y

1.6 Hz OUT Ø

VERTICAL RESPONSE SPECTRA



WYLE LABORATORIES

Customer JELCO

Job No. 54498

Report No. 54498-1

Page No. 22

Accel. No. 1

Response ()

Channel Identification: T/R 1 Trk. No. 1

Transducer S/N 1171 Control (X).

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen CONTROL PANEL

P/N 2CR-58,50,51 (#3)


Operator KMO/1

Axis of Test X-Y

Date 6-16-76 Polarity + 0.5%

1.6 Hz. IN Ø

HORIZONTAL RESPONSE SPECTRA



WYLE LABORATORIES

Customer JELCO

Job No. 54498

Report No. 54498-1

Page No. 23

Channel Identification: T/R 1 Trk. No. 2

Accel. No. 2

Transducer S/N 1034 Control (X)

Response ()

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen

CONTROL PANEL

Operator KANO 11


P/N 2CR-5E, 50.51 (6)

Date 6-16-76 Polarity + 0 5%

Axis of Test X-Y

1.6 Hz. in G

VERTICAL RESPONSE SPECTRA



Form Approval #

WYLE LABORATORIES

Customer JELCO

Job No. 54498

Report No. 54498-1

24

Page No.

Accel. No. 1

Response 1

Channel Identification: T/R 1 Trk. No. 1

Transducer S/N 1171 Control (X)

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen CONTROL PANEL

Operator Kno 11


P/N 2CR-58,50,51 (*3)

Date 6-16-76 Polarity + 0.5%

Axis of Test X-Y

2.0 FR. IN

HORIZONTAL RESPONSE SPECTRA



Form Approval #

WYLE LABORATORIES

Report No. 54498-1

Customer JELCO

Job No. 54498

Page No. 25

Channel Identification: T/R 1 Trk. No. 2

Accel. No. 2

Transducer S/N 1074 Control (X)

Response ()

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen CONTROL PANEL

Operator KNO 4


P/N ZCR-58, 50, 51

Date 6-16-76 Polarity + 0.5%

Axis of Test X-Y

2.0 Hz. in Ø

VERTICAL RESPONSE SPECTRA



Form Approval #

WYLE LABORATORIES

Customer JELCO Job No. 54498

Report No. 54498-1

Page No. 26

Channel Identification: T/R 1 Trk. No. 1 Accel. No. 1

Transducer S/N 1171 Control (X), Response ()


Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY Specimen CONTROL PANEL

Operator MNO 11 P/N 2CR-53 50.51 (#3)

Date 6-16-76 Polarity + 0.5% Axis of Test X-Y

HORIZONTAL RESPONSE SPECTRA



WYLE LABORATORIES

Customer JELCOJob No. 54498Report No. 54498-1Page No. 27Accel. No. 2

Response ()

Transducer S/N 1034

Control (X),

Full Scale 100

G Cal Voltage

500

MVPK/

1.0


G

Mode PRIMARY

Specimen

CONTROL PANELOperator KMO 11P/N 2CR-53, 50, 51Date 6-16-76Polarity + 0.5%

Axis of Test

X-Y2.0 Hz. OUT ØVERTICAL RESPONSE SPECTRA

Form Approval #

WYLE LABORATORIES

54498-1

Report No.

28

Page No.

Accel. No.

Response: ()

Customer JELCO Job No. 54498

Job No. 54498

Channel Identification: T/R 1 Trk. No. 1

Transducer S/N 4471 Control),

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen . CONTROL PANEL

Operator Kwoff

P/N 2CR-58,50,51 (*3)

Date 6-16-76 Polarity + Q 5%


Axis of Test $x - y$

408

25 43-

HORIZONTAL RESPONSE SPECTRA

2.5 43- OUT ♂



WYLE LABORATORIES

Customer JELCO

Job No. 54498

Report No. 24470-1

Page No. 29 3

Channel Identification: T/R 1 Trk. No. 2

Accel. No. 2

Transducer S/N 1034 Control (X)

Response ()

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen CONTROL PANEL

Operator KNO!!


P/N ZCR-58,50,51

Date 6-16-76 Polarity + 0.5%

Axis of Test X-Y

2.5 Hz. OUT Ø

VERTICAL RESPONSE SPECTRA



Form Approval *Bl*

WYLE LABORATORIES

Customer TELCO Job No. 54498

Channel Identification: T/R 1 Trk. No. 1

Transducer S/N 1171 Control

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen CONTROL PANEL

Operator Karo 11


Date 6-16-76 Polarity + 0.5%

P/N 2CR-58,50,51 (43)

Axis of Test X-Y

25Hz. in Ø.

HORIZONTAL RESPONSE SPECTRA



WYLE LABORATORIES

Report No. 54498-1

Customer JELCO

Job No. 54498

Page No. 31

Channel Identification: T/R 1 Trk. No. 2

Accel. No. 2

Transducer S/N 1034 Control (X),

Response ()

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen

CONTROL PANEL

Operator KWD/1


P/N 2CR-58, 50, 51 (

Date 6-16-76 Polarity + 0.5%

Axis of Test X-Y

2.5Hz. in Ø

VERTICAL RESPONSE SPECTRA



Form Approval #

0.1

WYLE LABORATORIES

Customer JELCO Job No. 54498

Report No. 54498-1

Channel Identification: T/R 1 Trk. No. 1

Page No. 32

Transducer S/N 1171 Control (X)

Accel. No. 1

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Specimen CONTROL PANEL

Karen Mode PRIMARY

P/N 2CR-58,50,51 (#3)


Operator 6-18-76

Axis of Test Z-Y

Date 6-18-76 Polarity + 0.5%

1.25 Hz. 1H 0

HORIZONTAL RESPONSE SPECTRA



WYLE LABORATORIES

Customer JELCO Job No. 54498

Channel Identification: T/R 1 Trk. No. 2

Transducer S/N 1024 Control (X).

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen

CONTROL PANEL

Operator KM 11


Date 6-18-76 Polarity + 0.5%

P/N 2CR-58,50,51

Axis of Test Z-Y

1.25Hz IN Ø

VERTICAL RESPONSE SPECTRA



Form Approval #

Report No. 54498-1

Page No. 33

Accel. No. 2

Response ()

G

WYLE LABORATORIES


Report No. 54498-1

Customer JELCOJob No. 54498Page No. 34Channel Identification: T/R 1 Trk. No. 1Accel. No. 1Transducer S/N 1171 Control (C),

Response ()

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 GMode PRIMARYSpecimen CONTROL PANELOperator Kao 11Date 6-18-76 Polarity + 0.5%P/N 2CR-58,50,51 (#3)Axis of Test Z-Y1.25Hz. CNT Ø

HORIZONTAL RESPONSE SPECTRA



WYLE LABORATORIES

Report No. 54498-1

Customer JELCO Job No. 54498

Page No. 35

Channel Identification: T/R 1 Trk. No. 2

Accel. No. 2

Transducer S/N 1034 Control (X),

Response ()

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen CONTROL PANEL

Operator Kno 11


P/N ZCR-58, 50, 51

Date 6-18-74 Polarity + 0.5%

Axis of Test Z-Y

VERTICAL RESPONSE SPECTRA

1.25 Hz OUT Ø



WYLE LABORATORIES

Report No. 54498-1

Customer TELCO

Job No. 54498

Page No. 36

Channel Identification: T/R 1 Trk. No. 1

Accel. No. 1

Transducer S/N 1171 Control (X),

Response ()

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen

CONTROL PANEL

Operator KWD 11


P/N 2CR-58,50,51 (#3)

Date 6-18-76 Polarity + 0.5%

Axis of Test Z-Y

1.6 Hz OUT Ø

HORIZONTAL RESPONSE SPECTRA



WYLE LABORATORIES

Customer JELCO

Job No. 54498

Report No. 54498-1

Page No. 37

Channel Identification: T/R 1 Trk. No. 2 Accel. No. 2

Transducer S/N 1034 Control (X), Response ()

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen CONTROL PANEL

Operator Kalo 11


P/N ZCR-58, 50, 51

Date 6-18-76 Polarity + 0.5%


Axis of Test Z-Y

1.6 Hz. OUT

VERTICAL RESPONSE SPECTRA



WYLE LABORATORIES

Customer JELCOJob No. 54498Report No. 54498-1Page No. 38Accel. No. 1Channel Identification: T/R 1 Trk. No. 1Transducer S/N 1171 Control Response Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 GMode PRIMARYSpecimen CONTROL PANELOperator Kvo 11P/N 2CR-58,50,51 (*3)Date 6-18-76 Polarity + 0.5%Axis of Test Z-Y1.642 IN Ø**HORIZONTAL RESPONSE SPECTRA**Form Approval No. 201

WYLE LABORATORIES

Customer JELCO Job No. 54498

Report No. 54498-1

Page No. 39

Channel Identification: T/R 1 Trk. No. 2

Accel. No. 2

Transducer S/N _____ Control (X)

Response ()

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen

CONTROL PANEL

Operator _____


P/N 2CR-58, 50, 51 (2)

Date _____ Polarity + 0 5%

Axis of Test Z-Y

VERTICAL RESPONSE SPECTRA

1.6HR. IN Ø



Form Approval Best

1.0

0.1

174

10

FREQUENCY Hz

100

200

1000


WYLE LABORATORIES

Customer JELCO Job No. 54498Report No. 54498-1Page No. 40Accel. No. 1

Response ()

Channel Identification: T/R 1 Trk. No. 1Transducer S/N 1C71 Control C(1)Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 GMode PRIMARYSpecimen CONTROL PANELOperator Kao HDate 6-18-71 Polarity + 0.5%P/N 2CR-53,50,51 (43)Axis of Test Z-Y2.0 Hz. IMP

HORIZONTAL RESPONSE SPECTRA



In Approval Pending

0.1

100

10

FREQUENCY Hz

100

200

WYLE LABORATORIES

Customer JELCO Job No. 54498

Channel Identification: T/R 1 Trk. No. 2

Transducer S/N 1034 Control (X),

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen

Operator KG 1011


P/N 265-71 S/N 51

Date 6-18-76 Polarity + 0 5%

Axis of Test Z

2.0 Hz

VERTICAL RESPONSE SPECTRA



WYLE LABORATORIES

Customer JELCO Job No. 54498

Report No. 54498-1

Page No. 42

Accel. No. 1

Response ()

Channel Identification: T/R _____ Trk. No. _____

Transducer S/N 1171 Control

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Mode PRIMARY

Specimen

Operator Kno!

Date 6-18-76 Polarity + Q 5%

P/N 2CR-53,50,51 (43)

Axis of Test


乙

HORIZONTAL RESPONSE SPECTRA

Test - 2.0 P. OUT


Axis of Test Z

TEST - 2042



Customer JELCO Job No. 54498Page No. 43Channel Identification: T/R 1 Trk. No. 2Accel. No. 2Transducer S/N 1034 Control (X), Response ()Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 GMode PRIMARY Specimen CONTROL PANELOperator KM00V P/N 2CR-58, 50, 51Date 6-18-76 Polarity + 0.5% Axis of Test Z.Y2.0 HZ. OUT

VERTICAL RESPONSE SPECTRA



WYLE LABORATORIES

Customer TELCO

Job No. 54493

54498-1

Report No. .

44

Accel. No.

Response ()

Response ()

Transduser S/N 1171 Control (C)

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 G

Operator "Kao"

Date 6-18-76 Polarity +/- q 5%


Specimen CONTROL PANEL

P/N 2CR-50,50,51 (-3)

Jan 2-4

Axis of Test $\pi = \frac{1}{2}$

HORIZONTAL RESPONSE SPECTRA




WYLE LABORATORIES

Customer JELCOJob No. 54498Report No. 54498-1Page No. 45Channel Identification: T/R 1 Trk. No. 2Accel. No. 2Transducer S/N 10745 Control (X)

Response ()

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 GMode PRIMARYSpecimen CONTROL PANELOperator KNOXP/N ZCR-58,50,51Date 6-18-76 Polarity \pm 0.5%Axis of Test Z-Y2.5 Hz. OUT

VERTICAL RESPONSE SPECTRA



WYLE LABORATORIES


Report No. 54498-1

Customer TELCO Job No. 54498

Page No. 46

Channel Identification: T/R 1 Trk. No. 1Accel. No. 1Transducer S/N 1171 Control C.

Response ()

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0 GMode PRIMARYSpecimen CONTROL PANELOperator KNO 11P/N 2CR-53,50,51 (23)Date 6-18-76 Polarity + 0.5%Axis of Test Z-YHORIZONTAL RESPONSE SPECTRA2.5Hz 1A 0

Form Approval #

0.1

WYLE LABORATORIES

Report No. 1

Customer JELCOJob No. 54498

Page No. 1

Channel Identification: T/R 1 Trk. No. 2


Accel. No. 1

Transducer S/N 1034 Control (K),

Response (1)

Full Scale 100 G Cal Voltage 500 MVPK/ 1.0Mode PRIMARYSpecimen SOOperator KWJ 11P/N 20Date 6-18-70 Polarity + 0.5%Axis of Test Z

VERTICAL RESPONSE SPECTRA

2.5 Hz

SPECIMEN

CUSTOMER

PART NO.

S/N

CONTROL CONSOLE

JOB NO.

DATE

TEST BY

WITNESS

5449

6-18-76

F. KNOLL

WYLE LABORATORIES

TEST: SEISMIC FREQUENCY SINE BEAT

EQUIPMENT	MANUFACTURER	MODEL NO.	RANGE	WYLE NO.	CALIBRATION		ACCY.
					LAST	DUE	
EXCITER	TEMA CORP	W 3000	12" DIS 30,000 FORGE LBS 10" DIS	-	-	-	N/A
EXCITER	TEMA CORP	W 1800	17,000 FORGE LBS 10" DIS	-	-	-	N/A
EXCITER	TEMA CORP	W 1800	16,000 FORGE LBS	-	-	-	N/A
SERVO CONTROLLER	MCFAADDEN	152-A	"	"	17-12-76	PRIOR TO USE	N/A
SERVO CONTROLLER	MCFAADDEN	152-A	-	-	PRIOR	PRIOR TO USE	N/A
SERVO CONTROLLER	MCFAADDEN	152-A	"	"	17-12-76	PRIOR TO USE	N/A
AMPLIFIER	MCFAADDEN	152-A	-	-	PRIOR	PRIOR TO USE	N/A
AMPLIFIER	MCFAADDEN	152-A	-	-	PRIOR	PRIOR TO USE	N/A
AMPLIFIER	MCFAADDEN	152-A	-	-	PRIOR	PRIOR TO USE	N/A
SHOCK SPECTRUM ANALYZER	DYNAMICS	13231	120 CHANNEL	7530	SYSTEM	CONFIGURATION	MFG. SPEC.
SPECTRUM SHAPER	KUJNER	123	12.5 TO 40 KHZ	31337	PRIOR	PRIOR TO USE	N/A
SPECTRUM SHAPER	KUJNER	123	12.5 TO 40 KHZ	31570	PRIOR	PRIOR TO USE	N/A
EQUALIZER SHAPER	TRACOR	822	1.25 TO 10 KHZ	31534	PRIOR	PRIOR TO USE	N/A
EQUALIZER SHAPER	TRACOR	822	1.25 TO 10 KHZ	31574	PRIOR	PRIOR TO USE	N/A
X-Y RECORDER	HEWLETT PACKARD	7005B	X = 30" / SEC Y = 20" / SEC	50889	PRIOR	PRIOR TO USE	MFG. SPEC.
OSCILLOSCOPE	HEWLETT PACKARD	122-AK	BNC TRACE	30626	5-17-76	11-21-76	+52.48
ELECTRONIC VOLTMETER	HEWLETT	2411B	0.01 TO 1000 VOLTS	6556	3-29-76	8-1-76	+49.44

SPECIMEN

Control CONSOCE

JOB NO.

54490

CUSTOMER

TELCO

DATE

6-18-76

PART NO.

2CR-58,50,51 (#5)

TEST BY

P. KNIOLL

S/N

WITNESS

WYLE LABORATORIES

TEST: SEISMIC RANDOM & SINE BEAT

EQUIPMENT	MANUFACTURER	MODEL NO.	RANGE	WYLE NO.	CALIBRATION		ACCY.
					LAST	DUUE	
SWEET OSCILLATOR	S.D.	14A-5	.005-50KHZ.	99987	2-19-76	8-22-76	± 2%
SERVO MONITOR	S.D.	105A		31306	PRIOR TO USE		N/A
ACCELEROMETER	UNIFOLC DICKIE	75D21	0-2000HZ	7567	4-9-76	7-9-76	± 2%
ACCELEROMETER	UNIFOLC DICKIE	75D21	0-2000HZ	7571	5-17-76	8-17-76	± 2%
ACCELEROMETER	UNIFOLC DICKIE	75D21	0-2000HZ	7144	6-10-76	9-10-76	± 2%
ACCELEROMETER	UNIFOLC DICKIE	75D21	0-2000HZ	7360	4-14-76	7-14-76	± 2%
ACCELEROMETER	UNIFOLC DICKIE	75D21	0-2000HZ	7320	6-15-76	9-15-76	± 2%
ACCELEROMETER	UNIFOLC DICKIE	75D21	0-1000HZ	7377	6-15-76	9-15-76	± 2%
ACCELEROMETER	UNIFOLC DICKIE	75D21	0-2000HZ	7398	6-15-76	9-15-76	± 2%
ACCELEROMETER	UNIFOLC DICKIE	75D21	0-2000HZ	7566	4-13-76	7-13-76	± 2%
ACCELEROMETER	UNIFOLC DICKIE	75D21	0-2000HZ	7523	6-15-76	9-15-76	± 2%
ACCELEROMETER	UNIFOLC DICKIE	75D21	0-2000HZ	7572	6-15-76	9-15-76	± 2%
ACCELEROMETER	UNIFOLC DICKIE	75D21	0-2000HZ	7300	6-15-76	9-15-76	± 2%
ACCELEROMETER	DICKIE	75D21	0-2000HZ	7557	4-27-76	7-27-76	± 2%
ACCELEROMETER	UNIFOLC DICKIE	75D21	0-2000HZ	7378	3-22-76	6-22-76	± 2%
ACCELEROMETER	UNIFOLC DICKIE	75D21	0-2000HZ	7143	3-22-76	6-22-76	± 2%
ACCELEROMETER	UNIFOLC DICKIE	75D21	0-2000HZ	7532	3-22-76	6-22-76	± 2%

Report No.

54490

SPECIMEN

Control Console

JOB NO.

54498

CUSTOMER

JELCO

DATE

6-18-76

PART NO.

2CR-58,50,51 (#3)

TEST BY

P. KNOLL

S/N

WITNESS

WYLE LABORATORIES

TEST: SEISMIC RANDOM + SINE BEAT

EQUIPMENT	MANUFACTURER	MODEL NO.	RANGE	WYLE NO.	CALIBRATION		ACCY.
					LAST	DUE	
ACCELEROMETER	UNIVOLTS DICKIE	25D21	0-2000Hz	7542	6-13-76	6-22-76	±2%
ACCELEROMETER	UNIVOLTS DICKIE	25D21	0-2000Hz	6225	6-18-76	9-15-76	±2%
ACCELEROMETER	Columbia	302-6	0-2000Hz	62018	5-4-76	8-4-76	±2%
ACCELEROMETER	Columbia	302-6	0-2000Hz	6619	5-4-76	8-4-76	±2%
ACCELEROMETER	ENDEVCO	2215	0-2000Hz	7540	6-15-76	9-15-76	±2%
ACCELEROMETER	ENDEVCO	2211	0-2000Hz	10794	6-15-76	9-15-76	±2%
ACCELEROMETER	ENDEVCO	2213	0-2000Hz	31023	6-10-76	9-10-76	±2%
CHARGE AMP	UNIVOLTS DICKIE	D22	0-1000g.	7341	1-13-76	7-11-76	±2%
CHARGE AMP	UNIVOLTS DICKIE	D22	0-1000g.	7342	1-13-76	7-11-76	±2%
CHARGE AMP	UNIVOLTS DICKIE	D22	0-1000g.	7343	1-13-76	7-11-76	±2%
CHARGE AMP	UNIVOLTS DICKIE	D22	0-1000g.	7344	1-13-76	7-11-76	±2%
CHARGE AMP	UNIVOLTS DICKIE	D22	0-1000g.	7338	1-27-76	7-25-76	±2%
CHARGE AMP	UNIVOLTS DICKIE	D22	0-1000g.	31407	3-16-76	9-19-76	±2%
CHARGE AMP	UNIVOLTS DICKIE	D22	0-1000g.	7335	1-27-76	7-11-76	±2%
CHARGE AMP	UNIVOLTS DICKIE	D22	0-1000g.	7336	1-27-76	7-11-76	±2%
CHARGE AMP	UNIVOLTS DICKIE	D22	0-1000g.	7337	1-27-76	7-11-76	±2%
CHARGE AMP	UNIVOLTS DICKIE	D22	0-1000g.	7340	1-27-76	7-11-76	±2%

SPECIMEN

CONTROL CONSOLE

JOB NO.

54498

CUSTOMER

TELCO

DATE

6-18-76

PART NO.

2CR-58,50,51 (#3)

TEST BY

P. KNOLL

S/N

WITNESS

WYLE LABORATORIES

TEST: SEISMIC RANDOM, SINE BEAT

EQUIPMENT	MANUFACTURER	MODEL NO.	RANGE	WYLE NO.	CALIBRATION		ACCY.
					LAST	DUE	
CHARGE AMP	UNHOLTB DICKIE	D22	0-1000 g.	7339	1-27-76	7-25-76	±2%
CHARGE AMP	UNHOLTB DICKIE	11	0-1000 g.	31488	1-13-76	7-11-76	±2%
CHARGE AMP	UNHOLTB DICKIE	8P	0-2000g	6723	5-12-76	11-14-76	±2%
CHARGE AMP	UNHOLTB DICKIE	8P	0-1000 g.	4541	1-8-76	7-4-76	±2%
CHARGE AMP	UNHOLTB DICKIE	8P	0-1000 g.	30990	5-12-76	11-14-76	±2%
CHARGE AMP	UNHOLTB DICKIE	8P	0-1000 g.	30997	5-12-76	11-14-76	±2%
CHARGE AMP	UNHOLTB DICKIE	8P	0-1000 g.	30770	5-12-76	11-14-76	±2%
CHARGE AMP	UNHOLTB DICKIE	8P	0-1000 g.	30880	5-12-76	11-14-76	±2%
CHARGE AMP	UNHOLTB DICKIE	11	0-1000 g.	31490	1-13-76	7-11-76	±2%
CHARGE AMP	UNHOLTB DICKIE	11	0-1000 g.	31493	3-15-76	9-9-76	±2%
CHARGE AMP	UNHOLTB DICKIE	11	0-1000 g.	31496	1-13-76	7-11-76	±2%
CHARGE AMP.	UNHOLTB DICKIE	11	0-1000 g.	31491	1-13-76	7-11-76	±2%
TAPE RECORDER	SANBORN	3924B	14 CHANNEL	31265	PRIOR TO USE	N/A	
TAPE RECORDER	SANBORN	3924B	14 CHANNEL	31266	PRIOR TO USE	N/A	
SINE BEAT GENERATOR sum or difference AMPLIFIER	McFADDEN	209A	.5-50Hz.	—	SYSTEM CALIB.	N/A	
	McFADDEN	200A4	4 CHANNEL	—	SYSTEM CALIB.	N/A	

Report No.

54498

Report No. 54498-1

Page No. 54


RECORDED

SEARCHED INDEXED SERIALIZED FILED

189

Report No. 54498-1

Page No. 53




PHOTOGRAPH 2

SCIENTIFIC TEST SETUP - X-Y BIAXIAL PLANE

188

Report No. 54498-1


Page No. 54



PHOTOGRAPH 3

SEISMIC TEST SETUP - Z-Y BIAXIAL PLANE

189




PHOTOGRAPH 4

SEISMIC TEST SETUP - Z-Y BIAXIAL PLANE

Report No. 54498-1

Page No. 56




PHOTOGRAPH 5

FULL PANEL VIEW

191

Report No. 54496-1


Page No. 57



PHOTOGRAPH 6

TYPICAL END SUPPORT
(L-Y Biwing Plane Only)


192 :



PHOTOGRAPH 7

ACCELEROMETER LOCATION

193



PHOTOGRAPH 8


ACCELEROMETER LOCATION
AND DURIV INSTRUMENT DETAIL

194

Report No.

5441

Page No.




PHOTOGRAPH 9

NEARBY LOCALITY
AND TYPICAL WILDFIRE LIFTING EYE

195

Report No. 50-11

Page No. 61



PHOTOGRAPH 10

ACCELERATION LOCATION


196



Report No.

54

Page No.




PHOTOGRAPH 12

ACCELEROMETRIC LOCATIONS

163

Report No. 54498-1

Page No. 64



PHOTOGRAPH 13

ACCELEROMETER LOCATIONS

199

Report No. 54498-1

Page No. 68




PHOTOGRAPH 14

ACCELEROMETER LOCATIONS

200

Report No. 54498-1

Page No. 66



PHOTOGRAPH 15

ACCELEROMETER LOCATIONS

201

TEST REPORT

REPORT NO. 54498
 OUR JOB NO. ND 54498
 YOUR P. O. NO. 7651
 CONTRACT —

5 - Page Addendum

DATE 29 June 1976

ADDENDUM I

1.0 REFERENCES

- 1.1 Jelco, Inc. Purchase Order No. 7651, dated 15 March 1976.
 1.2 Wyle Laboratories Test Report No. 7651, dated 31 March 1976.

2.0 PURPOSE

The purpose of this addendum is to incorporate four pages of test data sheets inadvertently omitted from Reference 1.2. These sheets furnish test information for resonance search and some random with sine beat tests on Shipping Section No. 7.

STATE OF CALIFORNIA } ss.
 COUNTY OF RIVERSIDE }

Ray C. Myrick, being duly sworn,
 deposes and says: That the information contained in this report is the result of
 complete and carefully conducted tests and is to the best of his knowledge true
 and correct in all respects.

Ray C. Myrick

SUBSCRIBED and sworn to before me this 29th day of June, 19 76.

Catherine C. Kelty
 Notary Public in and for the County of Riverside, State of California

130

EXPIRES 14 July, 19 79
 OFFICIAL SEAL
 CATHERINE C. KELTY
 NOTARY PUBLIC - CALIFORNIA

DEPARTMENT ELECTRONICS


DEPT. MGR. James J. Anderson

TEST ENGINEER W. K. Fenn

Registered
 Professional
 Engineer G. D. Shippy
George D. Shippy

DCAS-QAR VERIFICATION

M. H. Keeler



— NOTICE —

THE ATTACHED FILES ARE OFFICIAL RECORDS OF THE DIVISION OF DOCUMENT CONTROL. THEY HAVE BEEN CHARGED TO YOU FOR A LIMITED TIME PERIOD AND MUST BE RETURNED TO THE RECORDS FACILITY BRANCH 016. PLEASE DO NOT SEND DOCUMENTS CHARGED OUT THROUGH THE MAIL. REMOVAL OF ANY PAGE(S) FROM DOCUMENT FOR REPRODUCTION MUST BE REFERRED TO FILE PERSONNEL.

DEADLINE RETURN DATE

50-361

Control # 8103060541

DATE - 2/13/81

REGULATORY DOCKET FILE COPY
RECORDS FACILITY BRANCH