

Docket File Information

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Bothwell Regional Health Center 601 East 14th Street Sedalia, MO 65302-1706 REPORT NUMBER(S) 13-01	2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
--	---

3. DOCKET NUMBER(S) 030-10715	4. LICENSE NUMBER(S) 24-16275-01	5. DATE(S) OF INSPECTION November 20, 2013
--------------------------------------	---	---

6. INSPECTION PROCEDURES USED 87131, 87132	7. INSPECTION FOCUS AREAS 03.01-03.07
---	--

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02120	2. PRIORITY 3	3. LICENSEE CONTACT D. H. Roehrs, MD, RSO	4. TELEPHONE NUMBER (660) 827-9530
---------------------------------	----------------------	--	---

Main Office Inspection Next Inspection Date: 11/18/2016

Field Office Inspection _____

Temporary Job Site Inspection _____

PROGRAM SCOPE

The licensee was a medical institution authorized for 10 CFR 35.100, 35.200, 35.300, and 35.400 procedures pertaining to permanent seed implants at the location specified on the license. Since the previous inspection, an additional authorized location of use has been added to the license at 3700 W. 10th Street, however, the location was not inspected because the facility is closed on Wednesdays each week.

The licensee's Nuclear Medicine Department routinely conducts a daily average of 10-15 patient studies with a staff of 3 nuclear medicine technologists (NMTs). Iodine-123 is administered for uptake studies and can average none to two patient cases per month. Iodine-131 procedures requiring a written directive averaged 2-3 administrations per month. Thyroid carcinoma therapies averaged 1-2 each year. Palliative therapy administrations averaged 1-2 each year. The licensee receives unit doses from a local nuclear pharmacy as needed. All waste was held for decay-in-storage.

No low dose brachytherapy procedures using I-125 or palladium-103 seeds for prostate implant therapies have been performed since the previous inspection.

Performance Observations

The licensee's available NMTs and medical physicist demonstrated/discussed: (1) survey meter use and calibrations; (2) package check-in procedures; (3) unit dosage prep and safe use; (4) wipe test counting; (5) waste handling; (6) sealed source inventories and leak tests; (7) routine security of licensed material; (8) dose calibrator tests; (9) quarterly radiation safety program audits; (10) any contamination events (one event involving an incontinent I-131 patient); (11) written directives; (12) radiation safety committee meetings; and (13) dosimetry: for 2011: 211mR-DDE, 1770mR-SDE, and 2012: 251mR-DDE and 1840mR-SDE.

The inspector performed independent and confirmatory radiation measurements which indicated results consistent with licensee survey records and postings.