

U.S. NUCLEAR REGULATORY COMMISSION

Date: December 17, 2013
2:25 pm

TELEPHONE CONVERSATION RECORD

Mail Control No.
582350

License No. 06-00679-01

Docket No. 030-01243

Name of Licensee: Saint Mary's Hospital

Name of Participant(s): Paul DeGennaro, CNMT
Shawn Seeley, Health Physicist, RI DNMS

Telephone No. 203-709-6030

Subject: Clarification questions regarding license amendment dated October 12, 2013
(NOTE: This will be used as the Documents Title in ADAMS)

I called Mr. DeGennaro to ask him a couple of questions regarding the information contained in their amendment request.

1. Will the surgery center have access to a NaI well counter or other appropriately sensitive instrument for measuring wipes? **He replied that they will have access to a NaI well counter or equivalent detection instrument.**
2. Please confirm that the address for the surgery center is 160 Robbins Street and that no other NRC licensee occupies this same space. **He confirmed that he address was in fact 160 Robbins Street and that no other NRC licensed activity occurs in that space.**
3. Please identify the areas located above/below the hot lab. **This is a single story office building. There is nothing above or below the hot lab.**

I thank him for his time and ended the call.

Document Availability:

Publicly Available

Non-Publicly Available

Non-Sensitive
 Sensitive – Privacy Act/ PII
 Immediate Release

Non-Sensitive Copyright
 Sensitive – Internal
Normal Release Date:

Sensitive- Proprietary
 Sensitive – Security-Related
Delayed Release Date:

SUNSI Review Completed By: SSeeley / RA /

Document Accession #: _____