

From: [Weidner, Tara](#)
To: [Sharon Long \(slong@nationalphysics.com\)](mailto:slong@nationalphysics.com)
Subject: Request for additional information for Herbert J. Thomas Memorial Hospital (Renewal)
Date: Tuesday, December 24, 2013 9:16:00 AM

PLEASE CONFIRM RECEIPT OF THIS REQUEST FOR ADDITIONAL INFORMATION BY RETURN EMAIL

License No. 47-17746-01
Docket No. 03013267
Control No. 582513

REQUEST FOR ADDITIONAL INFORMATION CONCERNING THE RENEWAL OF THE HERBERT J. THOMAS MEMORIAL HOSPITAL NRC LICENSE

Dear Ms. Long,

This is in reference to the application dated November 3, 2013 requesting to renew Nuclear Regulatory Commission License Number 47-17746-01. In order to continue our review, we need the following additional information:

1. In order to facilitate future communications, please provide contact information (e-mail, telephone, and fax) for Daniel Lauffer . Please note that this information will be made public unless otherwise specified.
2. On your facility diagram, please note whether areas identified are considered to be restricted or unrestricted areas.
3. During the July 2013 inspection, the NRC inspector identified one violation of 10 CFR 20.2003(a), which involved discharging licensed material into sanitary sewer and not documenting the activity disposed. At that time, you committed to discontinuing this practice and retraining staff on the revised policy. Please confirm that the retraining has occurred and undocumented disposals to the sewer are no longer performed.

We will continue our review upon receipt of this information. Please reply to my attention at the Region I office and refer to Mail Control No. 582513. If you have any questions, feel free to contact me via e-mail or at 610-337-5272.

Thanks in advance,

Tara L. Weidner
Senior Health Physicist
U.S. Nuclear Regulatory Commission