

**From:** [Elliott, Robin](mailto:Elliott,Robin)  
**To:** [thuebner@rrmc.org](mailto:thuebner@rrmc.org)  
**Cc:** [jrademacher@rrmc.org](mailto:jrademacher@rrmc.org); [Paul Waite \(pwaite@rrmc.org\)](mailto:pwaite@rrmc.org)  
**Subject:** Renewal of License No. 44-14121-01  
**Date:** Monday, December 30, 2013 10:56:00 AM

---

Licensee Name: Rutland Regional Medical Center  
License No.: 44-14121-01  
Docket No.: 030-07587  
Mail Control No.: 581472

Dear Mr. Huebner,

As per my phone conversation with Paul Waite on your behalf today, additional information is needed to process the renewal application for Rutland Regional Medical Center.

- Please resubmit your December 5, 2013 communication in its entirety with the cover letter dated December 5, 2013.
- In order to facilitate communications please provide a business fax number.
- Please confirm that you do not possess nor intend to use PET radiopharmaceuticals.
- Provide the manufacturer and model number for any sealed sources that do not meet the criteria in 10 CFR 35.65 (e.g. greater than 30 millicuries).
- Please confirm that you wish to delete J. Christian Higgins, M.D. from your list of Authorized Users.
- Please confirm that by attaching a copy of Amendment 37 to your December 5, 2013 communication that you were requesting to maintain the following: your current authorizations for licensed material and use, your current Radiation Safety Officer, James N. Rademacher, M.D.; and your current authorized locations as listed in Condition 10.
- Please confirm that you will develop, implement and maintain written procedures for safe response to spills of licensed material in accordance with 10 Part 20.1101.
- Further additional information is needed for the current facility diagrams submitted as outlined in NUREG 1556 Volume 9 Revision 2 and can be found in section 8.16: <http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/sr1556v9r2-final.pdf#08-16>

As well as in NUREG Volume 9 Revision 2 Appendix E Figure E.1:  
<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/sr1556v9r2-final.pdf#app-e>

The additional information we are requesting is as follows:

- Provide information related to the security of the hot lab, storage of radiopharmaceuticals, preparation and dispensing of radiopharmaceuticals (e.g., lead glass L-block, etc.).
- Storage of radioactive waste, including decay-in-storage prior to disposal as nonradioactive waste. This area should be large enough to handle an accumulation of Tc-99m generators as well as other solid waste. If this area is not located within your main department, describe how you will secure the material.
- In addition, identify adjacent areas across (as well as above and below) the

walls from use and storage locations and show that adequate steps have been taken to assure that radiation levels in unrestricted areas will not result in doses to individual members of the public in excess of those specified in 10 CFR 20.1301.

- Drawings and diagrams that provide exact locations of materials or depict specific locations of safety or security equipment should be marked as “Security-related information – withhold under 10 CFR 2.390.”

Once we receive the additional information requested, we should be able to finalize the processing of your renewal application. You may respond to my attention in writing by letter or fax (610-337-5269), referencing mail control 581472. In order for us to continue a prompt review of your submittal, please provide a response within 15 calendar days from the date of this e-mail.

Regards,

*Robin L. Elliott*

Health Physicist

U.S. NRC, Region I

2100 Renaissance Blvd, Suite 100

King of Prussia, PA 19406-2713

(610) 337-5076 voice

(610) 337-5269 fax

[Robin.Elliott@nrc.gov](mailto:Robin.Elliott@nrc.gov)