



Region I Office  
 Division of Nuclear Materials Safety  
 2100 Renaissance Boulevard  
 Suite 100  
 King of Prussia, PA 19406-2713  
 (610)337-5000

## Telephone Conversation Record

Date: 1/29/14

License No.: 52-25430-03

Docket No. (no hyphens): 03036911

Mail Control/Report No.: 582862

Licensee Name: CRMI

Participant(s) Name/Title: David Rhoe, CEO

Work Telephone No.: 787-245-7248

Business Cellphone No.: 787-245-7248

NRC Representative Name/Title: T. Weidner, Sr. HP

Subject: RAI concerning I-131 inpatient treatment room

*(This is the title that will be used in ADAMS)*

Discussion: In the letter dated 12/30/13, CRMI requested the addition of an inpatient I-131 treatment room. A floor plan was provided that indicated that the treatment room shared a wall with an adjacent patient room. According to the diagram, the patient beds were arranged head-to-head on the common wall. On 1/29/14, D. Rhoe was contacted to discuss how CRMI was going to assure that the radiation levels in the adjacent patient room did not exceed the limits in 10 CFR 20.1301. According to D. Rhoe, the wall between the patient rooms is solid concrete and based on that he does not believe that the radiation levels in the adjacent room will exceed the limits. However, surveys will be performed to confirm his assumption. D. Rhoe will document this discussion in a letter.

Action Required: Wait for D. Rhoe's letter

<b>SUNSI REVIEW</b>		
<b>Document Availability:</b>	<input checked="" type="checkbox"/> Public	or <input type="checkbox"/> Non-Public
<b>Document Sensitivity:</b> (select "1" value to the right)	<input checked="" type="checkbox"/> Non-Sensitive	<input type="checkbox"/> MD 3.4 Non-Public B.1 (Non-Sensitive)
	<input type="checkbox"/> Non-Sensitive Copyright	<input type="checkbox"/> MD 3.4 Non-Public A.3 (Sensitive Security Related)
	<input type="checkbox"/> Sensitive – Proprietary	<input type="checkbox"/> MD 3.4 Non-Public A.4 (Sensitive Proprietary)
	<input type="checkbox"/> Sensitive – Privacy Act (includes Personally Identifiable Information (PII))	<input type="checkbox"/> MD 3.4 Non-Public A.5 (Sensitive-Privacy Act)
	<input type="checkbox"/> Sensitive – Internal, Periodic Review required (All Other Sensitive Internal Info.)	<input type="checkbox"/> MD 3.4 Non-Public A.6 (Sensitive-Federal, State, etc)
	<input type="checkbox"/> Sensitive – Security-Related-Periodic Review Required	<input type="checkbox"/> MD 3.4 Non-Public A.7 (Internal)
SUNSI Review Completed by: TLWeidner		

: If applicable, add OFFICIAL USE ONLY – SECURITY RELATED INFORMATION banners and change SUNSI review to A.3.