

David Rhoe  
CRMI  
Paseo de la Fuente  
D-4 Calle Tivoli  
San Juan, PR 00926-6459

January 29, 2014

USNRC Region I DNMS  
2100 Renaissance Blvd  
King of Prussia, PA 19406

RE: Amendment for CRMI license # 52-25430-03  
Addition of I-131 therapy room  
Additional Information

Dear Sir or Madam:

Please add this information to our amendment dated December 13, 2013 letter.

This is to confirm that the adjacent room will remain closed for the first I-131 patient therapy to verify that the radiation exposure through the wall is below 2 mR/hr. If the exposure is above 2 mR/hr, either the I-131 bed will be moved to the opposite wall or the adjacent room will remain closed when an I-131 treatment is performed.

If your need any further information, please contact David Rhoe at 787-245-7248.

Sincerely,



David Rhoe  
CEO