	NRC FORM 591M PART 1 U.S. NUCLEAR REGULATORY COMMISSION								
10 CFR 2.201 SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION									
1. LICENSEE/LOCATION INSPECTED: 2. NRC/REGIONAL OFFICE									
Jackson Cardiology Associates, P.C. 205 Page Avenue Jackson, Michigan 49201			Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352						
REPORT NUMBER(S		•	Le DATE(S) OF INCRESTION						
3. DOCKET NUMBER(S) 4.		4. LICENSE NUMBER	(S)	5. DATE(S) OF INSPECTION					
030-34118		21-26715-01	February 12		4				
LICENSEE: The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows: 1. Based on the inspection findings, no violations were identified. 2. Previous violation(s) closed. From IR 03034118 / 2012col (DNMS) The violations(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied. Non-cited violation(s) were discussed involving the following requirement(s):									
A. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11. (Violations and Corrective Actions) Statement of Corrective Actions I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.									
TITLE	PRINTED NAME		SIGNATURE		DATE				
LICENSEE'S REPRESENTATIVE									
NRC INSPECTOR	Geoffrey M. Warren		By W_		2/12/14				
BRANCH CHIEF	Aaron T. McCraw			1/	2/23/14				

NRC FORM 591M PART 1 (07-2012)

NRC FORM 591M PART 3 (07-2012)		Docket File Info		U.S. NUCLEAR REGULATORY COMMISSION					
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION									
1. LICENSEE/LOCATION INSP	ECTED:		2. NRC/REGIONAL OFFICE						
Jackson Cardiology As 205 Page Avenue Jackson, Michigan 492			Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352						
REPORT NUMBER(S) 20	14-001								
3. DOCKET NUMBER(S)		4. LICENSE NUMBER(S)		5. DATE(S) OF INSPECTION					
030-34118		21-26715-01	·	February 12, 2014					
6. INSPECTION PROCEDURES	SUSED	7. INSPECTION FOCUS AREAS							
87130		03.01 - 03.08, as relevant							
SUPPLEMENTAL INSPECTION INFORMATION									
1. PROGRAM CODE(S) 02201	2. PRIORITY	3. LICENSEE CONTAC		4. TELEPHONE NUMBER (5.1.7) 7.97, 2.577					
	5	Jashu R. Patel, N		(517) 787-3577					
✓ Main Office Ins	pection	Next Inspection	Date: No	o change					
Field Office Inspection									
Temporary Job Site Inspection									
PROGRAM SCOPE									
The purpose of this inspection was to follow up on escalated enforcement action.									
The licensee was cited (SL III) for the failure to either perform an evaluation to demonstrate that exposure would be below ten percent of the regulatory limit or provide dosimetry for several months and for providing information that was not complete and accurate to the NRC inspector. As corrective action, the licensee's RSO has (1) reinstated the contract to provide dosimetry and paid in advance yearly to assure dosimetry is available; (2) taken disciplinary action against the technologist and trained the technologist on providing correct information to the NRC (3) checked daily to ensure that the technologist is wearing dosimetry; (4) met with the physicist during each audit; and (5) taken personal control of the radiation safety program. The inspector verified that dosimetry has been provided since November 2012 and that quarterly reports of exposure are available. The RSO was not available at the time of the inspection, but the office manager verified this information. Based on this information, the violation is closed.									