

Parker, Bryan

From: Dan Dryden <ddryden@northernhealth.org>
Sent: Friday, March 07, 2014 10:45 AM
To: Parker, Bryan
Cc: Mary-Anne Ponti; Shelli Kritzman (mkritzman@mpcphysics.com)
Subject: RE: Info needed for NRC license renewal- Additional Info for CN 582735
Attachments: 2014 3-6 NRC additional info signed copy.pdf; ATT00001.txt

Mr. Parker,

McLaren Northern Michigan acknowledges receipt of RAI. Please see the attached signed pdf in response to your request. If you have questions or require additional information please call on me.

Sincerely,

Dan

Daniel Dryden, MS, RSO, DABR
Medical Physicist
McLaren Northern Michigan
416 Connable Ave
Petoskey, MI 49770
ddryden@northernhealth.org
phone: 231-487-4264
fax: 231-487-7840

From: Parker, Bryan [<mailto:Bryan.Parker@nrc.gov>]
Sent: Thursday, March 06, 2014 4:46 PM
To: Dan Dryden
Subject: FW: Info needed for NRC license renewal

Mr. Dryden,

Please ignore my previous attachment and use this one. I inadvertently left off an item.

I apologize for any confusion.

Bryan

From: Parker, Bryan
Sent: Thursday, March 06, 2014 4:25 PM
To: 'ddryden@northernhealth.org'
Subject: Info needed for NRC license renewal

Mr. Dryden,

I left you a voicemail on this earlier. I am performing the review of the renewal of your NRC License No. 21-16732-01 and have a need for additional information to continue that review.

Please review the attached Request for Additional Info and let me know if you have any questions. You may respond electronically by attaching a **signed PDF file of your response** to an email and sending it to me. Indicate that it is "**Additional Info for CN 582735.**"

I'm hoping this can be turned around quickly so I complete the review soon. Please provide a response within 7 days (by Thu March 13) if at all possible.

Again, if you have any questions, please let me know.

Also, please acknowledge your receipt of this RAI.

Thank you.
Bryan

Bryan A. Parker

Health Physicist

U. S. Nuclear Regulatory Commission - Region III

2443 Warrenville Road, Suite 210

Lisle, IL 60532-4352

bryan.parker@nrc.gov

678-828-7050

630-515-1078 (fax)





NORTHERN MICHIGAN

416 Cannable Ave.
Petoskey, Michigan
49770

tel 231 487 4000

northernhealth.org

3-07-2014

United States Nuclear Regulatory Commission
Region III, Materials Licensing
2443 Warrenville Road
Suite 210
Lisle, IL 60532-4352
630-829-9887

**Re: Additional information for license renewal No. 21-16732-01, McLaren
Northern Michigan
Control No. 582735**

- 1) Yes, we wish to keep the second location in Cheboygan, MI. The address is:


McLaren Northern Michigan Cheboygan Campus
748 South Main Street
Cheboygan, MI ~~49722~~
49721

The Cheboygan site is limited to 10 CFR 35.100, 35.200, 35.300 and 31.11 materials.

- 2) There will be no PET isotopes at any of the locations on this license. PET imaging is provided by a mobile company that has a separate NRC license.
- 3) Please set the total possession limit for all 35.300 materials as One Curie
- 4) Radiation monitory instruments are not calibrated by this license. We contract with a licensed vendor for all radiation monitory instrumentation calibrations. Radiation monitoring instruments will be calibrated by a person qualified to perform survey meter calibrations.

If you have questions or require additional information please contact me at 231-487-4264.

Sincerely,



Dan Dryden, MS, RSO, DABR
Medical Physicist
Radiation Oncology

Cc. Mary-Anne Ponti, COO
Michelle Kritzman, MS, DABR

2014 3-6 NRC additional info.doc