

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Complete Health Systems
 G-1128 S. Linden Road
 Suite 11
 Flint, MI 48532

REPORT NUMBER(S) 14-001

2. NRC/REGIONAL OFFICE

Region III
 U. S. Nuclear Regulatory Commission
 2443 Warrenville Road, Suite 210
 Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-36714

4. LICENSE NUMBER(S)

21-32543-01

5. DATE(S) OF INSPECTION

April 10, 2014

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
 (Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

| TITLE | PRINTED NAME | SIGNATURE | DATE |
|---------------------------|--------------------------------------|--|---------|
| LICENSEE'S REPRESENTATIVE | | | |
| NRC INSPECTOR | Deborah A. Piskura, Health Physicist |  | 4/10/14 |
| BRANCH CHIEF | Aaron T. McCraw, Chief, MIB |  | 4/16/14 |

Docket File Information

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

| | |
|---|---|
| 1. LICENSEE/LOCATION INSPECTED: Complete Health Systems G-1128 S. Linden Road Suite 11 Flint, MI 48532 REPORT NUMBER(S) 14-001 | 2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352 |
|---|---|

| | | |
|--------------------------------------|---|--|
| 3. DOCKET NUMBER(S) 030-36714 | 4. LICENSE NUMBER(S) 21-32543-01 | 5. DATE(S) OF INSPECTION April 10, 2014 |
|--------------------------------------|---|--|

| | |
|--|---|
| 6. INSPECTION PROCEDURES USED 87130 | 7. INSPECTION FOCUS AREAS 03.01- 03.07 |
|--|---|

SUPPLEMENTAL INSPECTION INFORMATION

| | | | |
|---------------------------------|----------------------|---|---|
| 1. PROGRAM CODE(S) 02220 | 2. PRIORITY 3 | 3. LICENSEE CONTACT Anthony Bennett, M.D., RSO | 4. TELEPHONE NUMBER (810) 720-3819 |
|---------------------------------|----------------------|---|---|

Main Office Inspection Next Inspection Date: April 2017

Field Office Inspection 5084 Villa Linde Parkway, Flint, Michigan

Temporary Job Site Inspection _____

PROGRAM SCOPE

This was routine inspection of a private practice physicians group authorized to use licensed material permitted by Sections 35.100 and 35.200. The nuclear medicine department was staffed with one full-time technologist and one PRN who performed approximately 100-120 diagnostic procedures monthly. The licensee conducted licensed activities at its base clinic and operated a mobile nuclear medicine service. Patient studies at the clinic were performed on Mondays and Thursdays. The mobile services were performed Monday through Thursday. The licensee received unit doses for imaging studies from a licensed radiopharmacy.

This inspection consisted of interviews with licensee personnel, a review of select records, a tour of the nuclear medicine department, and independent measurements. The inspector observed licensee personnel perform dose calibrator QA tests, radiation surveys, package surveys, security of byproduct material, and use of personnel monitoring. The inspector observed the licensee staff administer four diagnostic dosages. No mobile services were performed by the licensee on the day of this inspection.

This inspection reviewed the licensee's corrective actions for a violation of Section 35.80(a)(1) identified during the previous inspection. The inspector noted that the licensee maintained signed copies of agreements with each of its mobile service clients. No violations of NRC requirements were identified during this inspection.