

## Null, Kevin

---

**From:** demossd@aol.com  
**Sent:** Monday, April 21, 2014 7:33 PM  
**To:** Null, Kevin  
**Subject:** Renewal of License  
**Attachments:** NRC\_LETTER\_OF\_RESPONSE.pdf; Sources\_Certificate.pdf; Sources remaining to be disposed.docx

Kevin:

Please see the attached.

Thanks again for your assistance.

Don DeMoss



Date: April 17, 2014

Subj: Application for renewal of license # 13-03341-03

Attn: Kevin Null

The purpose of this transmittal is to summarize and to clarify outstanding items for NMC's renewal request.

Attached is the list of sources we want to maintain on our license at the present time. This is the same list of sources that I had previously sent. This time I have attached the documentation that was received with each source from the source manufacturer. To reiterate from my previous letter, these sources will probably be replaced sometime during this calendar year with the EAB series sources from Eckert & Zeigler.

To further simplify the renewal process, remove lead-210 from our renewal application. Should we require it in the future I will request an amendment.

As I also stated previously, the users of NMC's instruments often require the use of a Cesium-137 check source with an approximate activity of 5  $\mu$ Ci and a 2.0  $\mu$ Ci Barium-133 to verify correct operation of detectors. It is my understanding these sources are designated by Eckert & Zeigler as GF-137 and GF-133 respectively.

Additionally, NMC's customers often require calibration sources which are usually one or some combination of Cesium-137, Chlorine-36 or Barium-133. We are therefore requesting that our license reflect that we can redistribute Cesium-137, Chlorine-36 and Barium-133 calibration sources. They are designated as EAB-137, EAB-036 and EAB-133 series sources by Eckert & Zeigler.

50.8 mm OD x 45 mm AD  
Deposited onto polymeric Membrane,  
0.9 mg/cm<sup>2</sup> Aluminized Mylar Cover, NIST.  
Exempted Quantities

Internally we occasionally have a need to use Strontium-90, Cobalt-60, Americium-241, Krypton-85 and Xenon-133 for calibration purposes. These would also be purchased for Eckert & Zeigler and would be of the EAB series. Therefore, we have a need to maintain these isotopes on our license.



Finally, I have also attached a list of sources that are scheduled with Chase Environmental Group to be picked up and disposed of in May 2014.

All other line items in our current licenses in Item 6 can be removed. This would include 6B, 6F, 6J, 6L, 6M, 6N, 6O, 6P, and 6Q.

Thank you for your assistance in this matter.

Very truly yours,

A handwritten signature in black ink, appearing to read "Donald L. DeMoss", is written over a horizontal line.

Donald L. DeMoss

President

Cell Phone: 317-590-3824

Office: 317-546-2415

Email: [demosd@aol.com](mailto:demosd@aol.com)



Radiation standards and check sources  
2810 Siler Lane, Santa Fe, NM 87501  
(505)473-9538 FAX(505)473-5805

Ref. PO# 42521

## Certificate of Calibration (Gamma Sources)

The Cesium 137 gamma ray emission rate was compared with a NIST (formerly NBS) SRM 4200B-134. The comparison was completed using a Sodium Iodide or GeLi gamma detector.

The gamma activity of the standard on May 12, 1991 was 0.215uCi.

The uncertainty of the measurement and the activity of the source is  $\pm 10\%$  which is the sum of the uncertainty of the NBS standard and the random error of counting at the 99% confidence interval.

Serial NO. 91CS4702313

Active Diameter(or area) 42mm Mounting Material SS/Kapton/Al

Total Diameter(or area) 47mm Thickness 2mm

Calibrated by: Michael A. Ortiz Calibration Manager  
Date Signed: 6-5-91 Michael A. Ortiz

Approved by: Charles L. Gonzales Q.A. Manager Charles L. Gonzales



Radiation standards and check sources  
2810 Siler Lane Santa Fe, NM 87501  
(505)473-9538 FAX(505)473-5805

Ref. PO # 47702

## Certificate of Calibration (Gamma Source)

The Barium 133 gamma source was compared with a calibrated NIST gamma source 4241B-19 used in establishing traceability. The comparison was completed using a sodium iodide detector.

The activity of the source on 04-01-95 was 0.48uCi.

The overall uncertainty of the measurement and the activity of the source is  $\pm 10\%$ .

Serial No.	<u>95BA5002651</u>	Cover	<u>Mylar</u>	Thickness	<u>2mm</u>
Model	<u>S-Ba-50</u>	Mounting Material	<u>Plastic</u>		
Active Diameter (or area)	<u>44mm</u>	Total Diameter(or area)	<u>50mm</u>		

Michael A. Ortiz

Michael A. Ortiz  
Calibration Manager

Charles L. Gonzales

Charles L. Gonzales  
Quality Assurance Manager

less than 0.005

leak test results(uCi)

The overall uncertainty of the measurement is three times the value found from combining quadratically the sum of the overall uncertainty reported by NIST in the radioactive measurements assurance program; the standard deviation of the mean for the NIST standard as measured in the system used for calibration; and the standard deviation of the mean for the source measurements.

GC008-95



Radiation standards and check sources  
2810 Siler Lane Santa Fe, NM 87501  
(505)473-9538 FAX(505)473-5805

## Certificate of Calibration (Alpha Source)

The Americium 241 alpha source was measured in a gas proportional counter using P-10 as counting gas. The alpha emissions from the surface of the source were measured at its plateau voltage to determine its  $2\pi$  cpm rate. Corrections were applied for background, coincidence loss and backscatter factors when applicable. Alpha standard 93AM3204191 is our NIST referenced source used in establishing NIST traceability.

REF.PO# 48066  
Model S-Am-47

Active Diameter(or area)	<u>2.5mm</u>	Mounting Material	<u>SS</u>
Total Diameter(or area)	<u>4.7mm</u>	Thickness	<u>0.79mm</u>
<u>35,600</u>	cpm $\pm$	<u>1,780</u>	cpm $2\pi$
<u>70,200</u>	dpm $\pm$	<u>3,510</u>	dpm $4\pi$
<u>0.0316</u>	microcurie		
<u>08/02/95</u>	date of measurement		
<u>95AM4703621</u>	source serial number		
<u>5.0</u>	overall uncertainty(percent)		
<u>1.5</u>	backscatter (percent)		

  


Michael A. Ortiz  
Calibration Manager  
Charles L. Gonzales  
Quality Assurance Manager

less than 20 leak test results(dpm /source100cm<sup>2</sup>)

The overall uncertainty of the measurement is three times the value found from combining quadratically the sum of the overall uncertainty reported by NIST in the radioactive measurements assurance program; the standard deviation of the mean for the NIST standard as measured in the system used for calibration; and the standard deviation of the mean for the source measurements. AC006-95



Radiation standards and check sources  
2810 Siler Lane, Santa Fe, NM 87501  
(505)473-9538 FAX(505)473-5805

REF.PO# 43098

## Certificate of Calibration (Beta Sources)

The Strontium Yttrium 90 beta source was measured in a hemispherical 2pi windowless proportional counter using P-10 as counting gas. The beta emissions from the surface of the source were measured at its plateau voltage to determine its 2pi cpm rate. Corrections were applied for background, coincidence loss and backscatter factors when applicable. The source is referenced to NIST(formerly NBS) 4919E

Active Diameter(or area) 41mm Mounting Material SS/Al/Kapton  
Total Diameter(or area) 47mm Thickness 2mm

1,390,000 cpm +/- 69,500 cpm 2pi

1,990,000 dpm +/- 99,500 dpm 4pi

0.896 microcurie

07-10-91 date of measurement

Michael A. Ortiz Calibration Manager [Signature]

Charles L. Gonzales Q.A. Manager [Signature]

91SR4702793 source serial number

The uncertainty of the measurement at the 99% confidence interval is 5 percent.



Radiation standards and check sources  
2810 Siler Lane, Santa Fe, NM 87501  
(505)473-9538 FAX(505)473-5805

Ref. PO# 46403

## Certificate of Calibration (Gamma Source)

The Cobalt 60 gamma ray emission rate was compared with NIST referenced source SRM4203D-18. The comparison was completed using a Sodium Iodide or GeLi gamma detector.

The activity of the source on 12-01-93 1200EST was 0.015uCi.

The uncertainty of the measurement and the activity of the source is 10.0% which is the sum of the uncertainty of the NIST standard and the random error of counting at the 99.0% confidence interval.

Serial NO. 93CO4704147 Cover \_\_\_\_\_  
Active Diameter(or/area) 44mm Mounting Material SS  
Total Diameter(or/area) 47mm Thickness 0.79mm

  


Calibration Manager  
Michael A. Ortiz  
Q.A. Manager  
Charles L. Gonzales

The source was leak tested and was found to have <0.005uCi removeable activity.

GC007-93



### Isotopes Scheduled for Disposal in May

Isotope	Activity ( $\mu\text{Ci}$ )
Americium-241	2.16
Barium-133	6.91
Carbon-14	0.79
Chlorine-36	0.85
Cobalt-60	0.06
Cesium-137	18.23
Europium-152	0.14
Hydrogen-3	0.04
Potassium-40	0.0002
Lead-210	0.059
Strontium-90	0.01
Technicum-99	1.00
Thalium-204	0.01