


Community Hospital Anderson
License No. 13-32500-01
Docket No. 030-36525
Control No. 563446

Dear Ms. Mihail: I have been assigned the renewal of the subject license which is dated March 24, 2014 and signed by you. On March 27, 2014, I called you and explained that we did not have enough information to proceed with the renewal of this license. In your letter dated April 3, 2014, you provided additional information. However, you also requested the addition of 35.100, 35.200 and 35.300 material to be added to your license and for that material more information is needed. Please respond the following in writing:

1. Please provide authorized users (AU's) for the 35.100 and 200 material. None of the current AU's is authorized for this material.
2. Please provide a diagram(s) show the areas where 35.100, 200 and 300 radionuclides will be used. Drawings should be to scale and include room numbers.
3. Please indicate whether PET radionuclides will be used. If so, provide shielding calculations as well as diagrams showing use areas.
4. Please provide information concerning your dose calibrator or make the following statement: "Equipment used to measure dosages will be calibrated in accordance with nationally recognized standards or the manufacturer's instructions."
5. Please provide area survey procedures or make the following statement: " We have developed and will implement and maintain written procedures for area surveys in accordance with 10 CFR 20.1101 that meet the requirements of 10 CFR 20.1501 and 10 CFR 35.70."
6. Please provide procedures for safe use of unsealed licensed material or make the following statement: "We have developed and will implement and maintain written procedures for safe use of unsealed byproduct material that meet the requirements of 10 CFR 20.1101 and 10 CFR 20.1301."
7. Please provide spill/contamination procedures or make the following statement: "We have developed and will implement and maintain written procedures for safe response to spills of licensed material in accordance with 10 CFR 20.1101."
8. Please confirm that the HDR upgrade to Model 106.990 has been completed.
9. As a first time 35.300 radionuclide user, it is not clear to me why you would need a total possession of 9 curies. Please confirm and explain.
10. Please confirm that Drs. Han, Dugan, Marvel, Shaw and DesRosiers are no longer to be listed on your license.
11. Please confirm the physicists Slessinger, Dickow, Wang, and Longacre are no longer to be listed on your license.

Please respond to each of the items above in writing on letterhead signed by the RSO or a member of management. You may email your response to me or fax your response to 630-515-1078. Please provide your response as soon as possible but no later than May 31, 2014. If you have any questions, please call me at 630-829-9842.


Toye L. Simmons

Date

4/30/14

Simmons, Toye

From: Simmons, Toye
Sent: Wednesday, April 30, 2014 9:57 AM
To: 'amihail@inphysics.net'
Subject: RENEWAL OF COMMUNITY HOSPITAL ANDERSON'S LICENSE
Attachments: Scan001.PDF

Hi Ms. Mihail, I have reviewed the information that your sent regarding the subject license renewal and have questions. Please review the attached and provide the information requested in writing. Feel free to call me if you have questions.

Toye L. Simmons
Materials Licensing Branch
U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352
Office: (630) 829-9842
Fax: (630) 515-1078