

From: [Elliott, Robin](mailto:Robin.Elliott@nrc.gov)
To: bawheatley@srmedicalcenter.org
Cc: drudder@srmedicalcenter.org; [Weidner, Tara](mailto:Weidner.Tara@nrc.gov)
Subject: Renewal of USNRC License No. 55-17986-01
Date: Monday, May 12, 2014 10:53:00 AM

Licensee Name: Schneider Regional Medical Center
License No.: 55-17986-01
Docket No.: 030-13764
Mail Control No.: 582662

Dear Mr. Wheatley,

As per my phone conversation today regarding your April 23, 2014 letter with Damian Rudder, on your behalf, additional information is needed to process the license renewal for Schneider Regional Medical Center.

- Please provide a description of the adjacencies above and below the nuclear medicine facilities as well as what is behind the decay storage location.
- On page 7 you indicate that you have “developed and will implement and maintain procedures for safe use of unsealed byproduct material that meet the requirements of 10 CFR 20.1101 and 10 CFR 20.1301.” Please confirm that these procedures are written.
- Please provide technical specifications for the SWGM probe for the Eberline E-520 meter you referenced on page 5.
- Page 18 refers to performing a radiation survey with a survey meter calibrated with the last two years. Please explain the reason for the reference to the two year calibration period.

When submitting your response, please sign the letter transmitting the information. Once we receive the additional information requested, we should be able to finalize the processing of your license renewal. You may respond to my attention in writing by letter or fax (610-337-5269), referencing mail control 582662. If we do not receive a reply from you within 30 calendar days from the date of this e-mail, we will assume that you do not wish to pursue your renewal.

Regards,

Robin L Elliott

Health Physicist
U. S. Nuclear Regulatory Commission
Region I, Division of Nuclear Materials Safety
2100 Renaissance Blvd
King of Prussia, PA 19406-2713
(610) 337-5076 voice
(610) 337-5269 fax
Robin.Elliott@nrc.gov