

From: Eric Jameson
To: EPD-TT.SMTP("Jim.Hoey@vai-atl.ccmil.valmet.com")
Date: 5/23/96 9:30am
Subject: Re: Tapio device evaluation, Model BW-5h23 -Reply

Jim

In regard to the following excerpt of correspondence:

>>> <Jim.Hoey@vai-atl.ccmil.valmet.com> 05/22/96 05:24pm

. . . . Note that there is no 15mm aperture in the BW-5h23 device. It only contains a 2mm and a 5mm. The survey patterns were, and will be, performed with the 5mm aperture for this unit. . . . <<<

Could you, in your submittal of the referenced patterns, also include 1) a copy of the Aperture Disk drawing for the BW-5h23, similar to drawing # 3TBW2-04-4M for the BW-2h55; and 2) if there are any differences from the BW-2h55, a revised Functional Description describing and showing the operation of the solenoids/selection of different apertures (reference item D in Device Registry for BW-2h55, GA 596-D-111-G, letter & attachments to William Slocomb from Jukka Perento, dated Feb 25, 1993).

If you have any questions, feel free to contact me via phone or e-mail.

Sincerely,

Eric Jameson