

From Seppo Väisälä
Chief Inspector
Radioisotope Applications
FINNISH CENTRE FOR RADIATION AND NUCLEAR SAFETY

Date May 30. 1996

To Jim Hoey
Valmet Automation Inc

Re: Basis Weight Sensor Radiation Test Report

I have performed the radiation measurements for Tapio Technologies Inc for the sensor model BW-5h23. (at the date of may 24. 1996).
Test report page: "Strips in the upmost position" - shutter open and 7mg/cm2 window has been measured fully and totally by me. The blank spaces on this are the indication of the radiation levels less than 1 microSv/h, which I hereby testify and confirm.

Espoo May 30. 1996

Seppo Väisälä

TO: ERIC 404-362-2653

HERE IS THE CONFIRMING STATEMENT
WE REQUESTED FROM THE PERSON PERFORMING
THE RADIATION PATTERNS ON THE
TAPIO DEVICE.

REGARDS,

Jim Hoey