




NRC FORM 591M PART 1 <small>(07-2012)* 10 CFR 2.201</small>		U.S. NUCLEAR REGULATORY COMMISSION	
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION			
1. LICENSEE/LOCATION INSPECTED: Midland Trail Energy LLC 3301 Point Lick Drive and 110 Catenary Coal Road Charleston, West Virginia		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region I, 2100 Renaissance Blvd, Suite 100 King of Prussia, Pennsylvania 19406-2713	
REPORT NUMBER(S) 2014-001			
3. DOCKET NUMBER(S) 030-37777	4. LICENSE NUMBER(S) 47-31315-01	5. DATE(S) OF INSPECTION April 30, May 19 and 28, 2014	
LICENSEE: The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:			
<input type="checkbox"/> 1. Based on the inspection findings, no violations were identified.			
<input type="checkbox"/> 2. Previous violation(s) closed.			
<input type="checkbox"/> 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.			
Non-cited violation(s) were discussed involving the following requirement(s) and corrective action(s): _____			
<input checked="" type="checkbox"/> 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with the NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11. (Violations and Corrective Actions)			
See Form 591 Part 2			
Statement of Corrective Actions			
I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.			
Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE	David Kimbrell		5/28/2014
NRC INSPECTOR	Dennis Lawyer		5/28/2014
BRANCH CHIEF	Blake Welling		5/28/14

*NRC FORM 591M PART 1 (07-2012) (RI Rev. 09/12/2013) Part1.doc

G:\WordDocs\Current\Insp Record\47-31315-01.2014001.591M-

SUNSI Review Completed By: / RA / Dennis Lawyer

Public

Non-Sensitive

This document becomes an NRC Official Agency Record once it is signed by the Branch Chief

NRC FORM 591M PART 2 (07-2012) 10 CFR 2.201		U.S. NUCLEAR REGULATORY COMMISSION	
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION			
1. LICENSEE/LOCATION INSPECTED: Midland Trail Energy LLC 3301 Point Lick Drive and 110 Catenary Coal Road Charleston, West Virginia		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region I, 2100 Renaissance Blvd, Suite 100 King of Prussia, Pennsylvania 19406-2713	
REPORT NUMBER(S) 2014-001			
3. DOCKET NUMBER(S) 030-37777	4. LICENSE NUMBER(S) 47-31315-01	5. DATE(S) OF INSPECTION April 30, May 19 and 28, 2014	
<p>(Continued)</p> <p>License Condition 24 of License No. 47-31315-01 requires, in part, that the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures in the license application dated April 26, 2008.</p> <p>Item 10 of the license application dated April 26, 2008, states, in part, that each survey meter will be calibrated by the manufacturer or other person authorized by the NRC or an Agreement State to perform survey meter calibrations.</p> <p>Contrary to the above, the licensee did not possess a survey meter that was calibrated by the manufacturer or other person authorized by the NRC or an Agreement State to perform survey meter calibrations. Specifically, as of April 30, 2014, the licensee possessed a Radiation Alert Monitor Model 4EC, which was used for surveys of licensed material on February 12, 2014, and was not calibrated by the manufacturer or other persons authorized by the NRC or an Agreement State to perform survey meter calibrations.</p> <p>The licensee stated that the instrument would be sent for calibration and returned within 30 days. Methods would be employed to ensure the instrument will remain in calibration.</p>			